



महाराष्ट्र शासन, पशुसंवर्धन विभाग
महाराष्ट्र पशुधन विकास मंडळ, नागपूर, (अंतर्गत)

विभागीय व्यवस्थापक, शासकीय गोठित रेतमात्रा निर्मिती केंद्र खडकी, पुणे-03

ईमेल आयडी - rmfslpune@gmail.com

दुरध्वनी क्र.020-29910679

जा.क्र.शागोरेनिके/प्रयोगशाळा/ Air Shower with Biometry 960 / २०२६ खडकी, पुणे-०३, दिनांक ११/०३/२०२६

जाहीर सुचना

प्रति,

१. लागू असेल त्या,
सेवा पुरवठादार यांना.
२. सुचना बोर्ड/फलक

विषय:- शासकीय गोठित रेतमात्रा निर्मिती केंद्र पुणे-३ संस्थेकरीता Air Shower with Biometry खरेदी करणेसाठी दर पत्रके मागविण्याकरीता जाहीरात प्रसिध्द करणेबाबत...

उपरोक्त विषयास अनुसरून कळविण्यात येते की, या कार्यालयातील प्रयोगशाळा विभागाकरीता लागणाऱ्या संस्थेकरीता Air Shower with Biometry खरेदी करणेकरीता सेवा पुरवठादारांकडून खुल्या बाजारातून दर पत्रके मागविण्यात येत आहेत. तरी सदर खरेदी करीता लागणाऱ्या खर्चाचे सर्व करासहीत दरपत्रक या कार्यालयास दि.२४.०२.२०२६ रोजी सकाळी ११.०० वा. पर्यंत या कार्यालयास पोहचेल असे सादर करण्यात यावे. सोबत पुरवठादारांनी सादर करावयाचा दरपत्रकाचा नमुना यादीसह.

अटी शर्ती:-

१. दरपत्रके ही स्पेसीफीकेशन नुसारच असावी.
२. सदर उपकरण हे नमुद केलेल्या स्पेसीफीकेशन नुसारच आहे या बाबतचे अभियांत्रिकी महाविद्यालय, शिवाजीनगर, पुणे (सीओपी कॉलेज ऑफ इंजिनियरींग) यांचे प्रमाणपत्र आवश्यक आहे.
३. दरपत्रक संस्थेच्या लेटर हेड वर सही शिक्क्यासह हार्ड कॉपी द्वारे सादर करणे आवश्यक.
४. दरपत्रकासोबत संस्थेच्या पॅनकार्ड ची छायांकीत प्रत जोडणे आवश्यक.
५. ई-मेल द्वारे सादर केलेली दर पत्रके ग्राह्य धरण्यात येणार नाहीत.
६. दरपत्रकातील दर सर्व खर्चासहीत व सर्व करासहीत देण्यात यावेत.
७. दि.२४.०२.२०२६ रोजी सकाळी ११.०० वा. नंतर प्राप्त होणाऱ्या दर पत्रकांचा विचार केला जाणार नाही याची कृपया नोंद घ्यावी.

Rishal
11/2/26

विभागीय व्यवस्थापक,

शासकीय गोठित रेतमात्रा निर्मिती केंद्र,
खडकी, पुणे-3

Name of Supplier :-

Date:- / /

Quotation for Air Shower with Biometry
To,

Regional Manager,
Govt.Frozen Semen Station ,
Mula Road, Khadki, Pune-3.

Sr No	Manufacturing Company (Make)	Unit Quantity	Basic Rate Rs	GST Rs	Total Rate Including GST & Freight
1		1			

Authorized Signature

NB:- Quotation will not be accepted by mail.

Specifications

Overall Size	:- W 1100 mm x D 1000 mm x H 2800 mm
Dimension of Work Area	:- W 650 mm x D 1000 mm x H 2000 mm
Person Capacity	:- 1 To 2 Person at a time
Final Supply Filter	:- Mini-pleat box type 0.3 micron HEPA filter, efficiency 99.99 % down to 0.3 micron Particle size, Aluminum Anodized Casing. Size :- 610 mm x 610 mm x 116 mm Qty. - 01 No
Pre-Filter	:- Box type pleated efficiency 90 % down to 10 micron, HDPE + Non woven + HDPE mesh media, aluminum casing,
Motor	:- Hindustan Make, 0.75 HP, 2800 RPM, 440 Volt. 50 Hz., Floor mounted, Quantity - 02 Nos.
Casing	:- Fiber Glass - 02 Nos.
Blower	:- Aluminum Impeller - Quantity - 02 Nos.
Platform for Blower Motor Assembly.	:- Impermeable rexin with Flange.
Emergency off key	:- In side of 'Air Shower' for blower off & unlock of doors.
D-interlocking key	:- Outside of 'Air Shower' for unlock of doors.
Air Nozzles	:- Embossed Type nozzles on both & top sides. Left Side - 07 Nos , Right Side - 07 Nos, Top Side - 02 Nos.
Interlocking & other	:- Micro controller based controller
Controls	
Blower Timer	:- Max. 2 Minutes. Adjustable. Set timing by pot. Pot installed on PCB at component board.
Lighting Arrangement	:- Flushed type LED Fitting, 12 Watt x 01 No.
Power Supply Cable	:- 1.5 mm ² x 4 core x 3 Mtrs. Flexible
Required Supply Provision	:- 3 phase neutral, 4 pole MCB, 25 Amp
Maintenance Access	From in side of 'Air Shower'
Material of Construction OC	:- MSCRCA Powder Coated, 1.2 mm Thick Seven Tank Processing.
PERFORMANCE :-	
Air Cleanness	:- Class 100 (As per U.S. Fed, Std. 209E) Class ISO - 5 (As per ISO - 14644.1)
Efficiency of HEPA	:- 99.99 % down to 0.3 micron particle size.
Air Jet Velocity	:- Side Nozzles 26 M/Sec +/- 1 M/sec. Top Nozzles 23 M/Sec +/- 1 M/sec.
Noise Level	:- Less than 80 Db when ambient sound level is not greater 60 Db.
CCTV and Safety Buzzer	For ensuring security of personnel working in the BSL 3 Laboratory.


Regional Manager
 Frozen Semen Laboratory
 Khadki, Pune-03.