

Corrigendum Notice

Tender ID: 2026_UTDNH_7897_1

Subject: Purchase of Equipment's for Establishing of New Drugs Testing Laboratory for UT of DNH&DD.

The Department of Health & Family Welfare, DNH & DD, on behalf of the President of India, had invited online tender on <https://dnhtenders.gov.in/niceplapp> from the Manufacturers/Authorized Dealers/Suppliers for Purchase of Equipment's for Establishing of New Drugs Testing Laboratory for UT of DNH&DD, vide Tender ID :2026_UTDNH_7897_1

It is hereby notified that the provisions detailed below pertaining to specifications, warranty, and AMC/CMC for the purchase of equipment for the establishment of a new Drug Testing Laboratory for the UT of DNH & DD shall stand updated in the tender document. All other provisions of the tender shall remain unchanged.

Equipment Name	Specification Incorporated in tender document	Updated specification
Tablet Disintegration Apparatus	Weight: 20kg	Weight: 20kg + or - 5kg
	Temperature range: 2°C above room temp. to 40°C	Temperature range: 2°C above room temp. to 50°C (+ or - 10°C)
Tablet Dissolution Apparatus with Auto sampler	RPM – 25 to 200	RPM – 25 to 250 or more
	Large Graphic LCD (240×64) to indicate set parameter, like: RPM, Vessel Temperature, Time interval, Test time etc.	Large Graphic LCD / LED touch screen (240×64) to indicate set parameter, like: RPM, Vessel Temperature, Time interval, Test time etc.
	Temperature range: 30°C – 40°C with accuracy + or - 0.2°C	Temperature range: 30°C – 50°C with accuracy + or - 5°C
	Additional Specifications	High Pressure Pump: • 2 port surfactant pump

यू.टी. प्रशासन दादरा एवं नगर हवेली और दमन एवं दीव
 U.T. Administration of Dadra & Nagar Haveli and Daman & Diu
 स्वास्थ्य और परिवार कल्याण विभाग, डीएनएच और डीडी
 Department of Health and Family Welfare, DNH & DD
 केंद्रीय खरीद शाखा का कार्यालय
 Office of the Central Procurement Branch
 DNH & DD.

E-mail: cpbdaman106@gmail.com

Equipment Name	Specification Incorporated in tender document	Updated specification
		• Should be suitable for surfactant and SLS media
	Additional Specifications	Sample Volume: 0.5 to 25 ml or more
FTIR with ATR	Mirror Coating – Gold	Mirror Coating – Gold/Diamond
Cooling cabinet	Capacity: 200 ltr	Capacity: 200 ltr or more
Digital Photo Fluorometer	Overall size: 262 mm × 340 mm × 165 mm (approx) Weight: 5 kg (approx)	Overall size (W×D×H): (280 mm × 310 mm × 140 mm) + or – 20 mm Weight: 7 kg or less
Digital Refractometer	URL of the uploaded product catalogue: http://www.atago.net/	Removed
	Brix measurement range: 0 to 100%, 0 to 85%, 0 to 40%, 0 to 51%, NA	Brix measurement range: 0 to 100%, 0 to 85%, 0 to 40%, 0 to 51%
	Resolution of refractive index measurement range: 1.0e–06 percent	Resolution of refractive index measurement range: 0.1%

The warranty period for all equipment shall be **two (2) years.**

AMC/CMC – FTIR with ATR, HPLC with PDA detector, Gas Chromatography with Head space,

1. The AMC/CMC charges to be clearly mentioned by the Manufacturer on their letter head. If the Authorized dealer is going to carry out the service then they have to furnish the authority letter given by the manufacturer to sale/service the specified product in this Territory.
2. Rates for ***AMC / CMC Maintenance Contract*** for Eight Years (after warranty period of Two Years) should be mentioned separately in the Financial Bid. It should be clearly mentioned whether AMC / CMC will be done through company itself or its service franchise/dealers. In that case Manufacturing Company must give authority letter to such

यू.टी. प्रशासन दादरा एवं नगर हवेली और दमन एवं दीव
 U.T. Administration of Dadra & Nagar Haveli and Daman & Diu
 स्वास्थ्य और परिवार कल्याण विभाग, डीएनएच और डीडी
 Department of Health and Family Welfare, DNH & DD
 केंद्रीय खरीद शाखा का कार्यालय
 Office of the Central Procurement Branch
 DNH & DD.

E-mail: cpbdaman106@gmail.com

franchise/dealers on their letter head clearly mentioning service period and AMC / CMC for period of Eight Years.

Type of AMC / CMC	AMC Rates (excluding taxes) (Labor Charge + Repairing)	CMC Rates (excluding taxes) (Including -Labor charges +repairing, all spares, probes, software, printer, Display/screen)	Executed by (manufacturers/ authorized service dealers) Name and address to be specified here
1 st Year (after warranty)			
2 nd Year			
3 rd Year			
4 th Year			
5 th Year			
6 th Year			
7 th Year			
8 th Year			

Note:

- i) Quoted AMC / CMC price not more than cost of the system, otherwise offer will be outrightly rejected. The rates of AMC / CMC price should be quoted in Indian Rupees only.
- ii) The Rates quoted should be excluding taxes. Taxes shall be applicable extra as prevalent in the respective Year.
- iii) **The AMC/CMC rates should be mentioned on Manufacturer's letter head or Authorized Service dealer letter head accompanying the dealership letter, else outrightly rejected.**
- iv) Consideration of AMC/CMC, if any, for financial evaluation shall be subject to the decision of the parent department.

Sd/-
 In-charge
 Central Procurement Branch,
 DNH & DD
 E-mail Id: cpbdaman106@gmail.com