

SUPPLY ORDER

PHONE:- 0172-2570121, 2570123, 2570124
FAX: 0172-2570122

REGD.

From:

The Director General Supplies and Disposals,
Haryana, S.C.O. No.09, 1st & 2nd Floor, Sector-16,
PANCHKULA. E-mail: supplies@hry.nic.in

To

M/S Sterling Enterprises,,
Plot No.921,Industrial Area,
Phase-II, Ramdarbar,Chandigarh- 160002
E-Mail: sterling.chd@gmail.com

Memo. No. 01/HR/G-2/2023-24/
Dated Panchkula, the:-

Subject:- Purchase of Large size (having 226.40 Ltrs. Capacity) Ice Lined Refrigerators required by Haryana Livestock Development Board, Panchkula, tender opened on 05.02.2024.

With reference to your offer No. Nill dated 02.02.2024 and your letter No. NIL dated 04.03.2024, this office acceptance letter No.23234 dated 06.03.2024 and your letter No.NIL dated 15.03.2024, on the subject noted as above.

1. Please supply the following material within 45 days from the date of issue of this supply order, under the terms and conditions already accepted by you in the tender to the consignees to be intimated by the Managing Director, Haryana Livestock Development Board (HLDB), Panchkula.
2. The material may please be sent direct to the consignees to be intimated by the Managing Director, Haryana Livestock Development Board (HLDB), Panchkula for payment and the bill of the cost in triplicate to the consignees to be intimated by the Managing Director, Haryana Livestock Development Board (HLDB), Panchkula under advice to this office.

The Rates are F.O.R. destination basis per No. Large size Ice Lined Refrigerators (ILR) having capacity (226.4 Ltr.) including 2 years warranty & 3 years AMC after expiry of warranty period.

#	Descriptions of store	Type of ILR	Make and Model	Qty.	Gross rates with Taxes (GST @18% & Freight charges etc.) and F.O.R. destination basis in Rs.
1.	Large size of Ice Lined Refrigerators (ILR) having capacity 226.4 Ltrs.	Horizontal	Make: Godrej Model No. GHR200AC	300 Nos.	1,70,000.00

SPECIFICATION OF LARGE SIZE ILR:-

Sr.	Parameter	Values
1	Vaccine Storage Capacity	200-300 Litres
2	Cabinet Type	Vertical/Horizontal
3	Environment Requirement during Transport and storage (°C)	+30°C to +55°C when product is inactivated Ambient humidity- 5% to 95 % RH
4	Refrigerant	R134a/CFC free
5.	Temperature Range	2°C to + 8°C
6.	Temperature Controller	<p>a. Refrigerator Compartment: The entire vaccine load must remain within the acceptable temperature range during any continuous ambient temperature or day/night cycling temperature.</p> <p>b. Thermo stat: Programmable thermostats which are password protected. Temperature Display unit should be have Inbuilt Battery Backup.</p>
7.	Refrigerant	Compressors operating on R 134a refrigerant to be preferably used
8.	Temperature Display	Externally Readable cabinet-mounted electronic temperature display with alarm within built battery backup.
9.	Power Supply (V/Hz)	220-240V, 50 Hz
10	Minimum Holdover Time	More than 30 Hours

11.	Baskets/ Shelves (Minimum)	2
12.	Certificate Required	Electrical safety rating: Manufacturer to certify compliance with IEC 60335-1 and IEC 60335-2-24
13.	Certificate	Product must be tested and certified from WHO-PQS . The WHO-PQS certificate is to be provided along with other documents
14.	Freezer Protection Grade	'A'
15.	Minimum Warranty	2 years onsite warranty and followed by 3 years of onsite service commitment. Availability of spares parts for at least 10 years.
16.	Lock	The door or lid must be fitted with lock with minimum two keys
16.	Corrosion resistance	Internal and external cabinet, lid and frame protected against corrosion.
17	Physical Characteristics	Must be supplied with an appropriate device of equivalent performance to those voltage stabilizer

TERMS AND CONDITIONS:-

1. **F.O.R.:-** The above rates are F.O.R. destination basis to be supplied at Government Veterinary Hospitals/Dispensaries at various places in the State of Haryana at supplier's risk.
2. **GST:** - Inclusive in above rates.
3. **FREIGHT CHARGES:** Inclusive in above rates
4. **DELIVERY PERIOD:** The delivery should be within at least 45 days from the issuance of purchase/supply order.
5. **GUARANTEE/WARRANTY:** The Guarantee/Warranty shall be applicable as 2 years onsite warranty and followed by 3 years of onsite service commitment. Availability of spares parts for at least 10 years specified with the individual item. The warranty period will commence from the date of successful installation & running. If equipment is not repaired within 7 days after receipt of first call or complaint during guarantee / warranty period, the firm will have to pay a downtime penalty @ 0.2% of the net cost of equipment per day or a sum of Rs. 500/- per day, whichever is higher. If equipment is not repaired within 14 days without any cogent reasons, the department shall be at liberty to get the instrument/equipment repaired from local market at the cost of the firm.