

SUPPLY ORDER

PHONE:- 0172-2570121, 2570123, 2570124

FAX: 0172-2570122

REGD.

From:

The Director General Supplies and Disposals,
Haryana, S.C.O. No.09, 1st & 2nd Floor, Sector-16,
PANCHKULA. E-mail: supplies@hry.nic.in

To

M/S Newtek Instruments,
Block B (Ground Floor), Factory No.2,
DSIDC Industrial Complex, Jhilmil-Delhi- 110095
E-Mail: newtekinstruments@gmail.com

Memo. No. 01/HR/G-2/2023-24/
Dated Panchkula, the:-

Subject:- Purchase of Small size (having 120 Ltrs. capacity) Ice Lined Refrigerators required by Haryana Livestock Development Board, Panchkula, tender opened on 05.02.2024.

With reference to your offer No. Nil dated 31.01.2024 and your letter No. NIL dated 04.03.2024, this office acceptance letter No.23235 dated 06.03.2024 and your letter No.NIL dated 13.03.2024 received in this office on 14.03.2024, on the subject noted as above.

1. Please supply the following material within 45 days from the date of issue of this supply order, under the terms and conditions already accepted by you in the tender to the consignees to be intimated by the Managing Director, Haryana Livestock Development Board (HLDB), Panchkula.
2. The material may please be sent direct to the consignees to be intimated by the Managing Director, Haryana Livestock Development Board (HLDB), Panchkula for payment and the bill of the cost in triplicate to the consignees to be intimated by the Managing Director, Haryana Livestock Development Board (HLDB), Panchkula under advice to this office.

The Rates are F.O.R. destination basis per No. Small size Ice Lined Refrigerators (ILR) having capacity (120 Ltrs.) including 2 years warranty & 3 years AMC after expiry of warranty period.

#	Descriptions of store	Type of ILR	Make and Model	Qty.	Gross rates with Taxes (GST @18% & Freight charges etc.) and F.O.R. destination basis in Rs.
1.	Small size of Ice Lined Refrigerators (ILR) having capacity 120 Ltrs.	Horizontal	Make: Western Model No. I425H120	87 Nos.	1,40,000.00

SPECIFICATION OF LARGE SIZE ILR:-

Sr.	Parameter	Values
1	Vaccine Storage Capacity	75-150 Litres
2	Cabinet Type	Vertical/Horizontal
3	Environment Requirement during Transport and storage (.C)	+30.C to +55.C when product is inactivated Ambient humidity- 5% to 95 % RH
4	Refrigerant	R134a/CFC free
5.	Temperature Range	2.C to + 8.C
6.	Temperature Controller	Refrigerator Compartment: The entire vaccine load must remain within the acceptable temperature range during any continuous ambient temperature or day/night cycling temperature. Thermostat: Programmable thermostats which are password protected. Temperature Display unit should be have Inbuilt Battery Backup.
7.	Temperature Display	Externally Readable cabinet-mounted electronic temperature display with alarm.
8.	Refrigerant	Compressors operating on R 134a refrigerant to be preferably used
9.	Power Supply (V/Hz)	220-240V, 50 Hz
10	Minimum Holdover Time at 43.C	More than 30 Hours

11.	Baskets/ Shelves (Minimum)	2
12.	Certificate Required	Electrical safety rating: Manufacturer to certify compliance with IEC 60335-1 and IEC 60335-2-24
13.	Certificate	Product must be tested and certified from WHO-PQS. The WHO-PQS certificate is to be provided along with other documents
14.	Freezer Protection Grade	'A'
15.	Minimum Warranty	2 years onsite warranty and followed by 3 years of onsite service commitment. Availability of spares parts for at least 10 years.
16.	Lock	The door or lid must be fitted with lock with minimum two keys
17.	Corrosion resistance	Internal and external cabinet, lid and frame protected against corrosion.
18.	Physical Characteristics	Must be supplied with an appropriate device of equivalent performance to those voltage stabilizer

TERMS AND CONDITIONS:-

1. **F.O.R.:-** The above rates are F.O.R. destination basis to be supplied at Government Veterinary Hospitals/Dispensaries at various places in the State of Haryana at supplier's risk.
2. **GST:** - Inclusive in above rates.
3. **FREIGHT CHARGES:** Inclusive in above rates
4. **DELIVERY PERIOD:** The delivery should be within at least 45 days from the issuance of purchase/supply order.
5. **GUARANTEE/WARRANTY:** The Guarantee/Warranty shall be applicable as 2 years onsite warranty and followed by 3 years of onsite service commitment. Availability of spares parts for at least 10 years specified with the individual item. The warranty period will commence from the date of successful installation & running. If equipment is not repaired within 7 days after receipt of first call or complaint during guarantee / warranty period, the firm will have to pay a downtime penalty @ 0.2% of the net cost of equipment per day or a sum of Rs. 500/- per day, whichever is higher. If equipment is not repaired within 14 days without any cogent reasons, the department shall be at liberty to get the instrument/equipment repaired from local market at the cost of the firm.