

MAHARASHTRA MEDICAL GOODS PROCUREMENT AUTHORITY

Expression of Interest (EOI) for Market Assessment of Original Manufacturers, Authorized Importers for Supply of Consumables (Year 2026-27)

EOI reference No: E-294/MMGPA/ EOI for Consumables (2026-27)

1st Floor, Aarogya Bhawan St. George's Hospital Compound,
Near C.S.M.T. Railway Station, Mumbai - 400 001.
Maharashtra

Website: <http://mahatenders.gov.in>

Email: maha.mmgsa2023@gmail.com Phone : 022-22717527

Table of Contents

Table of Contents	2
Disclaimer.....	4
Abbreviations and Glossary.....	6
Notice and Schedule Inviting Expression of Interest (EOI)	8
EOI Notice.....	8
Section 1: Introduction	9
Section 2: General Definitions.....	10
Section 3: General Terms and Conditions	11
3.1 Responsibility of verification of contents of EOI.....	11
3.2 Acknowledgement of the Responder.....	11
3.3 Period of Validity of EOI	11
3.4 Nature of the Expression of Interest	11
3.5 Responsibility of EOI Respondents.....	12
3.6 Submission of EOI Responses	12
3.7 Clarifications and Amendments	12
3.8 Conflict of Interest.....	13
3.9 Fraudulent and Corrupt Practices	13
3.10 Confidentiality.....	13
3.11 Governing Law.....	13
3.12 Respondents Information.....	13
3.13 Manufacturing/Importing and Product Permissions	15
3.14 Delivery Capability	15
3.15 Respondent Classification.....	16
3.16 Eligibility Criteria.....	16
Appendix A: Schedule of Requirements	24
Appendix B: Checklist	100
Annexure 1: Technical Specifications and Compliance	102
Annexure-2: Undertaking by EoI Respondent.....	193
Annexure-3: Proforma for Production and Sale Statement	195
Annexure 4: Details of Manufacturing Unit	196
Annexure-4 (A): Proof of Distribution Network	199
Annexure-5: Details of Items Quoted	200

Annexure 6: Annual Turnover Statement for Three Years.....	201
Annexure-7: Format for Power of Attorney for signing of EOI	203
Annexure-8: Undertaking for Non-blacklisting	205
Annexure- 9: Litigation Affidavit	206
Annexure-10: Mandate Form.....	207
Annexure 11: Delivery and Supply Capability	208
Annexure 12: Respondent Classification Details	209

Disclaimer

This Expression of Interest (EOI) Document is not an agreement and is neither an offer nor invitation by the MMGPA to the prospective EOI Responders or any other person. The purpose of this Expression of Interest (EOI) Document is to invite interested and eligible entities to provide information relating to their regulatory status, manufacturing capability, quality systems, market presence, and related parameters, solely to enable MMGPA to undertake a market assessment and planning exercise in connection with any potential future procurement of consumables.

This Expression of Interest (EOI) Document contains information, statements, and indicative details prepared by MMGPA solely for the purpose of facilitating a broad market assessment and planning exercise in relation to the potential future procurement of consumables. Such information and statements are based on assumptions, preliminary assessments, and internal considerations and are not intended to be exhaustive or definitive. This EOI Document may not address the specific circumstances, operational capacities, or regulatory positions of every interested entity, and MMGPA, its employees, or advisors do not undertake to consider the individual requirements of each person or entity reviewing or responding to this document. The information provided herein may be subject to change and may not be complete, accurate, or current in all respects. Each EOI Respondent is therefore advised to conduct its own assessment, verification, and analysis of the information contained in this EOI Document and to seek independent professional or technical advice as it considers appropriate.

Information provided in this Expression of Interest (EOI) Document to the EOI Responder(s) is on a wide range of matters, some of which may depend upon interpretation of law. The information given is not intended to be an exhaustive account of statutory requirements and should not be regarded as a complete or authoritative statement of law. The MMGPA accepts no responsibility for the accuracy or otherwise for any interpretation or opinion on law expressed herein.

The MMGPA, its employees and advisors make no representation or warranty and shall have no liability to any person, including any EOI Responder(s) under any law, statute, rules or regulations or tort, principles of restitution or unjust enrichment or otherwise for any loss, damages, cost or expense which may arise from or be incurred or suffered on account of anything contained in this Expression of Interest (EOI) Document or otherwise, including the accuracy, adequacy, correctness, completeness or reliability of the Expression of Interest (EOI) Document and any assessment, assumption, statement or information contained therein or deemed to form part of this Expression of Interest (EOI) Document or arising in any way for participation in this Expression of Interest (EOI) Document.

The MMGPA may, in its absolute discretion but without being under any obligation to do so, update, amend or supplement the information, assessment or assumptions contained in this Expression of Interest (EOI) Document.

The EOI Responder(s) shall bear all its costs associated with or relating to the preparation and submission of its EOI Response including but not limited to preparation, copying, postage, delivery fees, expenses associated with any demonstrations or presentations which may be required by the MMGPA, or any other costs incurred in connection with or relating to its EOI Response. All such costs and expenses will remain with the EOI Responder(s) and the MMGPA shall not be liable in any manner whatsoever for the same or for any other costs or other expenses incurred by a EOI Responder(s) in preparation or submission of the EOI Response, regardless of the conduct or outcome of the responding process. The MMGPA also accepts no liability of any nature whether resulting from negligence or otherwise howsoever caused arising from reliance of

any EOI Responder(s) upon the statements contained in this Expression of Interest (EOI) Document. MMGPA shall not be responsible or liable for any costs, expenses, or losses incurred by EOI Respondents in connection with preparation or submission of EOI Responses, regardless of the outcome of the EOI process. Participation in this Expression of Interest (EOI) shall not confer any right or entitlement on the EOI Respondents to participate in any subsequent procurement process.

MMGPA reserves the right to independently determine the eligibility criteria, qualification requirements, and procurement strategy in any future tender process, without any reference to or obligation arising from this EOI.

Abbreviations and Glossary

Abbreviations	Description
BET	Bacterial Endotoxin Test
BIS	Bureau of Indian Standards
BP	British Pharmacopeia
CA	Chartered Accountant
CLA	Central Level Authority
COPP	Certificate of Pharmaceutical Products
CT	Clinical Trials
DCGI	Drug Controller General of India
DMER	Directorate of Medical Education & Research
DPCO	Drug Price Control Order
DPIIT	Department for Promotion of Industry and Internal Trade
ESIC	Employees State Insurance Corporation
FDA	Food and Drug Administration
FEMA	Foreign Exchange Management Act 1999
GMP	Good Manufacturing Practices
G.R.	Government Resolution
GST	Goods and Services Tax
GS1	Global Standards1
HBPCL	Haffkine Bio-Pharmaceutical Corporation Limited
IEC	Import Export Code
IP	Indian Pharmacopeia
ISI	Indian Standards Institute
I.V.	Intra Venous
LLP	Limited Liability Partnership
MCGM	Municipal Corporation of Greater Mumbai
MDR	Medical Device Rules
MMGPA	Maharashtra Medical Goods Procurement Authority
MRP	Maximum Retail Price
MSC	Market Standing Certificate
MSEs	Micro and Small Enterprises
NABL	National Accreditation Board for Laboratories

NSQ	Not of Standard Quality
NPPA	National Pharmaceutical Pricing Authority
OEM	Original Equipment Manufacturer
PAN	Permanent Account Number
PoA	Power of Attorney
PVC	Poly Vinyl Chloride
PVdC	Poly Vinylidene Chloride
SLA	State Level Authority
SSI	Small Scale Industries
UDIN	Unique Document Identification Number
USP	United States Pharmacopeia
WHO	World Health Organization

Notice and Schedule Inviting Expression of Interest (EOI)

EOI Notice

The Maharashtra Medical Goods Procurement Authority (MMGPA) invites Expressions of Interest (EOI) from eligible Original Manufacturers, Authorised Importers for assessing market capability, regulatory compliance, manufacturing capacity, and delivery capability for supply of consumables to public health institutions in the State of Maharashtra.

This EOI is a non-binding, pre-procurement exercise and does not constitute a tender or offer of contract. No price submissions, financial offers, or securities are required at this stage.

The objective of this EOI is to assess the number of eligible and capable OEMs and Authorized Importers available for the listed consumables and their readiness to participate in future procurement processes.

The information collected through this EOI will be used for planning future procurement strategies. MMGPA reserves the right to initiate or not initiate any subsequent tender or rate contract process without any obligation to the EOI respondents.

Only Original Manufacturers or their Authorized Importers are eligible to participate. Traders, distributors, and other intermediaries are not permitted.

Interested entities meeting the eligibility conditions are invited to submit their EOI responses within the timelines specified in this document.

EOI responses must be submitted within 15 working days from the date of publication of this EOI. Late submissions shall not be accepted.

All response-related activities, including downloading of documents and submission of responses, shall be governed by the schedule indicated below.

EOI Response Schedule

No.	Activity	Date & Time
1	Publication of Notice Inviting Expression of Interest (EOI)	From Dt. 01.06.2026, 12:00 PM
2	Last date and time for submission of EOI Responses	Dt. 15.06.2026, at 3:00 PM
3	Opening of EOI Responses (Technical & capability only, No price information shall be submitted)	Dt. 16.06.2026 at 3:00 PM
4	EoI Clarification Meeting	Dt.. 09.06.2026 , 03:00 PM
5	Address for communication	1st Floor, Aarogya Bhawan, St. Georges Hospital Compound, Near CSMT Railway Station, Mumbai- 400 001

Sd/-

CHIEF EXECUTIVE OFFICER,
MAHARASHTRA MEDICAL GOODS PROCUREMENT AUTHORITY
MUMBAI

Section 1: Introduction

Maharashtra Medical Goods Procurement Authority (MMGPA), Mumbai has been formed as per the Maharashtra Medical Goods Procurement Authority Act 2023 (Mah. Act No. XIII of 2023) (“Act”).

MMGPA works as a single point procurement authority of certain medical goods and execution of turnkey projects with highest standards of transparency, fairness, equity and for timely supply at optimum and uniform rates and of desired quality and quantity for the Public Health Department, Medical Education & Drug Department of Government of Maharashtra and other departments as mentioned in the said Act.

To meet the objective of the Act, MMGPA in exercise of its powers under section 36 of the Act issued the regulations named, Maharashtra Medical Goods Procurement Authority Regulations, 2025 as amended from time to time.

- 1.1 All related activities (process) like EOI downloading and response submission will be governed by the EOI Schedule.
- 1.2 All activities of this EOI are carried out online on website <https://mahatenders.gov.in>. The EOI is uploaded on Government of Maharashtra, (GoM) e-tendering website <https://mahatenders.gov.in> and has to be downloaded as well as filled up and submitted online only.
- 1.3 No commitment is made by MMGPA regarding procurement of any quantity pursuant to this EOI, and the Authority reserves the right to modify, expand, curtail, or altogether omit any items or quantities in any subsequent procurement process, if undertaken
- 1.4 EOI process related Queries can be sent on email – eproc.support@maharashtra.gov.in /Help: The 24 x 7 Toll Free Telephonic Help Desk Number 1800-3070-2232. / Mobile: +91-7878107985, +91-7878107986, +91-7878007972 and +91-7878007973.

Sd/-

CHIEF EXECUTIVE OFFICER,
MAHARASHTRA MEDICAL GOODS
PROCUREMENT AUTHORITY
MUMBAI

Section 2: General Definitions

For the purposes of this Expression of Interest (EOI) Document, and unless the context otherwise requires, the following terms shall have the meanings assigned to them below:

- 2.1 **“Applicable Laws”** Means all laws, statutes, rules, regulations, notifications, guidelines, and policies in force in India as issued by the Government of India or the Government of Maharashtra, including but not limited to the Drugs and Cosmetics Act, 1940 and Rules thereunder, Medical Device Rules, 2017, New Drugs and Clinical Trials Rules, 2019, NPPA/DPCO provisions, and other applicable regulatory requirements, as amended from time to time.
- 2.2 **“Authority” or “MMGPA”** Means the Maharashtra Medical Goods Procurement Authority, established under the Maharashtra Medical Goods Procurement Authority Act, 2023.
- 2.3 **“Expression of Interest” or “EOI”** Means this non-binding, pre-procurement process initiated by MMGPA to solicit information regarding market capability, regulatory compliance, manufacturing capacity, and quality systems from eligible entities.
- 2.4 **“EOI Document”** Means this document, including all sections, appendices, annexures (if any), and amendments issued by MMGPA in connection with the Expression of Interest.
- 2.5 **“EOI Response”** Means the information and declarations submitted by an EOI Respondent in response to this EOI Document, limited to technical, regulatory, manufacturing, and capability-related information.
- 2.6 **“EOI Respondent”** Means any manufacturer or importer who submits an EOI Response pursuant to this EOI Document.
- 2.7 **“Conflict of Interest”** Means any situation in which an EOI Respondent has interests that could improperly influence its participation in the EOI process, including but not limited to relationships that compromise fairness, transparency, or objectivity.
- 2.8 **“Fraudulent or Corrupt Practices”** Means misrepresentation, concealment of facts, submission of false or forged documents, or any attempt to influence the EOI process improperly.
- 2.9 **“Government”** Means the Government of India or the Government of Maharashtra, as applicable.

Interpretation Words and expressions used but not defined in this EOI Document shall have the meanings assigned to them under Applicable Laws.

Note: The words and expressions used in this EOI, but not defined, shall have the same meaning as respectively assigned to them under the prevailing Applicable Laws.

Section 3: General Terms and Conditions

This section deals with the general conditions of the EOI and contains the following terms & conditions governing the EOI.

3.1 Responsibility of verification of contents of EOI

- 3.1.1 It shall be the responsibility of the EOI Responders to read/examine all instructions, forms, terms, conditions, specifications, schedules, appendix, annexures and other provisions contained in the Expression of Interest (EOI) Document, including any addendum or corrigendum issued thereunder prior to submission of the response.
- 3.1.2 Responders shall ensure that all required documents as specified in **Appendix A (Checklist)** are duly uploaded.

3.2 Acknowledgement of the Responder

- 3.2.1 It shall be deemed that by submitting the EOI, the Responder has:
- i. examined the EOI Document and understood its contents.
 - ii. received all relevant information available from MMGPA
 - iii. acknowledged and agreed that inadequacy, lack of completeness or incorrectness of information provided in the Expression of Interest (EOI) Documents shall not be a basis for any claim for compensation, damages or extension of time.
 - iv. acknowledged that it does not have a Conflict of Interest as per Clause 3.9 and
 - v. agreed to be bound by the provisions and undertakings provided by it under and in terms hereof.

MMGPA shall not be liable for any omission, mistake or error in respect of any of the above or on account of any matter or thing arising out of or concerning or relating to the EOI or its Responding process, including any error or mistake therein or in any information or data given by MMGPA.

3.3 Period of Validity of EOI

- 3.3.1 The EOI Response submitted pursuant to this document shall remain valid for the purpose of this EOI process only until completion of the EOI exercise or initiation of any subsequent procurement process, whichever is earlier.

3.4 Nature of the Expression of Interest

- 3.4.1 The information received pursuant to this EOI shall be used by MMGPA only for:
- a) market assessment and planning,
 - b) identification of technically and regulatorily suitable entities, and
 - c) formulation of future procurement strategies, if any.
- 3.4.2 This Expression of Interest (EOI) is a non-binding, pre-procurement exercise issued solely for the purpose of assessing market capability, regulatory compliance, manufacturing capacity, and quality systems of eligible entities.

- 3.4.3 Submission of an EOI Response does not constitute a bid, tender, or offer, and does not create any contractual, financial, or legal obligation on either MMGPA or the EOI Respondent.
- 3.4.4 MMGPA reserves the right to initiate or not initiate any subsequent procurement process, including a tender or RFP, without assigning any reason.
- 3.4.5 The EOI is intended to assess the number of eligible and capable Original Manufacturers. Authorized Importers available for each consumables/item(s) and their readiness to participate in future procurement processes.

3.5 Responsibility of EOI Respondents

- 3.5.1 Each EOI Respondent shall:
- a) carefully examine the EOI Document and all instructions therein;
 - b) ensure that all information and documents submitted are true, accurate, and complete to the best of its knowledge;
 - c) satisfy itself regarding the scope, intent, and implications of the EOI.
 - d) Any misrepresentation, concealment, or submission of false information may result in rejection of the EOI Response.
- 3.5.2 Respondents shall provide accurate information regarding:
- (a) availability of consumables/item(s) against the provided list,
 - (b) manufacturing/import capacity,
 - (c) indicative delivery timelines.

3.6 Submission of EOI Responses

- 3.6.1 EOI Responses must be submitted within 7 working days from the date of publication of the EOI. Late submissions shall not be accepted.
- 3.6.2 The EOI Response shall be limited to technical, regulatory, manufacturing, and capability-related information only.
- 3.6.3 No price bids, financial offers, bid securities, earnest money deposits, or payments of any kind are required or permitted at the EOI stage.

3.7 Clarifications and Amendments

- 3.7.1 EOI Respondents may seek clarification, within the timelines specified in the EOI Schedule.
- 3.7.2 MMGPA may, at its discretion:
- a) **issue clarifications,**
 - b) **amend the EOI Document, or**
 - c) **modify the EOI Schedule.**

3.7.3 All such clarifications or amendments shall form an integral part of the EOI Document.

3.8 Conflict of Interest

3.8.1 EOI Respondents shall disclose any actual or potential conflict of interest that may affect their participation in the EOI process.

3.8.2 MMGPA reserves the right to reject an EOI Response where a conflict of interest is found to compromise fairness, transparency, or objectivity.

3.9 Fraudulent and Corrupt Practices

3.9.1 EOI Respondents shall observe the highest ethical standards during the EOI process.

3.9.2 Any attempt to:

- a) misrepresent facts,
- b) submit forged or misleading documents, or
- c) improperly influence the EOI process

3.9.3 may result in rejection of the EOI Response and may render the entity ineligible for participation in future EOIs issued by MMGPA.

3.10 Confidentiality

3.10.1 Information submitted by EOI Respondents shall be treated as confidential to the extent permitted under applicable laws.

3.10.2 MMGPA may disclose such information where required by law, regulation, or directions of statutory or judicial authorities.

3.11 Governing Law

This EOI process shall be governed by and construed in accordance with the laws of India.

3.12 Respondents Information

Sr. No.	Document Name / Requirement	Description
1	Manufacturing License	Tenderer shall be a manufacturer duly licensed in Form 25/Form 28/ Form 28D/Form MD5/ Form MD9 (as per products quoted) with current validity/ retention issued by the State Licensing Authority (SLA)/Central Licensing Authority as the case may be or a direct importer holding valid import license in Form 10 with Form 41/Form MD 15 issued by the Drugs Controller General of India (DCGI) accompanied with Licenses in Form 20B/21B with validity retention.

		Attested photocopy of valid manufacturing item(s) license with product list approved by Central/State Licensing Authority for each and every item quoted as per technical specification in the bid. The license must have been duly renewed up to date and the items quoted shall be clearly highlighted in the license . If quoted items are manufactured at different places, Manufacturing License & Performance certificate from all such places from respective Licensing Authority/ State Drug Authority should be enclosed. However, Loan Licensee/ third party licensee are not allowed.
2	Certificate	ISO 13485 (QMS) Quality Management System certificate issued by the competent authority. CE Certificate issued by the competent authority. Certificate of DPIIT (if applicable) Import Export Certificate (IEC Code) (if applicable).
4	Past Work Experience	The Bidder shall provide documentary evidence in support of its credentials such as agreement copy/ work order / Letter of Award, work completion certificate, customer satisfaction certificates with customer details and client certificate or statutory auditor's certificate or Chartered Accountant's certificate with his UDIN, as the case may be, for demonstrating the Technical Capacity. Such documentary evidence shall be duly signed by the authorized signatory of the Bidder. Past work experience valid for a period of 3 years (i.e., for financial year 2023-24,2024-25 and 2025-26) as a manufacturer/importer for each item(s) quoted in the tender within the Bid Due date. (As per Annexure 3)
5	GST Registration	GST Registration certificate issued by the GSTN authorities along with copy of recent paid Challan.
6	Incorporation/Registration Certificate	Certificate of Incorporation / Registration of the EOI Respondent .
7	Authorization Letter	Authorization nominating responsible person to attend EOI/tender related meetings.
8	Appendix-A	Schedule of Requirements. (List of Items)
9	Appendix-B	Checklist.
10	Annexure-1	Technical Specifications and Compliance.
11	Annexure-2	Letter Comprising Technical Bid.
12	Annexure-3	Proforma for Production and Sale Statement.
13	Annexure-4	Details of Manufacturing Unit.

13	Annexure-5	Details of Items Quoted with Drug Code.
15	Annexure-6	Annual Turnover Statement for three years (2022-23, 2023-24, and 2024-25) along with audited Balance Sheet & P&L Accounts for last 3 years certified by Statutory Auditor/Chartered Accountant with UDIN.
16	Annexure-7	Format of Power of Attorney for signing of Bid (not required for proprietorship).
17	Annexure-8	Affidavit for Non-Blacklisting.
18	Annexure-9	Litigation Affidavit.
19	Annexure-10	Mandate Form.
20	Schedule-4	Schedule for Packing of Items
21	Schedule-5	Bar Code & Advance Shipment Notification details.

3.13 Manufacturing/Importing and Product Permissions

- 3.13.1 EOI Respondents should be a manufacturer duly licensed in Form 25/Form 28/ Form 28D/Form 28E/Form MD5 and Form MD9 (as per drugs quoted) with current validity/ retention issued by the State Licensing Authority (SLA)/Central Licensing Authority (CLA) as the case may be or a direct importer holding valid import license in Form 10 or Form 10-A as applicable, with Form 41/Form MD 15 issued by the Drugs Controller General of India (DCGI) accompanied with Licenses in Form 20B/21B with validity retention.
- 3.13.2 EOI Respondent should have obtained permission to manufacture the product(s) quoted strictly as per specification indicated in the Tender Document and in accordance with the standards specified in the Drugs and Cosmetics Act, 1940 from the competent authority such as state and/or central licensing authority.
- 3.13.3 In case of new drugs, import permission in Form CT-20 and manufacturing permission in Form CT-23 should be furnished in accordance with Rule 76/81 of the New Drugs and Clinical Trials Rules 2019 and Drugs and Cosmetics Act, 1940.
- 3.13.4 To ensure timely and smooth distribution of consumables, respondent, if not having their registered office, manufacturing unit, or distribution facility within Maharashtra State, must have an established and operational distribution network in Maharashtra State at the time of bid submission. Proof of such distribution networks such as a valid drug selling/distribution license issued by the state's licensing authority, along with details (names, addresses, contact details)-must be submitted as part of the bid documentation

3.14 Delivery Capability

- 3.14.1 EOI Respondents shall provide indicative details on:
- (a) production lead time,
 - (b) dispatch timeline,
 - (c) emergency supply capability,

- (d) monthly production capacity,
- (e) warehousing/logistics arrangements in Maharashtra.

3.15 Respondent Classification

3.15.1 EOI Respondents shall provide information on:

- (a) MSME status,
- (b) SC/ST ownership,
- (c) location – within Maharashtra / outside Maharashtra.

3.16 Eligibility Criteria

Sr.No.	Basic Requirement	Specific Requirement	Documents required
1	Registered Legal Entity	The EOI Respondent may be a company or a firm, registered under applicable laws in India (“Respondent”). The Respondent shall be – a. A Private Limited Company or a section 8 Company registered under the Companies Act 1956/2013, b. A Limited Liability Partnership under the Limited Liability Partnership 2008 c. Registered with the GST Authorities. d. Should have a valid PAN/TAN number. Proprietorship firms are not allowed to participate in this EOI. The Respondent shall be original manufacturer of the medical goods as mentioned in EOI or should have importer license from original manufacturer in case importer goods.	Copy of certificate of incorporation/registration along with charter documents like copy of Memorandum and Articles of Association, and other registration documents according to the nature of entity. Copy of GST Registration certificate issued by GSTN authorities. Copy of PAN/TAN Card.

2	Licenses	The Respondent must have valid manufacturing license/import license for the drugs quoted as per technical specifications in the Tender Documents: - Only manufacturer/importer (i.e., authorized agent/subsidiary of the foreign manufacturer) will be allowed as Bidder. - Loan Licensee / third party licensee are not permitted to participate under this Tender Document.	Self-attested copies of original manufacturing license in Form 25, 28, 28-D, 28-E, MD-5, MD-9, MD-15 and Form 10 & 41. etc. approved by the Licensing authority along with retention Licenses Or Self-attested copies of original import license in Form-10 or Form 10-A (as applicable), Form-41, MD-15 etc., and licenses in Form 20B/21B approved by the corresponding licensing authority with retention licenses. Or Self-attested copies of licenses in Form CT-20, Form CT-23 in case of new drugs. Further, during supply of drugs, if the drugs are manufactured at more than one premises, the Bidder shall submit the applicable license such as the manufacturing license, for each such premises mandatorily. Authority letter of the original manufacturer for importing the product for which bid is offered or agreement between foreign manufacturer and importer. The drugs quoted shall be highlighted in the product
---	-----------------	---	--

			permission with their respective drug code as mentioned in the Appendix-A (Schedule of Requirement)
3	Certifications/Registration	WHO-GMP (WHO - Good Manufacturing Practices Certificate) Certificate issued by the licensing authority. The WHO-GMP certificate must not be older than one year from the Bid Due Date in case where validity is not mentioned in the certificate. The WHO-GMP certificate of all the manufacturing plants, of which products have been quoted, should be submitted. The Importer should produce WHO-GMP / COPP of the manufacturing firm or a certificate which is at par with WHO-GMP issued by exporting countries like US-FDA approval, etc. In the case of imported drugs, labels and product literature of all quoted products must be submitted. The manufacturing premises of the Bidder's products must comply with the requirements laid down under Schedule M of the Drugs and Cosmetics Rules, 1945. The bidder shall submit a valid compliance certificate/license issued by the State Licensing Authority unequivocally indicating Revised and upgraded Schedule M compliance as per GSR 922 (E), dated 28.12.2023. 2. ISO 13485 (QMS) Quality management System issued by Notified Body.	For manufacturers, WHO-GMP certificate with product list OR COPP, Quality management System (QMS) as per Medical Devices Rules, 2017 issued by the licensing authority wherever applicable. For importers, labels and product literature of all quoted drug(s) must be uploaded with WHO-GMP or certificate which is at par with WHO-GMP issued by the authorities of exporting countries of their principal manufacturing company or firm. Authority letter from OEM for the offered drug and valid import license issued by licensing authority. Import Export Certificate (IEC Code) for importer. Certificate as applicable. Certificate as applicable.

		3. CE certificate issued by Notified Body/USFDA.	
4	Past Experience	The Respondent must submit particulars of quantity of the past supplies of item(s) made as per Annexure 3. Note: At least 80% quantity for similar drug as specified in the Schedule of Requirements i.e. Appendix A' must have been supplied for any of the last 3 (three) financial years preceding the Bid Due Date, i.e., 2023-24, 2024- 25 & 2025-26. The Respondent shall submit the work done with satisfaction from the concerned authority. For the Purpose of quantification of 80% similar work only the Purchase order from Govt./Semi-Govt.& Hospitals registered under Maharashtra Public Trust Act 1950 shall be considered. Purchase Order from private hospitals shall not be considered.	Certificate issued by a statutory auditor/chartered accountant (as attached Annexure-3)
5	Average Annual Turnover	Average Annual Turnover (in last three financial years 2023-24, 2024- 25 & 2025-26. shall be as per mentioned in the Schedule of Requirements i.e. Appendix A.	Certificate issued by a statutory auditor/chartered accountant (as attached Annexure-6) along with audited financial statements confirming the average annual Turnover of the Bidder during the stated financial years must be submitted. The turnover should be as certified by the chartered accountant/statutory auditor (specifying

			UDIN) and having valid registration. Note: In case the annual accounts for the latest financial year are not audited and therefore the Bidder cannot make it available, the Bidder shall give an undertaking to this effect and the chartered accountant/statutory auditor shall certify the same. In such a case, the Bidder shall provide the audited financial statements for the financial year preceding the latest financial year for which the audited annual report is not being provided.
6	Net Worth	The net worth of the Bidder in the financial year immediately preceding the Bid Due date should be positive and should have not eroded by more than 30% (thirty percent) in the last three (3) years.	Certificate issued by a statutory auditor/chartered accountant (as attached Annexure-6). The Net Worth should be as certified by the chartered accountant/ statutory auditor (specifying UDIN) and having valid registration.
7	Production Capacity	Production capacity of the original drug manufacturer must be minimum Two (2) times of the quoted order quantity in last one financial year.	Certificate of Statutory Auditor/Chartered Accountant. As per Annexure 4.
8	Market Standing Certificate	Bidder should mandatorily possess 5 years Market Standing Certificate (i.e., for financial year (2021-22, 2022-23, 2023-24, 2024-25 & 2025-26) as a	Market Standing Certificate as issued by Central or State Licensing Authority under the Applicable Law.

		manufacturer/importer for each drug(s) quoted in the tender as on Bid Due Date. The period of Market Standing will be reckoned from the date of issue of product/drug permission. In case of an importer, their principal manufacturer located overseas should have 5 years market standing in India and the importer shall have 5 years market standing in the pharmaceutical field. In cases involving new item/items out of patent period it is sufficient to possess relevant market standing as applicable.	Bill of landing of a foreign manufacturer and bill of entry of the importer, mentioning the country of origin. Notarized/ certified copy of Drug Controller General of India, new Delhi, for permission for Items coming under, “New Drug and Fixed Dose Combinations” in form 45/46 as per Drugs & Cosmetic Act and Rules. Relevant period Market standing certificate for new item/drug/drugs out of patent period issued by licensing authority.
9	Proof of distribution network	Respondents, if not having their registered office, manufacturing unit, or distribution facility within Maharashtra State, must have an established and operational distribution network in Maharashtra State at the time of EOI submission. In case there is no registered office or established operational distribution network the respondent/bidder shall give undertaking for establishment of both within three (3) months from the last date of submission of EOI.	Proof of distribution network— such as a valid drug/product/item(s) selling/distribution license issued by the state’s licensing authority, along with details of authorized distributors (names, addresses, contact details) — must be submitted as part of the bid documentation Annexure 4 (A) and Copy of License of Sale / Distribution issued by State’s Licensing Authority

10	Blacklisting or banned	On the Bid Due Date, the Respondent should not be blacklisted or debarred by any ministry/department /attached offices/sub-ordinate offices under Government of India and any State government, Autonomous bodies (established by Central/State govt), any Central/State PSUs for unsatisfactory past performance, corrupt, fraudulent or any other Fraudulent and Corrupt business practices. The respondent, whose drug has been declared as misbranded, spurious or adulterated or any criminal case in respect of the above is pending in any court, as on Bid Due Date shall not be eligible to participate for that particular drug, in the bid. Similarly, a Bidder convicted by court of law shall not be eligible to participate in the bid.	Affidavit as per Annexure 8
11	Non-Conviction Certificate	The Respondent(s) quoting for this tender shall not have been convicted by any Court of Law in India/overseas in lieu of deficiency noticed in the any of the quoted Product(s) in the tender and tender should not be submitted for such products for which conviction was granted by any Court of Law for a period of Three years from date of such order(s).	Non-Conviction certificate Issued by licensing authority/State FDA.

12	Litigation	The Bidder should not be involved in any major litigation such as fraud, FEMA violations and any other civil or criminal proceedings that may have an impact of affecting or compromising the delivery of services as required under this contract.	Affidavit as per Annexure 9.
13	Conflict of Interest	On the date of submission of the Bid, the Bidder should not have any conflict-of-interest situation.	Undertaking by the authorized signatory as per Annexure 2
14	Quantity specified in EOI	The quantity mentioned in the EOI is indicative in nature. Detail requirement will be given in subsequent RFP.	-

Appendix A: Schedule of Requirements

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
1	Absorbent points 2 degree	ISO standardized paper points from number 15 onwards sizes (# 15-40) (#45-80) # 15, # 20, # 25 Should be supplied in assorted packets as per required numbers and quantity Should be colour coded should have 2 degree taper Should be length marked 120 points in 1 packet	consumables	85	28900
2	Absorbent points 4 degree	ISO standardized paper points from number 15 onwards sizes (# 15-40) # 15, # 20, # 25, #30, #35, #40 Should be supplied in assorted packets as per required numbers and quantity Should be colour coded should have 4 degree taper Should be length marked 60 points in 1 packet	consumables	103	28840
3	Absorbent points 6 degree	ISO standardized paper points from number 15 onwards sizes (# 15-40) #15, #20, #25, #30, #35, #40 Should be supplied in assorted packets as per required numbers and quantity Should be colour coded should have 6 degree taper Should be length marked 60 points in 1 packet	consumables	120	33600
4	Acrylic resin teeth (anatomic)	1) Shape (Mould) in assortment of shapes i.e. square avoid & tapering. 2) Shade ? Approx. of 6 shades from light medium to dark 3) Size ? Approx. 8 sizes ranging from small to large in each of the size & Shade 4) General ? The teeth must bear resemblances to natural teeth & be free of porosity. 5) The Teeth should have sufficient translucency to give shining after processing. After processing for denture along with denture base resin should not show structural changes. One complete set from each shade & shape & size for inspection should be supplied as sample with mould chart. 6) Multilayer shading minium 2 shading plus base material & cross linked with fluroscnt effect & one set of 28 teeth	consumables	900	216000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
5	ADDITION SILICONE	1.Should be supplied in the form of putty and light body as per ADA/ISO/ISI specification. 2.The light body should be available in the Automix cartridge and as ah and-mixed variant. 3. It should have high tear strength. 4. It should be dimensionally stable. 5. Expiry date should be atleast for 2 yrs from date of supply. Putty should have Mixing time-30-50 sec Working time-1:30min to 2min Setting time - 3-5min Light body should have Mixing time-30sec to 1 min Working time-2min to 2:30 min Setting time - 3-5 min kit contains: Minimum quantity putty- 900 ml per kit and light body 100 ml per kit along with minimum 10 automixing tips	consumables	120	576000
6	Adhesive Plate Sealer for 96 well Plates	Adhesive Plate Sealer for 96 well Plates	consumables	120	6168
7	Air Bed	Approx Size: 34*18*23cm N.W.: 3.0kgs G.W.: 3.5kgs Using Life: 30000hours Mattress Specification: 1. Material: PVC 3PLY 2. Thickness: 0.3-0.4mm 3. Size: 200CM*90CM*7CM(inflated) Pump Specification: 1. Size: 24*13*9cm 2. Weight: 1.0kgs 3. Voltage: AC 220/230V 4. Frequency: 50HZ/60HZ 5. Power: 10W 6. Pressure range: 15-20kpa (115-150mmHg) 7. Air output:4.5-5.5L/minute 8. Cycle time: 10 minutes 9. Controlled by microcomputer 10. Spare pcs: 10pcs of round material slices, 1bag of special glue	consumables	120	211200
8	Air rotor Cartridges	1) Should have the same company as Air Rotor supplied.	consumables	120	138240

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
9	Alginate	1) Supplied in the form powder with components as per the ADA/ISO/ISI specifications. 2) Approximate Composition ? a. Powder ? potassium alginate ? 15%, CaSo4 ? 16%, ZnO? 4%, Kltf ? 3%, Napo4 ? 2 %and Diatomaceous earth ? 60% 3) Material should not irritate the oral tissues, shall have pleasant color & flavour. Powder should have fine texture & should be free from foreign material. 4) On mixing ? When mixed in accordance with manufacturer instructions material should give a smooth plastic mass of workable consistency & should be free from lumps & granules. There shall be no separation of constituents any the material shall be sufficiently plastic for under cuts & anatomical; details to be accurately reproduced but shall not run out of such impression material. 5) Mixing Time ? 45 sec. ? 1 min 6) Setting time ? Not less than 2 minutes & 30 second 7) General strength not less than 50 lbs/sq. Inch. Should not rupture during removal from mouth. a) Packing ? Material should be available in tightly sealed, moisture proof Pkt. Information to be supplied by the manufacturer ? Date of Manufacture & date of expiry. Adequate instruction for proportioning & manipulation of the material suitable measures for powder & water for purpose of proportioning shall be supplied by the impression during the interval between its withdrawal & cast pouring. Weight of contents and setting times of material should be stated on the packing. b) There should be a colour indicator where in the colour changes when the alginate sets. 8) it should be 1 box should contain 12 alginate packets total demand is according to box. 9) one packet should contain 450 gms to 500 gms powder	consumables	221	689520

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
10	Ambu bag with face mask (Adult)	It should be self-inflatable and must have pop up valve (non-return valve), attachment for oxygen tube & oxygen reservoir. Bag must be made up of medical grade silicon, latex free double layered which retain sensitivity and it must be resistant to rough use. Inlet end of the bag must have separate port for Oxygen supplement. Outer port must be such that re-breathing valve or non-return valve can be attached. Must be supplied with Oxygen reservoir bag, Should have a bag volume of 1500ml. It should be autoclaveable. It should be adaptable to all type of face masks. Must be USFDA or European CE approved product	consumables	1656	947232
11	Ambu bag with face mask (Neonatal)	It should be self-inflatable and must have pop up valve (non-return valve), attachment for oxygen tube & oxygen reservoir. Bag must be made up of medical grade silicon, latex free double layered which retain sensitivity and it must be resistant to rough use. Inlet end of the bag must have separate port for Oxygen supplement. Outer port must be such that re-breathing valve or non-return valve can be attached. Must be supplied with Oxygen reservoir bag. Should have a bag volume of 250ml. It should be autoclaveable. It should be adaptable to all type of face masks. Must be USFDA or European CE approved product	consumables	62	35464
12	Ambu bag with face mask (Paediatric)	It should be self-inflatable and must have pop up valve (non-return valve), attachment for oxygen tube & oxygen reservoir. Bag must be made up of medical grade silicon, latex free double layered which retain sensitivity and it must be resistant to rough use. Inlet end of the bag must have separate port for Oxygen supplement. Outer port must be such that re-breathing valve or non-return valve can be attached. Must be supplied with Oxygen reservoir bag. Should have a bag volume of 500ml. It should be autoclaveable. It should be adaptable to all type of face masks. Must be USFDA or European CE approved product	consumables	161	92092

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
13	Applicator Tips	1) Should be made up of plastic material to apply 2) bonding agent or echant for dental useIt should contain pack of 100 tips	consumables	120	48000
14	Arch bars	Itshould be a flat, sturdy stainless steel bar on which fleatsorhooks are attached. Itshould be a type of the prefabricated Erich archbars. Should contain 1 feet arch bar It should be autoclavable. It should be mouldable to adapt it to the dental arches. It should be resistant to corrosion. The longitudinal edges of the strip should be continuous and smooth.	consumables	120	19200
15	Assorted teeth set for RPD various shades (ANTERIOR AND POSTERIOR)	1) It should be in anatomic or semi anatomic or monoplanic form as per the requirements 2) It should have good anatomic structure 3) It should be available in different shades 4) It should bond chemically with denture base resin 5) It should be dimensionally stable during curing process 6) It should have enough strength, abrasion resistance and resilience 7) Should be insoluble and impermeable to oral fluids 8) Should have low specific gravity (light in weight) 9) It should withstand flasking , deflasking , curing process 10) It should have coefficient of thermal expansion same as denture base resin 11) 1 anterior set should contain upper 6 anterior & lower 6 anteriorteeth 12) 1 posterior set should contain upper 8 posterior and lower 8 posterior(4 of each side)	consumables	100	21420
16	Autoclave indicator strip (label)	Label size 6 cm x 2.5 cm Chemical band : 4 cm x 1 cm with green colour & 5 indicator on one strip. 1 box contains 100 labels (strip) should have ISO & CE certification	consumables	53277	222485
17	B.P. Bulb with knob	Valve/knob:-Made of metal alloy hard and strong.Rust resistant can ensure long life.Rotating knob to hold and release air. Bulb:-Made of rubber soft material.Portable design and easy to hold.	consumables	1110	81074

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
18	Bain circuit, Adult	It should be corrugated,flexible,lightweight and transparent for easy monitoring of water build up or blockages.Should be provided with 22 mm corrugated tube, expiratory valve and 2 litre antistatic bag.The inner tubing should be non kinkable and non crackable.	consumables	380	112662
19	Bain circuit, paediatric	It should be corrugated,flexible,lightweight and transparent for easy monitoring of water build up or blockages.Should be provided with 10 and 15 mm corrugated tube, expiratory valve and 0.5 litre antistatic bag.The inner tubing should be non kinkable and non crackable.	consumables	280	153574
20	Band material 180 x 0.005	It should be of good quality, stianless steel non corrosive, 8 feet in length	consumables	120	19200
21	Barbor Linen Thread no. 40 Reel	Reel of 200 mtrs with individual fillaments of high quality linen with high tensile strength, low tissue reactivity, superior knot security & excellent handling characteristics.	consumables	710	568000
22	Barbor Linen Thread No. 60 Reel	Reel of 200 mtrs with individual fillaments of high quality linen with high tensile strength, low tissue reactivity, superior knot security & excellent handling characteristics.	consumables	675	712800
23	Barbor Linen Thread no. 80 Reel	Reel of 200 mtrs with individual fillaments of high quality linen with high tensile strength, low tissue reactivity, superior knot security & excellent handling characteristics.	consumables	65	69888
24	Barbour linen thread no.20	Reel of 200 mtrs with individual fillaments of high quality linen with high tensile strength, low tissue reactivity, superior knot security & excellent handling characteristics.	consumables	356	279104
25	Barbour linen thread no.30	Reel of 200 mtrs with individual filaments of high quality linen with high tensile strength, low tissue reactivity, superior knot security & excellent handling characteristics.	consumables	55	70400
26	Bed Pan (Plastic)	Bed Pan (Plastic)	consumables	320	56320

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
27	Biodentine	1) set in 10-12 minutes 2) versatile use in endodontic repair and restorative repair. 3) natural micromechanical anchorage for excellent sealing properties. 4) similar mechanical and behavior properties as human dentine 5) 3.5 mm aluminium radiopacity for long term follow up 6) 5 X capsule and 5 X single dose containers	consumables	120	816000
28	Biopsy Punches disposable 2.5mm	Biopsy Punches disposable 2.5mm	consumables	120	7637
29	Bisacrylamide	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	consumables	120	71520
30	Black Goggles for Ophthalmic Patients	Should be scratch resistant, plastic goggles with UV protection.	consumables	2500	35640
31	Blood Pressure Apparatus aneroid portable/Stand Model	Blood Pressure Apparatus aneroid portable/Stand Model	consumables	290	626400
32	Boro-Spirit Ear Drops 5ml	Boro-Spirit Ear Drops; Each 5 ml contains 0.183gm Boric Acid in 2.08 ml of Alcohol	consumables	33312	799488
33	Calcium hydroxide base for pulp capping	1) Should follow ISO/ISI standard composition. 2) It should have supply in two pastes in equal amount 3) It should be 13 gms base and 11 gms catalyst with mixing pad	consumables	120	144000
34	Calcium hydroxide in (Aqueous or water base) for endodontic use	1) Should be supplied as single paste system 2) Hydrated Lime with aqueous solution for dental use.	consumables	120	48000
35	Calcium Hydroxide Powder 10gm.	Calcium Hydroxide Powder for Dental Cement for Cavity Base, 10 gm Bottle	consumables	396	11088

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
36	Calcium hydroxide-idoform paste	1) It should have supply in two pastes in equal amount 2) It should be 5 to 10 gms base and 8 to 12 gms catalyst 3) It should be premixed paste for immediate use with minimal material waste and clean up 4) It should be syringe tip delivery/for excellent accessand control during placement , 5) It should be formulated with idoform for increased radioopacity and anti bacterial effect , 6) It should be used with gutta percha and root canal cements should have inuf quantity for about 40 to 50 fillings per syringe. 8) 1 box of 2 tubes of 2 ml each	consumables	55	100320
37	Carbide bur	TUNGSTEN CARBIDE . This bur is ideal for a variety of procedures including initial sectioning of impacted molars, cutting of teeth at root level, and root separation and amputation.After initial penetration, this long, tapered bur creates a funnel shape for easier access in the pulp chamber. The safe-ended, non-cutting tip prevents penetration of the pulp chamber floor or the root canal walls. Cutting surface length: 9 mm; total length: 21 mm.	consumables	120	144000
38	Ceramic finishing diamond stone	Stones with embedded diamond particles for smooth finishing composites, porcelain and precious alloys. Available as a pack of 12.	consumables	120	288000
39	Ceramic finishing stone	Stones for smooth finishing of indirect restorations like composites, porcelain and precious alloys. Available as a pack of 3 or 4. Maximum speed 30000 rpm	consumables	120	96000
40	Cheek retractor	Good plastic quality,Light should pass,transparent and in different size and shapes.	consumables	120	38400
41	Cold Cure Acrylic Power and Liquid	1. Should be pink in color, strength should be good, give glossy finish & clear translucency after polishing porosity 2. Working time :- 5-7 min 3. Compressive strength :- 75 Mpa 4. Tensile strength :- 48-62 Mpa 5. Hardness:- 16 knh 6. Volumetric shrinkage :-linear 0.26 % 7. Powder should be 400 gms and 400 ml liquid monomer in one lab pack	consumables	280	201600

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
42	Composite bonding agent 7th Generation	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) 1 Bottle should have minimum 3 - 5 ml bonding agent	consumables	164	590400
43	Composite Polishing Kit	1) Should contain various sizes and shapes of rotary/hand tips. 2) Coarse, medium fine and extra fine grit should be available with colour coding. 3) one kit should have 12 burs	consumables	120	288000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
44	Composite shade A1	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) Composite material should contain prepolymerised filler partical for more strength should have CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT.	consumables	120	115200

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
45	Composite shade A2	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations in anterior & posterior teeth. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user CE and US FDA approved from notified body. 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) Composite material should contain prepolymerised filler particles for more strength should have CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT	consumables	120	115200

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
46	Composite shade A3	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user CE and US FDA approved from notified body. 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply CE and US FDA approved from notified body. 7) Material should achieve clinically and biological properties after curing. 8) Composite material should contain prepolymerised filler particles for more strength should have CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT	consumables	120	115200

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
47	Composite shade B2	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user CE and US FDA approved from notified body. 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) Composite material should contain prepolymerised filler particles for more strength should have CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT	consumables	120	115200

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
48	Composite tooth coloured for restoration	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations in anterior & posterior teeth. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) Composite material should contain prepolymerised filler particles for more strength should have CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT	consumables	120	96000
49	Compostable Black Garbage Bag Size 18inch X 22inch	Compostable Black Garbage Bag Size 18inch X 22inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	4615	738400
50	Compostable Black Garbage Bag Size 24inch X 27inch	Compostable Black Garbage Bag Size 24inch X 27inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	305	48800
51	Compostable Black Garbage Bag Size 24inch X 36inch	Compostable Black Garbage Bag Size 24inch X 36inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	1200	192000
52	Compostable Black Garbage Bag Size 27inch X 33inch	Compostable Black Garbage Bag Size 27inch X 33inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	205	32800

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
53	Compostable Black Garbage Bag Size 30inch X 36inch	Compostable Black Garbage Bag Size 30inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	110	17600
54	Compostable Black Garbage Bag Size 33inch X 36inch	Compostable Black Garbage Bag Size 33inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	20500	3280000
55	Compostable Blue Garbage Bag Size 18inch X 22inch	Compostable Blue Garbage Bag Size 18inch X 22inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	3650	584000
56	Compostable Blue Garbage Bag Size 24inch X 27inch	Compostable Blue Garbage Bag Size 24inch X 27inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	100	16000
57	Compostable Blue Garbage Bag Size 24inch X 36inch	Compostable Blue Garbage Bag Size 24inch X 36inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	100	16000
58	Compostable Blue Garbage Bag Size 30inch X 33inch	Compostable Blue Garbage Bag Size 30inch X 33inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	120	19200
59	Compostable Blue Garbage Bag Size 33inch X 36inch	Compostable Blue Garbage Bag Size 33inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics. Compostable Blue Garbage Bag Size 33inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	3106	496960

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
60	Compostable Red Garbage Bag Size 18inch X 22inch	Compostable Red Garbage Bag Size 18inch X 22inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	2535	405600
61	Compostable Red Garbage Bag Size 24inch X 36inch	Compostable Red Garbage Bag Size 24inch X 36inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	120	19200
62	Compostable Red Garbage Bag Size 27inch X 33inch	Compostable Red Garbage Bag Size 27inch X 33inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	120	19200
63	Compostable Red Garbage Bag Size 33inch X 36inch	Compostable Red Garbage Bag Size 33inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	9110	1457600
64	Compostable Yellow Garbage Bag Size 18inch X 22inch	Compostable Yellow Garbage Bag Size 18inch X 22inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	4925	788000
65	Compostable Yellow Garbage Bag Size 24inch X 27inch	Compostable Yellow Garbage Bag Size 24inch X 27inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	100	16000
66	Compostable Yellow Garbage Bag Size 24inch X 36inch	Compostable Yellow Garbage Bag Size 24inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	160	25600
67	Compostable Yellow Garbage Bag Size 27inch X 33inch	Compostable Yellow Garbage Bag Size 27inch X 33inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	105	16800

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
68	Compostable Yellow Garbage Bag Size 30inch X 33inch	Compostable Yellow Garbage Bag Size 30inch X 33inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	100	16000
69	Compostable Yellow Garbage Bag Size 30inch X 36inch	Compostable Yellow Garbage Bag Size 30inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	120	19200
70	Compostable Yellow Garbage Bag Size 33inch X 36inch	Compostable Yellow Garbage Bag Size 33inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.	consumables	10000	1600000
71	conservative instrument kit	1) all the instruments in the kit should be made of high grade stainless steel 2) it should be light in weight, easy to grip. 3) it should have sharp working ends. 4)it should be corrosion resistant 5) it should be autoclavable 6) kit should contain williams probe, probe,explorer,mouth mirror top with handle,tweezer,cement spatula, spoon excavator,ball burnisher,parallelogram condensor, flat condensor,plastic filling instrument,GMT #28,GMT# 29, Chiesel, HOE,hallen beck carver, wards carver, frahm carver,chisel bent,	consumables	120	192000
72	Cord Clamp Sterile	Disposable Umbilical Cord Clamp are made from Non toxic Medical grade polymer. 1. Material: PVC, PE, PP, ABS etc. 2. Sterilize by ethylene oxide gas 3. Supplied sterile in individual peel-able polybag. Or Equivalent	consumables	59500	105196

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
73	Dental Amalgam Alloy	1) Composition should contain :- ? Silver ? 60-70% ? Copper ? 13-30% ? Fine cut of spherical alloy particles 2) Trituration time ? 1 min max 3) Consistency ? with the specified alloy and mercury ratio at the end of specified trituration time, mass should be of proper consistency allowing to form rope or ball 4) Carving ? At the end of 5-7 mins. It should allow carving without fracture or deformation 5) Creep ? restoration should not have more than 2% creep 6) Polishing ? after 24 hours, the restoration on polishing should give high luster 7) Compressive strength ? 60,000-70,000 psi. 8) Tensile strength ? 7000-9000 psi 9) 30 grm bottle	consumables	100	240000
74	Dental explorer	1) It should be made of stainless steel. 2) It should be double ended instrument 3) it should facilitate Dental Diagnostic Examination 4) Rust Free, high Quality Stainless Steel .	consumables	120	2880
75	Dental floss	1) it should help to remove plaque under gumline 2) it should be multitufted 3) it should be ideal to remove plaque around braces, bridges and wide spaces 4) it should have 50 Pieces per pack 5) It should be Provided With handle	consumables	120	14400
76	Dental Plaster	Type II gypsum(Plaster) should be white without any foreign particles suitable for Dental laboratory work. It should be free from lumps. Initial setting time ? should not be less than five minutes & not more than eight minutes. Final setting time ? 30 minutes. Compressive strength ? Not less than 1300 ib/sq. Inch after 24 hours. Packing ? it will be delivered from time to time as & when required in the appropriate quantity & supplied in small packages as per demand in moisture proof resin bags lined with thick quality polythene. Be marked with ? 1) Manufacturers registered name & or mark. 2) Weight in the Container 3) Setting time 4) Date of manufacturer & Expiry 5) it should have 5 packets; per packet 5 kg(seperate pouch to avoid moisture contamination) total 25 kg box	consumables	1225	117600

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
77	Dental Polishing Prophyl cup	1) it should be disposable screw-type cups and should work with contra-angle handpieces. 2) it should remove superficial white and brown stains. 3) Its Silicon carbide microparticles provide gentle mechanical abrasion 4) Less than 0.2mm enamel removal 5) it should be Autoclavable Cups and should minimize splatter. 6) it should have a curved rib interior within the cup which is intended to reduce splatter, which both the patient and operator. 7) Eliminate splatter of prophyl paste, blood and saliva 8) Disposable, Latex free, vibration free All screw-type cups work with contra-angle handpieces. 9) It should be ADA/ISO/ISI/CE/USFDA certified from notified body.	consumables	120	9600
78	Dental stone	Type III Gypsum (stone plaster) for dental laboratory work 1) Composition ? Powder ? (CaSO ₄) 2 H ₂ O, Gum Arabic + Calcium carbonate 2) Setting time ? 7 to 12 minutes. Final setting time ? 12 to 13 minutes. Colour ? Any light shade 3) Compressive Strength ? 3000 lb/inch after one hour of setting. Linear expansion ? 0.03% after one hour. Not more than 0.20% after two hours. 4) Packing ? In 25kg. Airtight the lined with thick quality polythene. Container should be marked with the following information. ? Manufacturer, Registered name & or mark ? Weight in container ? Setting time ? Linear expansion on setting ? Dt. Of manufacture and expiry 5) it should have 5 packets; per packet 5 kg (separate pouch to avoid moisture contamination) total 25 kg box	consumables	1380	165600

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
79	Dental X-Ray film IOPA Adult (Pack of 150 films)	1) Name ? Intra-Oral-Periapical films (Adult) 2) Size of the film should be 32 x 41 mm (approx) 3) It should be E-Speed film 4) It should have a protective plastic cover (water proof) with soft borders thus making it comfortable to the patients. 5) Individual X-Ray film packet should have a double emulsion E-Speed film with embossed dot for side identification 6) The film should be covered with light proof (black coloured) paper packet and should have a lead/tin foil for preventing the effect of back scatter and for radiation protection 7) The packet should be flexible and not very rigid 8) The film should provide good radiograph with adequate density, contrast and details and should not have any inherent artifacts 9) All the film packets should have sufficient shelf life and should be supplied in adult form only 10) It each box should contain 150 films	consumables	690	1656000
80	Dentulous non-perforated Impression Tray	Dentulous non-perforated Impression Tray. Made of corrosion resistant, German stainless steel. Pack of 8 pieces	consumables	120	38400
81	Dentulous Perforated Impression Tray	Dentulous Perforated Impression Tray. Made of corrosion resistant, German stainless steel. Pack of 8 pieces	consumables	120	38400
82	Denture Finishing Bur kit	1) Should contain 2 sand stone, 2 carbide & 2 steel burs, 1 abrasive separating disks with holder. 2) 8 to 12 fluted carbide burs or abrasive with particle size of 100 microns or larger for bulk reduction. 2 carbide & 2 abrasive stone bur of large [article size for bulk reduction. 3) 12 to 16 fluted 2 in number carbide burs & 30 ? 100 Åµm sized abrasive stone burs for contouring, particle for contouring. 4) kit contains: 18 ? 13 flute carbide burs 2 in number, fine & 2 super fine, 2 diamond burs & 2 abrasives between 8 & 20 Åµm size. Total burs kit contains 23	consumables	120	96000
83	Denture polishing buff	Should be of muslin cloth, should have long service life, should take optimum polishing agent, non sticky	consumables	120	38400

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
84	Developer	Name of the material ? developer powder pack contains : Powder to make 5 liter solution	consumables	120	76800
85	Diacetyl monoxime	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	consumables	120	230592
86	Diamond abrasive points FOR CROWN cutting commonly called as air rotor	1. Should be available in various sizes and shapes as per the requirements like round, straight fissure(flat and round end), tapered fissure (flat and round end), safe ended, disc or wheel shaped. 2. Should be available in friction type design. 3. Shank and neck should be made of steel of good quality.4. Diamond particles embedded on the working end in the matrix which is durable.5. Should not rust and corrode in common disinfecting solutions. 6. Should be autoclavable.7. Should not disastrously break during use. 8. Should not dull after short duration of use.	consumables	120	9600
87	Diamond Burs Inverted cone	1) Should be available in friction types design 2) Shank and neck should be made of steel of good quality. 3) Diamond particles embedded on the working end in the matrix which is durable. 4) Should not rust and corrode. 5) Should be autoclavable. 6) Should not disastrously break during use. 7) Should not dull after short duration of use.	consumables	51	30600
88	Diamond Burs Round	1) Should be available in friction types design 2) Shank and neck should be made of steel of good quality. 3) Diamond particles embedded on the working end in the matrix which is durable. 4) Should not rust and corrode. 5) Should be autoclavable. 6) Should not disastrously break during use. 7) Should not dull after short duration of use.	consumables	101	60600
89	Diamond Burs Straight	1) Should be available in friction types design 2) Shank and neck should be made of steel of good quality. 3) Diamond particles embedded on the working end in the matrix which is durable. 4) Should not rust and corrode. 5) Should be autoclavable. 6) Should not disastrously break during use. 7) Should not dull after short duration of use.	consumables	171	102600

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
90	Diamond Burs Tapered	1) Should be available in friction types design 2) Shank and neck should be made of steel of good quality. 3) Diamond particles embedded on the working end in the matrix which is durable. 4) Should not rust and corrode. 5) Should be autoclavable. 6) Should not disastrously break during use. 7) Should not dull after short duration of use.	consumables	81	48600
91	Diamond Pencil	useful for marking glass slide	consumables	690	16560
92	Diaper Adult disposable	Disposable Adult Diaper	consumables	3800	881600
93	Diaper neonatal	Diaper neonatal for New born baby soft cotton surface dry, breathable quick absorbent and water resistance	consumables	28700	229600
94	Drill bits for micromotor 0.8 mmx 3mmx65mm	stainless steel, Autoclavable	consumables	120	96000
95	Drill bits for micromotor 0.8 mmx 5mmx65mm	stainless steel, Autoclavable	consumables	120	96000
96	Edentulous Perforated Impression Tray	Edentulous perforated Impression Tray. Made of corrosion resistant, German stainless steel. Pack of 8 pieces	consumables	120	38400
97	Endo access bur	1) Should be available in friction types design 2) Shank and neck should be made of steel of good quality. 3) Diamond particles embedded on the working end in the matrix which is durable. 4) Should not rust and corrode. 5) Should be autoclavable. 6) Should not disastrously break during use. 7) Should not dull after short duration of use.	consumables	81	116640
98	Endodontic Broaches	1The Barbed Broach is designed for bulk removal of pulpal tissue and features an easy-to-grip, color-coded plastic handle. Packaged in a sterile blister pack. Should be ISO Standardized 2. Should be color coded 3. Six broaches in one pack	consumables	120	38400

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
99	Endodontic chelating gel	chelating agent and lubricant used to soften calcification and aid in cleansing the root canal system.10% carbamide peroxide; 15% EDTA. In conjunction with NaOCl, its effervescing action lifts dentinal shavings and debris for efficient removal.Disposable tip eliminates the possibility of cross-contamination by dispensing a clean, single dose.Smooth, gel-like consistency clings to files, making precise placement in the canal easy	consumables	120	28800
100	Endodontic files (45- 80)	Made up of stainless steel. Should follow international standardization for endodontic hand files ISO/ISI in all its features and colour coding. Should be supplied in various length slike21, 25mm and others as per the need of the user. Should have good cutting efficiency. Handle design should have comfortable and firm grip. Supplied in assorted sets from no 45 to 80as per the requirement of the users. Should be auto clavable. Should not rust and corrode in common disinfecting solutions and sodium hypochloride. Should show sigh of deformation before breakage. Endodontic files should be provided K files. should have good cutting efficacy minimum upto 30 canals single file means 1 pkt of 21 mm & 25 mm of same no size	consumables	120	28800

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
101	Endodontic files (8, 10, 15- 40)	Made up of stainless steel. Should follow international standardization for endodontic hand files ISO/ISI in all its features and colour coding. Should be supplied in various lengths like 21, 25 mm and others as per the need of the user. Should have good cutting efficiency. Handle design should have comfortable and firm grip. Supplied in assorted sets from no 8 and 10 also from 15 to 40 (1 packet per size should contain 6 no. files) Should be autoclavable. Should not rust and corrode in common disinfecting solutions and sodium hypochloride. Should show sign of deformation before breakage. Endodontic files should be provided K files. should have good cutting efficacy minimum upto 30 canals & 1 unit means 1 pkt of 21mm & 25mm of same no size	consumables	253	60720
102	Endodontic files H (10, 15- 40)	Made up of stainless steel. Should follow international standardization for endodontic hand files ISO/ISI in all its features and colour coding. Should be supplied in various lengths like 21, 25 mm and others as per the need of the user. Should have good cutting efficiency. Handle design should have comfortable and firm grip. Supplied in assorted sets from no 10 also from 15 to 40 (1 packet per size should contain 6 files) Should be autoclavable. Should not rust and corrode in common disinfecting solutions and sodium hypochloride. Should show sign of deformation before breakage. Endodontic files should be provided H files. should have good cutting efficacy minimum upto 30 canals & 1 unit means 1 pkt of 21mm & 25mm of same no size	consumables	215	51600

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
103	Endodontic files H 45-80	Made up of stainless steel. Should follow international standardization for endodontic hand files ISO/ISI in all its features and colour coding. Should be supplied in various lengths like 21, 25 mm and others as per the need of the user. Should have good cutting efficiency. Handle design should have comfortable and firm grip. Supplied in assorted sets from no 10 also from 45-80 (1 packet per size should contain 6 files) Should be autoclavable. Should not rust and corrode in common disinfecting solutions and sodium hypochlorite. Should show sign of deformation before breakage. Endodontic files should be provided H files. should have good cutting efficacy minimum upto 30 canals & 1 unit means 1 pkt of 21mm & 25mm of same no size	consumables	310	74400
104	Endodontic hand C files	Stainless steel C-Files ,. It should provide easier access to challenging, calcified canals. Heat-tempered steel for stiff performance. ? Twisted file design for greater strength. ? Eases penetration of calcified canals.	consumables	120	76800
105	Endodontic hand files super flex	Get all the benefits of nickel titanium files with a cut like stainless steel. Due to the flexibility of nickel titanium, SureFlex files are better at following and conforming to the curvature of the canal during instrumentation.	consumables	270	108000
106	Endodontic hand flexo files	? Greater working comfort. ? Improved tactile feedback. ? X-ray visible calibration rings for accuracy and improved diagnostics. (25 mm and 31 mm only) ? Flexible strength for preparing even the most curved and narrow canals. ? Pre-inserted silicone stops	consumables	120	57600

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
107	Endodontic niti hand files 4%	1. Made up of nickel titanium alloy. 2. Should have a greater taper than ISO standardized instruments as per the requirement of the user. E.g. 4, 6, 8, 10, 12. 3. Should be supplied in various lengths like 21, 25, 31 and others as per the need of the user. 4. Should have good cutting efficiency. 5. Latch type design for universal grip in various gear reduction hand pieces and endodontic motors. 6. Supplied in assorted sets of instruments of greater taper and tip dimension for thin medium and wide cana 12.Should be ISO/ADA/CE standardised /approved. 13.Should have publised scientific studies regarding the brand supplied.	consumables	120	336000
108	Endodontic niti rotary files 4%	1. Made up of nickel titanium alloy. 2. Should have a greater taper than ISO standardized instruments as per the requirement of the user. E.g. 4, 6, 8, 10, 12. 3. Should be supplied in various lengths like 21, 25, 31 and others as per the need of the user. 4. Should have good cutting efficiency. 5. Latch type design for universal grip in various gear reduction hand pieces and endodontic motors. 6. Supplied in assorted sets of instruments of greater taper and tip dimension for thin medium and wide canals. 12.Should be ISO/ADA/CE standardised /approved. 13.Should have publised scientific studies regarding the brand supplied.	consumables	143	400400
109	Endodontic screw post system kit	A fully vented parallel post means less chance of excess pressure at the apical section of the canal. Self-tapering threads make these posts ideal for reconstruction of remaining tooth structure without the excess stress of a fully-treaded post. The Radix-Anker core section has 20 undercuts that secure any core material tooth surface. Made of titanium alloy. kit contains Posts 12 penetration Drill:1 rooter Facer:1 spiral Bur:1 Gauge1 Mandre:1 Wrench:1	consumables	120	288000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
110	Eugenol free Impression Paste	1. It should contain orthoethoxy benzoic acid in place of eugenol. 2. It should not irritate, 3. should be supplied as base and accelerator paste, 4. Working time should be 3-4 min. 5. Setting time should be 5-10 mins.	consumables	120	52800
111	Extraction forceps kit	1)it should be made of high grade stainless steel. 2) it should have sharp beaks on tip-which should allow better adaptation to the root surface and should cut PDL Easily. 4) it should have large handles for better operator fit. 5) it should have long life.6)instruments should be light weight and should have comfortable grip. 6) each kit should have a set of 12: 1 X Upper anterior forceps , 1 X upper Premolar forceps, 1X Upper molar right forceps, 1X Upper molar left forceps , 1X Upper roots forceps, 1 X upper third molar forceps, 1 X upper molars right (cowhorn), 1 X upper molar left (cowhorn), 1 X Lower premolars, 1 X Lower molars forceps, 1 X lower root forceps, 1 X Lower molars (cowhorn) forceps.	consumables	120	720000
112	Fluoride Varnish	1) the varnish should be a protective against dental caries with fluoride in it. 2)It should dry quickly and show excellent adhesion to teeth. 3) it should be suitable for treating children, adolescents and adults.4)The protective action of should be based on t inhibition of demineralisation, stimulation of remineralisation, incorporation of fluoride into the lower layers of enamel, repair of initial caries lesions.5) It should have Calcium + fluoride + phosphate to protect the teeth against acid attacks.6) Optimum flow and wetting properties Low viscosity Clear, Colourless Excellent adhesion. 7) It should be available in the form of capsule having 0.5ml of varnish	consumables	120	24000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
113	Foam fluoride trays Blue- Large size	1)It should have a Anatomical design which allows a comfortable fit and effective delivery of gel to interproximal and occlusal areas. 2) it should have Contoured shape which should increase a comfort and should reduce gagging. 3) it should have a handle for easier and correct placement. 4)It should be available in Color coding by size. Large - blue, Available in pack of 50 Pairs.	consumables	120	2880
114	Foam fluoride trays white- Medium size	1)It should have a Anatomical design which allows a comfortable fit and effective delivery of gel to interproximal and occlusal areas. 2) it should have Contoured shape which should increase a comfort and should reduce gagging. 3) it should have a handle for easier and correct placement. 4)It should be available in Color coding by size. Medium - white. Available in pack of 50 Pairs.	consumables	120	2400
115	Foam fluoride trays Yellow- Small	1)It should have a Anatomical design which allows a comfortable fit and effective delivery of gel to interproximal and occlusal areas. 2) it should have Contoured shape which should increase a comfort and should reduce gagging. 3) it should have a handle for easier and correct placement. 4)It should be available in Color coding by size. Small - yellow. Available in pack of 50 Pairs.	consumables	120	2400
116	FP Commodities: PTK	FP Commodities: PTK (Pregnacy Test Kit)	consumables	3410310	122771160

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
117	Fumigation Disinfectant	It should have chemical components of potassium monopersulphate 40 to 50%, sodium C10-13-alkylbenzenesulphate 10-20%. sodium chloride 1-5% 1) It should have a property which can be used for aerial disinfection 2) it should be suitable for Hospital, operating and dental clinic use 3) it should be safe and should have no oral and dermal toxicity 4) it should be available in 500 g per jar 5) it should clean air against risk of airborne infections 6) it should contain no aldehyde and no toxic vapour phase. 7) it should have proven efficiency against viruses, bacteria, fungal pathogens. 8) it should be made of peroxygen compound, surfactant, organic acids, and inorganic buffer system. 9) it should be approved by environmental protection agency	consumables	120	336000
118	Gas Refill	1) Should contain Liquid petroleum gas. 2) Should have Light weight, Easy to operate. 3) Refillable gas tank	consumables	120	23040
119	gas torch	1) Should contain Liquid petroleum gas. 2) Should have Light weight, Easy to operate, Adjustable flame control. 3) Refillable gas tank 4) should have ignition switch with flow control valve.	consumables	118	141600
120	Gates glidden drills	Stainless-steel construction provides flexibility that helps prevent breakage. The drill bits should be compatible with any right-angle latch handpiece and typically should come in lengths of 28 mm, 32 mm, or 38 mm. A range of six sizes should allow the clinician to sequentially use the drills to achieve the appropriate taper and avoid excessive dentin removal.	consumables	120	62400
121	Gigli Saw wire	Made of martensitic steel (quenched, magnetic steel) and not stainless steel. Hooks at both ends. Should conform to ISO standard	consumables	1750	368480

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
122	Gingival retraction cord	1. Non impregnated and knitted cord 2. Should be 100% cotton, Should compress upon packing and then expand for optimal retraction 3. Should not entangle in diamond bur 4. Bright colors to facilitate easy location and removal 5. The bottle should have a high carbon steel blade at one end to cut desired length of cord 1 x Bottle of cord (8ft/244cm). 6. 1 x #000 cord, 1 x #00 cord, 1 x #0 cord	consumables	120	192000
123	Glass Ionomer cement (Luting/ Type 1)	1) Maximum particle size should be 15um 2) Powder/liquid ratio in the range of 1.8/1.0 Mixing time 3) (approx..) ? 20 sec working time (approx..) ? 2 min. (230 c) (from start of mixing) Final finishing commencing time 4 - 5 min (after seating the restoration 4) Film thickness of luting cement should be <20 um 5) Material should be streontium based gic with good track record of its use in dental care hospitals 6) ISO/ISI/CE/ADA specifications should be met 7) It should be 15 gms powder and 10 ml liquid.	consumables	152	401280
124	Glass Ionomer cement (Luting/ Type 2)	ADA specification No 66 Types Lutting Type I Restorative - Type II Liner Base Supplied as : Powder Liquid form	consumables	249	597600
125	Glass Ionomer cement type IX	Type IX GIC 1) It should be supplied in the form of powder & liquid with components as per ADA/ISO/ISI specification 2) Material should have good handling characteristics and shelf life of at least two years from supply 3) It should have compressive strength more than 180 MPA 4) Finishing of material should be possible immediately after setting 5) It should have adequate mixing & working time. It should have less setting time 6) It should be radioopaque. 7) It should be 15gms powder and 7.5ml liquid	consumables	106	315456
126	Glass Iononier Cement (GIC)	Glass Iononier Cement (GIC)	consumables	1068	153792

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
127	Green Stick	1) It should be available in stick form 2) Flow should be less than 6% at 37oC and 85% at 45oC 3) It should be homogeneous on heating on flame 4) It should be able to record fine details in mouth 5) It should be easy mold and handle 6) 10 stick per packet	consumables	600	115200
128	Gum Paint (Tannic Acid)	Gum Paint (Tannic Acid)	consumables	1068	34176
129	Gum tragacanth	SQ	consumables	120	50880
130	Gutta Percha (GP)	Gutta Percha (GP)	consumables	1068	512640
131	Gutta percha points 2 degree	1) ISO Standardized Gutta-percha points from number 15 onward sizes (#15-40) 2) (#45-80) 3) #15,#20,#25 4) Should be supplied in assorted packets as per required numbers and quantity 5) Should be colour coded 6) Have 2 degree taper 7) Should be accurate in tip diameter and taper as per ISO standards 8) Should not buckle down easily should be length marked 120 points in 1 packet	consumables	120	40800
132	Gutta percha points 4 degree	1) ISO Standardized Gutta-percha points from number 15 onward sizes (#15-40) 2) #15,#20,#25,#30.#35,#40 3) Should be supplied in assorted packets as per required numbers and quantity 4) Should be colour coded 5) Have 4 degree taper 6) Should be accurate in tip diameter and taper as per ISO standards 7) Should not buckle down easily should be length marked 60 points in 1 packet	consumables	120	33600
133	Gutta percha points 6 degree	1) ISO Standardized Gutta-percha points from number 15 onward sizes (#15-40) 2) #15,#20,#25,#30.#35,#40 3) Should be supplied in assorted packets as per required numbers and quantity 4) Should be colour coded 5) Have 6 degree taper 6) Should be accurate in tip diameter and taper as per ISO standards 7) Should not buckle down easily should be length marked 60 points in 1 packet	consumables	125	35000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
134	Hand held vein detector	The hand-held device to find veins for needle insertion powered by rechargeable battery.with minimum 30 Red & Orange LEDs for visualization of deeper as well as superficial veinsThe device should have the option to use Red and Orange LEDs separately /simultaneously.The viewing diameter = not be less than 33mm.The open area = not be less than 17mm .The device should show clear visualization of veins in adult/pediatric / geriatric/obese/dark skin patient with difficult vein access..power adaptor The device should be made by ABS material & not by plastic . EC REP	consumables	120	230400
135	Hard reliner	A hard, chairside, self-cure relining material with no/ minimal exothermic heat. It must completely set in the patients mouth in approximately 10 minutes. Dense, stable and color-fast. Easy to finish and polish	consumables	120	672000
136	Harris Haematoxiline	it should identify, and stain the cytological smears and tissue . It should stain the nuclear details. It should have a good shelf life .The stain retainability should be there for a longer period of time .	consumables	74	177600
137	Heat Cure Acrylic Powder and Liquid	1. It should be tasteless odorless , non toxic and non irritant 2. It should be available in different shades 3. It should bond chemically with denture acrylic teeth 4. It should be dimensionally stable during curing process 5. It should have enough strength, abrasion resistance and resilience 6. Should be insoluble and impermeable to oral fluids 7. Should have low specific gravity (light in weight) 8. It should withstand flasking , deflasking , curing process 9. It should have coefficient of thermal expansion same as denture acrylic teeth 10. It should be radio opaque so that can be detected on x ray, if swallowed accidentally 11. It should be easy to manipulate & repair 12. Powder should be 400 gms and 400 gms liquid monomer in one lab pack	consumables	232	167040
138	Hematoxyline Delafield	For microscopy ,owners quality control certificate	consumables	114	53078

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
139	Hematoxylen e Powder	For Staining, ce certified, owners quality control certificate	consumables	84	237888
140	Hemocytomet er (Square tube)	For estimation of Hemoglobin.	consumables	120	102720
141	Hot water bag Rubber for hot Fomentation	Should be made of rubber,which can resist the hot water pressure comfortably, Should be high quality, leak proof, easy to clean & keep water hot for longer time.Should be reusable.	consumables	127	10485
142	Humidifier bottle for oxygen flow-meter	Polycarbonate bottle,transparent,crack resistant,unbreakable,screw type,airtight	consumables	600	28800
143	Hydrofluoric gel	Low-viscosity buffered gel containing hydrofluoric acid at 9% for etching of porcelain. High visibility yellow hue Etch should be easy to control and place Should not stain composite or resin cement Should yield very good porcelain-to-resin bond strength	consumables	120	96000
144	Ideal model silicon mould (upper & lower set)	It should be made up of good quality silicone,It should have good anatomical landmarks,easily separatble,resistant to tearing	consumables	120	48000
145	Impression Compound	1) It should be available in cake form 2) It should be fuse in hot water and form homogeneous mass 3) Working temperature: 55oC ? 65oC 4) Manipulation time(Kneading time): 1-3 minutes 5) Fusion temperature:43.5oC 6) Flow should be less than 2% at 37oC and between 75-85% at 45oC 7) 5 cakes per packet with 200 grms in weight	consumables	400	80000
146	Impression disinfectant	99.9% isopropyl alcohol-based hospital grade disinfectant for dental impressions. Must be compatible with impression materials. Should not deform or distort the materials. Available in 500 ml spray bottles	consumables	120	115200
147	Impression Paste (ZOE)	ADA Specification no 16. should contain base and catalyst pastes Mixing time 1 min Final Setting time 15 mins Little or no dimensional change during setting It should record fine details in impression	consumables	100	48000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
148	Interdental brush	1) Interdental cleaning brush should be designed to clean various interdental spaces. 2) It should work by cleaning off braces, dental bridges and other spaces between the teeth. 3) It should be plastic-coated wire material and has a smooth inward-outward motion to improve movement. 4) It should be integrated with a rubber grip to ensure it stays intact and will not slip easily.	consumables	120	19200
149	Intermediate Restorative Material	ISO 9001:2008 Available in powder and liquid form (Dentsply Caulk) Powder contains Zinc oxide and polymethyl methacrylate Liquid contains Eugenol Mixing time - 5 to 10 secs Compressive strength 48 Mpa Setting time - 6 to 10 min	consumables	51	102000
150	Interproximal polishing strip for composite	should be color coded with matching grit size for easy identification	consumables	120	240000
151	Intra canal irrigation needle with lateral opening	1) Needle gauge size 27 or 30 2) Should have lateral opening & disposable. 3) It should be 100 needles per packet	consumables	120	144000
152	IUCD 375	IUCD 375	consumables	36600	2928000
153	IUCD 380A	IUCD 380A	consumables	53400	5126400
154	Lab Cleaning Solution	Laboratory glassware/Plastic Cleaning solution, owners quality control certificate, 1X1 Litre	consumables	778	168048
155	Lacto Phenol Cotton Blue	ISO certified, WHO/GMP confirmed to / CE notified	consumables	120	16992
156	Laryngoscope Adult	Should be LED laryngoscope set, adult, macintosh type, Stainless steel blades with removable led lamp. Handle and blades are made of 304 grade stainless steel. Dull/matt finish. Supplied with 4 blades size 1, 2, 3 and 4 and two batteries for handle. Should provide with 1 spare led lamp and batteries	consumables	1146	1552051
157	Laryngoscope Paediatric	Should be LED laryngoscope set, adult, Miller Type, Stainless steel blades with removable led lamp. Handle and blades are made of 304 grade stainless steel. Dull/matt finish. Supplied with 4 blades size 0,00,1 and 2, two batteries for handle. Should provide with 1 spare led lamp and batteries.	consumables	407	626780

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
158	Laser Tip	940 Fibreoptic Lez Tip 3-7 mm Single use, non sterile, Autoclavee before use	consumables	120	96000
159	lathe wheels	1) white in colour 2)should have rounded edges	consumables	120	48000
160	Lead equivalent protective apron with thyroid shield of minimum 0.5 mm thickness (96% protection) light weight adult size of all types foldable	Lead equivalent protective apron with thyroid shield of minimum 0.5 mm thickness (96% protection) light weight adult size of all types foldable	consumables	120	5952000
161	Light Cure Composite	Light Cure Composite	consumables	1068	4272000
162	Light Cure Resin Restorative material (Kit)	1) Light cure composite resin kit should consist of ? Various shades of the material white to yellowish and light to darker ? Preferably have Shade guide provided in the kit? Etchant gel ? Phosphoric acid 35-37% 3 ml ? Bonding agent packaged in a leak proof and good quality bottle ? 1 unit of Mixing pad should have 100 pad sheets) Applicator tips Requirements of the material: ? Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 3) Material should have optical properties like tooth and have life like appearance after making restorations. 4) Should have good Polishibility 5) The material should be freshly manufactured and should have a sufficient shelf life. 6) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without andue polymerization & SHOULD HAVE CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT 7) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user 8) Material	consumables	173	899600

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
		should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 9) Material should achieve clinically and biological properties after curing. 10) Composite material should contain prepolymerised filter partical for more strength kit contains: 5 composite syringes of 4-5gm each, 1 denting bonding agent 5ml, etchant 3-5ml, applicator tip s 100 sticks, mixing pad			
163	Lubricant spray for hand piece	1) It is used for maintenance of all dental hand pieces including high speed, slow speed and air motors. 2) it should have minimum quantity of 400 ml CE certified product	consumables	76	60800
164	Makintosh rubber sheet 110cm	Should be made of rubber with double colour, high grade polymer (55% grade A); Should be resistant to high temperature,chemical resistance and fair abrasion resistance.Should be water proof. Colour :Green,Blue, Red, Thickness-0.4mm, Width-110cm. Roll of 30 metres I.S.O.Certified,ISI Mark 4135: 1974	consumables	6150	1567512

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
165	Makintosh rubber sheet 90cm	Should be made of rubber with double colour, high grade polymer (55% grade A); Should be resistant to high temperature, chemical resistance and fair abrasion resistance. Should be water proof. Colour :Green, Blue, Red, Thickness-0.4mm, Width-90cm Roll of 30 mtr I.S.O. certified, Double colour, water proof ISI Mark 4135: 1974	consumables	6320	1610842
166	Malachite Green	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	consumables	120	483456
167	Male Condom	Male Condom	consumables	751200	6009600
168	Material for endodontic regeneration, healing & sealing.	Material should have excellent healing potential. Should be applicable in endodontic perforation repair, apexification, retrograde filling in endodontic surgeries. Should have scientifically proven biocompatible. Manipulation and handling characteristics should be good. Packaging should be compatible for multiple use of the material from the container.	consumables	120	288000
169	Matrix bands no 1	1) Metallic, stainless steel, Contour able, standard design 2) matrix bands no 1 3) Made up of Stainless Steel 4) should not rust and corrode in common disinfecting solution.	consumables	113	9040
170	Matrix bands no 8	1) Metallic, stainless steel, Contour able, standard design 2) matrix bands no 8 3) Made up of Stainless Steel 4) should not rust and corrode in common disinfecting solution.	consumables	120	9600
171	MBT bracket kit 0.018 Slot	It should be made up of good grade S.S. it should have whole arch (premolar to premolar) Brackets, Canine should have 0 degree Torque, Hooks with Canine and premolars.	consumables	300	240000
172	MBT ceramic bracket kit 0.018 slot	It should be of low profile. wings should not fracture. It should be of good quality international standard kit of 20 brackets	consumables	120	768000
173	Mercury for dental use	1) Should be 100% Pure 2) Should be specific for dental use. 3) Should be arsenic free. 4) Should be packaged in containers which are safe from accidental spillage and easy to dispense in required quantities.	consumables	100	400000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
174	Metal cutting Bur with air rotor attachment pack of 1 burs	1) Should not Rust. 2) Should Cut Metal & Ceramic crown effectively. 3) Should attach with Air rotor	consumables	120	57600
175	Micromotor diamond burs	1. Should be available in various sizes and shapes as per the requirements like round, straight fissure, tapered fissure, inverted cone, safe ended, disc or wheel shaped. 2. Should be available in friction type design. 3. Shank and neck should be made of steel of good quality. 4. Diamond particles embedded on the working end in the matrix which is durable. 5. Should not rust and corrode in common disinfecting solutions. 6. Should be autoclavable. 7. Should not disastrously break during use. 8. Should not dull after short duration of use.	consumables	120	57600
176	Microtome Disposable blade	Low Profile Blade for Microtome	consumables	173	1453200
177	Modelling wax	1) Should be soft & pink in colour. 2) It must be capable of holding denture teeth in position at room temp and mouth temp 3) It should cohere easily 4) It should not be flaky or adhering to finger 5) It should not cause irritation to the oral tissue 6) It should be easy to trim with a sharp instrument at 23 0c 7) 12 Sheet per packet or 225 grms in weight	consumables	250	36000
178	Mouth mirror & straight probe	It should be made of good stainless steel rust proof , with good reflective mirror changeable mirror top, sturdy handle. The straight probe should be of a good stainless steel rust proof material	consumables	1580	101120
179	Mouth mirror with handle	1. It should be made of good quality stainless steel rust proof material, 2. The mirror should be reflective and of a good quality	consumables	570	18240
180	Mucus extractor for neonatal deelee catheter Each	Transparent Tube mucus extractor for neonatal & infants. Sterile Pyrogen Free, Non Toxic The capacity of the Mucus Extractor is 25ml and a spare cap is also provided for the sealing. Soft and kink resistant tubing with frosted surface A traumatic, soft rounded, beveled open tip. Capacity: 25 ml.	consumables	1010	7190

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
181	Nichrome loop 1.3mm	Diameter 1.3mm,double wound, calibrated to 1.0 ul	consumables	120	24480
182	Nichrome loop 2mm	Diameter 2mm,double wound, calibrated to 0.005 ml	consumables	120	24480
183	Nichrome loop 4mm	Diameter 4mm,double wound, calibrated to 0.01 ml	consumables	153	31212
184	Nichrome straight wire	Suitable for stab inoculation	consumables	70	14504
185	NiTi Rotary endodontic instruments (Single file system)	1) Should follow international standardization ISO/ISI in all its features 2) Should be autoclavable 3) Should not rust & corrode in common disinfecting solutions & sodium Hypochlorite 4) Should not rust and corrode in common disinfecting solutions and sodium hypochloride. 5) Should show sign of deformation before breakage. 6) should have good cutting efficacy canal centring ability variable taper file pack of 6 or fixed taper	consumables	300	840000
186	Nitrous Oxide Gas for Inhalation (Store under pressure in metal cylinders of the type conforming to the appropriate safety regulations and at temperature not exceeding 37°C)	Nitrous Oxide Gas for Inhalation (Store under pressure in metal cylinders of the type conforming to the appropriate safety regulations and at temperature not exceeding 37°C)	consumables	1116	1339200
187	Orthodontic composite light cure material kit (etchant, primer, composite material syringe)	It should have long shelf life ,good bonding quality,available in syringe form ,with good working time of 3 min.of good viscosity, syringe of 4gm and bottle of 5ml primer, Primer should have good bonding quality.	consumables	120	192000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
188	Orthodontic mini implants with bracket head 1.4 x 8 mm, 1.6 x 6mm, 1.6 x 8 m, 2.0 x 8 mm	It should be made up of good quality, it should have different shape and type of head, it should be made up of good quality material, corrosion and tarnish resistant and it should have a good strength and resistant to deformation sizes : 1.4 x 8 mm, 1.6 x 6mm, 1.6 x 8 mm, 2.0 x 8 mm	consumables	55	66000
189	Oxygen Gas for Inhalation	Oxygen Gas for Inhalation	consumables	93516	52368960
190	PAP kit	Content nuclear stain solution 1 bottle (500ml), cytoplasmic staining solution 2A and 2B bottle (500ml) each, pack size 3X 500ml. the manufacturer should be CE notified, WHO GMP confirmed and should give certificate of analysis. Shelf life should be more than 2 years. time consumption should be less than 5 min.	consumables	327	384552
191	Peeso reamer	Peeso Reamer is excellent for enlarging the cervical portion of the canal, preparing the canal entrance, and removing gutta-percha for post space. Notched for easy size identification, they can be used in any right-angle handpiece. Available in standard (32 mm) and short (28 mm) lengths.	consumables	120	62400
192	Perio Dressing Pack	1) Easily mouldable 2) Elastic consistency 3) Patient friendly 4) Eugenol Free 5) Compatible with resins 6) Regular set 7) Two paste system should be provided. 8) pack should 90 gms of base and 90 gms catalyst 9) CE and US FDA approved from notified body	consumables	120	240000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
193	Personal Protection Equipment Kit(PPE Kit) Medium size XL	PPE Kit- (Personal protection equipment kit)- PPE kit should include the following safety gears: 1. Personal Protective coverall/Gown (Garment)- 1 no 2. Shoe Cover Upto knee ? 1 Pair 3. Eye Goggle ? 1 no 4. Face shield ? 1 no 5. N-95 Mask ? 1 no. NIOSH Certified 6. Nitrile Gloves ? 1 Pair 7. Bio Degradable Disposable Waste collection bag ? 1 no All the above safety gears shall be packed in a paper poly pouch. 1. N-95 Mask- One Unit ? Shape that will not collapse easily. ? NIOSH certified. ? High Filtration efficiency. ? Fluid Resistance, minimum 80mm Hg pressure based on ASTM G 1862, ISO 22609 or Equivalent. ? Quality compliant with standard for particulate respirator that can be worn with full face shield. ? Quality compliant with standards of surgical N-95 Mask 2. Face Shield ? One Unit ? Made of clear plastic and provides good visibility. ? Adjustable band to allow good fit around the head and snug fit against the forehead. ? Fog ? Resistant (preferable). ? Reusable (made of material that can be cleaned and disinfected) or disposable. ? Quality compliance with standards : EU Standard directive 86/686/EEC, EN 166/2002, ANSI/ISEA 287.1-2010 or equivalent. 3. Goggle ? One Unit ? With transparent Glass/plastic zero power, good fitting, covered from all sides with elastic band or adjustable holder. ? Good seal with the skin of the face ? Flexible Frame to easily fit all face contours without too much pressure. ? Covers the eyes & surrounding areas and accommodates for prescription glasses. ? Fog and scratch resistance ? Indirect venting to reduce fogging 4. Nitrile Gloves ? One Pair ? Nitrile ? Non-sterile, powder free. ? Outer gloves preferably reach fore-arm (min 225mm total length). Thickness 5.5 MIL and size (M/L) ? Quality compliance with below standard or equivalent. ? EU standard directive 93/42/EEC class I EN 455 or equivalent. ? EU standard directive 89/686/EEC category III EN 374 or equivalent. ? ANSI/SEA 105-2011 or equivalent. ASTM D6319-10 or	consumables	120	13344

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
		equivalent. 5. Disposable Coverall ? Along with One hood Drape OR One head Cover. (it should be 70 GSM and above ? Non Laminated ? One Unit ? Single use. ? The Fabric Garment/breathable Non Laminated Fabric ? Light colours to better detect possible contamination. ? Coverall Shall be designed to be Universal Fit. Soft Elastic to be fitted around front of hood, Wrist and ankles ? Coverall approved by SITRA (South Indian Textile Research Association), DRDE (Defence Research Development Establishment) and have valid Unique Covid Code (UCC). 6. Pairs of Shoe cover made of same fabric as of Coverall upto below Knee ? one pair 7. Bio Degradable Disposable Waste collection bag ? One Unit			
194	Pit and fissure sealant kit	1) Should have properties to flow in tooth developmental pit & fissures. 2) Should have tooth colour. 3) Should have anti carious properties. 4) This standard specifies requirements & test methods for resin based materials suitable for sealing pits & fissure in teeth. 5)This standard covers chemically cured/ LIGHT CURE 6) 1 kit contain 2 ml 2 syringes and 1 pit & 1 fissure small size carbide bur	consumables	120	528000
195	Plastic base formers	Plastic base former for dental models made of durable plastic. Available as set of maxillary and mandibular base formers	consumables	120	19200
196	Plastic tube 12 x 100 mm	12 x 100 mm	consumables	6110	4194
197	Plastic tube 12 x 75 mm	12 x 75 mm	consumables	28125	19305

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
198	Plastic Wedges	1) it should be pack of 70 2) It should be small , tapering , traingwar pieces of clear plastic about 1/2 nch length	consumables	120	48000
199	Polishing buff wet and dry	1) Should have fine bristles in a circular disk form. 2) Should have properties to polish denture surface.	consumables	120	6912
200	Polishing Cake	1) Should have abrasive grit suitable for polishing denture base resins. 2) Should not burn and stain the prosthesis 3) Should have properties of polishing. 4) Should not leave remnants on the surface of prosthesis.	consumables	120	17280
201	Polishing paper 90 & 100 no.	1) Useful for polishing acrylic denture surface or shellac base plate. 2) ISO Standard very fine (Macrogrit) of P150 grit of average particle diameter of 90/100 microns	consumables	100	400
202	Polishing Paste	1) Fine. Prophylaxis paste 2) Should contain fluoride 3) Should be available in various grit sizes and flavours 4) It should be 100 gms paste	consumables	120	34560
203	Powered Endo Cuuter reloadds 60 MM	Endoscopic white, Blue, Gold and Green 60 mm linear cutter GST cartridge with reload surface having pocket extensions to prevent tissue milking, staples with asymmetric bent tip for consistent staple formation with an option of closed staple height of 1mm, 1.5mm,1.8mm and 2mm 6 rows of staple line with 3 on each side of cut line compatible for articulating Gun. A) White - 1mm B) Blue - 1.5mm C)Gold - 1.8mm D) Green - 2mm*	consumables	120	1344000
204	Pre formed space maintainer	All in one Starter Kit with Acessories : 1) 37 universal bands with tube 2) 40 space components: straight,curved, distal shoe 3) 1 Intra oral crimper 4) 1 Intra oral cutter Specifications kit contains : Sizes per band: 24,24 1/2,25,25 1/2, 26???41 1/2, 42 Type of band: Universal Types of components: Straight, Curved, distal shoe Number of bands per size: 1/size Number of components: 40 Shape of bands : Universal Type of cutter: Wire Cutter with carbon attachment Type of crimper: Intra oral crimper Number of cutter: 1 Number of crimper: 1	consumables	120	3456000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
205	Prophy Powder	Abrasive Thixotropic Should be available in different grits and flavours Non splattering Homogenous and should not dry out.	consumables	120	11520
206	Provisional crown material	1.Should be supplied in the form of cartridge and automixing tips (in a packet) as per ADA/ISO/ISI specification. 2.Should be a self-curing composite for the fabrication of temporary crowns and bridges, inlays, onlays and veneers. 3. It should be free from methylmethacrylate and peroxides. 4. It should have a filler content: 40-45 wt.-% 5. Mixing time- 30-45 sec 6. Setting time- 2-3 min 7. Expiry should be at least for 2 yrs from the date of supply.	consumables	120	576000
207	Pumice Powder	Should follow ISO/ISI standards use in cleaning dentures use in cleaning dentures	consumables	77	5544
208	Rebreathing bags Size 1000ml	Should be made of non toxic, latex free material (Green/Black), with proper fix to the breathing circuit connector. Resistance: 5x104 / 5x106 ohms. 5602/BS 3353 standard .Capacity 1000ml.	consumables	210	8556
209	Rebreathing bags Size 1500ml	Should be made of non toxic, latex free material (Green/Black), with proper fix to the breathing circuit connector. Resistance: 5x104 / 5x106 ohms. 5602/BS 3353 standard .Capacity 1500ml.	consumables	300	13162
210	Rebreathing bags Size 2000ml	Should be made of non toxic, latex free material (Green/Black), with proper fix to the breathing circuit connector. Resistance: 5x104 / 5x106 ohms. 5602/BS 3353 standard .Capacity 2000ml.	consumables	160	10029
211	Rebreathing bags Size 500ml	Should be made of non toxic, latex free material (Green/Black), with proper fix to the breathing circuit connector. Resistance: 5x104 / 5x106 ohms. 5602/BS 3353 standard .Capacity 500ml.	consumables	120	4512
212	Resin Based Root canal Sealer	should have excellent radio-opacity, should have low shrinkage, low solubility and good flow, 1:1 paste to paste mixing system, silver free for no discoloration in teeth, biologically inert , thermoplastic ability for easy removal	consumables	120	412800

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
213	Resin luting cement	Dual cure resin luting cement. Should be supplied in two components It should have Translucent shade. Kit should contain: base, catalyst, etchant, 3gm primer, 5gm bond, applicator tips	consumables	52	249600
214	Retention mesh for RPD	Easy to adapt, tension-free wax retention meshes for RPD casting	consumables	120	384000
215	Retreatment Rotary File	pack of three files, 3 lengths, 3 progressive tapers and 3 apical diameter to remove material from coronal, middle and apical thirds of root canals	consumables	102	293760
216	RF/RA	Based on latex agglutination For the detection and semiquantitation of Anti-Streptolysin O in serum Sensitivity ? 200 IU/ml Specificity Purified Streptolysin O specific for Anti-Streptolysin O antibodies Latex reagent Uniform suspension of polystyrene latex particles coated with stabilized Streptolysin O GMP SCH M PRODUCT WISE AND TESTING FACILITY WISE, 50 test/kit	consumables	290	4176
217	Rnase away for molecular testing	Rnase away for molecular testing (Removes Rnase and DNA Contamination)	consumables	120	124608
218	Root Canal Conditioner EDTA 15%	Should contain 10% carbamide peroxide and 15 % EDTA, should have effervescing action in conjunction with sodium hypochlorite that will lift dentinal shavings and debris for efficient removal, lubricant used to soften calcification and aid in cleansing the root canal system. Dispense in syringe with disposable tips	consumables	100	40000
219	Rotary glide path files	flexibility and resistance to cyclic fatigue, less canal transportation than K files, suitable for curved canal, original canal anatomy is preserved	consumables	60	31200
220	Round bur surgical (straight micromotor hand piece)	HP8 stainless steel,	consumables	66	15840
221	Rubber Catheter Plain Size 5	Plain rubber catheter - soft pure rubber catheter , non sterile, non irritant to urethra , Length 16 inches, non irritant red rubber. Proximal funnel end & distal rounded atraumatic tip with side hole	consumables	200	3200

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
222	Rubber Catheter Plain Size 6	Plain rubber catheter - soft pure rubber catheter , non sterile, non irritant to urethra , Length 16 inches, non irritant red rubber. Proximal funnel end & distal rounded atrumatic tip with side hole	consumables	120	2112
223	Rubber Catheter Plain Size 7	Plain rubber catheter - soft pure rubber catheter , non sterile, non irritant to urethra , Length 16 inches, non irritant red rubber. Proximal funnel end & distal rounded atrumatic tip with side hole	consumables	300	5280
224	Rubber Catheter Plain Size 8	Plain rubber catheter - soft pure rubber catheter , non sterile, non irritant to urethra , Length 16 inches, non irritant red rubber. Proximal funnel end & distal rounded atrumatic tip with side hole	consumables	500	8800
225	Rubber dam Sheets	1. It should be available with dimensions 6x6inch. 2. It should be available with thickness 0.2mm heavy 0.02mm medium. 3. It should be available in a box containing 36 pieces. 4. 4.Should be green /blue colour	consumables	120	172800
226	Rubber teat Big	Big	consumables	635	2032
227	Rubber teat Small	Small	consumables	1350	27000
228	RVG senor sleeves disposable	Should be flexible and leak proof made up in polythene form For dental use It should have pack of 500 sleeves	consumables	91	43680
229	Sanitary Napkins with Belt ETO sterile Pack of 6	1) Overall Length (mm) - 540+/-10mm 2) Fluff core length (mm)-240+/-10mm 3) Fluff core width -75+/- 5mm 4) Thickness of a single pad 7-9 mm 5) weight of a single pad 9-11 gm 6) Absorbent ability- not less than 150ml	consumables	500	1800
230	Saponin	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	consumables	120	2307744
231	Screw expansion (jack screw)	It should be good quality S.S. expansion screw. For upper jaw ? dimention?12mm	consumables	120	14400

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
232	Sectional Matrix band system	Rings,matrices, wedges,forceps and pin tweezer for class II restoration , should create accurate contact and tight seal, minimal overhanging, rings should include silicone tips for better adjustment and clamps should be silicone protected. Should be non corrosive and autoclavable kit contains 2 rings, 1 ring holding forcep & minimum 20 bands	consumables	120	480000
233	Separators elastics	It should be of good quality . made up of medical grade latex, it should have long shelf life.	consumables	120	96000
234	Seperating Medium	1) It should have excellent separating properties in between plaster & denture base material and n between plaster and plaster. 2) Should contain water, sodium alginate, borax, glycerin and dye as per ISO/ISI Standards. 3) It should form thin layer on the cast 4) It should easy to separate denture from the cast after curing 5) It should not chemically react with denture resin or should not hamper polymerization of resin 6) It should be thin enough to apply with brush on the cast 7) It should be 5 litre	consumables	120	144000
235	Shoe Cover	Made of Nonwoven material which is water and alcohol repellent. The pair should be of standard size,should be long upto below knee.	consumables	485105	531675
236	Silane coupling agent	Silane coupling agent for adhesion with porcelain. Pre-activated Solvent based on ethanol. Available in 5 ml bottle.	consumables	120	115200
237	Silver Amalgam	Silver Amalgam	consumables	1068	427200
238	Silver solder wire 10 gm spool	It should be of minimum dimension of 6 feet,and weight 8 gms, it should be of good quality international standard.	consumables	120	192000
239	SKIN GRAFTING BLADE	Stainless steel blade, sharp blade edge should offer smooth shaving of healthy skin tissue,Individually sterile packed, european CE certified, Expiry date should be written.	consumables	850	189856
240	Slide Cabinet	Steel - 50,000 slides store cap.	consumables	120	3360000
241	Sodium barbitone	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	consumables	120	10272

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
242	Soft reliner	1. Secure bonding to the denture acrylic. 2. For manual mixing 3. Mixing ratio 1:1. 4. Bicompatible, tasteless and odourless. 5. Colour: similar to gingiva 6. Working time: 1 minute 30 seconds approx 7. Setting time: 5 minutes (in the mouth) approx	consumables	120	672000
243	Soyabean casein digest ager	ISO certified, WHO/GMP confirmed to and CE notified	consumables	101	281830
244	Spirometer Incentive 3 ball	Should have three color-coded balls in each chamber to provide a visual incentive for patients. Air should enter each chamber and raise the balls depending on the flow of air inhaled per second. Available in lung capacity : 600ml/sec, 900 ml/sec and 1,200ml/sec. The whole set should include EVA tubing, mouth piece 13M, tri-ball incentive spirometer. Single-patient use to minimize the risk of cross contamination. 12 mm OD connector Volume should be 600 mL/sec, 900 mL/sec, 1200 mL/sec Material: Should be AS for chamber, PE for ball and mouthpiece, and EVA for tubing	consumables	4860	438955
245	Sputum Container	Should be made of transparent plastic (polypropylene) with a screw cap. Should have optimal mechanical, thermal and chemical stability, be leak-proof, unbreakable and resistant against chemical influences. Standard size	consumables	2150	3922
246	Sputum Container ISO certified, WHO/GMP confirmed to / CE notified	Sputum Container ISO certified, WHO/GMP confirmed to / CE notified	consumables	11556	18767
247	stainless steel kidney tray large	1) It should be made of stainless steel 2)it should be rust free 3)It should be ideal for clinical use. 4) It should be ADA/ISO/ISI/CE/USFDA certified from notified body.	consumables	63	9828
248	Stainless steel kidney tray medium	1) It should be made of stainless steel 2)it should be rust free 3)It should be ideal for clinical use. 4) It should be ADA/ISO/ISI/CE/USFDA certified from notified body.	consumables	82	8528

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
249	stainless steel kidney tray small	1) it should be made of stainless steel 2)it should be rust free 3)It should be ideal for clinical use. 4)It should be ADA/ISO/ISI/CE/USFDA certified from notified body.	consumables	120	9600
250	Stainless steel surgical dressing drum large	1) It should be made of stainless steel 2)it should be rust free 3)It should be ideal for clinical use. 4) It should be ADA/ISO/ISI/CE/USFDA certified from notified body.	consumables	120	432000
251	Stainless steel surgical dressing drum medium	1) It should be made of stainless steel 2)it should be rust free 3)It should be ideal for clinical use. 4)It should be ADA/ISO/ISI/CE/USFDA certified from notified body.	consumables	120	336000
252	Stainless steel surgical dressing drum small	1) It should be made of stainless steel 2)it should be rust free 3)It should be ideal for clinical use. 4)It should be ADA/ISO/ISI/CE/USFDA certified from notified body.	consumables	120	240000
253	sterilized disposable wooden tongue depressor	1) it should be smooth ,clean and sturdy and should be taste free and odor free. 2) it should be sterile and packaed in a sterile condition. 3) they should be suitable for oral examination. 4) it should be available in a pack having 100 wooden depressors 5)it should have polished smooth edges.6) it should be 6 inch long.	consumables	266	95760
254	Sticky wax	A highly tenacious dental wax used to hold material or pieces in position,Should flow easily, Must set firmly. Rigid and brittle at room temperature Should melt when heated. Should give a clean burnout. Available in stick form. (10 sticks/pkt)Shelf life minimum 2 years	consumables	120	19200
255	Straight probe	1. It should be made of a good quality rust free stainless steel, 2. the probe should have a straight and sharp end	consumables	125	4000
256	suction tips	1. Should be made up of flexible plastic material for dental use. 2. Should have Smooth and fixed tips are securely fastened 3. It should be each box should contain 100 suction tips 4. Suction tip should have dimentions of 10 to 15 cm in length and 0.7mm to 1 cm diameter.	consumables	120	24000
257	Suture needle Straight Size	Surgical Needle Straight Cutting Needle size 6 to 10 6pcs in one Pkt	consumables	500	80000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
	6 to 10				
258	Temporary luting cement	Temporary luting cement for sealing of crowns and bridges Available in two tubes form- base and catalyst in contrasting colours. Should provide homogenous, smooth mix Base approx. 40 g and catalyst approx. 14 gm	consumables	120	192000
259	Titanium mini plates 2 mm , L shaped 90 degree , right and left	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	86400
260	Titanium mini plates 2 mm four hole with bar	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	86400
261	Titanium mini plates 2 mm, H shaped plates,	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	96000
262	Titanium mini plates 2 mm, two hole with bar	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	57600
263	Titanium mini plates 2 mm,double X shaped plates,	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	96000
264	Titanium mini plates1. mm , L shaped 90 degree , right and left	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	86400
265	Titanium mini plates1.2 mm, T shaped plates, small and big	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	67200
266	Titanium mini plates1.2 mm,double Y shaped plates, small and big	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	96000
267	Titanium mini plates1.5 mm six hole without bar	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	96000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
268	Titanium mini plates 1.5 mm, Z shaped plates,	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	96000
269	Titanium mini plates 1.5 mm, double Y shaped plates, small and big	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	96000
270	Titanium mini plates 1.5 mm, H shaped plates,	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	96000
271	Titanium orbita plates with bar 2 mm, 6 holes	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	288000
272	Titanium orbita plates with bar 2 mm, 8 holes	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	288000
273	Titanium screw 1.2 mm , size 4mm	Profile diameter 1.2 mm , core diameter 0.8 mm, titanium grade V	consumables	120	76800
274	Titanium screw 1.2 mm , size 6mm	Profile diameter 1.2 mm , core diameter 0.8 mm, titanium grade V	consumables	120	76800
275	Titanium screw 1.5 mm , size 4 mm	Profile diameter 1.5 mm , core diameter 1.0 mm, titanium grade V	consumables	120	76800
276	Titanium screw 2.0 mm , size 5 mm	profile diameter 2.0 mm , core diameter 1.5 mm, titanium grade V	consumables	120	96000
277	Titanium twisted plates, right and left 5 holes	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	288000
278	Titanium wire mesh plates, 1.5 mm	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product	consumables	120	192000
279	Transmetal Carbide bur	should be extra-fine cross cut shape , should be able to remove old amalgam filling, should be able to cut through porcelain fused to metal crowns, should be non corrosive, auto clavable	consumables	120	48000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
280	Triple trays	Disposable tray for functions of impression, counter impression and bite registration simultaneously. It must have ultra-fine net for precise bite impression. The Nylon net must be strong enough to avoid any tear during and after impression. It is very useful in close-bite registration cases. A set must consist of either Anterior, Posterior or Full arch.	consumables	120	96000
281	Tryptone water	ISO certified, WHO/GMP confirmed to and CE notified	consumables	120	288000
282	Tubal Ring	Tubal Ring	consumables	48000	1344000
283	Tubes for sputum collection	50 ml screw capped conical graduated and sterile	consumables	73	35647
284	Tubes for sputum collection 50ml	50 ml screw capped conical graduated and sterile.	consumables	2120	8480
285	Turpentine Oil I.P 5 litre can	Turpentine Oil I.P 5 litre can	consumables	120	86400
286	Turpentine Oil I.P 500ml	Turpentine Oil I.P 500ml	consumables	210	16800
287	Tweezer	1. It should be made of a good quality rust free stainless steel material	consumables	100	6400
288	Two-tone chewable disclosing tablets	1) It should Differentiate between old and new plaque; old plaque 2) Each pack contains multiple tablets (80/100)with full instructions 3) Available in fruit-it should be available in flavors.	consumables	120	1824
289	Ultrasonic Scaler Tip	high quality stainless steel compatible with major brand piezo and ultrasonic scaler, autoclavable and noncorrosive	consumables	120	576000
290	Umbilical Catheter Size 3.5 FR	Umbilical Catheter Size 3.5 FR	consumables	120	88800
291	Urine Pot plastic (M/F)	Urine Pot plastic (M/F) Female urine container size: 800ml color: whiteAND Male urine container size: 1000ml color: white	consumables	355	9841

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
292	Vacuum formed thermoplastic sheets	A clear transparent, hard, thermoplastic material for adaptation on dental models. It must be abrasion resistant and excellent form stability. The material must be highly biocompatible. Must be available in various thicknesses for bleaching trays, splints and crown and bridge work. A packet containing sheets of one thickness (either 0.5, 1, 1.5, 2, 3, 4 mm) amounting to total thickness of about 25 mm.	consumables	120	384000
293	Wedges wooden	1) It should be Small, Tapering, Triangular pieces of wood about 1/2 inch in length. 2) It should be each box should contain 50 wedges	consumables	120	51840
294	Weighting machine	Adult scale upto 200 kg capacity with 500gm marking with non slippery rubber mat, certified by weights and measures dept.	consumables	182	165984
295	White Petroleum Jelly 1kg conatiner	White Petroleum Jelly 1kg conatiner	consumables	301	30943
296	White soft paraffin 500gms	white soft paraffin 500gms	consumables	120	46080
297	White soft Paraffin 50gm	White soft Paraffin 50 gm	consumables	350	4480
298	Wire 0.016 australian	It should have good strength.It should be corrosion resistant. it should be heat treated and cold drawn to proper diameter.it should be of international quality.it should have good resilience.special plus type	consumables	120	192000
299	Wire 0.018 australian	It should have good strength.It should be corrosion resistant. it should be heat treated and cold drawn to proper diameter.it should be of international quality.it should have good resilience.special plus type	consumables	120	192000
300	Wire 0.032inch TMA round	It should be made up of titanium molybdenum. With smooth finish ,high elasticity excellent formabilty low friction nickel free.It should have good deflection property.It should be shape specified.	consumables	120	192000
301	Wire Nickel titanium 16X22 upper	It should be good quality niti wire , it should be ovoid arch arch form. It should be of uniform size through out.	consumables	120	57600

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
	& Lower arch				
302	Wire Stainless steel orthodontic 19 to 24 gauge	1) The wire should have Springiness, Stiffness, ability to bend without fracturing corrosion resistance and an ability to be simply joined by soldering or welding. 2) Circular in cross section. 1 packet should have length of 1 meter	consumables	120	144000
303	Wire TMA 16 x 22? upper and lower arch	It should have good strength.It should be corrosion resistant. it should be heat treated and cold drawn to proper diameter.It should be of international quality.it should have good resilience	consumables	120	336000
304	Wolf bottle	Wolf bottle	consumables	600	74880
305	zero degree teeth set (non anatomic)	1) Shape (Mould) in assortment of shapes i.e. square avoid & tapering. The teeth must bear resemblances to natural teeth & be free of porosity. 2) Shade ? Approx. of 6 shades from light medium to dark 3) Size ? Approx. 8 sizes ranging from small to large in each of the size & shade 4) General-The Teeth should bear resemblance to natural teeth and free of porosity. 5)teeth should not distort during or after processing, One complete set from each shade & shape & size for inspection should be supplied as sample with mould chart. 6) One set of 16 teeth (8 upper posterior and 8 lower posterior)	consumables	160	30720
306	Zinc Oxide	1) Should be specific for the dental use meeting ISO/ISI standard. 2) Should contain accelerator for rapid setting. 3) Should contain recommended percentages of other ingredients like magnesium oxide and others as per recommendations. 4) Should be usable as a temporary restorative material and also as root canal sealer. 5) It should be pure free of any impurities. 6) it should be 100 gms powder	consumables	120	19200

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
307	Zinc Phosphate Cement	ADA specification No 96 Available as Powder -90 gm Liquid - 39 ml Composition : Powder Zinc oxide, Magnesium oxide, Other oxides like Bismuth Trioxide, Calcium oxide, barium oxide) silica Liquid :Phosphoric acid , water , Aluminium Phosphate or sometimes zinc phosphate Powder liquid ratio 3:5:1 Mixing time 60-90 sec. Setting time is between 2.5 to 8 minutes	consumables	120	172800
308	Micro cuvettes For Digital haemoglobino meter (*FOR EXISTING MACHINES ONLY)	Microcuvettes / Strips Compatible with Digital Haemoglobinometer (Hemocue HB 201) contains CE Mark	consumables	1063830	28000006
309	Microslides	Glass Microscope Slides : 76 cm X 26 mm X 1.3 mm, plain, on epack of 50 slides	consumables	66627	2337808
310	Sputum Container	The sputum container should be made of special medical grade polypropylene, thin plastic translucent with diameter 4 cm, Height 5 cm and capacity 40 ml. It should have screwball cap and should be air tight and leak proof and have injection molding. Weight is not to be considered.	consumables	4185000	5591160
311	Steril Disp. Umbilical Cord Clamp	Disposable Umbilical Cord Clamp are made from Non toxic Medical grade polymer. 1. Material: PVC, PE, PP, ABS etc. 2. Sterilize by ethylene oxide gas 3. Supplied sterile in individual peelable polybag.	consumables	1806000	1863792
312	Transducer Protector individually packed and sterile,	Transducer Protector individually packed and sterile,	consumables	19000	107312

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
313	Urine Collection Bag	Urine Collection Bag - Disposable sterile with accurate calibration Non Toxic, Tube length 42"Manufactured from clinical grade PVC sheet. Super Smooth kink resistant tube, made from medical grade PVC. with connector. Non return valve and drainage outlet. Provided with hanging & carrying system. Capacity : 2000 ml. Approx. Sterile / Disposable / Individual packed.	consumables	25000	272600
314	Acid Concentrate Can	Acid concentrates for hemodialysis The electrolyte concentrations shown below are related to the ready to use dialysis solution in dilution: 1 liter of acid concentrate + 1,225 liter of bicarbonate concentrate 8,4% + 32,775 liters of purified water.	Consumables	15000	14400000
315	Airotor Diamond points (Commonly called as burs)	Airotor Diamond points (Commonly called as burs)	Consumables	16000	4480000
316	Airrotor diamond burs	1 Should be available in various sizes and shapes as per the requirements like round, straight fissure, tapered fissure, inverted cone, safe ended, disc or wheel shaped. 2. Should be available in friction type design. 3. Shank and neck should be made of steel of good quality. 4. Diamond particles embedded on the working end in the matrix which is durable. 5. Should not rust and corrode in common disinfecting solutions. 6. Should be autoclavable. 7. Should not disastrously break during use. 8. Should not dull after short duration of use.	Consumables	10000	1760000
317	Airway Plastic	Soft Plastic Nasal Airways are constructed of a soft, latex free plastic. The interior of the airway has striations for easy suctioning and insertion. Latex free- Small/Medium/Large size	consumables	90000	3240000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
318	Anti Infective Bed Cover Mat (AIBCM)	Anti Infective Bed Cover Mat (AIBCM) should be made of; Fabric:Denier 450 - 550 range, Core Material (Film):Polyvinylchloride, Grid (Construction):18*9 +/- 10% range, Size: 2mtr*1mtr (200cms x 100cms), Yarn:High Tenacity Polyester, Weight in GSM:320-350 GSM range, Adhesion: Non-separable, Color: Light color such as myrtle green etc.,	Consumables	96000	84480000
319	Autopsy Kit Clinical	Disposable Sterile with double packing. E.T.O. sterilized with sterile indicator. Each kit should contain: 1.Disposable Head cap 1 piece 2.Disposable Face Mask 1piece 3.Disposable Protective P.V.C. inner Apron without sleeves 1 piece 4.Disposable Surgical Gown with full sleeves & with elastic cuff non woven fabric-water repellent 1 piece 5. Disposable P.V.C. Hand gloves 1 pair free Sizes 6. Disposable P.V.C. Shoe cover 1 pair 7. Disposable Safety Goggles 1 pc 8.Plastic cover to wrap the body 1 pc Size: 80 cm x 120 cm.	Consumables	10000	48000000
320	Billigoggles	Yet to be determined	Consumables	15000	36000000
321	Black Goggles for Cata. Patients	Black Goggles for Cata. Patients Description 1.Frame Molding 2.Glass Bluish Gray 3.Cataract cuff goggle with side protection frame should be smooth and non poking	Consumables	36000	3744000
322	Bone Cement	Yet to be determined	Consumables	26000	83200000
323	Bone Cutters	Yet to be determined	Consumables	28000	224000000
324	Bonewax	4 mm & 6 mm Strip	Consumables	47000	41360000
325	Bovine serum Albumin 22%	storage temperature 2 to 8 degree 22% bovine albumin, 1 x 10 ml	Consumables	10000	80000000
326	C.T, Scan Films 14 X 17(Depending on the Camera)	C.T,Scan Films. (Wet & Dry) C.T. Scan Films. (Wet & Dry) (1) C T & M I R Films for Dry Camera Should have compatibility of Films to available camera (14" x 17") Should be quoted separately for Each camera. (2) C T & M I R Films for Wet Camera Should have compatibility of Films to available camera (14" x 17") Should be quoted separately for Each camera	Consumables	14000	1904000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
327	CBNaat Cartridges	Specifications for Xpert MTB/RIF cartridges: ➤ Cartridges should be self-contained and single use. Sample processing, amplification and detection should be carried out within this self-contained device ➤ Cartridge should be made up of Polypropylene material ➤ To avoid false negative results, cartridges should incorporate sample processing and PCR reaction controls. ➤ To avoid false negative results, cartridges should be monitored by the instrument for pressure and other error conditions. ➤ Kit should contain sufficient reagents to process 10 patient of quality control specimen. ➤ Reaction kit should contain all necessary reagents for real time PCR reaction (Primers, Probes, KCl, MgCl₂, HEPES, BSA (bovine serum albumin), Polymerase, dNTPs etc) ➤ Should have proper positive and negative control <u>Cartridge packaging:</u> Description: 5 x 10-pack pouched cartridges/one box of 50 Sample Reagent bottles/5 packs of 12x2ml pipettes/one CD-ROM. Each kit contains 5 pouches. Each pouch contains 10 cartridges. PHYSICAL DIMENSIONS/WEIGHT: 31 cm W x 19 cm H x 27 cm D/3.6 Kg.	Consumables	8500	8160000
328	Centrifuge Tube 50 ml (Falcon Tube)	Centrifuge Tube - 50 ml, made of propylene, conical shape bottom with tight screw cap, walls thick enough to allow centrifugation (maximum RCF) up to 5000 *g, solvent resistant, with white labelling area and black graduations, flat PE caps which can close with quick 3-4 turns, allowing one handed operation. Tubes should be radiation sterilized to ensure non-cytotoxicity, sold in packs of 500 tubes.	Consumables	46000	1104000
329	Cobbs Elevator	Yet to be determined	Consumables	90000	360000000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
330	Composite bonding agent 7th Generation	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) 1 Bottle should have minimum 3 - 5 ml bonding agent	Consumables	16000	38400000
331	Compressed Hydroxylating Polivinyl Acetate Nasal Tampon	Compressed Hydroxylating Polivinyl Acetate Nasal Tampon, Size 4.5 cm * 1.5 cm * 2 cm	Consumables	14000	5600000
332	CPAP Tubing	Disposable Silicone Tubings for CPAP	Consumables	36000	17280000
333	Dental X-ray films; IOPA xray film Pediatric, adult; Occlusal Xray film	Dental X-Ray films Dental X-ray Film Internal Peripheral (1 1/4" x 1 6/8 ") Packet of 150 films	Consumables	10000	640000
334	Device for Inhalation of Dry Powder Capsule	Each Act-O-Vial Contains ; (lower powder compartment) Methylprednisolone sodium succinate.... USP 500mg (eq to Methylprednisolone 500 mg) Monobasic sodium phosphate anhydrous.... USP (buffering agent) Dibasic Sodium Phosphate anhydrous USP (buffering agent) Sodium Hydroxide.... USP NF (for pH adjustment) (Upper Diluent Compartment) Sterile water for injection..... USP when reconstituted mix total volume 4 ml	Consumables	37000	22200000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
335	Diaper for New Born	Yet to be determined	Consumables	80000	960000
336	Dipsosable Sterile Drape	Yet to be determined	Consumables	47000	1880000
337	Disc Holding Forceps Straight & Angled	Yet to be determined	Consumables	50000	400000000
338	Disp. Negative suction drain set	Disposable Negative suction drain set comprising Bellow unit with connector. Connecting tube with clamp and "Y" connector. Curved needle with matching catheter. Spare perforated catheter. No.10, 12 G	Consumables	96000	61440000
339	Dural Retractors Straight & Angled	Yet to be determined	Consumables	89000	712000000
340	ECG Roll	Wax coated electro cardi graphic paper size 50 mm x 30mm or 60mmx20mm (superior quality)	Consumables	50000	2800000
341	ECG/EEG Gel	ECG Jelly Tube Electrode contact paste suitable for ECG, EEG, ECT Jar of 200 gm	Consumables	50000	1400000
342	Endobuttons with or without loop	Yet to be determined	Consumables	50000	280000000
343	External Fixator 3.5, 4.5	Yet to be determined	Consumables	56000	1568000000
344	External Fixator Imizalar	Yet to be determined	Consumables	50000	1400000000
345	Formaline Chamber 10"	Formalin chamber is sterilization system is lightweight transparent, corrosion proof and made from special grade of polymer with two sliding trays to ensure 100% sterilization of your surgical instruments.10",	Consumables	50000	600000000
346	Formaline Chamber 14"	Formalin chamber is sterilization system is lightweight transparent, corrosion proof and made from special grade of polymer with two sliding trays to ensure 100% sterilization of your surgical instruments. 14",	Consumables	50000	600000000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
347	Formaline Chamber 20"	Formalin chamber is sterilization system is lightweight transparent, corrosion proof and made from special grade of polymer with two sliding trays to ensure 100% sterilization of your surgical instruments., 20"	Consumables	60000	67200000
348	Formaline Chamber 23"	Formalin chamber is sterilization system is lightweight transparent, corrosion proof and made from special grade of polymer with two sliding trays to ensure 100% sterilization of your surgical instruments.23"	Consumables	40000	41600000
349	Formaline Chamber 26"	Formalin chamber is sterilization system is lightweight transparent, corrosion proof and made from special grade of polymer with two sliding trays to ensure 100% sterilization of your surgical instruments.26"	Consumables	45000	57600000
350	Glass Ionomer cement (Luting/ Type 1)	1) Maximum particle size should be 15um 2) Powder/liquid ratio in the range of 1.8/1.0 Mixing time 3) (approx..) ? 20 sec working time (approx..) ? 2 min. (230 c) (from start of mixing) Final finishing commencing time 4 - 5 min (after seating the restoration 4) Film thickness of luting cement should be <20 um 5) Material should be sreontium based gic with good track record of its use in dental care hospitals 6) ISO/ISI/CE/ADA specifications should be met 7) It should be 15 gms powder and 10 ml liquid.	Consumables	50000	80000000
351	HUB Cutter	HUB Cutter Safety portable Box of Disposal of Hypodermic Needles. Container with WHITE & BLUE casing. Puncture Resistant Hard and can be Locked. Size - Hub Cutter with 100 Hypodermic Needle Collection Capacity.	Consumables	10000	15200000
352	Hydrocollid Dressing Patches	Yet to be determined	Consumables	50000	280000000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
353	Impression Compound	1) It should be available in cake form 2) It should be fuse in hot water and form homogeneous mass 3) Working temperature: 55oC ? 65oC 4) Manipulation time(Kneading time): 1-3 minutes 5) Fusion temperature:43.5oC 6) Flow should be less than 2% at 37oC and between 75-85% at 45oC 7) 5 cakes per packet with 200 grms in weight	Consumables	110000	35200000
354	Interference Screws - Titanium	Yet to be determined	Consumables	50000	400000000
355	K- Wires 1, 1.5, 2, 2.5, 3 mm	Yet to be determined	Consumables	5000	20000000
356	Kangaru Mother Care Bag	Yet to be determined	Consumables	65000	20800000
357	Light Cure Resin Restorative material (Kit)	1) Light cure composite resin kit should consist of ? Various shades of the material white to yellowish and light to darker ? Preferably have Shade guide provided in the kit? Etchant gel ? Phosphoric acid 35-37% 3 ml ? Bonding agent packaged in a leak proof and good quality bottle ? 1 unit of Mixing pad should have 100 pad sheets) Applicator tips Requirements of the material: ? Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 3) Material should have optical properties like tooth and have life like appearance after making restorations. 4) Should have good Polishibility 5) The material should be freshly manufactured and should have a sufficient shelf life. 6) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without andue polymerization & SHOULD HAVE CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT 7) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user 8) Material should have optimum flow, good handling characteristics and shelf life of	Consumables	50000	200000000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
		atleast three years from supply. 9) Material should achieve clinically and biological properties after curing. 10) Composite material should contain prepolymerised filter partical for more strength kit contains: 5 composite syringes of 4-5gm each, 1 denting bonding agent 5ml, etchant 3-5ml, applicator tip s 100 sticks, mixing pad			
358	Makintosh Double Colour Water Proof Rubber 110cm Roll of 30 mtr	Rubber Makintosh double colour Rubber sheeting, High Polymer 55% & above , Colour (Green,Blue, Red), Thickness-0.4mm, Width-110cm Roll of 30 mtr I.S.certified, Double colour,water proof	Consumables	50000	14400000
359	Makintosh Double Colour Water Proof Rubber 90cm Roll of 30 mtr	Rubber Makintosh double colour Rubber sheeting, High Polymer 55% & above , Colour (Green,Blue, Red), Thickness-0.4mm, Width-90cm Roll of 30 mtr I.S.certified, Double colour,water proof	Consumables	45000	13320000
360	Mamography Films	Provides a 100-speed system with Min-R screens and a 150-speed system with Min-R 2000 screens Min-RS 18 x24 cm Mammography film & Min-RS 24x30 cm Mammography Film	Consumables	40000	156800000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
361	Microdrip I.V. sets with Volumetric chamber	Micro I.V. Infusion set for pediatric use for Infusion Fluid administration with volumetric chamber 60 drops per 1 ml. Non Toxic, Sterilized, Pyrogen Free, ISI Mark, Non vented needle Size 23 G x 1.5 inch, Self sealing latex tube for medication & easy flushing. Special beveled extra strong spike. Disc filter in drip chamber facilitate rapid adjustment fluid need of 23 G for smooth slow. Highly kink resistant tubing roller clamp for flow meter. Packed in one plastic bag with Label containing product name .Mfg.date, Exp.date mfg.by.	Consumables	50000	1800000
362	Mineral Trioxide Aggregate	1) set in 45 minutes 2) bioactive and radiopaque. 3) smooth consistency. 4) should contain gel and solution for mixing 5) 1GM POWDER	Consumables	40000	48000000
363	Modular Bipolar Prosthesis (37 to 53)	Yet to be determined	Consumables	50000	1200000000
364	Mucus extractor for neonatal deele catheter	Transperant Tube mucus extractor for neonatal & infants. Sterile Pyrogen Free, Non Toxic The capacity of the Mucus Extractor is 25ml and a spare cap is also provided for the sealing. Soft and kink resistant tubing with frosted surface A traumatic, soft rounded, beveled open tip. Capacity: 25 ml.	Consumables	70000	560000
365	Nasal Splint (External)	Yet to be determined	Consumables	60000	52800000
366	Nerve Root Retractors	Yet to be determined	Consumables	40000	64000000
367	Nitrous Oxide Refilling	Nitrous Oxide Refilling	Consumables	60000	240000000
368	Non DHEP IV Set	60" Pressure resistant IV Administration Set, Vented Spike 20 Drops/ml with Drip Chamber 60mm / Fluid Filter -15 micron/Flow Regulator/Male Luer Lock Adaptor with Cap/Back Check Valve/3.0mm x 4.1mm Tubing-Length 150cm/EO Sterile, DEHP and PVC Free	Consumables	50000	3600000
369	Oxygen Refilling	Oxygen Refilling	Consumables	60000	16800000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
370	Penta Blood Bags	Penta Blood Bag System with one PRIMARY Bag (containing Anticoagulant Solution) and four SATELLITE Bags Useful for paediatric and/or multiple transfusions Prevents wastage as one unit of blood can be given to 5 different recipients Reduces the donor related risks since the donor is the same Optional satellite bag of TOTM sheet for platelet storage upto 5 days Disposable, sterile and non-pyrogenic Sealed aluminium foil packing	Consumables	26000	8320000
371	Peritoneal Dialysis Solution IP 1 litre	Peritoneal dialysis solution IP 1 litre	Consumables	30000	9600000
372	Personal Protection Equipment Kit(PPE Kit) Medium size XL	PPE Kit- (Personal protection equipment kit)- PPE kit should include the following safety gears: 1. Personal Protective coverall/Gown (Garment)- 1 no 2. Shoe Cover Upto knee ? 1 Pair 3. Eye Goggle ? 1 no 4. Face shield ? 1 no 5. N-95 Mask ? 1 no. NIOSH Certified 6. Nitrile Gloves ? 1 Pair 7. Bio Degradable Disposable Waste collection bag ? 1 no All the above safety gears shall be packed in a paper poly pouch. 1. N-95 Mask- One Unit ? Shape that will not collapse easily. ? NIOSH certified. ? High Filtration efficiency. ? Fluid Resistance, minimum 80mm Hg pressure based on ASTM G 1862, ISO 22609 or Equivalent. ? Quality compliant with standard for particulate respirator that can be worn with full face shield. ? Quality compliant with standards of surgical N-95 Mask 2. Face Shield ? One Unit ? Made of clear plastic and provides good visibility. ? Adjustable band to allow good fit around the head and snug fit against the forehead. ? Fog ? Resistant (preferable). ? Reusable (made of material that can be cleaned and disinfected) or disposable. ? Quality compliance with standards : EU Standard directive 86/686/EEC, EN 166/2002, ANSI/ISEA 287.1-2010 or equivalent. 3. Goggle ? One Unit ? With transparent Glass/plastic zero power, good fitting, covered from all sides with elastic band or adjustable holder. ? Good	Consumables	60000	33600000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
		seal with the skin of the face ? Flexible Frame to easily fit all face contours without too much pressure. ? Covers the eyes & surrounding areas and accommodates for prescription glasses. ? Fog and scratch resistance ? Indirect venting to reduce fogging 4. Nitrile Gloves ? One Pair ? Nitrite ? Non-sterile, powder free. ? Outer gloves preferably reach fore-arm (min 225mm total length). Thickness 5.5 MIL and size (M/L) ? Quality compliance with below standard or equivalent. ? EU standard directive 93/42/EEC class I EN 455 or equivalent. ? EU standard directive 89/686/EEC category III EN 374 or equivalent. ? ANSI/SEA 105-2011 or equivalent. ASTM D6319-10 or equivalent. 5. Disposable Coverall ? Along with One hood Drape OR One head Cover. (it should be 70 GSM and above ? Non Laminated ? One Unit ? Single use. ? The Fabric Garment/breathable Non Laminated Fabric ? Light colours to better detect possible contamination. ? Coverall Shall be designed to be Universal Fit. Soft Elastic to be fitted around front of hood, Wrist and ankles ? Coverall approved by SITRA (South Indian Textile Research Association), DRDE (Defence Research Development Establishment) and have valid Unique Covid Code (UCC). 6. Pairs of Shoe cover made of same fabric as of Coverall upto below Knee ? one pair 7. Bio Degradable Disposable Waste collection bag ? One Unit			
373	PICC Lines	PICC Line Number 23, 24	Consumables	45000	432000000
374	PPE Kit (Large Size)	Attached Separately	Consumables	25000	18000000
375	PPE Kit (Medium Size)	Attached Separately	Consumables	15000	10800000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
376	Pre formed space maintainer	All in one Starter Kit with Accessories : 1) 37 universal bands with tube 2) 40 space components: straight,curved, distal shoe 3) 1 Intra oral crimper 4) 1 Intra oral cutter Specifications kit contains : Sizes per band: 24,24 1/2,25,25 1/2, 26???.41 1/2, 42 Type of band: Universal Types of components: Straight, Curved, distal shoe Number of bands per size: 1/size Number of components: 40 Shape of bands : Universal Type of cutter: Wire Cutter with carbon attachment Type of crimper: Intra oral crimper Number of cutter: 1 Number of crimper: 1	Consumables	45000	72000000
377	Pricking Lancet	Pricking Lancet - Disposable/ Single-use , Materials: stainless steel ,The disposable lancet is sterilized by gamma radiation or E T O sterilized . The lancet comes with smooth tribevel points. Quality in accordance with the CE and ISO standards	Consumables	150200	841120
378	Pricking Lancet Needles	Disposable/Single use Pricking needle no. 24 The disposable lancet needle is sterilized by gamma radiation. Made by stainless steel CE & ISO cerificate	Consumables	120000	480000
379	Provisional crown material	1.Should be supplied in the form of cartridge and automixing tips (in a packet) as per ADA/ISO/ISI specification. 2.Should be aself-curing composite for the fabrication of temporary crownsand bridges,inlays,onlaysandveneers. 3. It should be free from methylmethacrylate and peroxides. 4. It should have a fillercontent:40-45wt.-% 5. Mixing time-30?45 sec 6.Setting time- 2?3min 7.Expiry should be atleast for 2 yrs from the date of supply.	Consumables	50000	100000000
380	Replanisher 5 ltr jar	An aqueous photographic silver halide developer/replenisher solution particularly useful in low throughput machine processing of exposed X-ray silver halide film contains 1-phenyl-3-pyrazolidone, hydroquinone, bromide ion, organic antifoggant and film speed restrainer and is buffered with an alkaline material to a pH of 10 ± 0.3, replenisher having a lower pH than developer.- 5 Ltr	Consumables	10000	12000000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
381	Reusable Capacitive Coupling Current Quality Monitoring Electrode	1. Reusable Capacitive Coupling Current Quality Monitoring Electrode made of Viscoelastic Polymer, 2. Should have Current limiting nature & hence eliminate Patient pad site burns, 3. Should be based on Capacitive coupling principle, 4. Should made of aktom polymers, 5. Can be used for all patient's weight >350 grams, 6. Should be Radiolucent & latex free, no adhesive related irritation to patient skin. 7. Should be US FDA approved, 8. Should be compatible to any Electrosurgical generator, 9. Size: 36" L * 20"W * 1/8" thickness.	Consumables	45000	252000000
382	RF/RA	Based on latex agglutination For the detection and semiquantitation of Anti-Streptolysin O in serum Sensitivity ? 200 IU/ml Specificity Purified Streptolysin O specific for Anti-Streptolysin O antibodies Latex reagent Uniform suspension of polystyrene latex particles coated with stabilized Streptolysin O GMP SCH M PRODUCT WISE AND TESTING FACILITY WISE, 50 test/kit	Consumables	50000	200000000
383	Ring and Cup Curretes (Angled)	Yet to be determined	Consumables	25000	100000000
384	Rod Cutter (Jumbo Cutter)	Yet to be determined	Consumables	15000	180000000
385	Roungers 1, 2, 3, 4 mm	Yet to be determined	Consumables	89000	640800000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)															
386	Sanitary Napkin	Description : 1. Covering – The covering of the absorbent filler shall be made of good quality perforated Polyurethrene film sleeve which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. This shall be made of a product that is non-allergenic and bio – compatible. The sanitary napkins shall have a non-absorbent barrier on one side which shall have an identifying mark indicating clearly the side of the barrier. 2. Absorbent Filler – The filler material shall consist of cellulose pulp (either based on wood or paper or poly-acrylate gels). This shall be free from lumps, oil spots, dirt or foreign material etc. 3. Back Strip – A back strip for sticking the sanitary napkin onto the underwear should be there using good quality adhesive material. 4. Absorbency – The sanitary napkin should be able to absorb not less than 30 ml of normal saline (I.P.) @ 15 ml/minute. 5. Size – The size of absorbent section of the Sanitary Napkin shall be as follows: <table><tr><td>Pad Length</td><td>Core Length</td><td>Overall</td></tr><tr><td>Width</td><td>Core Width</td><td>Pad</td></tr><tr><td>Thickness</td><td></td><td></td></tr><tr><td>235+_10mm</td><td>210+_10mm</td><td></td></tr><tr><td>155+_5mm</td><td>65+_5mm</td><td>7-9mm</td></tr></table> The thickness shall be measured by stacking 10 complete pads and measuring the stack height. The average thickness for the 10 pads shall be used as the pad thickness. 6. Weight – The weight of one full sanitary napkin shall not be more than 7.5 – 9 grams. 7. pH Value – The sanitary napkin shall be free from acids and alkali and the pH of the absorbent material shall be 6 – 8.5 when tested by the method given in IS : 1390 – 1961 (methods for determination of pH value of aqueous extracts of textile materials.)	Pad Length	Core Length	Overall	Width	Core Width	Pad	Thickness			235+_10mm	210+_10mm		155+_5mm	65+_5mm	7-9mm	Consumables	200000	2400000
Pad Length	Core Length	Overall																		
Width	Core Width	Pad																		
Thickness																				
235+_10mm	210+_10mm																			
155+_5mm	65+_5mm	7-9mm																		

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
387	Silastic Band	EVE'S Ring is the first silicone band.(also known as the Silastic* band)A nonreactive rubber band that is used for closing Each fallopian tube as a procedure for female sterilization. The ring is placed around a 2-3 cm segment of fallopian tube, after being drawn inside a narrow cone-shaped applicator. Silastic medical graded silicon tubing seamless coil in 10 mtr length - 3 mm ID x 6 mm OD wall thickness 1 mm/ 4 mm ID x 6 mm OD 1.50 mm thickness/5.00 mm ID x 8 mm OD 1.50 mm thickness /6 mm ID x 9.00 mmOD 2.00 mm wall thickness / 6 mm ID x 10 mm OD x 2.00 mm wall thickness/ 9.00 mm ID x 12.00 mm OD 2.00 mm wall thickness / 9.00 mm ID x 14.0 mm OD 2.50 mm wall thickness10 mm ID 15.0 m OD 2.50 mm wal thickness	Consumables	56000	112000000
388	Silicon Oil	Silicon gel 15 gm	Consumables	15000	96000000
389	Sonography Jelly 250 gm	All purpose sonography jelly Contact jelly for ultra sonography - Bottle of 250 gm	Consumables	60000	2160000
390	Sterile Plastic Sheets (Adult)	Yet to be determined	Consumables	96000	1382400
391	Strips / Cuvettes for Haemoglobin ometer	Microcuvettes/ Strips Compatible with Digital Haemoglobinometer contains CE Mark	Consumables	50000	1400000
392	Suction Catheter,Sterile	Designed for tracheal/bronchial suction.Smooth clear PVC ensures efficient mucus suction while minimizing potential damage to the delicate tracheal mucous membranes.Specially designed for easy passage through plastic endotracheal and tracheotomy tubes with minimum irritation to patient.Open distal end with elliptical eye for easy and efficient suction.Color-coded connector at proximal end for instant identification of catheter size.Available with thumb control device for positive control during suction.size -16- 50 cm length.	Consumables	120000	2880000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
393	Suction tube	Sterile PVC Latex free Tube for Suction.Reusable	Consumables	150000	26400000
394	Suture Ancil Preloader	Yet to be determined	Consumables	80000	640000000
395	Suturing needles Straight	Surgical Needle(Suture Needles) Straight Cutting Needles Different sizes 6 to 10 6 to 10 , IS Certification 6 No. in One Pkt	Consumables	80000	4864000
396	TBW Spools 16, 18, 20 G	Yet to be determined	Consumables	90000	108000000
397	Thermocol Boxes (2 Sample Size)	Thermocol Transport Box size for 2 falcon Tubes (In standing Position) One Piece Thermocol molded box in Dye Thickness of Box - 2.5 cm Outer Dimensions : Length 18.5 cm, Breadth - 13 cm, Height -12 cm (without Lid) & Height - 14 cm (with Lid) Inner Dimensions : Length 14.5 cm, Breadth - 8 cm, Height -12.5 cm (without Lid) & Height - 13 cm (with inner part of Lid)., No. of Gel Pack required 2 Weight of Fully Packed Consignment Box 400 gm	Consumables	80000	10240000
398	Thermocol Boxes (6 Sample Size)	Yet to be determined	Consumables	80000	12160000
399	Thermometer Clinical ISI mark Centigrade & Fahrenheit	1. Measuring Range 32.0~44.0Â°C(89.6~111.2Â°F); 2. Accuracy: +/-0.1Â°C(+/-0.2Â°F) for 35.0~42.0Â°C(95.0~107.6Â°F); +/- 0.2Â°C (+/-0.4Â°F) for remaining range at an ambient of 22Â°C(71.6Â°F); 3. Display: 3 1/2 LCD digits (Unit of display: 0.1Â°C/0.1Â°F); 4. Battery: LR41 alkaline; Battery life: About 250 hours or 1000 times; 5.Fever alarm; Waterproof case; 6.Memory for last measurement; 6. Low battery indication; Automatic power off; 7. Unit size: 120*20*12mm; Weight: 10.6g.	Consumables	80000	25600000
400	Thymol	IP	Consumables	80000	51200000
401	Topical anaesthetic spray	Should contain topical anaesthetic lignocaine in a metal packed bottle. Should follow ISO/ISI standards	Consumables	90000	57600000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
402	Topical Fluoride Gel	1) It should be a Fluoride Gel with A.P.F. formula. 2) it should Contain 1.23% fluoride ion at pH 3.5. 3) it should have Thixotropic property. 4) It should be available in any of the below mentioned Flavors: Chocolate-Vanilla (dye free)/ Orange (dye free)/mango (dye free) / Strawberry (dye free). 5) It should be available in a quantity of 500 ml /pack. 6)The Gel should have a property which can be dispensed into a fluoride tray. 7)It should be ADA/ISO/ISI/CE/USFDA certified from notified body.	Consumables	80000	44800000
403	Total Hip Prosthesis (Cemented & Uncemented)	Yet to be determined	Consumables	90000	10800000000
404	Total Knee Prosthesis (Modular)	Yet to be determined	Consumables	80000	9600000000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
405	TrueNaat Chips	Rapid Diagnostic MTB chip specifications- 1. MTB chip should function on Real Time Polymerase Chain Reaction (PCR) test for the detection of Mycobacterium tuberculosis (MTB) in human pulmonary (sputum/non sputum) specimen. 2. MTB Chip should be able to run on the Real Time micro PCR Analyzer Machine with samples extracted using Auto sample prep device. 3. The MTB chip should be stable at temperature (20-30°C) 4. The remaining shelf life at the time of delivery of items should not be less than 5/6th of the total assigned shelf life. 5. The MTB chip should comprise of a reaction well containing dried down MgCl₂ 6. The dried down PCR reagents should be available in microtube for performing Real Time PCR test for detection of MTB. 7. The target sequence for MTB should be part of the ribonucleoside-diphosphate reductase gene specific to the MTB complex. 8. The MTB chip should have flash memory to retain lot information and standard curve values for automatic quantitative determination 9. MTB chip should not be available for reuse. The reuse and post expiry use should be detectable by flash memory. 10. The time taken for the diagnostic test result should be lower or equal to one hour. 11. The MTB chip should be amenable for disinfection and disposable as per applicable Biomedical Waste Management Rules. The chip should not be infective after use. 12. The consumables should be provided in sufficient quantity to cover 10% wastage <u>The following contents should be provided with the MTB chip:-</u> A} AUTO Universal Cartridge Based Sample Prep Kit containing 1. Reagent pack containing ready to use buffers for automatic sample extraction 2. Cartridge pack with a fluidic cartridge	Consumables	90000	79200000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
		to move sample and buffers through a proprietary matrix for extracting nucleic acids from the sample. 3. Disposable transfer pipette B} AUTO MTB Sample pre-treatment pack containing acer 3. Disposable transfer pipette (graduated) C} Individually sealed pouches, each containing 1. MTB micro PCR chip. 2. Microtube with freeze dried PCR reagents 3. DNase & RNase free filter pipette tip. 4. Desiccant pouch.			
406	Tympanic Grommet	Yet to be determined	Consumables	90000	108000000
407	Umbilical Cannula	Umbilical Cannula Number 3, 4, 5	Consumables	80000	44800000
408	Urine Pot plastic (M/F)	Urine Pot plastic (M/F) Female urine container size: 800ml color: whiteAND Male urine container size: 1000ml color: white	Consumables	90000	2880000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
409	Vaginal pH Test Kit	Technical Specifications for Vaginal pH test kit having Test strips, mobile application for detection, Software System with Dashboard. Digital Image Colorimetry, Vaginal pH Level 4 – 7 pH, On-screen patients result on mobile screen & dashboard required, Sample material - Vaginal Fluid, Measuring Time - Less than 1 Minute, Auto / Self calibration, Sensitivity - More than 90%, Specificity - More than 90%. Memory to store data- Minimum 100000 test data with date, time, location and personnel details, Data auto transfer from application to backend server. Strip (working environment)- Disposable and stable at specified environmental condition (Working temperature 10-40°C, working humidity 5-95%), Strip (storage environment)- Should be stable at temperature 25-30°C and humidity of 5-95%, Intra-sample variation -Should be less than 5% & Accuracy -More than 95%. Certification & Standards - 1. ISO 13485 : 2016 for medical device, 2. CDSCO for medical device, Shelf Life: Two years.	Consumables	50000	12000000
410	X-ray Developer	X-ray developer powder to make solution of 13.5 Ltr	Consumables	60000	72000000
411	X-ray Developer	X-ray developer powder to make solution of 22.5 Ltr	Consumables	60000	72000000
412	X-ray films Blue Base 10 x 12	X-Ray films 10x12 Blue Sensitive Polyester base round corner X-ray film Blue sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 10 x 12	Consumables	80000	480000000
413	X-ray films Blue Base 12 x 15	X-Ray films 12x15 Blue Sensitive Polyester base round corner X-ray film Blue sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 12 x 15	Consumables	90000	540000000

Sr No.	Item Name	Specification	Sub Category	Proposed Quantity	Required Turnover (last 3 Years)
414	X-ray films Blue Base 6.5 x 8.5	X-Ray films 6.5 x 8.5 Blue Sensitive Polyester base round corner X-ray film Blue sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 6.5 x 8.5	Consumables	60000	360000000
415	X-ray films Blue Base 8 x 10	X-Ray films 8x10 Blue Sensitive Polyester base round corner X-ray film Blue sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 8 x 10	Consumables	50000	300000000
416	X-ray films Green base 10 x 12	X-Ray films 10x12 Green Sensitive Polyester base round corner X-ray film Green sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 10 x 12	Consumables	90000	648000000
417	X-ray films Green base 12 x 15	X-Ray films 12x15 Green Sensitive Polyester base round corner X-ray film Green sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 12 x 15	Consumables	60000	432000000
418	X-ray films Green base 6.5 x 8.5	X-Ray films 6.5 x 8.5 Green Sensitive Polyester base round corner X-ray film Green sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 6.5 x 8.5	Consumables	90000	648000000
419	X-ray films Green base 8 x 10	X-Ray films 8x10 Green Sensitive Polyester base round corner X-ray film Green sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 8 x 10	Consumables	80000	576000000
420	X-ray fixer 22.5 ltr, 13.5 ltr	X-ray fixer powder to make solution of 13.5 Ltr	Consumables	80000	160000000
421	X-ray fixer 22.5 ltr, 13.5 ltr	X-ray fixer powder to make solution of 22.5 Ltr	Consumables	60000	120000000

Appendix B: Checklist

(Mandatory Documents to be uploaded online in the EOI)

For the avoidance of any confusion, scanned copies of the above-mentioned documents shall be uploaded online on the Official Website on or prior to the EOI Due Date.

DOCUMENTS.			Page No. (as per the Bid)
1	Copy of certificate of incorporation/registration along with charter documents, and other registration documents according to the nature of entity such as MOA/Article of Association.	Yes / No	
2	PAN/TAN Card Details	Yes / No	
3	GST Registration certificate along with copy of the GST return of last quarter.	Yes / No	
4	Copy of manufacturing license with product list duly approved by the Licensing Authority/ State drug authority for each and every product quoted as per specification in the bid. The license must have been duly renewed up to date and the items quoted shall be clearly highlighted in the license/CDSCO Certificate (MD-5/MD-9/MD-15)	Yes / No	
5	QMS Certificate (13485:2016)	Yes / No	
6	Registration Certificate to sale, stop, Exhibit or offer for sale or distribute medical device (Form 42)	Yes / No	
7	Market Standing Certificate as issued by Central or State Licensing Authority for relevant period against quoted Drugs	Yes / No	
8	Non-Conviction Certificate issued by Licensing Authority.	Yes / No	
9	Annexure-1 (Technical Specifications and Compliance)	Yes / No	
11	Annexure-2 (Undertaking by EOI Respondent)	Yes / No	

12	Annexure-3 (Proforma for Production and Sale Statement)	Yes / No	
13	Annexure-4 (Details of Manufacturing Unit)	Yes / No	
14	Annexure-5 (Details of Items Quoted with Drug Code)	Yes / No	
15	Annexure-6 (Annual Turnover Statement for Three Years) along with Copies of Balance Sheet and Profit and Loss Accounts for last three years i.e., (2023-24, 2024-25 and 2025-26) certified by the Statutory Auditor or Chartered Accountant.	Yes / No	
16	Annexure-7 (Format for Power of Attorney for signing of EOI)	Yes / No	
17	Annexure-8 (Affidavit for Blacklisting)	Yes / No	
18	Annexure-9 (Litigation Affidavit)	Yes / No	
19	Annexure-10 (Mandate Form)	Yes / No	
20	Annexure -11 Delivery and Supply Capability	Yes / No	
21	Annexure-12 Respondent Classification	Yes / No	
22	Import Export Certificate (IEC Code) for Importer	Yes / No	
23	GST Registration certificate along with copy of the GST return of last quarter.	Yes / No	

Annexure 1: Technical Specifications and Compliance

Item Name: Following are the minimum requirements. Products offered must meet these parameters herein.

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generi c Name / Brand Name (only for Import er)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS M E / SSI	Re m ark s, if any
	A	B	C	D	E	F	G	H
1	Absorbent points 2 degree	ISO standardized paper points from number 15 onwards sizes (# 15-40) (#45-80) # 15, # 20, # 25 Should be supplied in assorted packets as per required numbers and quantity Should be colour coded should have 2 degree taper Should be length marked 120 points in 1 packet						
2	Absorbent points 4 degree	ISO standardized paper points from number 15 onwards sizes (# 15-40) # 15, # 20, # 25, #30, #35, #40 Should be supplied in assorted packets as per required numbers and quantity Should be colour coded should have 4 degree taper Should be length marked 60 points in 1 packet						
3	Absorbent points 6 degree	ISO standardized paper points from number 15 onwards sizes (# 15-40) #15, #20, #25, #30, #35, #40 Should be supplied in assorted packets as per required numbers and quantity Should be colour coded should have 6 degree taper Should be length marked 60 points in 1 packet						
4	Acrylic resin teeth (anatomic)	1) Shape (Mould) in assortment of shapes i.e. square avoid & tapering. 2) Shade ? Approx. of 6 shades from light medium to dark 3) Size ? Approx. 8 sizes ranging from small to large in each of the size & Shade 4) General ? The teeth must bear resemblances to natural teeth & be free of porosity. 5) The Teeth should have sufficient translucency to give						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		shining after processing. After processing for denture along with denture base resin should not show structural changes. One complete set from each shade & shape & size for inspection should be supplied as sample with mould chart. 6) Multilayer shading minimum 2 shading plus base material & cross linked with fluorescent effect & one set of 28 teeth						
5	ADDITION SILICONE	1.Should be supplied in the form of putty and light body as per ADA/ISO/ISI specification. 2.The light body should be available in the Automix cartridge and as a hand-mixed variant. 3. It should have high tear strength. 4. It should be dimensionally stable. 5. Expiry date should be at least for 2 yrs from date of supply. Putty should have Mixing time-30-50 sec Working time- 1:30min to 2min Setting time - 3-5min Light body should have Mixing time- 30sec to 1 min Working time-2min to 2:30 min Setting time - 3-5 min kit contains: Minimum quantity putty- 900 ml per kit and light body 100 ml per kit along with minimum 10 automixing tips						
6	Adhesive Plate Sealer for 96 well Plates	Adhesive Plate Sealer for 96 well Plates						
7	Air Bed	Approx Size: 34*18*23cm N.W.: 3.0kgs G.W.: 3.5kgs Using Life: 30000hours Mattress Specification: 1. Material: PVC 3PLY 2. Thickness: 0.3-0.4mm 3. Size: 200CM*90CM*7CM(inflated)						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SSI	Remarks, if any
	A	B	C	D	E	F	G	H
		Pump Specification: 1. Size: 24*13*9cm 2. Weight: 1.0kgs 3. Voltage: AC 220/230V 4. Frequency: 50HZ/60HZ 5. Power: 10W 6. Pressure range: 15-20kpa (115-150mmHg) 7. Air output:4.5-5.5L/minute 8. Cycle time: 10 minutes 9. Controlled by microcomputer 10. Spare pcs: 10pcs of round material slices, 1bag of special glue						
8	Airotor Cartridges	1) Should have the same company as Air Rotor supplied.						
9	Alginate	1) Supplied in the form powder with components as per the ADA/ISO/ISI specifications. 2) Approximate Composition ? a. Powder ? potassium alginate ? 15%, CaSo4 ? 16%, ZnO? 4%, Kltf ? 3%, Napo4 ? 2 %and Diatomaceous earth ? 60% 3) Material should not irritate the oral tissues, shall have pleasant color & flavour. Powder should have fine texture & should be free from foreign material. 4) On mixing ? When mixed in accordance with manufacturer instructions material should give a smooth plastic mass of workable consistency & should be free from lumps & granules. There shall be no separation of constituents any the material shall be sufficiently plastic for under cuts & anatomical; details to be accurately reproduced but shall not run out of such impression material. 5) Mixing Time ? 45 sec. ? 1 min 6) Setting time ? Not less than 2 minutes & 30 second 7) General strength not less than 50 lbs/sq. Inch. Should not rupture during removal from mouth. a) Packing ?						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		Material should be available in tightly sealed, moisture proof Pkt. Information to be supplied by the manufacturer ? Date of Manufacture & date of expiry. Adequate instruction for proportioning & manipulation of the material suitable measures for powder & water for purpose of proportioning shall be supplied by the impression during the internal between its withdrawal & cast pouring. Weight of contents and setting times of material should be stated on the packing. b) There should be a colour indicator where in the colour changes when the alginate sets. 8) it should be 1 box should contain 12 alginate packets total demand is according to box. 9) one packet should contain 450 gms to 500 gms powder						
10	Ambu bag with face mask (Adult)	It should be self-inflatable and must have pop up valve (non-return valve), attachment for oxygen tube & oxygen reservoir. Bag must be made up of medical grade silicon, latex free double layered which retain sensitivity and it must be resistant to rough use. Inlet end of the bag must have separate port for Oxygen supplement. Outer port must be such that re-breathing valve or non-return valve can be attached. Must be supplied with Oxygen reservoir bag, Should have a bag volume of 1500ml. It should be autoclaveable. It should be adaptable to all type of face masks. Must be USFDA or European CE approved product						
11	Ambu bag with face	It should be self-inflatable and must have pop up valve (non-return						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
	mask (Neonatal)	valve), attachment for oxygen tube & oxygen reservoir. Bag must be made up of medical grade silicon, latex free double layered which retain sensitivity and it must be resistant to rough use. Inlet end of the bag must have separate port for Oxygen supplement. Outer port must be such that re-breathing valve or non-return valve can be attached. Must be supplied with Oxygen reservoir bag. Should have a bag volume of 250ml. It should be autoclaveable. It should be adaptable to all type of face masks. Must be USFDA or European CE approved product						
12	Ambu bag with face mask (Paediatric)	It should be self-inflatable and must have pop up valve (non-return valve), attachment for oxygen tube & oxygen reservoir. Bag must be made up of medical grade silicon, latex free double layered which retain sensitivity and it must be resistant to rough use. Inlet end of the bag must have separate port for Oxygen supplement. Outer port must be such that re-breathing valve or non-return valve can be attached. Must be supplied with Oxygen reservoir bag. Should have a bag volume of 500ml. It should be autoclaveable. It should be adaptable to all type of face masks. Must be USFDA or European CE approved product						
13	Applicator Tips	1) Should be made up of plastic material to apply 2) bonding agent or echant for dental useIt should contain pack of 100 tips						
14	Arch bars	Itshould be a flat, sturdy stainless steel bar on which fleatsorhooks are						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		attached. It should be a type of the prefabricated Erich archbars. Should contain 1 feet arch bar It should be autoclavable. It should be mouldable to adapt it to the dental arches. It should be resistant to corrosion. The longitudinal edges of the strip should be continuous and smooth.						
15	Assorted teeth set for RPD various shades (ANTERIOR AND POSTERIOR)	1) It should be in anatomic or semi anatomic or monoplanic form as per the requirements 2) It should have good anatomic structure 3) It should be available in different shades 4) It should bond chemically with denture base resin 5) It should be dimensionally stable during curing process 6) It should have enough strength, abrasion resistance and resilience 7) Should be insoluble and impermeable to oral fluids 8) Should have low specific gravity (light in weight) 9) It should withstand flasking , deflasking , curing process 10) It should have coefficient of thermal expansion same as denture base resin 11) 1 anterior set should contain upper 6 anterior & lower 6 anterior teeth 12) 1 posterior set should contain upper 8 posterior and lower 8 posterior (4 of each side)						
16	Autoclave indicator strip (label)	Label size 6 cm x 2.5 cm Chemical band : 4 cm x 1 cm with green colour & 5 indicator on one strip. 1 box contains 100 labels (strip) should have ISO & CE certification						
17	B.P. Bulb with knob	Valve/knob:-Made of metal alloy hard and strong. Rust resistant can ensure long life. Rotating knob to hold and release air. Bulb:-Made of rubber soft material. Portable design and easy to hold.						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
18	Bain circuit, Adult	It should be corrugated,flexible,lightweight and transparent for easy monitoring of water build up or blockages.Should be provided with 22 mm corrugated tube, expiratory valve and 2 litre antistatic bag.The inner tubing should be non kinkable and non crackable.						
19	Bain circuit, paediatric	It should be corrugated,flexible,lightweight and transparent for easy monitoring of water build up or blockages.Should be provided with 10 and 15 mm corrugated tube, expiratory valve and 0.5 litre antistatic bag.The inner tubing should be non kinkable and non crackable.						
20	Band material 180 x 0.005	It should be of good quality, stainless steel non corrosive, 8 feet in length						
21	Barbor Linen Thread no. 40 Reel	Reel of 200 mtrs with individual fillaments of high quality linen with high tensile strength, low tissue reactivity, superior knot security & excellent handling characteristics.						
22	Barbor Linen Thread No. 60 Reel	Reel of 200 mtrs with individual fillaments of high quality linen with high tensile strength, low tissue reactivity, superior knot security & excellent handling characteristics.						
23	Barbor Linen Thread no. 80 Reel	Reel of 200 mtrs with individual fillaments of high quality linen with high tensile strength, low tissue reactivity, superior knot security & excellent handling characteristics.						
24	Barbour linen thread no.20	Reel of 200 mtrs with individual fillaments of high quality linen with high tensile strength, low tissue reactivity, superior knot security & excellent handling characteristics.						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
25	Barbour linen thread no.30	Reel of 200 mtrs with individual filaments of high quality linen with high tensile strength, low tissue reactivity, superior knot security & excellent handling characteristics.						
26	Bed Pan (Plastic)	Bed Pan (Plastic)						
27	Biodentine	1) set in 10-12 minutes 2) versatile use in edodontic repair and restorative repair. 3) natural micromechanical anchorage for excellent sealing properties. 4) similar mechanical and behavior properties as human dentine 5) 3.5 mm aluminium radiopacity for long term follow up 6) 5 X capsule and 5 X single dose containers						
28	Biopsy Punches disposable 2.5mm	Biopsy Punches disposable 2.5mm						
29	Bisacrylam ide	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified						
30	Black Goggles for Ophthalmic Patients	Should be scratch resistant, plastic goggles with UV protection.						
31	Blood Pressure Apparatus aneroid portable/St and Model	Blood Pressure Apparatus aneroid portable/Stand Model						
32	Boro-Spirit Ear Drops 5ml	Boro-Spirit Ear Drops; Each 5 ml contains 0.183gm Boric Acid in 2.08 ml of Alcohol						
33	Calcium hydroxide base for pulp	1) Should follow ISO/ISI standard composition. 2) It should have supply in two pastes in equal amount 3) It should be 13 gms base						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
	capping	and 11 gms catalyst with mixing pad						
34	Calcium hydroxide in (Aqueous or water base) for endodontic use	1) Should be supplied as single paste system 2) Hydrated Lime with aqueous solution for dental use.						
35	Calcium Hydroxide Powder 10gm.	Calcium Hydroxide Powder for Dental Cement for Cavity Base, 10 gm Bottle						
36	Calcium hydroxide- idoform paste	1) It should have supply in two pastes in equal amount 2) It should be 5 to 10 gms base and 8 to 12 gms catalyst 3) It should be premixed paste for immediate use with minimal material waste and clean up 4) It should be syringe tip delivery/for excellent access and control during placement , 5) It should be formulated with idoform for increased radioopacity and anti bacterial effect , 6) It should be used with gutta percha and root canal cements should have inuf quantity for about 40 to 50 fillings per syringe. 8) 1 box of 2 tubes of 2 ml each						
37	Carbide bur	TUNGSTEN CARBIDE . This bur is ideal for a variety of procedures including initial sectioning of impacted molars, cutting of teeth at root level, and root separation and amputation.After initial penetration, this long, tapered bur creates a funnel shape for easier access in the pulp chamber. The safe-ended, non- cutting tip prevents penetration of the pulp chamber floor or the root						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		canal walls. Cutting surface length: 9 mm; total length: 21 mm.						
38	Ceramic finishing diamond stone	Stones with embedded diamond particles for smooth finishing composites, porcelain and precious alloys. Available as a pack of 12.						
39	Ceramic finishing stone	Stones for smooth finishing of indirect restorations like composites, porcelain and precious alloys. Available as a pack of 3 or 4. Maximum speed 30000 rpm						
40	Cheek retractor	Good plastic quality, Light should pass, transparent and in different size and shapes.						
41	Cold Cure Acrylic Power and Liquid	1. Should be pink in color, strength should be good, give glossy finish & clear translucency after polishing porosity 2. Working time :- 5-7 min 3. Compressive strength :- 75 Mpa 4. Tensile strength :- 48-62 Mpa 5. Hardness:- 16 knh 6. Volumetric shrinkage :- linear 0.26 % 7. Powder should be 400 gms and 400 ml liquid monomer in one lab pack						
42	Composite bonding agent 7th Generation	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		the requirements of the user 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) 1 Bottle should have minimum 3 - 5 ml bonding agent						
43	Composite Polishing Kit	1) Should contain various sizes and shapes of rotary/hand tips. 2) Coarse, medium fine and extra fine grit should be available with colour coding. 3) one kit should have 12 burs						
44	Composite shade A1	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) Composite material should contain prepolymerised filler partical for more strength should have CHAMELEON EFFECT						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		WITH INTERNAL REFLECTION OF LIGHT.						
45	Composite shade A2	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations in anterior & posterior teeth. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user CE and US FDA approved from notified body. 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) Composite material should contain prepolymerised filler partical for more strength should have CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT						
46	Composite shade A3	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas.						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		2) Material should have optical properties like tooth and have life like appearance after making restorations. 3) Should have good Polishibility 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without andue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user CE and US FDA approved from notified body. 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply CE and US FDA approved from notified body. 7) Material should achieve clinically and biological properties after curing. 8) Composite material should contain prepolymerised filler partical for more strength should have CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT						
47	Composite shade B2	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations. 3) Should have good Polishibility 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without andue polymerization. 5) A good variety of						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		shades suitable for Indian population should be supplied as per the requirements of the user CE and US FDA approved from notified body. 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) Composite material should contain prepolymerised filler partical for more strength should have CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT						
48	Composite tooth coloured for restoration	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations in anterior & posterior teeth. 3) Should have good Polishibility 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without andue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) Composite material						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		should contain prepolymerised filler partical for more strength should have CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT						
49	Compostable Black Garbage Bag Size 18inch X 22inch	Compostable Black Garbage Bag Size 18inch X 22inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.						
50	Compostable Black Garbage Bag Size 24inch X 27inch	Compostable Black Garbage Bag Size 24inch X 27inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.						
51	Compostable Black Garbage Bag Size 24inch X 36inch	Compostable Black Garbage Bag Size 24inch X 36inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.						
52	Compostable Black Garbage Bag Size 27inch X 33inch	Compostable Black Garbage Bag Size 27inch X 33inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
53	Compostable Black Garbage Bag Size 30inch X 36inch	Compostable Black Garbage Bag Size 30inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
54	Compostable Black	Compostable Black Garbage Bag Size 33inch X 36inch Thickness : 70						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	Garbage Bag Size 33inch X 36inch	micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
55	Compostable Blue Garbage Bag Size 18inch X 22inch	Compostable Blue Garbage Bag Size 18inch X 22inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.						
56	Compostable Blue Garbage Bag Size 24inch X 27inch	Compostable Blue Garbage Bag Size 24inch X 27inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.						
57	Compostable Blue Garbage Bag Size 24inch X 36inch	Compostable Blue Garbage Bag Size 24inch X 36inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.						
58	Compostable Blue Garbage Bag Size 30inch X 33inch	Compostable Blue Garbage Bag Size 30inch X 33inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
59	Compostable Blue Garbage Bag Size 33inch X 36inch	Compostable Blue Garbage Bag Size 33inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics. Compostable Blue Garbage Bag Size 33inch X 36inch Thickness : 70 micron Quality :						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
60	Compostable Red Garbage Bag Size 18inch X 22inch	Compostable Red Garbage Bag Size 18inch X 22inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.						
61	Compostable Red Garbage Bag Size 24inch X 36inch	Compostable Red Garbage Bag Size 24inch X 36inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.						
62	Compostable Red Garbage Bag Size 27inch X 33inch	Compostable Red Garbage Bag Size 27inch X 33inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
63	Compostable Red Garbage Bag Size 33inch X 36inch	Compostable Red Garbage Bag Size 33inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
64	Compostable Yellow Garbage Bag Size 18inch X 22inch	Compostable Yellow Garbage Bag Size 18inch X 22inch Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.						
65	Compostable Yellow Garbage Bag Size 24inch X 27inch	Compostable Yellow Garbage Bag Size 24inch X 27inch Thickness : 70 micron Quality : Double sealing						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	Bag Size 24inch X 27inch	chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics.						
66	Compostable Yellow Garbage Bag Size 24inch X 36inch	Compostable Yellow Garbage Bag Size 24inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
67	Compostable Yellow Garbage Bag Size 27inch X 33inch	Compostable Yellow Garbage Bag Size 27inch X 33inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
68	Compostable Yellow Garbage Bag Size 30inch X 33inch	Compostable Yellow Garbage Bag Size 30inch X 33inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
69	Compostable Yellow Garbage Bag Size 30inch X 36inch	Compostable Yellow Garbage Bag Size 30inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
70	Compostable Yellow Garbage Bag Size 33inch X 36inch	Compostable Yellow Garbage Bag Size 33inch X 36inch Thickness : 70 micron Quality : Double sealing, chlorine free, leak proof and complaint with Government of Maharashtra notifications of plastics.						
71	conservative instrument	1) all the instruments in the kit should be made of high grade stainless steel 2) it should be light in						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
	kit	weight, easy to grip. 3) it should have sharp working ends. 4) it should be corrosion resistant 5) it should be autoclavable 6) kit should contain williams probe, probe, explorer, mouth mirror top with handle, tweezer, cement spatula, spoon excavator, ball burnisher, paralelogram condensor, flat condensor, plastic filling instrument, GMT #28, GMT# 29, Chiesel, HOE, hallen beck carver, wards carver, frahm carver, chisel bent,						
72	Cord Clamp Sterile	Disposable Umbilical Cord Clamp are made from Non toxic Medical grade polymer. 1. Material: PVC, PE, PP, ABS etc. 2. Sterilize by ethylene oxide gas 3. Supplied sterile in individual peel-able polybag. Or Equivalent						
73	Dental Amalgam Alloy	1) Composition should contain :- ? Silver ? 60-70% ? Copper ? 13-30% ? Fine cut of spherical alloy particles 2) Trituration time ? 1 min max 3) Consistency ? with the specified alloy and mercury ratio at the end of specified trituration time, mass should be of proper consistency allowing to form rope or ball 4) Carving ? At the end of 5-7 mins. It should allow carving without fracture or deformation 5) Creep ? restoration should not have more than 2% creep 6) Polishing ? after 24 hours, the restoration on polishing should give nigh luster 7) Compressive strength ? 60,000-70,000 psi. 8) Tensile strength ? 7000-9000 psi 9) 30 grm bottle						
74	Dental explorer	1) It should be made of stainless steel. 2) It should be double ended						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		instrument 3) it should facilitate Dental Diagnostic Examination 4) Rust Free, high Quality Stainless Steel .						
75	Dental floss	1) it should help to remove plaque under gumline 2) it should be multitufted 3)it should be ideal to remove plaque around braces, bridges and wide spaces 4) it should have 50 Pieces per pack 5) It should be Provided With handle						
76	Dental Plaster	Type II gypsum(Plaster) should be white without any foreign particles suitable for Dental laboratory work. It should be free from lumps. Initial setting time ? should not be less than five minutes & not more than eight minutes. Final setting time ? 30 minutes. Compressive strength ? Not less than 1300 ib/sq. Inch after 24 hours. Packing ? it will be delivered from time to time as & when required in the appropriate quantity & supplied in small packages as per demand in moisture proof resin bags lined with thick quality polythene. Be marked with ? 1) Manufacturers registered name & or mark. 2) Weight in the Container 3) Setting time 4) Date of manufacturer & Expiry 5) it should have 5 packets; per packet 5 kg(seperate pouch to avoid moisture contamination) total 25 kg box						
77	Dental Polishing Prophy cup	1) it should be disposable screw-type cups and shouldn't work with contra-angle handpieces. 2) it should remove superficial white and brown stains. 3) Its Silicon carbide microparticles provide gentle mechanical abrasion 4) Less than 0.2mm enamel removal 5) it should						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		be Autoclavable Cups and should minimize splatter. 6) it should have a curved rib interior within the cup which is intended to reduce splatter, which both the patient and operator. 7) Eliminate splatter of prophylaxis paste, blood and saliva 8) Disposable, Latex free, vibration free All screw-type cups work with contra-angle handpieces. 9) It should be ADA/ISO/ISI/CE/USFDA certified from notified body.						
78	Dental stone	Type III Gypsum (stone plaster) for dental laboratory work 1) Composition ? Powder ? (CaSO ₄) ₂ H ₂ O ₂ , Gum Arabic + Calcium carbonate 2) Setting time ? 7 to 12 minutes. Final setting time ? 12 to 13 minutes. Colour ? Any light shade 3) Compressive Strength ? 3000 lb/inch after one hour of setting. Linear expansion ? 0.03% after one hour. Not more than 0.20% after two hours. 4) Packing ? In 25kg. Airtight the lined with thick quality polythene. Container should be marked with the following information. ? Manufacturer, Registered name & or mark ? Weight in container ? Setting time ? Linear expansion on setting ? Dt. Of manufacture and expiry 5) it should have 5 packets; per packet 5 kg (separate pouch to avoid moisture contamination) total 25 kg box						
79	Dental X-Ray film IOPA Adult (Pack of 150 films)	1) Name ? Intra-Oral-Periapical films (Adult) 2) Size of the film should be 32 x 41 mm (approx) 3) It should be E-Speed film 4) It should have a protective plastic cover (water proof) with soft borders thus						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		making it comfortable to the patients. 5) Individual X-Ray film packet should have a double emulsion E-Speed film with embossed dot for side identification 6) The film should be covered with light proof (black coloured) paper packet and should have a lead/tin foil for preventing the effect of back scatter and for radiation protection 7) The packet should be flexible and not very rigid 8) The film should provide good radiograph with adequate density, contrast and details and should not have any inherent artifacts 9) All the film packets should have sufficient shelf life and should be supplied in adult form only 10) It each box should contain 150 films						
80	Dentulous non-perforated Impression Tray	Dentulous non-perforated Impression Tray. Made of corrosion resistant, German stainless steel. Pack of 8 pieces						
81	Dentulous Perforated Impression Tray	Dentulous Perforated Impression Tray. Made of corrosion resistant, German stainless steel. Pack of 8 pieces						
82	Denture Finishing Bur kit	1) Should contain 2 sand stone, 2 carbide & 2 steel burs, 1 abrasive separating disks with holder. 2) 8 to 12 fluted carbide burs or abrasive with particle size of 100 microns or larger for bulk reduction. 2 carbide & 2 abrasive stone bur of large [article size for bulk reduction. 3) 12 to 16 fluted 2 in number carbide burs & 30 ? 100 µm sized abrasive stone burs for contouring. particle for contouring. 4) kit contains: 18 ? 13 flute carbide burs 2 in number,						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		fine & 2 super fine, 2 diamond burs & 2 abrasives between 8 & 20 µm size. Total burs kit contains 23						
83	Denture polishing buff	Should be of muslin cloth, should have long service life, should take optimum polishing agent, non sticky						
84	Developer	Name of the material ? developer powder pack contains : Powder to make 5 liter solution						
85	Diacetyl monoxime	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified						
86	Diamond abrasive points FOR CROWN cutting commonly called as air rotor	1. Should be available in various sizes and shapes as per the requirements like round, straight fissure (flat and round end), tapered fissure (flat and round end), safe ended, disc or wheel shaped. 2. Should be available in friction type design. 3. Shank and neck should be made of steel of good quality. 4. Diamond particles embedded on the working end in the matrix which is durable. 5. Should not rust and corrode in common disinfecting solutions. 6. Should be autoclavable. 7. Should not disastrously break during use. 8. Should not dull after short duration of use.						
87	Diamond Burs Inverted cone	1) Should be available in friction types design 2) Shank and neck should be made of steel of good quality. 3) Diamond particles embedded on the working end in the matrix which is durable. 4) Should not rust and corrode. 5) Should be autoclavable. 6) Should not disastrously break during use. 7) Should not dull after short duration of use.						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
88	Diamond Burs Round	1) Should be available in friction types design 2) Shank and neck should be made of steel of good quality. 3) Diamond particles embedded on the working end in the matrix which is durable. 4) Should not rust and corrode. 5) Should be autoclavable. 6) Should not disastrously break during use. 7) Should not dull after short duration of use.						
89	Diamond Burs Straight	1) Should be available in friction types design 2) Shank and neck should be made of steel of good quality. 3) Diamond particles embedded on the working end in the matrix which is durable. 4) Should not rust and corrode. 5) Should be autoclavable. 6) Should not disastrously break during use. 7) Should not dull after short duration of use.						
90	Diamond Burs Tapered	1) Should be available in friction types design 2) Shank and neck should be made of steel of good quality. 3) Diamond particles embedded on the working end in the matrix which is durable. 4) Should not rust and corrode. 5) Should be autoclavable. 6) Should not disastrously break during use. 7) Should not dull after short duration of use.						
91	Diamond Pencil	useful for marking glass slide						
92	Diaper Adult disposable	Disposable Adult Diaper						
93	Diaper neonatal	Diaper neonatal for New born baby soft cotton surface dry, breathable quick absorbent and water resistance						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
94	Drill bits for micromotor 0.8 mmx 3mmx65mm	stainless steel, Autoclavable						
95	Drill bits for micromotor 0.8 mmx 5mmx65mm	stainless steel, Autoclavable						
96	Edentulous Perforated Impression Tray	Edentulous perforated Impression Tray. Made of corrosion resistant, German stainless steel. Pack of 8 pieces						
97	Endo access bur	1) Should be available in friction types design 2) Shank and neck should be made of steel of good quality. 3) Diamond particles embedded on the working end in the matrix which is durable. 4) Should not rust and corrode. 5) Should be autoclavable. 6) Should not disastrously break during use. 7) Should not dull after short duration of use.						
98	Endodontic Broaches	1) The Barbed Broach is designed for bulk removal of pulpal tissue and features an easy-to-grip, color-coded plastic handle. Packaged in a sterile blister pack. Should be ISO Standardized 2. Should be color coded 3. Six broaches in one pack						
99	Endodontic chelating gel	chelating agent and lubricant used to soften calcification and aid in cleansing the root canal system. 10% carbamide peroxide; 15% EDTA. In conjunction with NaOCl, its effervescent action lifts dentinal shavings and debris for efficient removal. Disposable tip eliminates						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		the possibility of cross-contamination by dispensing a clean, single dose. Smooth, gel-like consistency clings to files, making precise placement in the canal easy						
100	Endodontic files (45-80)	Made up of stainless steel. Should follow international standardization for endodontic hand files ISO/ISI in all its features and colour coding. Should be supplied in various lengths like 21, 25mm and others as per the need of the user. Should have good cutting efficiency. Handle design should have comfortable and firm grip. Supplied in assorted sets from no 45 to 80 as per the requirement of the users. Should be autoclavable. Should not rust and corrode in common disinfecting solutions and sodium hypochloride. Should show sign of deformation before breakage. Endodontic files should be provided K files. should have good cutting efficacy minimum upto 30 canals single file means 1 pkt of 21 mm & 25 mm of same no size						
101	Endodontic files (8, 10, 15- 40)	Made up of stainless steel. Should follow international standardization for endodontic hand files ISO/ISI in all its features and colour coding. Should be supplied in various lengths like 21, 25 mm and others as per the need of the user. Should have good cutting efficiency. Handle design should have comfortable and firm grip. Supplied in assorted sets from no 8 and 10 also from 15 to 40 (1 packet per size should contain 6 no. files) Should be autoclavable. Should not rust and corrode in common disinfecting solutions and sodium						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		hypochloride. Should show sign of deformation before breakage. Endodontic files should be provided K files. should have good cutting efficacy minimum upto 30 canals & 1 unit means 1 pkt of 21mm & 25mm of same no size						
102	Endodontic files H (10, 15- 40)	Made up of stainless steel. Should follow international standardization for endodontic hand files ISO/ISI in all its features and colour coding. Should be supplied in various lengths like 21, 25 mm and others as per the need of the user. Should have good cutting efficiency. Handle design should have comfortable and firm grip. Supplied in assorted sets from no 10 also from 15 to 40 (1 packet per size should contain 6 files) Should be autoclavable. Should not rust and corrode in common disinfecting solutions and sodium hypochloride. Should show sign of deformation before breakage. Endodontic files should be provided H files. should have good cutting efficacy minimum upto 30 canals & 1 unit means 1 pkt of 21mm & 25mm of same no size						
103	Endodontic files H 45- 80	Made up of stainless steel. Should follow international standardization for endodontic hand files ISO/ISI in all its features and colour coding. Should be supplied in various lengths like 21, 25 mm and others as per the need of the user. Should have good cutting efficiency. Handle design should have comfortable and firm grip. Supplied in assorted sets from no 10 also from 45- 80 (1 packet per size						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		should contain 6 files) Should be autoclavable. Should not rust and corrode in common disinfecting solutions and sodium hypochlorite. Should show sign of deformation before breakage. Endodontic files should be provided H files. should have good cutting efficacy minimum upto 30 canals & 1 unit means 1 pkt of 21mm & 25mm of same no size						
104	Endodontic hand C files	Stainless steel C-Files . It should provide easier access to challenging, calcified canals. Heat-tempered steel for stiff performance. ? Twisted file design for greater strength. ? Eases penetration of calcified canals.						
105	Endodontic hand files super flex	Get all the benefits of nickel titanium files with a cut like stainless steel. Due to the flexibility of nickel titanium, SureFlex files are better at following and conforming to the curvature of the canal during instrumentation.						
106	Endodontic hand flexo files	? Greater working comfort. ? Improved tactile feedback. ? X-ray visible calibration rings for accuracy and improved diagnostics. (25 mm and 31 mm only) ? Flexible strength for preparing even the most curved and narrow canals. ? Pre-inserted silicone stops						
107	Endodontic niti hand files 4%	1. Made up of nickel titanium alloy. 2. Should have a greater taper than ISO standardized instruments as per the requirement of the user. E.g. 4, 6, 8, 10, 12. 3. Should be supplied in various lengths like 21, 25, 31 and others as per the need of the user. 4. Should have good cutting efficiency. 5. Latch type design for universal grip in various gear reduction hand						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		pieces and endodontic motors. 6. Supplied in assorted sets of instruments of greater taper and tip dimension for thin medium and wide cana 12.Should be ISO/ADA/CE standardised /approved. 13.Should have published scientific studies regarding the brand supplied.						
108	Endodontic niti rotary files 4%	1. Made up of nickel titanium alloy. 2. Should have a greater taper than ISO standardized instruments as per the requirement of the user. E.g. 4, 6, 8, 10, 12. 3. Should be supplied in various lengths like 21, 25, 31 and others as per the need of the user. 4. Should have good cutting efficiency. 5. Latch type design for universal grip in various gear reduction hand pieces and endodontic motors. 6. Supplied in assorted sets of instruments of greater taper and tip dimension for thin medium and wide canals. 12.Should be ISO/ADA/CE standardised /approved. 13.Should have published scientific studies regarding the brand supplied.						
109	Endodontic screw post system kit	A fully vented parallel post means less chance of excess pressure at the apical section of the canal. Self- tapering threads make these posts ideal for reconstruction of remaining tooth structure without the excess stress of a fully-treaded post. The Radix-Anker core section has 20 undercuts that secure any core material tooth surface. Made of titanium alloy. kit contains Posts 12 penetration Drill:1 rooter Facer:1 spiral Bur:1 Gauge1 Mandre:1 Wrench:1						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
110	Eugenol free Impression Paste	1. It should contain orthoethoxy benzoic acid in place of eugenol. 2. It should not irritate, 3. should be supplied as base and accelerator paste, 4. Working time should be 3-4 min. 5. Setting time should be 5-10 mins.						
111	Extraction forceps kit	1)it should be made of high grade stainless steel. 2) it should have sharp beaks on tip-which should allow better adaptation to the root surface and should cut PDL Easily. 4) it should have large handles for better operator fit. 5) it should have long life.6)instruments should be light weight and should have comfortable grip. 6) each kit should have a set of 12: 1 X Upper anterior forceps , 1 X upper Premolar forceps, 1X Upper molar right forceps, 1X Upper molar left forceps , 1X Upper roots forceps, 1 X upper third molar forceps, 1 X upper molars right (cowhorn), 1 X upper molar left (cowhorn), 1 X Lower premolars, 1 X Lower molars forceps, 1 X lower root forceps, 1 X Lower molars (cowhorn) forceps.						
112	Fluoride Varnish	1) the varnish should be a protective against dental caries with fluoride in it. 2)It should dry quickly and show excellent adhesion to teeth. 3) it should be suitable for treating children, adolescents and adults.4)The protective action of should be based on t inhibition of demineralisation, stimulation of remineralisation, incorporation of fluoride into the lower layers of enamel, repair of initial caries lesions.5) It should have Calcium + fluoride + phosphate to protect the						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		teeth against acid attacks.6) Optimum flow and wetting properties Low viscosity Clear, Colourless Excellent adhesion. 7) It should be available in the form of capsule having 0.5ml of varnish						
113	Foam fluoride trays Blue- Large size	1)It should have a Anatomical design which allows a comfortable fit and effective delivery of gel to interproximal and occlusal areas. 2) it should have Contoured shape which should increase a comfort and should reduce gagging. 3) it should have a handle for easier and correct placement. 4)It should be available in Color coding by size. Large - blue, Available in pack of 50 Pairs.						
114	Foam fluoride trays white- Medium size	1)It should have a Anatomical design which allows a comfortable fit and effective delivery of gel to interproximal and occlusal areas. 2) it should have Contoured shape which should increase a comfort and should reduce gagging. 3) it should have a handle for easier and correct placement. 4)It should be available in Color coding by size. Medium - white. Available in pack of 50 Pairs.						
115	Foam fluoride trays Yellow- Small	1)It should have a Anatomical design which allows a comfortable fit and effective delivery of gel to interproximal and occlusal areas. 2) it should have Contoured shape which should increase a comfort and should reduce gagging. 3) it should have a handle for easier and correct placement. 4)It should be available in Color coding by size. Small - yellow. Available in pack of 50 Pairs.						
116	FP Commoditi	FP Commodities: PTK (Pregnacy Test Kit)						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	es: PTK							
117	Fumigation Disinfectant	It should have chemical components of potassium monopersulphate 40 to 50%, sodium C10-13-alkylbenzenesulphate 10-20%. sodium chloride 1-5% 1) It should have a property which can be used for aerial disinfection 2) it should be suitable for Hospital, operating and dental clinic use 3) it should be safe and should have no oral and dermal toxicity 4) it should be available in 500 g per jar 5) it should clean air against risk of airborne infections 6) it should contain no aldehyde and no toxic vapour phase. 7) it should have proven efficiency against viruses, bacteria, fungal pathogens. 8) it should be made of pyroxygen compound, surfactant, organic acids, and inorganic buffer system. 9) it should be approved by environmental protection agency						
118	Gas Refill	1) Should contain Liquid petroleum gas. 2) Should have Light weight, Easy to operate. 3) Refillable gas tank						
119	gas torch	1) Should contain Liquid petroleum gas. 2) Should have Light weight, Easy to operate, Adjustable flame control. 3) Refillable gas tank 4) should have ignition switch with flow control valve.						
120	Gates glidden drills	Stainless-steel construction provides flexibility that helps prevent breakage. The drill bits should be compatible with any right-angle latch handpiece and typically should come in lengths of 28 mm, 32 mm, or 38 mm. A range of six sizes						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation on how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		should allow the clinician to sequentially use the drills to achieve the appropriate taper and avoid excessive dentin removal.						
121	Gigli Saw wire	Made of martensitic steel (quenched, magnetic steel) and not stainless steel. Hooks at both ends. Should conform to ISO standard						
122	Gingival retraction cord	1. Non impregnated and knitted cord 2. Should be 100% cotton, Should compress upon packing and then expand for optimal retraction 3. Should not entangle in diamond bur 4. Bright colors to facilitate easy location and removal 5. The bottle should have a high carbon steel blade at one end to cut desired length of cord 1 x Bottle of cord (8ft/244cm). 6. 1 x #000 cord, 1 x #00 cord, 1 x #0 cord						
123	Glass Ionomer cement (Luting/ Type 1)	1) Maximum particle size should be 15um 2) Powder/liquid ratio in the range of 1.8/1.0 3) Mixing time (approx..) ? 20 sec working time (approx..) ? 2 min. (230 c) (from start of mixing) 4) Final finishing commencing time 4 - 5 min (after seating the restoration) 5) Film thickness of luting cement should be <20 um 6) Material should be silicate based gic with good track record of its use in dental care hospitals 7) ISO/ISI/CE/ADA specifications should be met 8) It should be 15 gms powder and 10 ml liquid.						
124	Glass Ionomer cement (Luting/ Type 2)	ADA specification No 66 Types Luting Type I Restorative - Type II Liner Base Supplied as : Powder Liquid form						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
125	Glass Ionomer cement type IX	Type IX GIC 1) It should be supplied in the form of powder & liquid with components as per ADA/ISO/ISI specification 2) Material should have good handling characteristics and shelf life of at least two years from supply 3) It should have compressive strength more than 180 MPA 4) Finishing of material should be possible immediately after setting 5) It should have adequate mixing & working time. It should have less setting time 6) It should be radioopaque. 7) It should be 15gms powder and 7.5ml liquid						
126	Glass Ionomer Cement (GIC)	Glass Ionomer Cement (GIC)						
127	Green Stick	1) It should be available in stick form 2) Flow should be less than 6% at 37oC and 85% at 45oC 3) It should be homogeneous on heating on flame 4) It should be able to record fine details in mouth 5) It should be easy mold and handle 6) 10 stick per packet						
128	Gum Paint (Tannic Acid)	Gum Paint (Tannic Acid)						
129	Gum tragacanth	SQ						
130	Gutta Percha (GP)	Gutta Percha (GP)						
131	Gutta percha points 2 degree	) ISO Standardized Gutta-percha points from number 15 onward sizes (#15-40) 2) (#45-80) 3) #15,#20,#25 4) Should be supplied in assorted packets as per required numbers and quantity 5) Should be colour coded						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		6) Have 2 degree taper 7) Should be accurate in tip diameter and taper as per ISO standards 8) Should not buckle down easily should be length marked 120 points in 1 packet						
132	Gutta percha points 4 degree	1) ISO Standardized Gutta-percha points from number 15 onward sizes (#15-40) 2) #15,#20,#25,#30.#35,#40 3) Should be supplied in assorted packets as per required numbers and quantity 4) Should be colour coded 5) Have 4 degree taper 6) Should be accurate in tip diameter and taper as per ISO standards 7) Should not buckle down easily should be length marked 60 points in 1 packet						
133	Gutta percha points 6 degree	1) ISO Standardized Gutta-percha points from number 15 onward sizes (#15-40) 2) #15,#20,#25,#30.#35,#40 3) Should be supplied in assorted packets as per required numbers and quantity 4) Should be colour coded 5) Have 6 degree taper 6) Should be accurate in tip diameter and taper as per ISO standards 7) Should not buckle down easily should be length marked 60 points in 1 packet						
134	Hand held vein detector	The hand-held device to find veins for needle insertion powered by rechargeable battery.with minimum 30 Red & Orange LEDs for visualization of deeper as well as superficial veinsThe device should have the option to use Red and Orange LEDs separately /simultaneously.The viewing diameter = not be less than 33mm.The open area = not be less than 17mm .The device should show clear visualization of veins in						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		adult/pediatric / geriatric/obese/dark skin patient with difficult vein access..power adaptor The device should be made by ABS material & not by plastic . EC REP						
135	Hard reliner	A hard, chairside, self-cure relining material with no/ minimal exothermic heat. It must completely set in the patients mouth in approximately 10 minutes. Dense, stable and color-fast. Easy to finish and polish						
136	Harris Haematoxiline	it should identify, and stain the cytological smears and tissue . It should stain the nuclear details. It should have a good shelf life .The stain retainability should be there for a longer period of time .						
137	Heat Cure Acrylic Powder and Liquid	1. It should be tasteless odorless , non toxic and non irritant 2. It should available in different shades 3. It should bond chemically with denture acrylic teeth 4. It should be dimensionally stable during curing process 5. It should have enough strength, abrasion resistance and resilience 6. Should be insoluble and impermeable to oral fluids 7. Should have low specific gravity (light in weight) 8. It should withstand flasking , deflasking , curing process 9. It should have coefficient of thermal expansion same as denture acrylic teeth 10. It should be radio opaque so that can be detected on x ray, if swallowed accidentally 11. It should be easy to manipulate & repair 12. Powder should be 400 gms and 400 gms liquid monomer in one lab pack						
138	Hematoxiline	For microscopy ,owners quality control certificate						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
	Delafield							
139	Hematoxylene Powder	For Staining, ce certified, owners quality control certificate						
140	Hemocytometer (Square tube)	For estimation of Hemoglobin.						
141	Hot water bag Rubber for hot Fomentation	Should be made of rubber, which can resist the hot water pressure comfortably, Should be high quality, leak proof, easy to clean & keep water hot for longer time. Should be reusable.						
142	Humidifier bottle for oxygen flow-meter	Polycarbonate bottle, transparent, crack resistant, unbreakable, screw type, airtight						
143	Hydrofluoric gel	Low-viscosity buffered gel containing hydrofluoric acid at 9% for etching of porcelain. High visibility yellow hue Etch should be easy to control and place Should not stain composite or resin cement Should yield very good porcelain-to-resin bond strength						
144	Ideal model silicon mould (upper & lower set)	It should be made up of good quality silicone, It should have good anatomical landmarks, easily separable, resistant to tearing						
145	Impression Compound	1) It should be available in cake form 2) It should be fuse in hot water and form homogeneous mass 3) Working temperature: 55oC ? 65oC 4) Manipulation time (Kneading time): 1-3 minutes 5) Fusion temperature: 43.5oC 6) Flow should be less than 2% at 37oC and						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		between 75-85% at 45oC 7) 5 cakes per packet with 200 grms in weight						
146	Impression disinfectant	99.9% isopropyl alcohol-based hospital grade disinfectant for dental impressions. Must be compatible with impression materials. Should not deform or distort the materials. Available in 500 ml spray bottles						
147	Impression Paste (ZOE)	ADA Specification no 16. should contain base and catalyst pastes Mixing time 1 min Final Setting time 15 mins Little or no dimensional change during setting It should record fine details in impression						
148	Interdental brush	1) Interdental cleaning brush should be designed to clean various interdental spaces. 2) It should work by cleaning off braces, dental bridges and other spaces between the teeth. 3) It should be plastic- coated wire material and has a smooth inward-outward motion to improve movement. 4) It should be integrated with a rubber grip to ensure it stays intact and will not slip easily.						
149	Intermediate Restorative Material	ISO 9001:2008 Available in powder and liquid form (Dentsply Caulk) Powder contains Zinc oxide and polymethyl methacrylate Liquid contains Eugenol Mixing time - 5 to 10 secs Compressive strength 48 Mpa Setting time - 6 to 10 min						
150	Interproximal polishing strip for composite	should be color coded with with matching grit size for easy identification						
151	Intra canal irrigation needle	1) Needle gauge size 27 or 30 2) Should have lateral opening & disposable. 3) It should be 100						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	with lateral opening	needles per packet						
152	IUCD 375	IUCD 375						
153	IUCD 380A	IUCD 380A						
154	Lab Cleaning Solution	Laboratory glassware/Plastic Cleaning solution,owners quality control certificate, 1X1 Litre						
155	Lacto Phenol Cotton Blue	ISO certified, WHO/GMP confirmed to / CE notified						
156	Laryngosc ope Adult	Should be LED laryngoscope set, adult, macintosh type, Stainless steel blades with removable led lamp. Handle and blades are made of 304 grade stainless steel.Dull/matt finish. Supplied with 4 blades size 1, 2, 3 and 4 and two batteries for handle. Should provide with 1 spare led lamp and batteries						
157	Laryngosc ope Peadiatric	Should be LED laryngoscope set, adult, Miller Type, Stainless steel blades with removable led lamp. Handle and blades are made of 304 grade stainless steel.Dull/matt finish. Supplied with 4 blades size 0,00,1 and 2, two batteries for handle. Should provide with 1 spare led lamp and batteries.						
158	Laser Tip	940 Fibreoptic Lez Tip 3-7 mm Single use, non sterile, Autoclavee before use						
159	lathe wheels	1) white in colour 2)should have rounded edges						
160	Lead equivalent protective apron with thyroid	Lead equivalent protective apron with thyroid shield of minimum 0.5 mm thickness (96% protection) light weight adult size of all types foldable						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
	shield of minimum 0.5 mm thickness (96% protection) light weight adult size of all types foldable							
161	Light Cure Composite	Light Cure Composite						
162	Light Cure Resin Restorative material (Kit)	1) Light cure composite resin kit should consist of ? Various shades of the material white to yellowish and light to darker ? Preferably have Shade guide provided in the kit? Etchant gel ? Phosphoric acid 35-37% 3 ml ? Bonding agent packaged in a leak proof and good quality bottle ? 1 unit of Mixing pad should have 100 pad sheets) Applicator tips Requirements of the material: ? Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 3) Material should have optical properties like tooth and have life like appearance after making restorations. 4) Should have good Polishibility 5) The material should be freshly manufactured and should have a sufficient shelf life. 6) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without andue						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		polymerization & SHOULD HAVE CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT 7) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user 8) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 9) Material should achieve clinically and biological properties after curing. 10) Composite material should contain prepolymerised filter partical for more strength kit contains: 5 composite syringes of 4-5gm each, 1 denting bonding agent 5ml, etchant 3-5ml, applicator tip s 100 sticks, mixing pad						
163	Lubricant spray for hand piece	1) It is used for maintenance of all dental hand pieces including high speed, slow speed and air motors. 2) it should have minimum quantity of 400 ml CE certified product						
164	Makintosh rubber sheet 110cm	Should be made of rubber with double colour, high grade polymer (55% grade A); Should be resistant to high temperature, chemical resistance and fair abrasion resistance. Should be water proof. Colour :Green, Blue, Red, Thickness-0.4mm, Width-110cm. Roll of 30 metres I.S.O. Certified, ISI Mark 4135: 1974						
165	Makintosh rubber sheet 90cm	Should be made of rubber with double colour, high grade polymer (55% grade A); Should be resistant to high temperature, chemical resistance and fair abrasion resistance. Should be water proof. Colour :Green, Blue, Red,						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		Thickness-0.4mm, Width-90cm Roll of 30 mtr I.S.O.certified, Double colour,water proof ISI Mark 4135: 1974						
166	Malachite Green	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified						
167	Male Condom	Male Condom						
168	Material for endodontic regeneration, healing & sealing.	Material should have excellent healing potential Should be applicable in endodontic perforation repair, apexification, retrograde filling in endodontic surgeries. Should have scientifically proven biocompatible. Manipulation and handling characteristics should be good. Packaging should be compatible for multiple use of the material from the container.						
169	Matrix bands no 1	1) Metallic, stainless steel, Contourable, standard design 2) matrix bands no 1 3)Madeup of Stainless Steel 4) should not rust and corrode in common disinfecting solution.						
170	Matrix bands no 8	1) Metallic, stainless steel, Contourable, standard design 2) matrix bands no 8 3)Madeup of Stainless Steel 4) should not rust and corrode in common disinfecting solution.						
171	MBT bracket kit 0.018 Slot	It should me made up of good grade S.S. it should have whole arch (premolar to premolar) Brackets, Canine should have 0 degree Torque, Hooks with Canine and premolars.						
172	MBT ceramic bracket kit 0.018 slot	It should be of low profile. wings should not fracture. It should be of good quality international standard kit of 20 brackets						
173	Mercury for dental	1) Should be 100% Pure 2) Should be specific for dental use. 3) Should						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	use	be arsenic free. 4) Should be packaged in containers which at safe from accidental spillage and easy to dispense in required quantities.						
174	Metal cutting Bur with air rotor attachment pack of 1 burs	1) Should not Rust. 2) Should Cut Metal & Ceramic crown effectively. 3) Should attach with Air rotor						
175	Micromotor diamond burs	1. Should be available in various sizes and shapes as per the requirements like round, straight fissure, tapered fissure, inverted cone, safe ended, disc or wheel shaped. 2. Should be available in friction type design. 3. Shank and neck should be made of steel of good quality. 4. Diamond particles embedded on the working end in the matrix which is durable. 5. Should not rust and corrode in common disinfecting solutions. 6. Should be autoclavable. 7. Should not disastrously break during use. 8. Should not dull after short duration of use.						
176	Microtome Disposable blade	Low Profile Blade for Microtome						
177	Modelling wax	1) Should be soft & pink in colour. 2) It must be capable of holding denture teeth in position at room temp and mouth temp 3) It should cohere easily 4) It should not be flaky or adhering to finger 5) It should not cause irritation to the oral tissue 6) It should be easy to trim with a sharp instrument at 23°C 7) 12 Sheet per packet or 225 grms in weight						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
178	Mouth mirror & straight probe	It should be made of good stainless steel rust proof , with good reflective mirror changeable mirror top, sturdy handle. The straight probe should be of a good stainless steel rust proof material						
179	Mouth mirror with handle	1. It should be made of good quality stainless steel rust proof material, 2. The mirror should be reflective and of a good quality						
180	Mucus extractor for neonatal deele catheter Each	Transparent Tube mucus extractor for neonatal & infants.Sterile Pyrogen Free,Non Toxic The capacity of the Mucus Extractor is 25ml and a spare cap is also provided for the sealing.Soft and kink resistant tubing with frosted surface A traumatic, soft rounded, beveled open tip. Capacity: 25 ml.						
181	Nichrome loop 1.3mm	Diameter 1.3mm,double wound, calibrated to 1.0 ul						
182	Nichrome loop 2mm	Diameter 2mm,double wound, calibrated to 0.005 ml						
183	Nichrome loop 4mm	Diameter 4mm,double wound, calibrated to 0.01 ml						
184	Nichrome straight wire	Suitable for stab inoculation						
185	NiTi Rotary endodontic instruments (Single file system)	1) Should follow international standardization ISO/ISI in all its features 2) Should be autoclavable 3) Should not rust & corrode in common disinfecting solutions & sodium Hypochlorite 4) Should not rust and corrode in common disinfecting solutions and sodium hypochloride. 5) Should show sign of deformation before breakage. 6) should have good cutting efficacy canal centring ability variable taper file pack of 6 or fixed taper						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
186	Nitrous Oxide Gas for Inhalation (Store under pressure in metal cylinders of the type conforming to the appropriate safety regulations and at temperature not exceeding 37°C)	Nitrous Oxide Gas for Inhalation (Store under pressure in metal cylinders of the type conforming to the appropriate safety regulations and at temperature not exceeding 37°C)						
187	Orthodontic composite light cure material kit (etchant, primer, composite material syringe)	It should have long shelf life ,good bonding quality,available in syringe form ,with good working time of 3 min.of good viscosity, syringe of 4gm and bottle of 5ml primer, Primer should have good bonding quality.						
188	Orthodontic mini implants with bracket head 1.4 x 8 mm, 1.6 x 6mm, 1.6 x 8 m, 2.0 x 8 mm	It should be made up of good quality, it should have different shape and type of head, it should be made up of good quality material, corrosion and tarnish resistant and it should have a good strength and resistant to deformation sizes : 1.4 x 8 mm, 1.6 x 6mm, 1.6 x 8 mm, 2.0 x 8 mm						
189	Oxygen Gas for	Oxygen Gas for Inhalation						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	Inhalation							
190	PAP kit	Content nuclear stain solution 1 bottle (500ml), cytoplasmic staining solution 2A and 2B bottle (500ml) each, pack size 3X 500ml. the manufacturer should be CE notified, WHO GMP confirmed and should give certificate of analysis. Shelf life should be more than 2 years. time consumption should be less than 5 min.						
191	Peeso reamer	Peeso Reamer is excellent for enlarging the cervical portion of the canal, preparing the canal entrance, and removing gutta-percha for post space. Notched for easy size identification, they can be used in any right-angle handpiece. Available in standard (32 mm) and short (28 mm) lengths.						
192	Perio Dressing Pack	1) Easily mouldable 2) Elastic consistency 3) Patient friendly 4) Eugenol Free 5) Compatible with resins 6) Regular set 7) Two paste system should be provided. 8) pack should 90 gms of base and 90 gms catalyst 9) CE and US FDA approved from notified body						
193	Personal Protection Equipment Kit (PPE Kit) Medium size XL	PPE Kit- (Personal protection equipment kit)- PPE kit should include the following safety gears: 1. Personal Protective coverall/Gown (Garment) - 1 no 2. Shoe Cover Upto knee ? 1 Pair 3. Eye Goggle ? 1 no 4. Face shield ? 1 no 5. N-95 Mask ? 1 no. NIOSH Certified 6. Nitrile Gloves ? 1 Pair 7. Bio Degradable Disposable Waste collection bag ? 1 no All the above safety gears shall be packed in a paper poly pouch. 1. N-95 Mask-						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generi c Name / Brand Name (only for Import er)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS M E / SSI	Re m ark s, if any
	A	B	C	D	E	F	G	H
		One Unit ? Shape that will not collapse easily. ? NIOSH certified. ? High Filtration efficiency. ? Fluid Resistance, minimum 80mm Hg pressure based on ASTM G 1862, ISO 22609 or Equivalent. ? Quality compliant with standard for particulate respirator that can be worn with full face shield. ? Quality compliant with standards of surgical N-95 Mask 2. Face Shield ? One Unit ? Made of clear plastic and provides good visibility. ? Adjustable band to allow good fit around the head and snug fit against the forehead. ? Fog ? Resistant (preferable). ? Reusable (made of material that can be cleaned and disinfected) or disposable. ? Quality compliance with standards : EU Standard directive 86/686/EEC, EN 166/2002, ANSI/ISEA 287.1-2010 or equivalent. 3. Goggle ? One Unit ? With transparent Glass/plastic zero power, good fitting, covered from all sides with elastic band or adjustable holder. ? Good seal with the skin of the face ? Flexible Frame to easily fit all face contours without too much pressure. ? Covers the eyes & surrounding areas and accommodates for prescription glasses. ? Fog and scratch resistance ? Indirect venting to reduce fogging 4. Nitrile Gloves ? One Pair ? Nitrile ? Non-sterile, powder free. ? Outer gloves preferably reach fore-arm (min 225mm total length). Thickness 5.5 MIL and size (M/L) ? Quality compliance with below standard or equivalent. ? EU standard directive 93/42/EEC class I EN 455 or equivalent. ? EU standard directive						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		89/686/EEC category III EN 374 or equivalent. ? ANSI/SEA 105-2011 or equivalent. ASTM D6319-10 or equivalent. 5. Disposable Coverall ? Along with One hood Drape OR One head Cover. (it should be 70 GSM and above ? Non Laminated ? One Unit ? Single use. ? The Fabric Garment/breathable Non Laminated Fabric ? Light colours to better detect possible contamination. ? Coverall Shall be designed to be Universal Fit. Soft Elastic to be fitted around front of hood, Wrist and ankles ? Coverall approved by SITRA (South Indian Textile Research Association), DRDE (Defence Research Development Establishment) and have valid Unique Covid Code (UCC). 6. Pairs of Shoe cover made of same fabric as of Coverall upto below Knee ? one pair 7. Bio Degradable Disposable Waste collection bag ? One Unit						
194	Pit and fissure sealant kit	1) Should have properties to flow in tooth developmental pit & fissures. 2) Should have tooth colour. 3) Should have anti carious properties. 4) This standard specifies requirements & test methods for resin based materials suitable for sealing pits & fissure in teeth. 5) This standard covers chemically cured/ LIGHT CURE 6) 1 kit contain 2 ml 2 syringes and 1 pit & 1 fissure small size carbide bur						
195	Plastic base formers	Plastic base former for dental models made of durable plastic. Available as set of maxillary and mandibular base formers						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
196	Plastic tube 12 x 100 mm	12 x 100 mm						
197	Plastic tube 12 x 75 mm	12 x 75 mm						
198	Plastic Wedges	1) it should be pack of 70 2) It should be small , tapering , traingwar pieces of clear plastic about 1/2 nch length						
199	Polishing buff wet and dry	1) Should have fine bristles in a circular disk form. 2) Should have properties to polish denture surface.						
200	Polishing Cake	1) Should have abrasive grit suitable for polishing denture base resins. 2) Should not burn and stain the prosthesis 3) Should have properties of polishing. 4) Should not leave remnants on the surface of prosthesis.						
201	Polishing paper 90 & 100 no.	1) Useful for polishing acrylic denture surface or shellac base plate. 2) ISO Standard very fine (Macrogrit) of P150 grit of average particle diameter of 90/100 microns						
202	Polishing Paste	1) Fine. Prophylaxis paste 2) Should contain fluoride 3) Should be available in various grit sizes and flavours 4) It should be 100 gms paste						
203	Powered Endo Cuuter reloaeds 60 MM	Endoscopic white, Blue, Gold and Green 60 mm linear cutter GST cartridge with reload surface having pocket extensions to prevent tissue milking, staples with asymmetric bent tip for consistent staple formation with an option of closed staple height of 1mm, 1.5mm,1.8mm and 2mm 6 rows of staple line with 3 on each side of cut line compatible for articulating Gun. A) White - 1mm B) Blue - 1.5mm						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		C)Gold - 1.8mm D) Green - 2mm*						
204	Pre formed space maintainer	All in one Starter Kit with Accessories : 1) 37 universal bands with tube 2) 40 space components: straight,curved, distal shoe 3) 1 Intra oral crimper 4) 1 Intra oral cutter Specifications kit contains : Sizes per band: 24,24 1/2,25,25 1/2, 26???.41 1/2, 42 Type of band: Universal Types of components: Straight, Curved, distal shoe Number of bands per size: 1/size Number of components: 40 Shape of bands : Universal Type of cutter: Wire Cutter with carbon attachment Type of crimper: Intra oral crimper Number of cutter: 1 Number of crimper: 1						
205	Prophy Powder	Abrasive Thixotropic Should be available in different grits and flavours Non splattering Homogenous and should not dry out.						
206	Provisional crown material	1.Should be supplied in the form of cartridge and automixing tips (in a packet) as per ADA/ISO/ISI specification. 2.Should be aself- curing composite for the fabrication of temporary crownsand bridges,inlays,onlaysandveneers. 3. It should be free from methylemethacrylate and peroxides. 4. It should have a fillercontent:40- 45wt.-% 5. Mixing time-30?45 sec 6.Setting time- 2?3min 7.Expiry should be atleast for 2 yrs from the date of supply.						
207	Pumice Powder	Should follow ISO/ISI standards use in cleaning dentures use in cleaning dentures						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
208	Rebreathing bags Size 1000ml	Should be made of non toxic, latex free material (Green/Black), with proper fix to the breathing circuit connector. Resistance: 5x104 / 5x106 ohms. 5602/BS 3353 standard .Capacity 1000ml.						
209	Rebreathing bags Size 1500ml	Should be made of non toxic, latex free material (Green/Black), with proper fix to the breathing circuit connector. Resistance: 5x104 / 5x106 ohms. 5602/BS 3353 standard .Capacity 1500ml.						
210	Rebreathing bags Size 2000ml	Should be made of non toxic, latex free material (Green/Black), with proper fix to the breathing circuit connector. Resistance: 5x104 / 5x106 ohms. 5602/BS 3353 standard .Capacity 2000ml.						
211	Rebreathing bags Size 500ml	Should be made of non toxic, latex free material (Green/Black), with proper fix to the breathing circuit connector. Resistance: 5x104 / 5x106 ohms. 5602/BS 3353 standard. Capacity 500ml.						
212	Resin Based Root canal Sealer	should have excellent radio-opacity, should have low shrinkage, low solubility and good flow, 1:1 paste to paste mixing system, silver free for no discoloration teeth, biologically inert , thermoplastic ability for easy removal						
213	Resin luting cement	Dual cure resin luting cement. Should be supplied in two components It should have Translucent shade. Kit should contain: base, catalyst, etchant, 3gm primer, 5gm bond, applicator tips						
214	Retention mesh for RPD	Easy to adapt, tension-free wax retention meshes for RPD casting						
215	Retreatment Rotary	pack of three files, 3 lengths, 3 progressive tapers and 3 apical						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
	File	diameter to remove material from coronal, middle and apical thirds of root canals						
216	RF/RA	Based on latex agglutination For the detection and semiquantitation of Anti-Streptolysin O in serum Sensitivity ? 200 IU/ml Specificity Purified Streptolysin O specific for Anti-Streptolysin O antibodies Latex reagent Uniform suspension of polystyrene latex particles coated with stabilized Streptolysin O GMP SCH M PRODUCT WISE AND TESTING FACILITY WISE, 50 test/kit						
217	Rnase away for molecular testing	Rnase away for molecular testing (Removes Rnase and DNA Contamination						
218	Root Canal Conditioner EDTA 15%	Should contain 10% carbamide peroxide and 15 % EDTA, should have effervescing action in conjunction with sodium hypochlorite that will lifts dentinal shavings and debris for efficient removal, lubricant used to soften calcification and aid in cleansing the root canal system. Dispense in syringe with disposable tips						
219	Rotary glide path files	flexibility and resistance to cyclic fatigue, less canal transportation than K files, suitable for curved canal, original canal anatomy is preserved						
220	Round bur surgical (straight micromotor hand piece)	HP8 stainless steel,						
221	Rubber Catheter	Plain rubber catheter - soft pure rubber catheter , non sterile, non						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	Plain Size 5	irritant to urethra , Length 16 inches, non irritant red rubber. Proximal funnel end & distal rounded atrumatic tip with side hole						
222	Rubber Catheter Plain Size 6	Plain rubber catheter - soft pure rubber catheter , non sterile, non irritant to urethra , Length 16 inches, non irritant red rubber. Proximal funnel end & distal rounded atrumatic tip with side hole						
223	Rubber Catheter Plain Size 7	Plain rubber catheter - soft pure rubber catheter , non sterile, non irritant to urethra , Length 16 inches, non irritant red rubber. Proximal funnel end & distal rounded atrumatic tip with side hole						
224	Rubber Catheter Plain Size 8	Plain rubber catheter - soft pure rubber catheter , non sterile, non irritant to urethra , Length 16 inches, non irritant red rubber. Proximal funnel end & distal rounded atrumatic tip with side hole						
225	Rubber dam Sheets	1. It should be available with dimensions 6x6inch. 2. It should be available with thickness 0.2mm heavy 0.02mm medium. 3. It should be available in a box containing 36 pieces. 4. 4.Should be green /blue colour						
226	Rubber teat Big	Big						
227	Rubber teat Small	Small						
228	RVG senor sleeves disposable	Should be flexible and leak proof made up in polythene form For dental use It should have pack of 500 sleeves						
229	Sanitary Napkins with Belt ETO sterile	1) Overall Length (mm) - 540+/- 10mm 2) Fluff core length (mm)- 240+/-10mm 3) Fluff core width - 75+/- 5mm 4) Thickness of a single pad 7-9 mm 5) weight of a single						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	Pack of 6	pad 9-11 gm 6) Absorbent ability- not less than 150ml						
230	Saponin	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified						
231	Screw expansion (jack screw)	It should be good quality S.S. expansion screw. For upper jaw ? dimention?12mm						
232	Sectional Matrix band system	Rings,matrices, wedges,forceps and pin tweezer for class II restoration , should create accurate contact and tight seal, minimal overhanging, rings should include silicone tips for better adjustment and clamps should be silicone protected. Should be non corrossive and autoclavable kit contains 2 rings, 1 ring holding forcep & mimimum 20 bands						
233	Separators elastics	It should be of good quality . made up of medical grade latex, it should have long shelf life.						
234	Seperating Medium	1) It should have excellent separating properties in between plaster & denture base material and n between plaster and plaster. 2) Should contain water, sodium alginate, borax, glycerin and dye as per ISO/ISI Standards. 3) It should form thin layer on the cast 4) It should easy to separate denture from the cast after curing 5) It should not chemically react with denture resin or should not hamper polymerization of resin 6) It should be thin enough to apply with brush on the cast 7) It should be 5 litre						
235	Shoe Cover	Made of Nonwoven material which is water and alcohol repellent. The pair should be of standard size,should be long upto below knee.						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
236	Silane coupling agent	Silane coupling agent for adhesion with porcelain. Pre-activated Solvent based on ethanol. Available in 5 ml bottle.						
237	Silver Amalgam	Silver Amalgam						
238	Silver solder wire 10 gm spool	It should be of minimum dimension of 6 feet, and weight 8 gms, it should be of good quality international standard.						
239	SKIN GRAFTING BLADE	Stainless steel blade, sharp blade edge should offer smooth shaving of healthy skin tissue, Individually sterile packed, European CE certified, Expiry date should be written.						
240	Slide Cabinet	Steel - 50,000 slides store cap.						
241	Sodium barbitone	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified						
242	Soft reliner	1. Secure bonding to the denture acrylic. 2. For manual mixing 3. Mixing ratio 1:1. 4. Biocompatible, tasteless and odourless. 5. Colour: similar to gingiva 6. Working time: 1 minute 30 seconds approx 7. Setting time: 5 minutes (in the mouth) approx						
243	Soyabean casein digest agar	ISO certified, WHO/GMP confirmed to and CE notified						
244	Spirometer Incentive 3 ball	Should have three color-coded balls in each chamber to provide a visual incentive for patients. Air should enter each chamber and raise the balls depending on the flow of air inhaled per second. Available in lung capacity : 600ml/sec, 900 ml/sec and 1,200ml/sec. The whole set should include EVA tubing, mouth piece 13M, tri-ball incentive						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		spirometer. Single-patient use to minimize the risk of cross contamination. 12 mm OD connector Volume should be 600 mL/sec, 900 mL/sec, 1200 mL/sec Material: Should be AS for chamber, PE for ball and mouthpiece, and EVA for tubing						
245	Sputum Container	Should be made of transparent plastic (polypropylene) with a screw cap. Should have optimal mechanical, thermal and chemical stability, be leak-proof, unbreakable and resistant against chemical influences. Standard size						
246	Sputum Container ISO certified, WHO/GMP confirmed to / CE notified	Sputum Container ISO certified, WHO/GMP confirmed to / CE notified						
247	stainless steel kidney tray large	1) It should be made of stainless steel 2) it should be rust free 3) It should be ideal for clinical use. 4) It should be ADA/ISO/ISI/CE/USFDA certified from notified body.						
248	Stainless steel kidney tray medium	1) It should be made of stainless steel 2) it should be rust free 3) It should be ideal for clinical use. 4) It should be ADA/ISO/ISI/CE/USFDA certified from notified body.						
249	stainless steel kidney tray small	1) it should be made of stainless steel 2) it should be rust free 3) It should be ideal for clinical use. 4) It should be ADA/ISO/ISI/CE/USFDA certified from notified body.						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
250	Stainless steel surgical dressing drum large	1) It should be made of stainless steel 2)it should be rust free 3)It should be ideal for clinical use. 4) It should be ADA/ISO/ISI/CE/USFDA certified from notified body.						
251	Stainless steel surgical dressing drum medium	1) It should be made of stainless steel 2)it should be rust free 3)It should be ideal for clinical use. 4)It should be ADA/ISO/ISI/CE/USFDA certified from notified body.						
252	Stainless steel surgical dressing drum small	1) It should be made of stainless steel 2)it should be rust free 3)It should be ideal for clinical use. 4)It should be ADA/ISO/ISI/CE/USFDA certified from notified body.						
253	sterilized disposable wooden tongue depressor	1) it should be smooth ,clean and sturdy and should be taste free and odor free. 2) it should be sterile and packaed in a sterile condition. 3) they should be suitable for oral examination. 4) it should be available in a pack having 100 wooden depressors 5)it should have polished smooth edges.6) it should be 6 inch long.						
254	Sticky wax	A highly tenacious dental wax used to hold material or pieces in position,Should flow easily, Must set firmly. Rigid and brittle at room temperature Should melt when heated. Should give a clean burnout. Available in stick form. (10 sticks/pkt)Shelf life minimum 2 years						
255	Straight probe	1. It should be made of a good quality rust free stainless steel, 2. the probe should have a straight and sharp end						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
256	suction tips	1. Should be made up of flexible plastic material for dental use. 2. Should have Smooth and fixed tips are securely fastened 3. It should be each box should contain 100 suction tips 4. Suction tip should have dimensions of 10 to 15 cm in length and 0.7mm to 1 cm diameter.						
257	Suture needle Straight Size 6 to 10	Surgical Needle Straight Cutting Needle size 6 to 10 6pcs in one Pkt						
258	Temporary luting cement	Temporary luting cement for sealing of crowns and bridges Available in two tubes form- base and catalyst in contrasting colours. Should provide homogenous, smooth mix Base approx. 40 g and catalyst approx. 14 gm						
259	Titanium mini plates 2 mm , L shaped 90 degree , right and left	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
260	Titanium mini plates 2 mm four hole with bar	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
261	Titanium mini plates 2 mm, H shaped plates,	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
262	Titanium mini plates 2 mm, two hole with bar	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
263	Titanium mini plates 2 mm, double X shaped plates,	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
264	Titanium mini plates 1. mm , L shaped 90 degree , right and left	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
265	Titanium mini plates 1.2 mm, T shaped plates, small and big	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
266	Titanium mini plates 1.2 mm, double Y shaped plates, small and big	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
267	Titanium mini plates 1.5 mm six hole without bar	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
268	Titanium mini plates 1.5 mm, Z shaped plates,	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
269	Titanium mini plates 1.5 mm, double Y shaped plates, small and big	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
270	Titanium mini plates 1.5 mm, H shaped plates,	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
271	Titanium orbital plates with bar 2 mm, 6 holes	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
272	Titanium orbital plates with bar 2 mm, 8 holes	Titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
273	Titanium screw 1.2 mm , size 4mm	Profile diameter 1.2 mm , core diameter 0.8 mm, titanium grade V						
274	Titanium screw 1.2 mm , size 6mm	Profile diameter 1.2 mm , core diameter 0.8 mm, titanium grade V						
275	Titanium screw 1.5 mm , size 4 mm	Profile diameter 1.5 mm , core diameter 1.0 mm, titanium grade V						
276	Titanium screw 2.0 mm , size 5 mm	profile diameter 2.0 mm , core diameter 1.5 mm, titanium grade V						
277	Titanium twisted	Titanium grade V, colour coded, supplied in food grade polythene						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
	plates, right and left 5 holes	pouch with label bearing all the details of product						
278	Titanium wire mesh plates, 1.5 mm	titanium grade V, colour coded, supplied in food grade polythene pouch with label bearing all the details of product						
279	Transmetal Carbide bur	should be extra-fine cross cut shape , should be able to remove old amalgam filling, should be able to cut through porcelain fused to metal crowns, should be non corrosive, auto clavable						
280	Triple trays	Disposable tray for functions of impression, counter impression and bite registration simultaneously. It must have ultra-fine net for precise bite impression. The Nylon net must be strong enough to avoid any tear during and after impression. It is very useful in close-bite registration cases. A set must consist of either Anterior, Posterior or Full arch.						
281	Tryptone water	ISO certified, WHO/GMP confirmed to and CE notified						
282	Tubal Ring	Tubal Ring						
283	Tubes for sputum collection	50 ml screw capped conical graduated and sterile						
284	Tubes for sputum collection 50ml	50 ml screw capped conical graduated and sterile.						
285	Turpentine Oil I.P 5 litre can	Turpentine Oil I.P 5 litre can						
286	Turpentine Oil I.P 500ml	Turpentine Oil I.P 500ml						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
287	Tweezer	1. It should be made of a good quality rust free stainless steel material						
288	Two-tone chewable disclosing tablets	1) It should Differentiate between old and new plaque; old plaque 2) Each pack contains multiple tablets (80/100)with full instructions 3) Available in fruit-it should be available in flavors.						
289	Ultrasonic Scaler Tip	high quality stainless steel compatible with major brand piezo and ultrasonic scaler, autoclavable and noncorrosive						
290	Umbilical Catheter Size 3.5 FR	Umbilical Catheter Size 3.5 FR						
291	Urine Pot plastic (M/F)	Urine Pot plastic (M/F) Female urine container size: 800ml color: whiteAND Male urine container size: 1000ml color: white						
292	Vacuum formed thermoplastic sheets	A clear transparent, hard, thermoplastic material for adaptation on dental models. It must be abrasion resistant and excellent form stability. The material must be highly biocompatible. Must be available in various thicknesses for bleaching trays, splints and crown and bridge work. A packet containing sheets of one thickness (either 0.5, 1, 1.5, 2, 3, 4 mm) amounting to total thickness of about 25 mm.						
293	Wedges wooden	1) It should be Small, Tapering, Triangular pieces of wood about 1/2 inch in length. 2) It should be each box should contain 50 wedges						
294	Weighting machine	Adult scale upto 200 kg capacity with 500gm marking with non slippery rubber mat, certified by weights and measures dept.						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
295	White Petroleum Jelly 1kg container	White Petroleum Jelly 1kg container						
296	White soft paraffin 500gms	white soft paraffin 500gms						
297	White soft Paraffin 50gm	White soft Paraffin 50 gm						
298	Wire 0.016 australian	It should have good strength.It should be corrosion resistant. it should be heat treated and cold drawn to proper diameter.it should be of international quality.it should have good resilience.special plus type						
299	Wire 0.018 australian	It should have good strength.It should be corrosion resistant. it should be heat treated and cold drawn to proper diameter.it should be of international quality.it should have good resilience.special plus type						
300	Wire 0.032inch TMA round	It should be made up of titanium molybdenum. With smooth finish ,high elasticity excellent formability low friction nickel free.It should have good deflection property.It should be shape specified.						
301	Wire Nickel titanium 16X22 upper & Lower arch	It should be good quality niti wire , it should be ovoid arch arch form. It should be of uniform size through out.						
302	Wire Stainless steel orthodontic 19 to 24 gauge	1) The wire should have Springiness, Stiffness, ability to bend without fracturing corrosion resistance and an ability to be simply joined by soldering or welding. 2) Circular in cross section.						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		1 packet should have length of 1 meter						
303	Wire TMA 16 x 22? upper and lower arch	It should have good strength. It should be corrosion resistant. it should be heat treated and cold drawn to proper diameter. It should be of international quality. it should have good resilience						
304	Wolf bottle	Wolf bottle						
305	zero degree teeth set (non anatomic)	1) Shape (Mould) in assortment of shapes i.e. square avoid & tapering. The teeth must bear resemblances to natural teeth & be free of porosity. 2) Shade ? Approx. of 6 shades from light medium to dark 3) Size ? Approx. 8 sizes ranging from small to large in each of the size & shade 4) General-The Teeth should bear resemblance to natural teeth and free of porosity. 5) teeth should not distort during or after processing. One complete set from each shade & shape & size for inspection should be supplied as sample with mould chart. 6) One set of 16 teeth (8 upper posterior and 8 lower posterior)						
306	Zinc Oxide	1) Should be specific for the dental use meeting ISO/ISI standard. 2) Should contain accelerator for rapid setting. 3) Should contain recommended percentages of other ingredients like magnesium oxide and others as per recommendations. 4) Should be usable as a temporary restorative material and also as root cancel sealer. 5) It should be pure free of any impurities. 6) it should be 100 gms powder						
307	Zinc Phosphate Cement	ADA specification No 96 Available as Powder -90 gm Liquid - 39 ml Composition : Powder Zinc oxide,						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		Magnesium oxide, Other oxides like Bismuth Trioxide, Calcium oxide, barium oxide) silica Liquid :Phosphoric acid , water , Aluminium Phosphate or sometimes zinc phosphate Powder liquid ratio 3:5:1 Mixing time 60-90 sec. Setting time is between 2.5 to 8 minutes						
308	Micro cuvettes For Digital haemoglobinometer (*FOR EXISTING MACHINES ONLY)	Microcuvettes / Strips Compatible with Digital Haemoglobinometer (Hemocue HB 201) contains CE Mark						
309	Microslides	Glass Microscope Slides : 76 cm X 26 mm X 1.3 mm, plain, on epack of 50 slides						
310	Sputum Container	The sputum container should be made of special medical grade polypropylene, thin plastic translucent with diameter 4 cm, Height 5 cm and capacity 40 ml. It should have screwball cap and should be air tight and leak proof and have injection molding. Weight is not to be considered.						
311	Steril Disp. Umbilical Cord Clamp	Disposable Umbilical Cord Clamp are made from Non toxic Medical grade polymer. 1. Material: PVC, PE, PP, ABS etc. 2. Sterilize by ethylene oxide gas 3. Supplied sterile in individual peel-able polybag.						
312	Transducer Protector individually packed and sterile,	Transducer Protector individually packed and sterile,						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
313	Urine Collection Bag	Urine Collection Bag - Disposable sterile with accurate calibration Non Toxic, Tube length 42"Manufactured from clinical grade PVC sheet. Super Smooth kink resistant tube, made from medical grade PVC. with connector. Non return valve and drainage outlet. Provided with hanging & carrying system. Capacity : 2000 ml. Approx. Sterile / Disposable / Individual packed.						
314	Acid Concentrat e Can	Acid concentrates for hemodialysis The electrolyte concentrations shown below are related to the ready to use dialysis solution in dilution: 1 liter of acid concentrate + 1,225 liter of bicarbonate concentrate 8,4% + 32,775 liters of purified water.						
315	Airotor Diamond points (Commonl y called as burs)	Airotor Diamond points (Commonly called as burs)						
316	Airrotor diamond burs	1 Should be available in various sizes and shapes as per the requirements like round, straight fissure, tapered fissure, inverted cone, safe ended, disc or wheel shaped. 2. Should be available in friction type design. 3. Shank and neck should be made of steel of good quality. 4. Diamond particles embedded on the working end in the matrix which is durable. 5. Should not rust and corrode in common disinfecting solutions. 6. Should be						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		autoclavable. 7. Should not disastrously break during use. 8. Should not dull after short duration of use.						
317	Airway Plastic	Soft Plastic Nasal Airways are constructed of a soft, latex free plastic. The interior of the airway has striations for easy suctioning and insertion. Latex free- Small/Medium/Large size						
318	Anti Infective Bed Cover Mat (AIBCM)	Anti Infective Bed Cover Mat (AIBCM) should be made of; Fabric:Denier 450 -550 range, Core Material (Film):Polyvinylchloride, Grid (Construction):18*9 +/- 10% range, Size: 2mtr*1mtr (200cms x 100cms), Yarn:High Tenacity Polyester, Weight in GSM:320-350 GSM range, Adhesion: Non- separable, Color: Light color such as myrtle green etc.,						
319	Autopsy Kit Clinical	Disposable Sterile with double packing. E.T.O. sterilized with sterile indicator. Each kit should contain: 1.Disposable Head cap 1 piece 2.Disposable Face Mask 1piece 3.Disposable Protective P.V.C. inner Apron without sleeves 1 piece 4.Disposable Surgical Gown with full sleeves & with elastic cuff non woven fabric-water repellent 1 piece 5. Disposable P.V.C. Hand gloves 1 pair free Sizes 6. Disposable P.V.C. Shoe cover 1 pair 7. Disposable Safety Goggles 1 pc 8.Plastic cover to wrap the body 1 pc Size: 80 cm x 120 cm.						
320	Billigoggles	Yet to be determined						
321	Black Goggles	Black Goggles for Cata. Patients Description						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	for Cata. Patients	1.Frame Molding 2.Glass Bluish Gray 3.Cataract cuff goggle with side protection frame should be smooth and non poking						
322	Bone Cement	Yet to be determined						
323	Bone Cutters	Yet to be determined						
324	Bonewax	4 mm & 6 mm Strip						
325	Bovine serum Albumin 22%	storage temperture 2 to 8 degree 22% bovine albumin, 1 x 10 ml						
326	C.T, Scan Films 14 X 17(Depend ing on the Camera)	C.T,Scan Films. (Wet & Dry) C.T. Scan Films. (Wet & Dry) (1) C T & M I R Films for Dry Camera Should have compatibility of Films to available camera (14" x 17") Should be quoted separately for Each camera. (2) C T & M I R Films for Wet Camera Should have compatibility of Films to available camera (14" x 17") Should be quoted separately for Each camera						
327	CBNaat Cartridges	Specifications for Xpert MTB/RIF cartridges: ➤ Cartridges should be self- contained and single use. Sample processing, amplification and detection should be carried out within this self-contained device ➤ Cartridge should be made up of Polypropylene material ➤ To avoid false negative results, cartridges should incorporate sample processing and PCR reaction controls. ➤ To avoid false negative results,						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		cartridges should be monitored by the instrument for pressure and other error conditions. ➤ Kit should contain sufficient reagents to process 10 patient of quality control specimen. ➤ Reaction kit should contain all necessary reagents for real time PCR reaction (Primers, Probes, KCl, MgCl₂, HEPES, BSA (bovine serum albumin), Polymerase, dNTPs etc) ➤ Should have proper positive and negative control <u>Cartridge packaging:</u> Description: 5 x 10-pack pouched cartridges/one box of 50 Sample Reagent bottles/5 packs of 12x2ml pipettes/one CD-ROM. Each kit contains 5 pouches. Each pouch contains 10 cartridges. PHYSICAL DIMENSIONS/WEIGHT: 31 cm W x 19 cm H x 27 cm D/3.6 Kg.						
328	Centrifuge Tube 50 ml (Falcon Tube)	Centrifuge Tube - 50 ml, made of propylene, conical shape bottom with tight screw cap, walls thick enough to allow centrifugation (maximum RCF) up to 5000 *g, solvent resistant, with white labelling area and black graduations, flat PE caps which can close with quick 3-4 turns, allowing one handed operation. Tubes should be radiation sterilized to ensure non-cytotoxicity, sold in packs of 500 tubes.						
329	Cobbs Elevator	Yet to be determined						
330	Composite bonding agent 7th Generation	1) Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO ₂ , and Organo sealants and other						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		common ingredients. Suitable for user in esthetically sensitive areas. 2) Material should have optical properties like tooth and have life like appearance after making restorations. 3) Should have good Polishability 4) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without undue polymerization. 5) A good variety of shades suitable for Indian population should be supplied as per the requirements of the user 6) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 7) Material should achieve clinically and biological properties after curing. 8) 1 Bottle should have minimum 3 - 5 ml bonding agent						
331	Compressed Hydroxylating Poliviny Acetate Nasal Tampon	Compressed Hydroxylating Poliviny Acetate Nasal Tampon, Size 4.5 cm * 1.5 cm * 2 cm						
332	CPAP Tubing	Disposable Silicone Tubings for CPAP						
333	Dental X-ray films; IOPA xray film Pediatric, adult; Occlusal Xray film	Dental X-Ray films Dental X-ray Film Internal Peripheral (1 1/4" x 1 6/8 ") Packet of 150 films						
334	Device for Inhalation of Dry	Each Act-O-Vial Contains ; (lower powder compartment) Methylprednisolone sodium						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	Powder Capsule	succinate.... USP 500mg (eq to Methylprednisolone 500 mg) Monobasic sodium phosphate anhydrous.... USP (buffering agent) Dibasic Sodium Phosphate anhydrous USP (buffering agent) Sodium Hydroxide.... USP NF (for pH adjustment) (Upper Diluent Compartment) Sterile water for injection..... USP when reconstituted mix total volume 4 ml						
335	Diaper for New Born	Yet to be determined						
336	Dipsosable Sterile Drape	Yet to be determined						
337	Disc Holding Foreceps Straight & Angled	Yet to be determined						
338	Disp. Negative suction drain set	Disposable Negative suction drain set comprising Bellow unit with connector. Connecting tube with clamp and "Y" connector. Curved needle with matching catheter. Spare perforated catheter. No.10, 12 G						
339	Dural Retractors Straight & Angled	Yet to be determined						
340	ECG Roll	Wax coated electro cardi graphic paper size 50 mm x 30mm or 60mmx20mm (superior quality)						
341	ECG/EEG Gel	ECG Jelly Tube Electrode contact paste suitable for ECG, EEG, ECT Jar of 200 gm						
342	Endobutto ns with or without	Yet to be determined						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation n how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	loop							
343	External Fixator 3.5, 4.5	Yet to be determined						
344	External Fixator Imizalar	Yet to be determined						
345	Formaline Chamber 10"	Formalin chamber is sterilization system is lightweight transparent, corrosion proof and made from special grade of polymer with two sliding trays to ensure 100% sterilization of your surgical instruments.10",						
346	Formaline Chamber 14"	Formalin chamber is sterilization system is lightweight transparent, corrosion proof and made from special grade of polymer with two sliding trays to ensure 100% sterilization of your surgical instruments. 14",						
347	Formaline Chamber 20"	Formalin chamber is sterilization system is lightweight transparent, corrosion proof and made from special grade of polymer with two sliding trays to ensure 100% sterilization of your surgical instruments., 20"						
348	Formaline Chamber 23"	Formalin chamber is sterilization system is lightweight transparent, corrosion proof and made from special grade of polymer with two sliding trays to ensure 100% sterilization of your surgical instruments.23"						
349	Formaline Chamber 26"	Formalin chamber is sterilization system is lightweight transparent, corrosion proof and made from special grade of polymer with two sliding trays to ensure 100% sterilization of your surgical						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		instruments.26"						
350	Glass Ionomer cement (Luting/ Type 1)	1) Maximum particle size should be 15um 2) Powder/liquid ratio in the range of 1.8/1.0 Mixing time 3) (approx..) ? 20 sec working time (approx..) ? 2 min. (230 c) (from start of mixing) Final finishing commencing time 4 - 5 min (after seating the restoration 4) Film thickness of luting cement should be <20 um 5) Material should be sreontium based gic with good track record of its use in dental care hospitals 6) ISO/ISI/CE/ADA specifications should be met 7) It should be 15 gms powder and 10 ml liquid.						
351	HUB Cutter	HUB Cutter Safety portable Box of Disposal of Hypodermic Needles. Container with WHITE & BLUE casing. Puncture Resistant Hard and can be Locked. Size - Hub Cutter with 100 Hypodermic Needle Collection Capacity.						
352	Hydrocolli d Dressing Patches	Yet to be determined						
353	Impression Compound	1) It should be available in cake form 2) It should be fuse in hot water and form homogeneous mass 3) Working temperature: 55oC ? 65oC 4) Manipulation time(Kneading time): 1-3 minutes 5) Fusion temperature:43.5oC 6) Flow should be less than 2% at 37oC and between 75-85% at 45oC 7) 5 cakes per packet with 200 grms in weight						
354	Interferenc e Screws -	Yet to be determined						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SSI	Remarks, if any
	A	B	C	D	E	F	G	H
	Titanium							
355	K- Wires 1, 1.5, 2, 2.5, 3 mm	Yet to be determined						
356	Kangaru Mother Care Bag	Yet to be determined						
357	Light Cure Resin Restorative material (Kit)	1) Light cure composite resin kit should consist of ? Various shades of the material white to yellowish and light to darker ? Preferably have Shade guide provided in the kit? Etchant gel ? Phosphoric acid 35-37% 3 ml ? Bonding agent packaged in a leak proof and good quality bottle ? 1 unit of Mixing pad should have 100 pad sheets) Applicator tips Requirements of the material: ? Composition of material should be as per the ADA specifications ? Composition of the resin : BISGMA, TEGDMA, UDMA, SIO2, and Organo sealants and other common ingredients. Suitable for user in esthetically sensitive areas. 3) Material should have optical properties like tooth and have life like appearance after making restorations. 4) Should have good Polishability 5) The material should be freshly manufactured and should have a sufficient shelf life. 6) Material should be supplied in syringe form, should be easy to dispense and storage during use should be safe without andue polymerization & SHOULD HAVE CHAMELEON EFFECT WITH INTERNAL REFLECTION OF LIGHT 7) A good variety of shades suitable for Indian population						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		should be supplied as per the requirements of the user 8) Material should have optimum flow, good handling characteristics and shelf life of atleast three years from supply. 9) Material should achieve clinically and biological properties after curing. 10) Composite material should contain prepolymerised filter partical for more strength kit contains: 5 composite syringes of 4-5gm each, 1 denting bonding agent 5ml, etchant 3-5ml, applicator tip s 100 sticks, mixing pad						
358	Makintosh Double Colour Water Proof Rubber 110cm Roll of 30 mtr	Rubber Makintosh double colour Rubber sheeting, High Polymer 55% & above , Colour (Green,Blue, Red), Thickness-0.4mm, Width-110cm Roll of 30 mtr I.S.certified, Double colour,water proof						
359	Makintosh Double Colour Water Proof Rubber 90cm Roll of 30 mtr	Rubber Makintosh double colour Rubber sheeting, High Polymer 55% & above , Colour (Green,Blue, Red), Thickness-0.4mm, Width-90cm Roll of 30 mtr I.S.certified, Double colour,water proof						
360	Mamography Films	Provides a 100-speed system with Min-R screens and a 150-speed system with Min-R 2000 screens Min-RS 18 x24 cm Mammography film & Min-RS 24x30 cm Mammography Film						
361	Microdrip I.V. sets with Volumetric	Micro I.V.Infusion set for pediatric use for Infusion Fluid administration with volumetric chamber 60 drops per 1 ml. Non						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
	chamber	Toxic, Sterilized, Pyrogen Free, ISI Mark, Non vented needle Size 23 G x 1.5 inch, Self sealing latex tube for medication & easy flushing. Special beveled extra strong spike. Disc filter in drip chamber facilitate rapid adjustment fluid need of 23 G for smooth flow. Highly kink resistant tubing roller clamp for flow meter. Packed in one plastic bag with Label containing product name .Mfg.date, Exp.date mfg.by.						
362	Mineral Trioxide Aggregate	1) set in 45 minutes 2) bioactive and radiopaque. 3) smooth consistency. 4) should contain gel and solution for mixing 5) 1GM POWDER						
363	Modular Bipolar Prosthesis (37 to 53)	Yet to be determined						
364	Mucus extractor for neonatal deele catheter	Transperant Tube mucus extractor for neonatal & infants. Sterile Pyrogen Free, Non Toxic The capacity of the Mucus Extractor is 25ml and a spare cap is also provided for the sealing. Soft and kink resistant tubing with frosted surface A traumatic, soft rounded, beveled open tip. Capacity: 25 ml.						
365	Nasal Splint (External)	Yet to be determined						
366	Nerve Root Retractors	Yet to be determined						
367	Nitrous Oxide Refilling	Nitrous Oxide Refilling						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
368	Non DHEP IV Set	60" Pressure resistant IV Administration Set, Vented Spike 20 Drops/ml with Drip Chamber 60mm / Fluid Filter - 15 micron/Flow Regulator/Male Luer Lock Adaptor with Cap/Back Check Valve/3.0mm x 4.1mm Tubing-Length 150cm/EO Sterile, DEHP and PVC Free						
369	Oxygen Refilling	Oxygen Refilling						
370	Penta Blood Bags	Penta Blood Bag System with one PRIMARY Bag (containing Anticoagulant Solution) and four SATELLITE Bags Useful for paediatric and/or multiple transfusions Prevents wastage as one unit of blood can be given to 5 different recipients Reduces the donor related risks since the donor is the same Optional satellite bag of TOTM sheet for platelet storage upto 5 days Disposable, sterile and non- pyrogenic Sealed aluminium foil packing						
371	Peritoneal Dialysis Solution IP 1 litre	Peritoneal dialysis solution IP 1 litre						
372	Personal Protection Equipment Kit(PPE Kit) Medium size XL	PPE Kit- (Personal protection equipment kit)- PPE kit should include the following safety gears: 1. Personal Protective coverall/Gown (Garment)- 1 no 2. Shoe Cover Upto knee ? 1 Pair 3. Eye Goggle ? 1 no 4. Face shield ? 1 no 5. N-95 Mask ? 1 no. NIOSH Certified 6. Nitrile Gloves ? 1 Pair 7. Bio Degradable Disposable Waste collection bag ? 1 no All the above						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SSI	Remarks, if any
	A	B	C	D	E	F	G	H
		safety gears shall be packed in a paper poly pouch. 1. N-95 Mask- One Unit ? Shape that will not collapse easily. ? NIOSH certified. ? High Filtration efficiency. ? Fluid Resistance, minimum 80mm Hg pressure based on ASTM G 1862, ISO 22609 or Equivalent. ? Quality compliant with standard for particulate respirator that can be worn with full face shield. ? Quality compliant with standards of surgical N-95 Mask 2. Face Shield ? One Unit ? Made of clear plastic and provides good visibility. ? Adjustable band to allow good fit around the head and snug fit against the forehead. ? Fog ? Resistant (preferable). ? Reusable (made of material that can be cleaned and disinfected) or disposable. ? Quality compliance with standards : EU Standard directive 86/686/EEC, EN 166/2002, ANSI/ISEA 287.1-2010 or equivalent. 3. Goggle ? One Unit ? With transparent Glass/plastic zero power, good fitting, covered from all sides with elastic band or adjustable holder. ? Good seal with the skin of the face ? Flexible Frame to easily fit all face contours without too much pressure. ? Covers the eyes & surrounding areas and accommodates for prescription glasses. ? Fog and scratch resistance ? Indirect venting to reduce fogging 4. Nitrile Gloves ? One Pair ? Nitrile ? Non-sterile, powder free. ? Outer gloves preferably reach fore-arm (min 225mm total length). Thickness 5.5 MIL and size (M/L) ? Quality compliance with below standard or equivalent. ? EU standard directive						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation on how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		93/42/EEC class I EN 455 or equivalent. ? EU standard directive 89/686/EEC category III EN 374 or equivalent. ? ANSI/SEA 105-2011 or equivalent. ASTM D6319-10 or equivalent. 5. Disposable Coverall ? Along with One hood Drape OR One head Cover. (it should be 70 GSM and above ? Non Laminated ? One Unit ? Single use. ? The Fabric Garment/breathable Non Laminated Fabric ? Light colours to better detect possible contamination. ? Coverall Shall be designed to be Universal Fit. Soft Elastic to be fitted around front of hood, Wrist and ankles ? Coverall approved by SITRA (South Indian Textile Research Association), DRDE (Defence Research Development Establishment) and have valid Unique Covid Code (UCC). 6. Pairs of Shoe cover made of same fabric as of Coverall upto below Knee ? one pair 7. Bio Degradable Disposable Waste collection bag ? One Unit						
373	PICC Lines	PICC Line Number 23, 24						
374	PPE Kit (Large Size)	Attached Separately						
375	PPE Kit (Medium Size)	Attached Separately						
376	Pre formed space maintainer	All in one Starter Kit with Accessories : 1) 37 universal bands with tube 2) 40 space components: straight, curved, distal shoe 3) 1 Intra oral crimper 4) 1 Intra oral cutter Specifications kit contains : Sizes per band: 24,24 1/2,25,25 1/2,						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		26???41 1/2, 42 Type of band: Universal Types of components: Straight, Curved, distal shoe Number of bands per size: 1/size Number of components: 40 Shape of bands : Universal Type of cutter: Wire Cutter with carbon attachment Type of crimper: Intra oral crimper Number of cutter: 1 Number of crimper: 1						
377	Pricking Lancet	Pricking Lancet - Disposable/ Single-use , Materials: stainless steel ,The disposable lancet is sterilized by gamma radiation or E T O sterilized . The lancet comes with smooth tribevel points. Quality in accordance with the CE and ISO standards						
378	Pricking Lancet Needles	Disposable/Single use Pricking needle no. 24 The disposable lancet needle is sterilized by gamma radiation. Made by stainless steel CE & ISO cerificate						
379	Provisional crown material	1.Should be supplied in the form of cartridge and automixing tips (in a packet) as per ADA/ISO/ISI specification. 2.Should be aself- curing composite for the fabrication of temporary crownsand bridges,inlays,onlaysandveneers. 3. It should be free from methylemethacrylate and peroxides. 4. It should have a fillercontent:40- 45wt.-% 5. Mixing time-30?45 sec 6.Setting time- 2?3min 7.Expiry should be atleast for 2 yrs from the date of supply.						
380	Replanisher 5 ltr jar	An aqueous photographic silver halide developer/replenisher solution particularly useful in low						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		throughput machine processing of exposed X-ray silver halide film contains 1-phenyl-3-pyrazolidone, hydroquinone, bromide ion, organic antifoggant and film speed restrainer and is buffered with an alkaline material to a pH of 10 ± 0.3 , replenisher having a lower pH than developer.- 5 Ltr						
381	Reusable Capacitive Coupling Current Quality Monitoring Electrode	1. Reusable Capacitive Coupling Current Quality Monitoring Electrode made of Viscoelastic Polymer, 2. Should have Current limiting nature & hence eliminate Patient pad site burns, 3. Should be based on Capacitive coupling principle, 4. Should made of akton polymers, 5. Can be used for all patient's weight >350 grams, 6. Should be Radiolucent & latex free, no adhesive related irritation to patient skin. 7. Should be US FDA approved, 8. Should be compatible to any Electrosurgical generator, 9. Size: 36" L * 20"W * 1/8" thickness.						
382	RF/RA	Based on latex agglutination For the detection and semiquantitation of Anti-Streptolysin O in serum Sensitivity ? 200 IU/ml Specificity Purified Streptolysin O specific for Anti-Streptolysin O antibodies Latex reagent Uniform suspension of polystyrene latex particles coated with stabilized Streptolysin O GMP SCH M PRODUCT WISE AND TESTING FACILITY WISE, 50 test/kit						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SSI	Remarks, if any
	A	B	C	D	E	F	G	H
383	Ring and Cup Curretes(Angled)	Yet to be determined						
384	Rod Cutter (Jumbo Cutter)	Yet to be determined						
385	Roungers 1, 2, 3, 4 mm	Yet to be determined						
386	Sanitary Napkin	Description : 1. Covering – The covering of the absorbent filler shall be made of good quality perforated Polyurethrene film sleeve which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. This shall be made of a product that is non-allergenic and bio – compatible. The sanitary napkins shall have a non-absorbent barrier on one side which shall have an identifying mark indicating clearly the side of the barrier. 2. Absorbent Filler – The filler material shall consist of cellulose pulp (either based on wood or paper or poly-acrylate gels). This shall be free from lumps, oil spots, dirt or foreign material etc. 3. Back Strip – A back strip for sticking the sanitary napkin onto the underwear should be there using good quality adhesive material. 4. Absorbency – The sanitary napkin should be able to absorb not less than 30 ml of normal saline (I.P.) @ 15 ml/minute. 5. Size – The size of absorbent section of the Sanitary Napkin shall be as follows:						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		Pad Length Core Length Overall Width Core Width Pad Thickness 235+_10mm 210+_10mm 155+_5mm 65+_5mm 7-9mm The thickness shall be measured by stacking 10 complete pads and measuring the stack height. The average thickness for the 10 pads shall be used as the pad thickness. 6. Weight – The weight of one full sanitary napkin shall not be more than 7.5 – 9 grams. 7. pH Value – The sanitary napkin shall be free from acids and alkali and the pH of the absorbent material shall be 6 – 8.5 when tested by the method given in IS : 1390 – 1961 (methods for determination of pH value of aqueous extracts of textile materials.)						
387	Silastic Band	EVE'S Ring is the first silicone band.(also known as the Silastic* band)A nonreactive rubber band that is used for closing Each fallopian tube as a procedure for female sterilization. The ring is placed around a 2-3 cm segment of fallopian tube, after being drawn inside a narrow cone-shaped applicator. Silastic medical graded silicon tubing seamless coil in 10 mtr length - 3 mm ID x 6 mm OD wall thickness 1 mm/ 4 mm ID x 6 mm OD 1.50 mm thickness/5.00 mm ID x 8 mm OD 1.50 mm thickness /6 mm ID x 9.00 mmOD 2.00 mm wall thickness / 6 mm ID x 10 mm OD x 2.00 mm wall thickness/ 9.00 mm ID x 12.00 mm OD 2.00 mm wall thickness / 9.00						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
		mm ID x 14.0 mm OD 2.50 mm wall thickness10 mm ID 15.0 m OD 2.50 mm wal thickness						
388	Silicon Oil	Silicon gel 15 gm						
389	Sonography Jelly 250 gm	All purpose sonography jelly Contact jelly for ultra sonography - Bottle of 250 gm						
390	Sterile Plastic Sheets (Adult)	Yet to be determined						
391	Strips / Cuvettes for Haemoglobinometer	Microcuvettes/ Strips Compatible with Digital Haemoglobinometer contains CE Mark						
392	Suction Catheter, Sterile	Designed for tracheal/bronchial suction.Smooth clear PVC ensures efficient mucus suction while minimizing potential damage to the delicate tracheal mucous membranes.Specially designed for easy passage through plastic endotracheal and tracheotomy tubes with minimum irritation to patient.Open distal end with elliptical eye for easy and efficient suction.Color-coded connector at proximal end for instant identification of catheter size.Available with thumb control device for positive control during suction.size -16- 50 cm length.						
393	Suction tube	Sterile PVC Latex free Tube for Suction.Reusable						
394	Suture Prelo	Yet to be determined						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
395	Suturing needles Straight	Surgical Needle(Suture Needles) Straight Cutting Needles Different sizes 6 to 10 6 to 10 , IS Certification 6 No. in One Pkt						
396	TBW Spools 16, 18, 20 G	Yet to be determined						
397	Thermocol Boxes (2 Sample Size)	Thermocol Transport Box size for 2 falcon Tubes (In standing Position) One Piece Thermocol molded box in Dye Thickness of Box - 2.5 cm Outer Dimensions : Length 18.5 cm, Breadth - 13 cm, Height -12 cm (without Lid) & Height - 14 cm (with Lid) Inner Dimensions : Length 14.5 cm, Breadth - 8 cm, Height -12.5 cm (without Lid) & Height - 13 cm (with inner part of Lid),. No. of Gel Pack required 2 Weight of Fully Packed Consignment Box 400 gm						
398	Thermocol Boxes (6 Sample Size)	Yet to be determined						
399	Thermome ter Clinical ISI mark Centigrade & Fahrenheit	1. Measuring Range 32.0~44.0Â°C(89.6~111.2Â°F); 2. Accuracy: +/-0.1Â°C(+/-0.2Â°F) for 35.0~42.0Â°C(95.0~107.6Â°F); +/- 0.2Â°C (+/-0.4Â°F) for remaining range at an ambient of 22Â°C(71.6Â°F); 3. Display: 3 1/2 LCD digits (Unit of display: 0.1Â°C/0.1Â°F); 4. Battery: LR41 alkaline; Battery life: About 250 hours or 1000 times; 5.Fever alarm; Waterproof case; 6.Memory for last measurement; 6. Low battery indication; Automatic power off; 7. Unit size: 120*20*12mm; Weight: 10.6g.						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
400	Thymol	IP						
401	Topical anaesthetic spray	Should contain topical anaesthetic lignocaine in a metal packed bottle. Should follow ISO/ISI standards						
402	Topical Fluoride Gel	1) It should be a Fluoride Gel with A.P.F. formula. 2) it should Contain 1.23% fluoride ion at pH 3.5. 3) it should have Thixotropic property. 4) It should be available in any of the below mentioned Flavors: Chocolate-Vanilla (dye free)/ Orange (dye free)/mango (dye free) / Strawberry (dye free). 5) It should be available in a quantity of 500 ml /pack. 6)The Gel should have a property which can be dispensed into a fluoride tray. 7)It should be ADA/ISO/ISI/CE/USFDA certified from notified body.						
403	Total Hip Prosthesis (Cemented & Uncemented)	Yet to be determined						
404	Total Knee Prosthesis (Modular)	Yet to be determined						
405	TrueNaat Chips	Rapid Diagnostic MTB chip specifications- 1. MTB chip should function on Real Time Polymerase Chain Reaction (PCR) test for the detection of Mycobacterium tuberculosis (MTB) in human pulmonary (sputum/non sputum) specimen. 2. MTB Chip should be able to run on the Real Time micro PCR Analyzer Machine with samples extracted using Auto sample prep						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SSI	Remarks, if any
	A	B	C	D	E	F	G	H
		device. 3. The MTB chip should be stable at temperature (20-30°C) 4. The remaining shelf life at the time of delivery of items should not be less than 5/6th of the total assigned shelf life. 5. The MTB chip should comprise of a reaction well containing dried down MgCl₂ 6. The dried down PCR reagents should be available in microtube for performing Real Time PCR test for detection of MTB. 7. The target sequence for MTB should be part of the ribonucleoside-diphosphate reductase gene specific to the MTB complex. 8. The MTB chip should have flash memory to retain lot information and standard curve values for automatic quantitative determination 9. MTB chip should not be available for reuse. The reuse and post expiry use should be detectable by flash memory. 10. The time taken for the diagnostic test result should be lower or equal to one hour. 11. The MTB chip should be amenable for disinfection and disposable as per applicable Biomedical Waste Management Rules. The chip should not be infective after use. 12. The consumables should be provided in sufficient quantity to cover 10% wastage <u>The following contents should be provided with the MTB chip:-</u> A} AUTO Universal Cartridge Based Sample Prep Kit containing 1. Reagent pack containing ready to						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re marks, if any
	A	B	C	D	E	F	G	H
		use buffers for automatic sample extraction 2. Cartridge pack with a fluidic cartridge to move sample and buffers through a proprietary matrix for extracting nucleic acids from the sample. 3. Disposable transfer pipette B} AUTO MTB Sample pre-treatment pack containing acer 3. Disposable transfer pipette (graduated) C} Individually sealed pouches, each containing 1. MTB micro PCR chip. 2. Microtube with freeze dried PCR reagents 3. DNase & RNase free filter pipette tip. 4. Desiccant pouch.						
406	Tympanic Grommet	Yet to be determined						
407	Umbilical Cannula	Umbilical Cannula Number 3, 4, 5						
408	Urine Pot plastic (M/F)	Urine Pot plastic (M/F) Female urine container size: 800ml color: white AND Male urine container size: 1000ml color: white						
409	Vaginal pH Test Kit	Technical Specifications for Vaginal pH test kit having Test strips, mobile application for detection, Software System with Dashboard. Digital Image Colorimetry, Vaginal pH Level 4 – 7 pH, On-screen patients result on mobile screen & dashboard required, Sample material - Vaginal Fluid, Measuring Time - Less than 1 Minute, Auto / Self calibration, Sensitivity - More than						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
		90%, Specificity - More than 90%. Memory to store data- Minimum 100000 test data with date, time, location and personnel details, Data auto transfer from application to backend server. Strip (working environment)- Disposable and stable at specified environmental condition (Working temperature 10-40°C, working humidity 5-95%), Strip (storage environment)- Should be stable at temperature 25-30°C and humidity of 5-95%, Intra-sample variation -Should be less than 5% & Accuracy -More than 95%. Certification & Standards - 1. ISO 13485 : 2016 for medical device, 2. CDSCO for medical device, Shelf Life: Two years.						
410	X-ray Developer	X-ray developer powder to make solution of 13.5 Ltr						
411	X-ray Developer	X-ray developer powder to make solution of 22.5 Ltr						
412	X-ray films Blue Base 10 x 12	X-Ray films 10x12 Blue Sensitive Polyester base round corner X-ray film Blue sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 10 x 12						
413	X-ray films Blue Base 12 x 15	X-Ray films 12x15 Blue Sensitive Polyester base round corner X-ray film Blue sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 12 x 15						
414	X-ray films Blue Base 6.5 x	X-Ray films 6.5 x 8.5 Blue Sensitive Polyester base round corner X-ray film Blue sensitive						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Re mark s, if any
	A	B	C	D	E	F	G	H
	8.5	Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 6.5 x 8.5						
415	X-ray films Blue Base 8 x 10	X-Ray films 8x10 Blue Sensitive Polyester base round corner X-ray film Blue sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 8 x 10						
416	X-ray films Green base 10 x 12	X-Ray films 10x12 Green Sensitive Polyester base round corner X-ray film Green sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 10 x 12						
417	X-ray films Green base 12 x 15	X-Ray films 12x15 Green Sensitive Polyester base round corner X-ray film Green sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 12 x 15						
418	X-ray films Green base 6.5 x 8.5	X-Ray films 6.5 x 8.5 Green Sensitive Polyester base round corner X-ray film Green sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 6.5 x 8.5						

S r. N o	Item Name	Technical specification s/ composition of tender enquiry	Compliance on each parameter with detailed substantiation on how the offered product meets the requirement.	If Column B=C (write Yes or No)	Generic Name / Brand Name (only for Importer)	Drug Mfg. License (Form 25 or 28 & 26)/ Medical devices/ Import License (Form 10 & 41)	MS ME / SS I	Remarks, if any
	A	B	C	D	E	F	G	H
419	X-ray films Green base 8 x 10	X-Ray films 8x10 Green Sensitive Polyester base round corner X-ray film Green sensitive Blue base with base fog less than or equal to 0.23 & Dmax not less than 3.10 interleaved The density values should be met at developer temperature 32 to 35 degree = 8 x 10						
420	X-ray fixer 22.5 ltr, 13.5 ltr	X-ray fixer powder to make solution of 13.5 Ltr						
421	X-ray fixer 22.5 ltr, 13.5 ltr	X-ray fixer powder to make solution of 22.5 Ltr						

Note:

- Pharmacopoeia standards IP/BP/USP etc. should be clearly mentioned against each drug/constituent of the formulation quoted as per the provisions of the Applicable Laws.
- Active ingredient used in formulation of item quoted shall be of mentioned Pharmacopoeia quality & Specifications.

Sign

Stamp

Date

Annexure-2: Undertaking by EOI Respondent

To,
Chief Executive Officer,
Maharashtra Medical Goods Procurement Authority,
1st Floor, Aarogya Bhawan,
Near CSMT Railway Station, Mumbai 400001 (Maharashtra)

Subject: EOI for the [*****]

Dear Sir,

With reference to your EOI dated _____, I, having examined the document and understood its contents, hereby submit my/our EOI for the aforesaid [*****]. The EOI is submitted voluntarily and without reservation, solely for market assessment purposes.

1. I/ We acknowledge that MMGPA will be relying on the information provided in the EOI and the documents accompanying such EOI, and we certify that all information provided therein is true and correct; nothing has been omitted which renders such information misleading; and all documents accompanying such EOI are true copies of their respective originals.
2. I/ We shall make available to MMGPA any additional information it may find necessary or require supplementing or authenticate the EOI.
3. I/ We declare that:
 - a) I/ We have examined and have no reservations to the EOI, including any Addendum / Corrigendum issued by MMGPA.
 - b) I/ We do not have any conflict of interest in accordance with the EOIs; and
 - c) I/We have not directly or indirectly or through an agent engaged or indulged in any unethical practice, as defined in the EOI, in respect of any tender or request for proposal issued by or any agreement entered into with MMGPA or any other public sector enterprise or any government, Central or State.
4. I/ We understand that this EOI is purely for market analysis purpose and you may not issue any following tender or empanelment process after this.
5. I/ We certify that we the Annual Turnover and Net Worth given by us is correct to the best of our knowledge.
6. I/ We certify that in regard to matters other than security and integrity of the country, we or any of our associates have not been convicted by a Court of Law or indicted or adverse orders passed by a regulatory authority which could cast a doubt on our ability to undertake the Contract or which relates to a grave offence that outrages the moral sense of the community.
7. I/ We further certify that in regard to matters relating to security and integrity of the country, we or any of our associates have not been charge-sheeted by any agency of the Government or convicted by a Court of Law.
8. I/ We further certify that no investigation which materially affects regulatory compliance or manufacturing/import operations is pending either against us or against our associates or against our CEO or any of our directors/ managers/ employees / Partners /Trustees.
9. I/We further certify that we or any of our Associates are not barred by the Central Government/ State Government or any entity controlled by it, from participating in any supply of drugs contract, and no bar subsists as on the EOI Due Date.

10. The power of attorney for signing of EOI, as per format provided at Annexure-6 of the EOI, is also enclosed.

In witness thereof, I/we submit this EOI under and in accordance with the terms of the EOI.

Yours faithfully,

Date:

(Signature, Name and designation of the Authorised signatory)

Place:

(Name and seal of Responder)

Annexure-3: Proforma for Production and Sale Statement

Name of the EOI Respondent entity:

Sr. No.	Year	Name and full Address of the Purchaser	Purchasing Entity (Gov./Semi Gov./Reg. under Trust)	Name of the Product	Purchase Order No. & Date	Purchase Order Quantity	Batch No.		PO Copy enclosed on Pg. No.
							Manufactured Qty	Sold Qty	
1	2023-24								
2	2024-25								
3	2025-26								

Add rows as per requirement.

Note:

1. In support of above statement, enclose the copies of supply orders and client's satisfactory certificates. All purchase orders should be enclosed in the serial as per the data provided in table above.
2. In case of importer, sold quantity shall be provided, production and manufactured good details may not be provided-
3. All the data of the bidding entity, as provided in the above table has been verified by undersigned Chartered Accountant/Statutory Auditor.
4. The issuer of this certificate must ensure that the above information/details are related to the bidding entity only.

Name, Membership number and signature of the Chartered Accountant/Statutory Auditor: UDIN

Name and seal of the firm:

Location, Date:

Authorized Signature (PoA holder)

[In full and initials with Seal]:

Name and Title of Signatory:

Name of Bidder (Firm/ Organization's name): Address:

Telephone:

Email:

(Name and seal of the Bidder)

[Location, Date]

Annexure 4: Details of Manufacturing Unit

(The details of manufacturing facility to be furnished)

- 1 Name of the EOI Respondent :
- 2 Registered Office Address :
- 3 Manufacturing Unit Address :
- 4 PAN :
- 5 GST No. :
- 6 Manufacturing License No.:
- 7 Date of Issue :
- 8 Valid up to / Retention :
- 9 Issued by (SLA / CLA) :
- 10 WHO-GMP / Equivalent Certification:
 - a. Certifying Authority
 - b. Certificate Validity
- 11 Details of Installed Manufacturing Capability (Indicative)

Type	Indicative Capability Details
Consumable	

Declaration

I/We hereby declare that the information furnished above is true and correct to the best of my/our knowledge and belief.

Signature of Authorized Signatory :

Name :

Designation :

Date :

Place :

Details of Installed Production Capacity for 30 days
(In Terms of Unit Packs)

Consumables :

General :

Beta-Lactum :

Injection

Ampoules :

Vials :

I.V.Fluids :

Sterile Powder:

Liquids

Suspension :

Syrups :

Drops :

Ointment :

Powders :

Antiseptics/

Disinfectants :

Name & designation of the authorized signatory:

Specimen signature of the authorized Signatory:

*The details of manufacturing unit should be for the premises where drugs quoted are actually manufactured.

THE DETAILS OF FACTORY PREMISES

Person In-charge of Factory

Name :

Phone No. :

Mobile No. :

Nearest Landmark of Factory:

Layout

Km from Airport :

Name of the Airport and City:

Km from Railway Station :

Name of the Railway Station :

Km from Bus Stand :

Name of the Bus Stand :
And City

All the data provided in the above table has been verified by undersigned Chartered Accountant/Statutory Auditor.

Name, Membership number and signature of the Chartered Accountant/Statutory Auditor: UDIN

Name and seal of the firm:

Location, Date:

Authorized Signature (PoA holder)

[In full and initials with Seal]:

Name and Title of Signatory:

Name of Bidder (Firm/ Organization's name): Address:

Telephone:

Email:

(Name and seal of the Bidder)

[Location, Date]

Annexure-4 (A): Proof of Distribution Network

A. Distribution / Logistical Details

S.No	Particulars	Details
1	Whether distribution/logistics arrangement exists	Yes / No
2	Nature of arrangement (Own / Third-party / Proposed)	
3	Broad geographic coverage	
4	Presence in Maharashtra (if any)	Yes / No
5	Any other relevant details	

B. Authorised Logistic Partners (If Applicable)

Sr. No.	Name	Address
1		
2		

Declaration

I/We hereby declare that the information furnished above is true and correct to the best of my/our knowledge and belief.

Signature of Authorized Signatory :

Name :

Designation :

Date :

Place :

Annexure-5: Details of Items Quoted

1. Name of the firm:
2. Address as given in manufacturing license:
3. Mfg. License No.:
4. Import License No. :
5. Date of issue:
6. Validity:
7. Non-conviction Certificate obtained on:
8. Market standing certificate obtained on:
9. Details of endorsement for all products:

Sr. No.	Item name	Specifications	Date of endorsement obtained from Central/State Drugs Commissioner	Whether Endorsement is in Generic or Brand Name

Add As Many Rows as possible You Want to Add

(Additional column should be inserted asking date of permission from CDSCO, in case of all newly introduced drugs and Fixed dose combinations)

Annexure 6: Annual Turnover Statement for Three Years

M/s. is participating in Tender No..... Items details as per Below Table.

Sr. No. of Item (as per Appendix A)	Name of Item	Required Turnover (In Rs)
-------------------------------------	--------------	---------------------------

Total Required Turnover

The Average Annual Turnover and Net Worth details of M/s_____ for participation under the Tender Document are given below and certified that the statement is true and correct.

Sr. No.	Year	Turnover (In Rs.)
1		
2		
3		
4		
Net Worth		
5	Net worth in the latest financial year preceding the Bid Due Date (positive/negative)	

“Net Worth” in case of Company shall mean the aggregate value of the paid-up share capital and all reserves created out of the profits and securities premium account, after deducting the aggregate value of the accumulated losses, deferred expenditure and miscellaneous expenditure not written off, as per the audited balance sheet, but does not include reserves created out of revaluation of assets, write-back of depreciation and amalgamation.

For other eligible entities, the Net Worth shall mean the amount derived by subtracting the liabilities from the assets as certified by the chartered accountant/statutory auditor having valid registration.

Note:

1. Certificate issued by a statutory auditor/chartered accountant along with audited financial statements confirming the average annual turnover of the Bidder during the stated financial years must be submitted on the letterhead of the statutory auditor/chartered accountant.
2. Provide supporting audited financial statements (Balance Sheets, Profit and Loss Statements, etc.) of the bidding entity.

3. The Net Worth of the bidder in the financial year immediately preceding the Bid Due Date should be positive.
4. “Turnover” for the purposes of this Tender Document shall mean the monetary value of goods sold by the Bidder.

Name, Membership number and signature of the Chartered Accountant:

UDIN

Name and seal of the firm:

Location, Date:

Authorized Signature of Bidder (PoA holder)

[In full and initials with Seal]:

Name and Title of Signatory:

Name of Bidder (Firm/ Organization’s name):

Address:

Telephone:

Email:

(Name and seal of the Bidder)

[Location, Date]

Annexure-7: Format for Power of Attorney for signing of EOI

(To be executed as an Affidavit on a Stamp paper of appropriate value)

Know all men by these presents, We, (name of the firm and address of the registered office) do hereby irrevocably constitute, nominate, appoint and authorize Mr. / Ms (Name), son/daughter/wife of..... and presently residing at, who is presently employed with us and holding the position of, as our true and lawful attorney (hereinafter referred to as the “Attorney”) to do in our name and on our behalf, all such acts, deeds and things as are necessary or required in connection with or incidental to submission of our EOI for the [*****] ,proposed or being developed by the [*****] (the “Purchaser”) including but not limited to signing and submission of all applications, EOIs and other documents and writings, participate in Responders' and other conferences and providing information / responses to MMGPA, representing us in all matters before MMGPA, signing and execution of all contracts including the Contract and undertakings consequent to acceptance of our EOI, and generally dealing with MMGPA in all matters in connection with or relating to or arising out of our EOI for the said Contract and/or upon award thereof to us and/or till the entering into of the Contract with MMGPA.

AND we hereby agree to ratify and confirm and do hereby ratify and confirm all acts, deeds and things done or caused to be done by our said Attorney pursuant to and in exercise of the powers conferred by this Power of Attorney and that all acts, deeds and things done by our said Attorney in exercise of the powers hereby conferred shall and shall always be deemed to have been done by us.

IN WITNESS WHEREOF WE, ..., THE
ABOVE-NAMED PRINCIPAL HAVE EXECUTED THIS POWER OF ATTORNEY ON
THIS
DAY OF, 20.....

For

(Signature, name, designation and address of person authorized by Board Resolution
in case of Firms/Company)/Partner in case of
Partnership Firms

Witnesses:

1.

2.

Notarised Person identified by me/personally appeared before me

/Signed before me/Attested/Authenticated* (*Notary to specify as applicable) (Signature, Name and Address of the Notary)

Seal of the Notary Registration Number of the Notary

Date____

Signature, name, designation and address of the Attorney)

Notes:

- The mode of execution of the Power of Attorney should be in accordance with the procedure, if any, laid down by the applicable law and the charter documents of the executants (s) and when it is so required, the same should be under common seal affixed in accordance with the required procedure.

Wherever required, the Responder should submit for verification the extract of the charter documents and documents such as a board or shareholders resolution/power of attorney in favour of the person executing this Power of Attorney for the delegation of power hereunder on behalf of the Responder.

- For a Power of Attorney executed and issued overseas, the document will also have to be legalized by the Indian Embassy and notarized in the jurisdiction where the Power of Attorney is being issued. However, the Power of Attorney provided by Responders from countries that have signed the Hague Legislation Convention, 1961 are not required to be legalized by the Indian Embassy if it carries a conforming Apostille certificate issued by the designated competent authority and has been notarized by the public notary.

Annexure-8: Undertaking for Non-blacklisting

(on Letterhead of the Respondent)

EOI reference No:294/MMGPA/ EOI for Consumables (2026-27)

1. I/We, (Name of the Responder), the Responder herein, has not been found guilty of malpractices, misconduct or Blacklisted/Debarred/ deregistered for the quoted product by any department of Govt. of Maharashtra or by any local authority and semi Govt. organization and other State Government/Central Government's organizations/ procurement corporation as on the date of submission EOI for the quoted items.

4. I/We undertake that I/we are not involved in any major litigation such as fraud, FEMA violations and any other civil or criminal proceedings.

Seal

Signature of Authorized Signatory

Date

Place

Verification

I, the above named [*Name of the Responder*], do hereby solemnly verify that the contents of the above Affidavit are true and correct to my knowledge and belief. Nothing false has been stated therein or material concealed therefrom.

Verified at {*location*} on {*Date*}

Annexure- 9: Litigation Affidavit

(on Non-judicial Stamp Paper of Rs.500/-)

EOI reference No:294/MMGPA/ EOI for Consumables (2026-27)

I _____ age _____ address(authorized signatory to sign the contract), hereby submit, vide this affidavit in truth, that I am the owner/authorized signatory of the bidding entity and I am submitting the documents in Envelope no.1 for the purpose of security of the contract. I hereby agree to the conditions mentioned below:-

1. I am liable for action under Bharatiya Nyaya Sanhita (BNS) for submission of any false/ fraudulent paper/information submitted in Envelope no.1
2. I am liable for action under Bharatiya Nyaya Sanhita (BNS) if during contract period and any false information, false bill of purchases supporting proof of purchase, proof of testing submitted by my staff, subletting company or by myself, I will be liable for action under Bharatiya Nyaya Sanhita (BNS).
3. I am liable for action under Bharatiya Nyaya Sanhita (BNS) if any paper is found false / fraudulent during contract period and even after the completion of contract (finalization of final bill).

Authorised Signature of Bidder

Seal of Company

Verification

I, the above named [*Name of the Bidder*], do hereby solemnly verify that the contents of the above Affidavit are true and correct to my knowledge and belief. Nothing false has been stated therein or material concealed therefrom.

Verified at {*location*} on {*Date*}

Annexure-10: Mandate Form

01	Company Name	
02	Postal Address of the company with Telephone No., Fax No. and Mail address	
03	Name of the Managing Director/ Director/Manager Mobile No. /Phone No. E-mail address	
04	Name and designation of the authorized company official Mobile No. /Phone No. E-mail address	

Bank Details

01	Name of the Bank Branch Name & Address: Branch Code No. Branch Manager Mobile No. Branch Telephone no. Branch E-mail ID	
02	9-digit MICR code number of the bank and branch appearing on the MICR cheque issued by the bank.	
03	IFSC code of the Branch	
04	Type of Account (Current/Savings)	
05	Account Number (as appear in cheque book)	

(in lieu of the bank certificate to be obtained, please **attach the original cancelled cheque** issued by your bank for verification of the above particulars)

I/We hereby declare that the particulars given above are correct and complete. If the transaction is delayed or not effected at all for reasons of incomplete or incorrect information, I would not hold Maharashtra Medical Goods Procurement Authority, Mumbai. I have read the conditions of tender/ agreement entered and agree to discharge the responsibility expected of me/from the company as a tenderer/ successful bidder.

Date

Company Seal

Signature

Place

(Name of the person signing & designation)

CERTIFIED THAT THE PARTICULARS FURNISHED ABOVE BY THE COMPANY ARE CORRECT AS PER OUR RECORDS.

Bank Seal with address

Signature of the Authorized Official of the bank

Annexure 11: Delivery and Supply Capability

(To be submitted on the letterhead of the EOI Respondent)

This Annexure is intended to capture indicative information on the delivery readiness and supply capability of the EOI Respondent for the purpose of market assessment.

A. General Information

Name of the EOI Respondent: _____

Type (OEM / Authorized Importer): _____

Sr. No.	Parameter	Details
1	Average Production Lead Time (in days)	
2	Average Dispatch Time after production readiness (in days)	
3	Expected Delivery Timeline to Maharashtra (in days)	
4	Emergency Supply Capability (Yes/No, give details)	
7	Warehousing Facility in Maharashtra (Yes/No, location)	
8	Cold Chain Facility (if applicable)	
9	Logistics Arrangement (Own / Third-party / Hybrid)	
10	Any other relevant information	

Note: Submission of a physical Delivery Challan at the time of delivery is mandatory. The challan must bear the signature and seal of the authorized accepting authority; without such acknowledgment, the delivery shall not be treated as complete.

Declaration

I/We hereby declare that the information furnished above is true and correct to the best of my/our knowledge and belief.

Authorized Signatory

Name: _____

Designation: _____

Date: _____

Place: _____

Annexure 11 (a) : Delivery Acceptance Certificate

FINAL ACCEPTANCE CERTIFICATE

Date: / /2026.

To,
Chief Executive Officer,
Maharashtra Medical Goods Procurement Authority,
1st Floor, Aarogya Bhawan,
St. George's Hospital Compound, Mumbai 400001
Maha.mmga2023@gmail.com

Subject: Acceptance Certificate for the supply of

Reference: Purchase Order No.

As per above mentioned reference purchase order, the said material has been received by this office / hospital in good condition & as per terms & condition of the purchase order. The details of the same are as follows;

Sr. No.	Particular(s)	Specific information
1	Name of Material Received	
2	Quantity as per Purchase Order	
3	Invoice Number with Date	
4	Batch Number(s) with Quantity	
5	Total Quantity Received	
6	Date of Receipt of Material	
7	Entry Taken in Stock Book	
8	Stock Book No.	
9	Page Number	
10	Name of the Supplier	
11	Name of the Manufacturer	
12	Address of the Manufacturer as per Invoice	
13	Material Received as per Technical Specifications mentioned in Purchase Order (YES/NO)?	

****This is certifying that, the above information is true on the basis of office record.**

Name of Pharmacist/ Store In-charge;

Name of Office/ Hospital In-charge;

(Signature with Stamp)

(Signature with Stamp)

Date:

Place:

Mail id for Acceptance: maha.mmgsa2023@gmail.com

Annexure 12: Respondent Classification Details

(To be submitted on the letterhead of the EOI Respondent)

This Annexure captures classification details for MSME status, social category, and geographical presence.

A. Basic Details

Sr. No.	Parameter	Details
1	Name of EOI Respondent	
2	Type of Entity (Company / LLP)	
3	Nature (OEM / Authorized Importer)	

B. Classification Details

Sr. No.	Parameter	Details
1	MSME Status (Yes/No)	
2	MSME Registration Number	
3	SC/ST Ownership (Yes/No)	
4	Certificate Details	

C. Classification

Sr. No.	Classification Parameter	Select One
1	Location Classification	☐ Within Maharashtra ☐ Outside Maharashtra
2	MSME Category	☐ MSME (Within Maharashtra) ☐ MSME (Outside Maharashtra) ☐ Non-MSME
3	SC/ST Category	☐ SC/ST (Within Maharashtra) ☐ SC/ST (Outside Maharashtra) ☐ Not Applicable

Declaration

I/We hereby declare that the information furnished above is true and correct to the best of my/our knowledge and belief.

Authorized Signatory

Name: _____

Designation: _____

Date: _____

Place: _____