



Maharashtra Medical Goods Procurement Authority

Office- 1st Floor, Aarogya Bhawan, Commissionerate of Health Services, Mumbai, D'Mello Road,
St. Georges Hospital Compound, Mumbai 400001

Contact Number- 022-22611471, 022-22717500 E-mail - maha.mmgsa2023@gmail.com

Purchase Order No 1008/EQ/MMPGA/2025

Dt: 02-04-25

To,

M/S. GENITRACK E-PROCUREMENT
PRIVATE LIMITED

315, Indralok Arcade Complex,

Maharshi Arvind Marg, V Mart,

Talab Darwaja, Junagadh,

Junagadh, Gujarat, 362001

Email - genitrackeprocurement@gmail.com

Subject: Purchase Order for Supply of Power Bone Drill

Reference : 1. Govt. Resolution Dated 01.12.2016

2. Administrative Approval. No: 224. Dated 27.12.2023

3. Revised Administrative Approval. No: 224. Dated 20.06.2024

4. Tender No : E-105 - Power Bone Drill

5. Your offer in response to above mentioned Tender

6. Sanction of Tender Approval Committee. Meeting dated 30.09.2024

7. Notification of Award of Contract No. NOA. 721 A/MMGPA/Equipment/2024.
Dated 09/10/2024

8. Your acceptance of NOA dated 09.10.2024

With reference to the above reference no.6, you have been awarded for the supply & installation of **Power Bone Drill** w.r.t approval as per reference no.5. Accordingly you are requested to supply the following goods to the consignee as per details mentioned below (**Consignee list mentioned separately in this order**)

Sr.No.	Item Description	Total Qty	Total Unit Rate incl. of GST in Rs.	Total Amount Incl. of GST in Rs.
1	Power Bone Drill (Make Biotech multifunction Model BOS-MF2nd SY -121) Warranty: 3 years from Installation	10	1049250	10492500
Amount in Words: INR One Crore Four lakhs Ninety Two Thousand and Five Hundred Only (Inclusive of GST)				

Terms & Conditions:

1. Performance Security: As mention in Contract NOA.721A/MMGPA/Equipment/2024. Dated 09/10/2024

2. The above price mentioned is accepted on FOR delivery and Installation at destination basis and is inclusive of all other charges.
3. The Warranty period is 3 years from date of installation. During the Warranty Period, the supplier shall visit all the user institutions/Consignee once in every 3 months (4 visits in a year) as preventive maintenance or as required any time for attending repairs/break down calls.
4. The technical specification of the above item should be as per Technical Specification mentioned at Annexure X.
5. Acceptance and Receipt In prescribed format. It should be submitted in Original Certificate copy to the purchasing authority along with triplicate copies of the Invoice.
6. The Consignee upon Receipt of the material should be issue acceptance certificate and after installation. Installation Certificate should be issued within 7 days of receipt of material in two copies. One copy should handover to Supplier and one copy should be sent Maharashtra Medical Goods Procurement Authority, Aarogya Bhawan St. George's Hospital Compound, Mumbai by E-mail.(Email- ID mahamgpa2023@gmail.com)
7. Invoice Copy- Should be sent in triplicate on the Name of Chief Executive Officer, Maharashtra Medical Goods Procurement Authority, Aarogya Bhawan St. George's Hospital Compound Mumbai
8. The Supplies must be completed within 90 days from date of issue of this Purchase Order.
9. If the supplier fails to deliver any or all the goods or fails to perform the services within the time frame (3) prescribed in the Contract, MMGPA will deduct from the Contract Price, as liquidated damages, a sum equivalent to 0.5% of the equipment to be supplied per week of delay or part thereof on delayed supply of goods and/or services until actual delivery or performance subject to 8. This Purchase Order is being issued on the anticipation that the Performance Security as mentioned maximum of 10% of the Contract Price.
10. Payments terms will be as mentioned in NOA, 721 A /MMGPA/Equipment /2024, Dated 09/10/2024
11. Consignee as per List Attached
12. The tentative delivery schedule should be shared with MMGPA within 07 days from date of issue of this Purchase Order.
13. All other terms and conditions of the above-mentioned RFP, Notification of Award of Contract and corrigendum thereof will be applicable.

(As per approval of
Hon. Chief Executive Officer
MMGPA, Mumbai)

JADHAV
SANJEEVKUMAR
MANOHARRAO
Digitally signed by
JADHAV SANJEEVKUMAR
MANOHARRAO
Date: 2025.04.02 17:15:28
+05'30'

(Dr. Sanjeevkumar jadhav)
General Manager (Technical)

**Maharashtra Medical Goods Procurement Authority
Mumbai**

Copy to: 1) Hon. Commissioner, Health Services & Mission Director, NHM, Mumbai
2) Hon. Director of Health Services, Mumbai
3) JDHS (Hospital Cell) Mumbai
4) Chief Accounts and Finance Officer, Medical Goods Procurement Authority, Mumbai
5) Office File

Copy to Consignee(s) All:

They should accept Equipment as per order & entry of the said Equipment is to be taken in dead stock register immediately within 24hrs of actual delivery

Copy Submitted to: 1) Hon. Addl. Chief Secretary, Public Health Department, Mantralaya, Mumbai.
2) Hon. Secretary-2, Public Health Department, Mantralaya, Mumbai.

Maharashtra Medical Goods and Procurement Authority

Tender No : E-105/ Power Bone Drill

	Gmdn name	Power bone drill
1	Clinical purpose	For fixation of fracture e.g. plating & nailing
2	Used by clinical department/ward	Orthopedic operation theater
3	Technical characteristics.(specific to this type pf device) ➤ <u>Drill and reamer hand piece</u> ➤ Multifunction Handpiece : Drilling and Reaming Single trigger R/F Switch DC brush motor 360 degree Rotate . should have DC brush motor for low maintenance with appropriate adapters for drilling ,reaming and pin placement and wire placement and wire placement future up gradation option for navigation interface for joint replacement surgeries weight of the hand piece with battery should be<3.5lbs fully cumulated 4.0mm hand piece should have piston grip hand piece tool less 360deg.attachments insertion should be autoclavable dedicated forward and reverse switch with safe mode. ➤ <u>Sternum saw hand piece</u> 1) Should have safe mode 2) Should have min.15000 CPM 3) Should have DC brush less motor for maintenance weight of the hand piece with battery should be<3.5lbs 4) Should have piston grip hand piece 5) Shouls have tool less mounting of accessories 6) Saw noise level should not more than 93db 7) Should have autoclavable 8) Should have min.13500CPM 9) Should have option of sternum guard. ➤ <u>Sagital saw hand piece</u> Should have variable speed single controls with standard and fast mode.free speed of 10000-12000 cycles per minute Weight of the hand piece with battery should be<3.5lbs Blade mount should be adjustable to different angles with 360 deg.rotation Should have tool less mounting of accessories Should have DC brush motor Should have autoclavable Should have safe mode ➤ Reciprocating saw hand pieces Should have safe mode Should have min.15000CPM Weight of the hand piece with battery should be<3.5lbs Should have piston grip hand piece Should have tool less mounting of accessories Saw noise level should not more than 93db Should be autoclavable with different blades it should have max.speed of 13500 CPM	
4	User's interface, setting	manual

5	Weight(lbs.kg)	<2.5 kg with stand and accessories
6	Noise (in dBA)	<93dBA
7	Heat dissipation	Heat dissipation :should maintain nominal temp
8	Mobility.portability	Yes
9	Power requirements	220-240v- 15%,60.60 Hz=3%
10	battery operated 1. Battery charger 220-240 v charger and should have the feature to count the charging cycle for a particular battery Should have capability to identify the worn out battery Should have to charge four batteries at a time Should have an indicator to provide battery cycle(no of time and total time) Should have reconditioning futures for battery Should be able to charge different batteries with same charger 2)battery kit Ni mh batteries with low internal impedance to deliver higher current than other battery types Ni,mh cell with capacity to produce more torque and non autoclavable with life of 300 approximate charging cycles Should have a run time of min.21 minutes Should include autoclavable outer housing Shield to protect battery from the housing 180 deg opening of battery housing for easy insertion of battery Should have option for the autoclave batteries QTY 4 nos	
11	Accessories, mandatory ➤ <u>Drill and reaming attachment</u> 1)1/4"drill attachment with key 2)keyless chuck 3)hudson modified tinkle attachment 4)pin collect 5)k wire collect ➤ <u>Sterilization case</u> Should be accommodate all hand piece, attachment and accessories for autoclave	
12	Atmosphere/ambiance(air conditioning ,humidity, dust,...)	1) operating room temp:-up to 40°C 2) Storage room temp :- up to 60°C 3) Relative humidity :- up to 90% non condensing
13	User's care, cleaning,disinfection&sterility issues	The case is to be cleanable with alcohol
14	Certificates performance and safety standards	USFDA/CE/BIS/CDSCO/ISO/AERB Approved
15	Requirements for sign-off	The company should be ISO certified and products should be ISI mark.
16	Training of staff (medical,paramedical,technicians)	Training of users in operation and basic maintenance shall be provided Advanced maintenance tasks required shall be documented

17	warranty	Three year with free servicing (min.3 year)during warranty &provide technical support and required spaces and consumable for 7 yrs after warranty period.
18	Maintenance tasks	Supplier to perform installation, safety and operation checks before handover local clinical staff to affirm completion of installation
19	Service contract clauses. Including prices	Training of users in operation and basic maintenance shall be provided advanced maintenance tasks required shall be documented
20	Operating manuals, service manuals, other manuals	Operating manuals, service manuals,other manuals
21	Other accompanying documents	Maintenance manual detailing complete maintaining schedule
22	Recommendations or warnings	Any recommendations for best use and supplementary warning for safety should be declared

Consignee List for Power Bone Drill

Consignee List of Power Bone Drill		
Sr. No.	Name of Institute	Unit
1	Trauma Care Unit Deori, Dist. Gondia	1
2	Trauma Care Unit Karanja, Dis. Wardha	1
3	Trauma Care Unit, Mahad, Dist- Raigad	1
4	Trauma Care Unit. Deulgaon Raja, Dist. Buldhana	1
5	Trauma Care Unit, Sangameshwar, Dist Ratnagiri	1
6	Trauma Care Unit, Naigaon, Dist. Nanded	1
7	Trauma Care Unit, Jintur, Dist. Parbhani	1
8	Trauma Care Unit, Kankavali, Dist. Sindhudurg	1
9	Trauma Care Unit, Kannad, Dist. Chatrapati Sambhajinagar	1
10	Trauma Care Unit, Georai, Dist. Beed	1
	Total	10