

**बिड दस्तावेज़ / Bid Document**

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	19-02-2026 15:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	19-02-2026 15:30:00
बिड पेशकश वैधता (बद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Social Justice And Empowerment
विभाग का नाम/Department Name	Department Of Empowerment Of Persons With Disabilities
संगठन का नाम/Organisation Name	Swami Vivekanand National Institute Of Rehabilitation Training And Research (svnirtar)
कार्यालय का नाम/Office Name	Cuttack
कुल मात्रा/Total Quantity	47
वस्तु श्रेणी /Item Category	ICU Bed , Bed side Locker , Bed side Table , Saline Stand as per spec , Bed side stool , Biomedical Waste Bin Small Set , Multipara Monitor , Syringe pump , Laryngoscope , Infrared Thermometer , Ambu Bag Adult and Paediatric , Android BP Apparatus , Electronic Weigh Scale , Electrical Suction Apparatus , Stretcher Trolley SS , Foldable wheelchair , Crash Cart , Dressing Trolley , Drug Trolley , Oxygen Trolley , Glucometer , Stethoscope , Nebulizer , X Ray Display , Air Bed , Bipap Machine , Convective Warmer
GeMARPTS में खोजी गई स्ट्रिंग्स / Searched Strings used in GeMARPTS	monitor
GeMARPTS में खोजा गया परिणाम / Searched Result generated in GeMARPTS	Contamination monitor, Multipara Monitor - Low End, Aerosol monitor, Advance Fire Monitor (V2), Voltage Monitor Relay, Holter Monitor (V2), Computer Monitor (V2), High End Multipara Monitor (V2), Ferrous debris monitor, Veterinary Multipara Monitor (V2)
अधिसूचना के लिए चयनित प्रासंगिक श्रेणियाँ / Relevant Categories selected for notification	<ul style="list-style-type: none"> <li>Styling Chair / Salon Chair / Swiveling Barber Chair</li> </ul>
बीओक्यू शीर्षक /BOQ Title	PMRD ICU BID III
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) /Minimum Average Annual Turnover of the bidder (For 3 Years)	5 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	2 Year (s)

बिड विवरण/Bid Details	
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है/MSE Relaxation for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	<p>Experience Criteria,Bidder Turnover,Certificate (Requested in ATC),Compliance of BoQ specification and supporting document</p> <p>*In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer</p>
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	1
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	3
ऑटो एक्सेंटेशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
बिड से रिवर्स नीलामी सक्रिय किया जाना है। / Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	H1-Highest Priced Bid Elimination
बिड का प्रकार/Type of Bid	Two Packet Bid
प्राथमिक उत्पाद श्रेणी/Primary product category	ICU Bed
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्ण पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
अनुमानित बिड मूल्य /Estimated Bid Value	1167000
मूल्यांकन पद्धति/Evaluation Method	Item wise evaluation
मध्यस्थता खंड/Arbitration Clause	No

**बिड विवरण/Bid Details**

**सुलह खंड/Mediation Clause**

No

**ईएमडी विवरण/EMD Detail**

एडवाईजरी बैंक/Advisory Bank	State Bank of India
Schedule 1 ईएमडी राशि/EMD Amount (In INR)	1600
Schedule 2 ईएमडी राशि/EMD Amount (In INR)	400
Schedule 3 ईएमडी राशि/EMD Amount (In INR)	200
Schedule 4 ईएमडी राशि/EMD Amount (In INR)	400
Schedule 5 ईएमडी राशि/EMD Amount (In INR)	400
Schedule 6 ईएमडी राशि/EMD Amount (In INR)	200
Schedule 7 ईएमडी राशि/EMD Amount (In INR)	2000
Schedule 8 ईएमडी राशि/EMD Amount (In INR)	400
Schedule 9 ईएमडी राशि/EMD Amount (In INR)	200
Schedule 10 ईएमडी राशि/EMD Amount (In INR)	40
Schedule 11 ईएमडी राशि/EMD Amount (In INR)	120
Schedule 12 ईएमडी राशि/EMD Amount (In INR)	80
Schedule 13 ईएमडी राशि/EMD Amount (In INR)	200
Schedule 14 ईएमडी राशि/EMD Amount (In INR)	1600
Schedule 15 ईएमडी राशि/EMD Amount (In INR)	300
Schedule 16 ईएमडी राशि/EMD Amount (In INR)	200
Schedule 17 ईएमडी राशि/EMD Amount (In INR)	500
Schedule 18 ईएमडी राशि/EMD Amount (In INR)	300
Schedule 19 ईएमडी राशि/EMD Amount (In INR)	300
Schedule 20 ईएमडी राशि/EMD Amount (In INR)	60
Schedule 21 ईएमडी राशि/EMD Amount (In INR)	40
Schedule 22 ईएमडी राशि/EMD Amount (In INR)	600
Schedule 23 ईएमडी राशि/EMD Amount (In INR)	200
Schedule 24 ईएमडी राशि/EMD Amount (In INR)	600

Schedule 25 ईएमडी राशि/EMD Amount (In INR)	400
Schedule 26 ईएमडी राशि/EMD Amount (In INR)	3000
Schedule 27 ईएमडी राशि/EMD Amount (In INR)	3000

### ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%)/ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	26

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). The EMD Amount will be applicable for each schedule/group selected during Bid creation.

(c).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

### लाभार्थी /Beneficiary :

Director

CUTTACK, Department of Empowerment of Persons with Disabilities, Swami Vivekanand National Institute of Rehabilitation Training and Research (SVNIRTAR), Ministry of Social Justice and Empowerment (Dr P P Mohanty)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

### एमआईआई खरीद वरीयता / MII Purchase Preference

एमआईआई खरीद वरीयता / MII Purchase Preference	Yes
मेक इन इंडिया विक्रेताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ तक की सीमा में है / Purchase Preference to MII sellers available upto price within $L1+X\%$	20
मेक इन इंडिया खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MII purchase preference	50

<p>सार्वजनिक खरीद (मेक-इन-इंडिया को प्राथमिकता) आदेश 2017 के अनुसार केवल क्लास 1/क्लास 2 के स्थानीय आपूर्तिकर्ताओं को ही भागीदारी की अनुमति है दिनांक 16.09.2020 (समय-समय पर संशोधित एवं लागू) / Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time)</p>	<p>Yes, in compliance with the MII ORDER : DPIIT Order(as amended and applicable time to time)</p>
--	--

#### एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ तक की सीमा में हो / Purchase Preference to MSE OEMs available upto price within $L1+X\%$	15
सूक्ष्म और लघु उद्यम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference	25

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

3. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023. [OM No.1\\_4\\_2021\\_PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

4. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are

excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

5. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

6. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- i. If number of technically qualified bidders are only 2 or 3.
- ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

#### **मूल्यांकन विधि(मदवार मूल्यांकन विधि) / Evaluation Method ( Item Wise Evaluation Method )**

Contract will be awarded schedulewise and the determination of L1 will be done separately for each schedule. The details of item-consignee combination covered under each schedule are as under:

मूल्यांकन अनुसूचियां / Evaluation Schedules	अनुमानित मूल्य / Estimated Value	वस्तु/श्रेणी / Item/Category	मात्रा / Quantity
Schedule 1	80000	Icu Bed	2
Schedule 2	20000	Bed Side Locker	2
Schedule 3	10000	Bed Side Table	2
Schedule 4	20000	Saline Stand As Per Spec	4
Schedule 5	20000	Bed Side Stool	4
Schedule 6	10000	Biomedical Waste Bin Small Set	1
Schedule 7	400000	Multipara Monitor	2
Schedule 8	20000	Syringe Pump	2
Schedule 9	10000	Laryngoscope	1
Schedule 10	2000	Infrared Thermometer	2
Schedule 11	6000	Ambu Bag Adult And Paediatric	2
Schedule 12	4000	Android Bp Apparatus	2
Schedule 13	10000	Electronic Weigh Scale	2

Schedule 14	80000	Electrical Suction Apparatus	2
Schedule 15	15000	Stretcher Trolley Ss	1
Schedule 16	10000	Foldable Wheelchair	1
Schedule 17	25000	Crash Cart	1
Schedule 18	15000	Dressing Trolley	1
Schedule 19	15000	Drug Trolley	1
Schedule 20	3000	Oxygen Trolley	1
Schedule 21	2000	Glucometer	2
Schedule 22	30000	Stethoscope	2
Schedule 23	10000	Nebulizer	2
Schedule 24	30000	X Ray Display	1
Schedule 25	20000	Air Bed	2
Schedule 26	150000	Bipap Machine	1
Schedule 27	150000	Conventive Warmer	1

### ICU Bed

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

### तकनीकी विशेषियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

### Bed Side Locker

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

### तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

### Bed Side Table

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

### तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

### Saline Stand As Per Spec

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local

**Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)****तकनीकी विशिष्टियाँ /Technical Specifications**

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	4	45

**Bed Side Stool**

(कमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	4	45

**Biomedical Waste Bin Small Set**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

#### तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

#### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

#### Multipara Monitor

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

#### तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

#### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

## Syringe Pump

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

### तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

## Laryngoscope

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

### तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

### Infrared Thermometer

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

#### तकनीकी विशेषियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

### Ambu Bag Adult And Paediatric

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

#### तकनीकी विशेषियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

**Android BP Apparatus**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

**Electronic Weigh Scale**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

**Electrical Suction Apparatus**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

**Stretcher Trolley SS**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

**Foldable Wheelchair**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

**Crash Cart**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

**Dressing Trolley**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशेषियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

**Drug Trolley**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशेषियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

#### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

#### Oxygen Trolley

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

#### तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

#### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

#### Glucometer

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

#### तकनीकी विशिष्टियाँ /Technical Specifications

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

**Stethoscope**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

**Nebulizer**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

<b>Specification Document</b>	<a href="#">View File</a>
-------------------------------	---------------------------

**BOQ Detail Document**[View File](#)

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

**X Ray Display**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

**Air Bed**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	2	45

**Bipap Machine**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

**Conventive Warmer**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

<b>Specification Document</b>	<a href="#">View File</a>
<b>BOQ Detail Document</b>	<a href="#">View File</a>

Advisory-Please refer attached BOQ document for detailed consignee list and delivery period.

#### परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasanta Rout	754010,At-Olatpur, Po-Bairoi, Dist-Cuttack, Pin-754010, Odisha	1	45

#### क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

##### 1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be  $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$ , subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

##### 2. Forms of EMD and PBG

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

Director SVNIRTAR  
payable at  
Olatpur Cuttack

Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

##### 3. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GEM GTC). DD should be made in favour of

Director SVNIRTAR  
payable at  
Olatpur Cuttack  
. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to

ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

#### 4. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

## **SWAMI VIVEKANAND NATIONAL INSTITUTE OF REHABILITATION TRAINING AND RESEARCH**

(Ministry of Social Justice & Empowerment, Govt. of India)

Department of Empowerment of Persons with Disabilities, (Divya ngjan)

OLATPUR, P.O: BAIROI, DIST: CUTTACK (ODISHA) Pin: 754010.

Website: [www.svnirtar.nic.in](http://www.svnirtar.nic.in), E-mail: dir.nirtar@nic.in & svnirtar@gmail.com

Ref-AD-5/BOQ/Equipment/ICU/PMRD//25-26

Date-13.01.2026

### **GEM BOQ BID DOCUMENT FOR PROCUREMENT OF ICU BED ITEMS of PMRD**

### **GENERAL INSTRUCTIONS AND TERMS & CONDITIONS OF TENDER DOCUMENT**

**Bidders can participate in GEM BOQ Bid in GEM site those are registered seller of GEM.**

**All rules & regulations of bid are as per default conditions and procedure of GEM.**

### **Ø FOLLOWING DOCUMENTS TO BE UPLOADED.**

SI	Name of documents
1	GST Certificate

2	Pan Card
3	Aadhar card
4	Last quarter/month GST 3B receipt.
5	IT return for 24-25/25-26
6	Bidder turnover certificate mentioning UDIN Number of CA.
7	Experience certificate in GEM and past service in similar items.
8	MSME certificate if any & MII certificate (mandatory).
9	Technical Bid details as per Annexure-I (mandatory)
10	Compliance Statement as per Annexure-II (mandatory)
11	Declaration as per Annexure-III (mandatory)
12	Bank Account details, (mandatory)
14	Leaflet/ specification of items to be supplied. Mandatory
15	Warranty declaration to be supplied. Mandatory
16	Declaration for another 5 yrs of CMC after normal warranty period.
17	OEM declaration if any.

—  
—  
P.S-I/C PMRD Administrative Officer/OIC-Accounts D.D P&A-Offg. D.D T  
ech-Offg

Director

—  
—  
—  
**Annexure-I (Technical bid details format).**

Sl. N O.	Information required	Information to be pr ovided by the bide r
1	Name of the firm/ organization	
2	Whether the organization is pr iotorship/ partnership/ Pvt. Ltd.,/ Ltd. Company	

3	Local/present Address,	
4	Mobile No. & Whatsap No. and E-mail. Id of the proprietor/firm/contact person..	
4	If the head office other than in (3) above, then give the address for the same	

**Annexure-II (FORMAT OF COMPLIANCE STATEMENT) - MANDATORY**

Name of Items	Required Specification	Whether complied to the required specification YES/NO	Name of brand/model to be supplied
1. ICU Bed			
2. Bedside Locker			
3. Bedside Table			
4. Saline Stand			
5. Bedside stool			
6. Biomedical Waste Bin Small Set of 3			
7. Multipara Monitor			

8. Syringe pump			
9. Laryngoscope			
10. Infrared Thermometer			
11. Ambu Bag Adult and Paediatric			
12. Android BP Apparatus			
13. Electronic Weigh Scale			
14. Electrical Suction Apparatus			
15. Stretcher Trolley- SS			
<b>16. Foldable wheelchair</b>			
<b>17. Crash Cart</b>			
18. Dressing Trolley			
19. Drug Trolley			
20. Oxygen Trolley			
21. Glucometer			
22. Stethoscope			
23. Nebulizer			
24. X-Ray Display			
25. Air Bed			
26. Bipap Machine			
27. Conventive Warmer			

### **Annexure-III (DECLARATION FORMAT)**

I/we ..... having  
my/our office at .....  
..... do hereby declare that I/we have carefully read all the terms & conditions of GEM BOQ BID of the SVNIRTAR, Olatpur, Bairoi, Dist-Cuttack, Odisha-754010. I will abide by all terms and conditions set forth in the tender paper and GEM for the period of agreement.

Signature of Bidder

#### **Ø Terms & Conditions**

#### **Ø As per default terms & condition of GEM.**

- Ø Equipment's will be delivered and installed at SVNIRTAR, PMRD during working hours and on working days only. If the consignee is changed, it will be informed by mail or clarification/ATC in the bid.

- Ø If any items are not accepted, they may be replaced within seven days otherwise they will be rejected.
- Ø If the price of any items or total amount of order is more than our estimate value, excess amount items may be rejected, or total bid may be cancelled.
- Ø Work will be treated as ok if it is accepted by the indenting department.
- Ø Installation/demonstration is the responsibility of supplier. No extra cost will be paid for it.
- Ø Warranty certificates may be submitted along with materials. Minimum warranty period is 02 years. Materials or any parts required for repair during warranty period will be provided by the supplier. CMC of the equipment if required will be provided by the supplier after warranty period for minimum 05 years. Declaration of CMC must be submitted.
- Ø Supplier must attend the call for maintenance of any defect of equipment/materials within warranty period.
- Ø Defects in equipment/materials must be replaced/repaired within seven days and equipment must be ok.
- Ø Bidders may submit leaflets and details specification of materials to be supplied. If required Virtual Demo by link must be given at the time of technical bid evaluation and acceptance. If any bidder fails to show demo, his bid or quoted items may not be accepted.
- Ø Bidders may also show physical demo of quoted items as per intimation.
- Ø MSME certificate will be accepted for exemption of EMD if bidder is the manufacturer of the bid items otherwise bidder will deposit EMD. If OEM, quoted item name must be mentioned in the MSME certificate.**

Ø **The authority reserves the right to accept/reject and cancel any bid.**

## **SPECIFICATIONS**

### **1. ICU Bed**

#### **(Hospital ICU Bed Electric 6 Function X-Ray Compatible)**

Frame work made of Rectangular M.S Tube, 4 Section Top made of perforated M.S sheet | Back Section X-Ray Translucent, Size 2 120\*1000\*460- 720mm (LxWxH) S.S Telescopic I.V Rod with 4 locations

Holder for Urine Bag | ABS Duck type safety Railings Heavy ABS Head and Foot Panel with locking Positions Obtained by Hand and Remote : Hi-Low, Backrest elevation, Footrest elevation, Trendelenburg/Reverse Trendelenburg Electric CPR for zero positioning of bed with cable remote control Positions obtained/Functions on Nurse Control System(ACP)- Hi-Low with locking function for handset, Backrest elevation with locking function for handset, Knee rest elevation with locking function for handset, Trendelenburg/Reverse Trendelenburg with locking function for handset, One button Electric CPR Position and Manual CPR function for any Emergency case, One Touch Chair Position | Finish-Pre-Treated & Epoxy powder coated frame mounted on 125mm Luxury twin castors with central locking system

### **2. Ward care Bedside Locker (S.S. Top) with drawer and cabi**

#### **net door**

Over all approx size: 410 mm x 410 mm x 820 mm H without buffer ( $\pm 10$  mm Engineering Variation)

Over all approx size: 410 mm x 420 mm x 820 mm with Buffer. ( $\pm 10$  mm Engineering Variation)

Body consisting of 2 sides and back are made from one piece made of 1.0 mm (20 G) CRCA MS sheet.

Top made from 12 mm mdf shall be fitted with superimposed stainless steel ( SS 304 ) sheet having thickness 0.6 mm (24 G) with raised edges 13 mm height on three sides.

One drawer is provided below the top of size 100 mm H x 350 mm W x 390 mm D approx fitted with very smooth slides. ( $\pm 10$  mm Engineering Variation)

Under the drawer is an open storage space and below it is a closed-door cabinet.

The size of open storage is 390 mm L x 400 mm W x 190 mm H. ( $\pm 10$  mm Engineering Variation)

The size of closed door cabinet is 390 mm L x 400 mm W x 360 mm H ( $\pm 10$  mm Engineering Variation)

Drawer and cabinet is made of 1.0 mm (20 G) CRCA MS sheet.

Doors of the cabinet is made from MS sheet having thickness 1.00 and by press bend operation. Door is provided with ball lock. Door of the cabinet box is pivoted at top with Pin and bottom with Screw.

Base of the Locker is fitted with twin wheel four swivel castors of 50 mm dia made from synthetic material without brake.

Two buffers shall be provided at rear side of the locker box. Safe Working load is 30 Kgs.

M.S. tubular parts, linkages, flats are to be In-house, pre-treated and Epoxy powder coated with coating thickness 50 to 100 microns.

### **3. Height Adjustable Bedside Table**

Overall approx size of top should be 960 mm L x 420 mm W from handle to base ( $\pm 10$  mm Engineering Variation)

Top size is 810 mm x 355 mm. ( $\pm 10$  mm Engineering Variation)

M.S. Tubular telescopic stem made from MS CRCA sheet 2.0 mm (14 G) with hardened EN grade material gear and S S handle with HDP bush for grip for height adjustment from approx.: minimum 715 mm to maximum 1080 mm. ( $\pm$  10 mm Engineering Variation)

M.S. Tubular stem is connected with base with the help flat size 57 mm x 57 mm x 6 mm thick having thread size M8 & bolts.

MS flat size 100 mm L x 50 mm W x 5 mm thick is welded at the bottom of the base at telescopic stem area for strength.

Two sections top shall be fixed on frame work made from 19 mm Square ERW tube and MS HR flat 20 mm x 5 mm and 170 mm L x 50 mm W x 6 mm thickness. ( $\pm$  10 mm Engineering Variation)

Size of Fixed section is 200 mm x 355 mm. Size of adjustable section is 610 mm x 355 mm ( $\pm$  10 mm Engineering Variation)

Bigger section of the table top should be hinged & could be inclined to three raised position options. The raised position is achieved by three slot provided in MS flat size 265 mm x 6 mm and lever made from MS rod diameter 6 mm. The one end of the rod is covered with red color PVC sleeve.

PVC Extrusion on one side of bigger section.

Railing S.S (304 grade) on small section covering three corner.

Small section is mounted on the telescopic stem with help of MS "U" shape external size 120 mm x 96 mm x 6 mm with 19 mm square tube.

Weighing bearing capacity 20 Kgs.

Base of the adjustable table made from M.S rectangular tube of size 60 mm x 30 mm x 1.6 mm and 40 mm x 20 mm x 1.6 mm and fitted with four castor wheel dia 50 mm made from anti-rusting material, two with brake and two without brake.

M.S. tubular parts, linkages, flats are to be In-house, pre-treated and Epoxy powder coated with coating thickness 50 to 100 microns.

#### **4. Saline Stand with SS Rod & Castors**

Adjustable height - 1395 mm to 2290 mm. ( $\pm$  10 mm Engineering Variation)

Strong M.S tube diameter 31.75 mm x 1.2 mm (18 G) thick mounted on five pronged rectangular 20 mm X 40 mm X 1.6 mm (16 G) thick.

Saline Stand fitted with five swivel rust proof synthetic body castor 50mm diameter w/o brake.

Saline Stand is provided with S.S 304 rod 12 mm diameter with two hooks, one-star Knob at Top and PVC cap at bottom.

The stand distance between the ages of the saline rod and any of the castor wheel bush is 235 mm.

Weighing bearing capacity 5 kgs per Hook Load bearing capacity: 20 kgs

M.S. tubular parts, linkages, flats are to be In-house, pretreated and Epoxy powder coated with coating thickness 50 to 100 microns.

All SS parts are of 304 grade with buffed and matt finished.

## 5. Four-legged Bedside stool

Overall Approx. Size:460mm to 670 mm H. ( $\pm$  10 mm Engineering Variation), Tubular Quadrilateral base of 25.4 x 14 g ERW M.S tube with revolving SS top. Height adjustment by screw. 300mm dia

The legs fitted with 4 high quality PVC shoes with nylon reinforcement.

S. S Ring for footrest. All SS parts made of 304 grade with buffed and matt finished. M.S. tubular parts, linkages, flats are to be In-house, pretreated and Epoxy powder coated with coating thickness 50 to 100 microns.

Load bearing capacity 135 Kgs

## **6. Bio-medical waste Bin Small Set of 3.**

The trolley is made of CRCA Tubes on 100 mm Castor wheels

Foot Operating System for touch less opening of lid. 30-liter Capacity of each bin.

Bins can be interchanged as and when required. For 3 Bins Trolley: 1330 mm L X 320mm W X 700mm H.

## **7. Technical Specifications for Multipara Monitor**

### **(Integrated automated charting system with 12 Inches monitor for Wards)**

#### **Specifications**

1. Should have ECG, SpO2(Masimo), NIBP, Respiration and Temperature. Should have facility to display ECG, SpO2, NIBP, Respiration and temperature simultaneously
2. Monitor must be upgradable to minimally invasive continuous Cardiac Output, ETCO2, 2 x invasive Pressure, BIS & AGM
3. Should display at least 8 waveforms of selected parameters simultaneously.
4. Should have an inbuilt continuous battery backup of minimum 3 hours;
5. Display: Color TFT touch display size of 12" or more. The operation should be through full Touchscreen and Trim Knob.
6. Should have dual temperature monitoring either in Celsius or Fahrenheit.
7. Should have facility for displaying multi-screen configurations
8. Should be able to store & display at least **240 hours** of tabular & graphical trends of all parameters.
9. Should be suitable for monitoring adult & paediatric patients & neonate patients

10. The SpO<sub>2</sub> technology should be **Masimo** Rainbow SET & should sense hypotension, shivering & motion. The Monitor should also be upgradeable to display the following parameters in addition to SPO2: Sphb/Spoc and RRa.
11. Should have oscillometric Technology for measurement of NIBP with Auto, STAT and Manual modes.
12. The monitor must be upgradable to display early warning scores and PPV, should have a measuring range of 0-300 mm Hg, should have auto, manual and stat modes of operation, should have different patient type selection. The respiration rate should be calculated through Impedance method. Should be able to analyse advance arrhythmias & ST segment changes. Monitor should be upgradable to monitor 12 lead ECG with ST maps.
13. Monitor Should Be Upgradable to display 12 lead ECG through 5/6/10 Lead ECG Cable, should be able to give visual & audible alarms with three levels of volume adjustment.
14. Memory should not wipe off in case of monitor switch off/power loss Should have connectivity to Central station through Ethernet card or Wireless Connectivity. Monitor must be upgradable to connect with the Central monitoring station. Should be European CE/USFDA/BIS approved, Prices must be quoted with 5 years' comprehensive warranty and 5 year CMC.
15. Equipment should be supplied with compatible 5 and 3 lead ECG cable (1 Nos.), SpO<sub>2</sub> (Adult) (1 Nos.), NIBP Cuff with NIBP hose of Large Adult, Adult & Paediatric (2 each), Temperature probe (1 No.)
16. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model in the Technical Bid along with the compliance sheet as per technical specifications.
17. Patient monitor manufacturing companies must have integrated automated charting systems for general care monitors of the same make.
18. The charting system must have the facility to record and display data on a spot check or 5-6 hourly basis on a PC or Laptop It must display All vitals connected to a patient and their trend for the complete stay period regardless of the number of days.
19. The charting system must have the option to enter Physical and nursing Notes The charting system Must generate an Early warning score for any patients whose vitals are deteriorati

ng.

20. The charting system Must be upgradable to connect with the ventilator, monitor & syringe pump to convert ward /ICU/OT to completely paperless charting.
21. The firm must have a proven track record of more than 200 beds of charting system installation in India. List to enclosed while quoting The charting system must be USFDA /BIS approved A firm should mention all the pre-installation requirements in the technical bid.
22. Equipment Must be demonstrated with charting solution
23. The installation will be done by the supplier free of cost.

## **8. SPECIFICATION FOR SYRINGE PUMP**

- 1** Flow rate range should be 0.1ml/hr to 1200 ml/h with accuracy of  $\pm 1\%$  on mechanism  
Should work on the following syringe capacities 5, 10, 20, 30/35, 50/60 CC and should be compatible with min 100types of syringe
- 2** Infusion modes to be present ml/hr mode, Dose mode, Volume / time: 0.1 - 99.9 mL; 00h01 - 96h00.  
Should have the following Dose rate mode units: ng/h, ng/kg/min, ng/kg/h, microg/min, microg/h, microg/kg/min, microg/kg/h, mg/min, mg/h, mg/24h, mg/kg/min, mg/kg/h, mg/kg/24h, mg/m<sup>2</sup>/h, mg/m<sup>2</sup>/24h, g/h, g/kg/min, g/kg/h, g/kg/24h, mmol/h, mmol/kg/h, mmol/kg/24h, mU/min, mU/kg/min, mU/kg/h, U/min, U/h, U/kg/min, U/kg/h, kcal/h, kcal/24h, kcal/kg/h, mEq/min, mEq/h, mEq/kg/min, mEq/kg/h.
- 3** Should have drug library of minimum 3800 drugs categorized in user defined 19 categories with facility to set all infusion parameters like soft limit, hard limit, bolus dose etc.
- 4** Should have Soft and Hard limit for max. or min. Flow/Dosage rate that cannot be exceeded and is rejected by the pump.

7 Should have Drug Error Reduction System (DERS) for safe infusion

8 Should have facility to upload drug library simultaneously through a single interface in the station with up to 4 infusion pumps in a system with an external hardware

9 Should have direct bolus option with flow rate 50ml /hr to 1200ml/hr with increment of 50ml/hr along with programmable bolus with settable dose or volume/time

10 Should have settable KVO option ranging from 0.1 to 5ml/hr with feature to keep it off

11 3 modes Priming is required (Mandatory, not mandatory, or advised) with max flow rate: 1200 ml/hr.

12 Should have induction dose facility with setting of Dose / time: 0.1 - 99.9 units / 1 second - 24 hours rate auto-calculation

13 Night mode programmed manually or automatically in a variable time range is must to decrease the brightness of the screen.

14 Fast start option is mandatory with pause option programmable from 1min to 24hrs

15 Variable and 3 pre-set levels pressure mode is must. Range from 50 to 900 mmHg. (25 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 250 to 900 mmHg). Can be enabled / disabled and adjusted with facility to display real-time inline pressure in both digital and analog form.

16 The Dynamic Pressure System with maximum and minimum threshold setting is mandatory.

17 Anti-bolus system is required

18 On Screen Graphical display of the following history "Volume / dose infused, pressure, flow rate" must be present.

**19** Should save 1500 data log events in real time and should have graphical history of Volume/ dose infused, pressure and flow rate

**20** The device should have Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection.

The device should have the following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, patient line disconnection, end of infusion pre-alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or key padlock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, plunger disengaged alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad), preventive maintenance.

**21** The device should have pusher guard for syringe protection.

**22** Should have LCD display of size equal or more than 66 mm x 33 mm

**23** Should have Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole

**24** Should have Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 11 h at 5 mL/h.

**25** The device should have RS232 Communication port with HL7 compatibility.

**26** Should have option of self-stack ability of min 3 pumps

The device should have Docking station to fit in 4 or 6 pump

**28**

Docking station should be able to communicate to HIS and should be HL7 compliant

HL7 compatibility should be verified online.

**29**

### **The device should Compliance with the following**

- a** Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24
- b** Medical Device Regulation - CE (Pursuant to Regulation (EU) 2017/745 on Medical Devices)
- c** Electrical Compliance - Protection against leakage current: Defibrillation-proof type CF and Protection against electric shocks: class II in accordance with IEC 60601-1
- d** Alarm system - IEC 60601-1-8
- e** Usability Engineering - IEC 60601-1-6 and IEC 6236 6
- f** The device should be European CE as per new MDR or USFDA certified

### **9. Laryngoscope**

LED Laryngoscope 4 different size blade Adult

## **10. Infrared Thermometer**

Precious Measurement, Contact -free measurement of forehead LED Display light, Red Light Indicates Fever Above 38 ° C

Yellow Light Indicates Low fever (37° C -37.9° C) Automatic Shutdown Function, Power DC3v battery operated (AAAX2).

## **11. Ambu Bag Adult & Paediatric**

- a) Ambu Bag 1600 ML silicon with mask
- b) Ambu Bag 500 ML silicon with mask

## **12. Aneriod BP Apparatus**

## **13. Electronic Weight Scale (Adult)**

Maximum Capacity 180 Kg 70X28 mm LCD Size, LCD Backlight Low /Error Indicators. SWS201

## **14. Electrical Suction Apparatus**

## **15. STRETCHER ON TROLLEY (Fully S.S.)**

Overall size: 2050 mm L x 570 mm W x 820 mm H. ( $\pm 10$  mm Engineering Variation) Mattress Platform: 1810 mm L x 560 mm W. ( $\pm 10$  mm Engineering Variation)

SS Frame work of Trolley is consisting of vertical tube size diameter 31.75 mm x 1.2 mm (18 G) thick, with reinforced at bottom with diameter 34.92 mm x 1.2 mm (18 G) thick tube for fitting castors. The Frame work is mounted on 200 mm castors two with brakes and two without brakes.

All horizontal stays are made of SS tube diameter 25.4 mm X 1.2 mm (18 G). SS flat size 32 mm x 5 mm is welded to frame work to support stretcher.

Removable Stretcher Top frame made from SS tube diameter 25.4 mm x 1.6 mm (16 G)

thick and diameter 15.8 mm x 1.2 mm (18 G) thick and top made from 1.2 mm (18 G) SS 304 sheet.

Handle is made of SS tube size diameter 25.4 mm x 1.2 mm (18 G) covered with PVC black color sleeve.

Three additional flat supports made from SS flat size 32 mm x 5 mm should be welded to support SS top from underneath widthwise.

Four stump legs made of 25.4 mm x 1.2 mm (18 G) ERW tube shall be welded at the bottom of the removable stretcher frame and should be provided with PVC material having nylon reinforced.

All SS parts made from SS 304 grade material and finished with Matt polish. Safe working load of 135 kgs and patient load bearing capacity of 130 kgs Supplied in SKD condition.

All Process Parameters to be as per documented IMS Procedures for Quality Assurance (ISO 9001:2015, ISO 14001:2015, ISO 45001:2018 & ISO 13485:2016 Quality Management Systems).

### **Accessories:**

Safety belts

Rexene covered Mattress 50 mm (2") thick with single section. Mattress Dimension: 1830mm L X 560 mm W ( $\pm$  5 mm)

S.S I.V. Rod 12 mm diameter with 2 Hooks

### **16. FOLDABLE WHEELCHAIR**

Overall dimensions: 1070 mm L x 635 mm W x 935 mm H. ( $\pm$  10 mm Engineering Variation) Seat width: 460 mm W ( $\pm$  10 mm Engineering Variation). Folded width: 290 mm ( $\pm$  10 mm Engineering Variation) Provided with Ergonomic handgrip.

Provided with Comfortable backrest

Provided tubeless Polyurethane tyres for durability, firm grip & safety brake with lever. Provided Lightweight aluminum propelling ring for enhanced grip.

Provided detachable armrest with cushion. Provided Antibacterial Fabric.

Provided Non Corrosive smooth finishing metal frame.

Provided Detachable and adjustable footrest with 90° swing on either side for easy patient. Provided adjustable and folda

ble footplates.

Rear wheel size 610mm.

Patient bearing and Weighing bearing capacity 120 Kgs. Provided with patient safety belt.

## **17. Crash Cart**

Overall Size: 1090 mm L x 400 mm W X 1650 mm H.

The crash cart should have 25.4 mm x 1.2 mm (18 G) Mild steel tubular frame work.

The emergency equipment cart should have the following facilities: 12 Nos. hand out bins to keep important supplies easily accessible of size approx 110 mm W x 125 mm D x 75 mm H. Two lockable box units made of high impact polystyrene with 3 drawers should have dimension 305 mm L x 380 mm D x 320 mm H.

The three drawers each to hold emergency medicines, ambu. bags, IV solutions, etc. separately and methodically in boxes of sizes approx. - 85 mm W x 85 mm L x 50 mm H - 6 nos., - 85 mm W x 175 mm L x 50 mm H - 3 nos. and 85 mm W x 255 mm L x 50 mm H - 4 nos.

Facility to carry monitors, ECG, suction apparatus on open areas at top center and bottom shelves.

Stainless steel saline rod made of 10 mm dia. 304 grade S. S. approx. 750 mm long and bent at top to have an arm of 400 mm approx. at the end of which of 6 mm dia. S.S. hook shall be welded with TIG process.

Crash cart with 125 mm dia non-rusting castor two with brakes and two without. Castor made from high grade non floor-staining synthetic materials with integrated thread guards. Wheel center having precision ball bearing to run smoothly. Provided with two round nylon stopper at the backside.

The size of middle shelf is 620 mm L x 390 mm W made from SS 304 0.9 mm (20 G) sheet. Bottom shelf is 670 mm L x 380 mm W made from sheet M.S. 1.2 mm (18 G). The shelves are provided with railing on three sides. This railing is made from M.S. Rod of diameter 9.5 mm. The size of top shelf is 670 mm L x 235 mm W made from M.S. 1.2 mm (18 G) sheet. Two side support size 237 mm W and 75 mm made from M.S. 1.2 mm (18 G) sheet provided to top shelf.

Pullout cardiac massage board made of laminated formica

of minimum size 675 mm x 330 mm x 8 mm. it shall have water resistance property and it should be made from ecofriendly material.

Oxygen cylinder cage epoxy powder coated, on one side.

Handle for pushing the crash cart is made from M.S. tube size 22.22 mm x 1.2 mm (18 G) and M.S. tube size 19.02 mm x 1.2 mm (18 G), provided on other side. Safe Working Load & Patient bearing capacity - 50 kg.

Supplied in KDC condition.

All stainless steel wherever used should be 304 grade. S.S parts finished with Matt Polish. M.S. tubular parts, linkages, flat s aluminium base are to be In-house, pretreated and Epoxy powder coated with coating thickness 50 to 100 microns.

## **18. Dressing Trolley**

Overall size 76x48cmsx 81cms. Stainless steel Tubular construction. Stainless steel top & shelf.

Four side guard railing on both shelves. Mounted on 100 mm dia. Castors. with stainless steel bucket & basin & push handle

## **19. Drug Trolley**

Overall Size: 680 mm L x 410 mm W X 1570 mm H, 27" x 16" x 62" (L x W X H) ( $\pm$  10 mm Engineering Variation)

The trolley has the following facilities: 30 Nos. multicolored removable bins to keep important supplies easily accessible in which 12 nos. of large and 18 nos. of small bins Large bin dimension: 250 mm x 165 mm x 125 mm, 10" x 6" x 5" (L x W x H) ( $\pm$  10 mm Engineering Variation). Small bin dimension: 150 mm x 110 mm x 75 mm, 6" x 4" x 3" (L x W x H) ( $\pm$  10 mm Engineering Variation).

The size of top shelf is 670 mm L x 235 mm W made from SS 304 1 mm sheet. Two side support size 237 mm W and 75 mm H made from SS 304 2.0 mm (14 G) sheet provided to top shelf.

MS tray of size 680 mm L x 410 mm W at the bottom for storage. Tray frame made of MS SQ tube 30 mm x 30 mm x 2mm and MS Round tube of size 25.4 mm x 1.2 mm and MS sheet of size 620 mm x 350 mm x 2 mm.

Trolley frame made of MS SQ tube 30 mm x 30 mm x 2 mm, Fl

at size of 32 mm x 6 mm and MS round tube 15.87 mm 1.2 m m.

Trolley with 125 mm dia non-rusting castor two with brakes and two without. Castor made from high grade non floor-staining synthetic materials with integrated thread guards. Wheel centre having precision ball bearing to run smoothly. Safe Working Load - 50 kg. Supplied in KDC condition.

All stainless steel wherever used should be 304 grade. S.S parts finished with Matt Polish. M.S. tubular parts, flats & sheet are to be In-house, pretreated and Epoxy powder coated with coating thickness 50 to 100 microns.

## **20. Oxygen Trolley (Push Type)**

Suitable for 1320-liter gas cylinder.

MS tubular framework made of approx 25.4 mm x 1.2 mm (18 G) MS. Round tube fitted with two rubber wheels 125 mm dia.

Trolley provided with locking provision made of 6mm x 25mm thick M.S. flats. Trolley fitted with SS 304 base of 2.0 mm (14 G) sheet & epoxy powder finish.

M.S. tubular parts, linkages, flats are to be In-house, pretreated and Epoxy powder coated with coating thickness 50 to 100 microns.

## **21. Glucometer**

20-600 mg/dL test range 5second test time Memory storage 300 Results 35-50% Haematocrit Range With 25 strip.

## **22. Nebulizer**

Piston Type Compressor Maximum 2.5 Bar Pressure Maximum 8LPM Airflow 220V/90Hz,50-60 Hz Power supply <65w Power Consumption

ISO & CE CERTIFICATE AVAILABEL

## **NEBULISER Technical specifications**

Nebulization pressure	$\geq 45$ kPa
Particle size	1 - 5 $\mu\text{m}$ > 60 %
Nebulization performance	$\geq 0,2$ ml/min
Medication container capacity	10 ml
Noise Level	$\leq 60$ dB
Usage	Inhalation of medicinal liquid for cold, bronchitis, asthma and allergies
Power supply	power adapter
Weight of the main unit	215 g
Size of the main unit	125 mm x 100 mm x 62 mm
Scope of delivery	main unit, nebulizer kit, child and adult mask, mouthpiece, tube, 5x air filter, power supply, storage bag, instruction manual

## 23. Stethoscope (Littman)

Class-III

## 24. X Ray View Box

### LED X-RAY VIEW BOX - SPECIFICATIONS :

Parameter	DUAL SCREEN
Dimensions (mm)	Frame (L x H x T)
	Viewing Area (W x H)
Power Consumption	30W
Light Source	LED
Power Supply Input	AC100-240V 50/60Hz
Brightness (LUX)	>10,000
Net Weight (Kgs)	

## 25. Air Bed

220-230v/50HZ power supply 5-7 liter/Min Air output capacity 70-110 mm Hg pressure range 7W power consumption. ISO & CE CERTIFICATE AVAILABEL

## 26. Technical Specification of Bipap Machine

S.No	Item Name and tender description
1.	<b>Intended Use:</b>

	Should be used for Non-Invasive Positive Pressure ventilation
	Should be utilized for both paediatric and adult patients >18 Kgs with OSA & Respiratory Insufficiency patients
	Back-up rate option should be available with pressure support modes.
<b>2.</b>	<b>Detailed Description &amp; Technical Specifications:</b>
	Should be Light Weight, Portable, Low maintenance, Easy and Simple to operate
	Should be Electrically powers, Microprocessor controlled Blowers
	Should have a maximum pressure limit up to 30 cm of H20
	Should have default algorithm for automatic triggering and cycling sensitivity. No need to set trigger and cycle sensitivity manually.
	Provision of Hybrid ventilation options and comfort feature option available.
	Integrated Humidifier should be Heated Humidifier & single unit chamber
	Should have Provider Lock options for Clinical settings to avoid accidental change of clinical settings.
	Should have provision of Advanced features for improving Patient Adherence therapy.
	<b>Flex Therapy Options-</b> Exhalation Pressure relief option should be available
	<b>Check Mask fit</b> Option – to ensure proper mask fit to improve better clinical outcome.
	Provision for both <b>Adaptive</b> and <b>classic humidity</b> for better compliance.
	Should display the information of Usage hours, Mask fit and AHI for better adherence to therapy.

	Incorporated with SD provision for easy extraction of the therapy compliance data.
	<b>Performance Check feature should be available for easy trouble shooting.</b>
<b>3.</b>	<b>Modes of ventilation</b>
	<b>The following modes should be available in the device</b>
.	Cpap mode:
.	S or Spontaneous Mode
.	S/T or Spontaneous Timed Mode
.	PC or Pressure controlled Mode
.	T or Timed mode
.	<b>Hybrid Mode: Optional :VAPS with AutoEpap algorithm</b>
.	<b>VAPS - Volume assured Pressure Support can be enabled as a feature on all Bi-pap therapy Modes</b>
.	Should be recommended to use from 18 kg onwards
	<ul style="list-style-type: none"> <li><b>Auto- Epap Algorithm</b> – a default feature to treat patients with Overlap syndrome patients</li> <li>In Bi-Level Therapy Modes, Auto – Epap feature can be enabled or Disabled</li> </ul>
<b>4.</b>	<b>Monitoring</b> : Measured and display parameters
.	Pressure
.	Exhaled Tidal volume
.	Exhaled Minute Ventilation
.	Respiratory Rate
.	Estimated Leak Rate
.	Patient trigger indicator
<b>5.</b>	<b>Controls:</b>

	<ul style="list-style-type: none"> <li>· Cpap : 4 to 20 cm H20</li> <li>· Ipap : 4 to 30 cm of H20</li> <li>· Epap: 4 to 25 cm of H20</li> <li>· Breath rate - 0 to 30 cm of H20</li> <li>· Ti - 0.5 sec to 3 sec</li> <li>· Target Tidal Volume - 200 to 1500 ml (If AVAPS enabled )</li> <li>· Rise time should be minimum 100 ms to 600 ms</li> <li>· Ramp Time and Ramp starting pressure should be available</li> <li>· Avaps Feature can be enabled.</li> </ul>
<b>6.</b>	<b>Alarms &amp; System Messages:</b>
	<ul style="list-style-type: none"> <li>· Patient Disconnection alarm</li> <li>· Apnea alarm</li> <li>· Low Minute Ventilation alarm</li> <li>· Low Tidal Volume alarm</li> </ul>
<b>7.</b>	<b>Accessories to be supplied:</b>
	<ol style="list-style-type: none"> <li>1. Power cord with Power adapter</li> <li>2. Disposable or Reusable Tubing - 22mm</li> <li>3. Disposable Exhalation port</li> <li>4. Oxygen Enrichment port</li> </ol>
	<b>5. Silicon or Gel Reusable Full face masks</b>
	One or Two Standard size(S/M/L) Full Facemask
	<b>6. Nasal mask</b>
	<ol style="list-style-type: none"> <li>a. 2 medium size</li> <li>b. 2 large size</li> </ol>
	7. Upgradeable to SpO2 Module & Data Module

	8. Should have SD card or USB card slot for recording information for clinical study.
	9. Should be supplied with integrated Heated Humidifier with heated tube option.
	10. 2 Pollen air filters
<b>8. Standards:</b>	
	2 years comprehensive warranty
	CMC for 5 years after completion of warranty for spare parts / accessories used during maintenance
	The company should give the certificates that the model quoted is the latest and not obsolete; and spares will be easily available for next 5-7 years.
	Should have safety certificate from a competent authority CE issued by a notified body registered in European commission & FDA (US)

## **27. CONVENTIVE WARMER**

The Convective Warming Level 1 specifications include: a temperature range of ambient, 36°C, 40°C, and 44°C, a digital display showing the selected temperature, over/under temperature alarms, a high-efficiency air filter, a secure hose-end blanket connector, and the ability to be mounted on an IV pole, table, rolling cart, or floor.

Key points about the Equator Level 1 Convective Warmer:

Temperature control:

Offers multiple temperature settings, typically including ambient, 36°C, 40°C, and 44°C.

Safety features:

Equipped with over and under temperature alarms to alert healthcare providers if the temperature deviates from the set point.

Air filtration:

Incorporates a high-efficiency air filter to minimize airborne particulates delivered to the patient.

Hose connection:

Features a secure connection point for the warming blanket hose to prevent accidental disconnection.

Mounting options:

Can be mounted on various surfaces like an IV pole, table, rolling cart, or floor for flexibility in placement.

## अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for attached categories, trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ

with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.**

**For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.**

**The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:**

- All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and
- All operative provisions of the erstwhile Labour Laws until their complete substitution.

**All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.**

यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस नियिदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---धन्यवाद/Thank You---**