

दरपत्रक सुची

राष्ट्रीय आरोग्य अभियान, जिल्हा रुग्णालय सिंधुदुर्ग - राष्ट्रीय अंधत्व व दृष्टि क्षीणता नियंत्रण कार्यक्रमांतर्गत DH व SDH करिता उपकरणे साहित्य खरेदीकरिता दर निश्चित करावयाचे आहेत. तरी खरेदी करावयाच्या उपकरणे व साहित्याची माहिती खालीलप्रमाणे आहे.

Ophthalmic Equipment For DH

Sr. No	Name of Item & Specification	Purchase Qty	Per Qty Rate
1	Tonometer SPECIFICATION: High Quality Agate Bearing For An Extremely Long Service Life, Precision Measurement On A Scale Of 0 To 20 Subdivision & 0.1 Sub-Division , Perfect Reading Of The Scale With Red Pointer , All Vital Parts Are Made Of Stainless Steel Other Parts Are Chrome-Plated, Complete In A Black Deluxe Case With Velvet Look Inserts , Supplied With Three Weights 95.5g, 7.5g, 10g) & Conversion Table .	2	
2	Vision Drum - Wall Hanging Plastic Eye Testing Drum Chart Distance Vision Testing Dori Vision Drum Doori Drum Vision Eye Testing Chart Eye Vision Test Chart For Eye Check-Up.	10	
3	Indirect Ophthalmoscope TECHNICAL SPECIFICATION: Focus Distance-300 -800 mm, Inter Pupillary Distance 54-75 mm, Pupil size optics 1 mm, Illumination LED illumination , Field Good Clear circular, ILLUMINATION Spot size - 04.4mm, 030mm 012mm, Illumination area 080mm 060 mm 025mm illumination control- liner intensity control from 0-100%, Filters_ Cobalt blue gree (Redfree)	1	
4	Volk lens + 20 D Volk lens + 20 D - used for better evaluation of Fundus on IDO.	1	
5	Volk lens + 90 D Volk lens + 90 D- used for better evaluation of Eye on Slit lamp microscope.	1	
6	TOURCH STANDARD size LED RECHARGEABLE	4	
7	Simcoe No.22, No.23, No.24 (20 Each) (Stainless steel , ISO Standard)	60	
8	Irrigation cannula No.26 & No.27 (20 Each) (Stainless steel , ISO Standard)	40	
9	Utrata forceps (Stainless steel , ISO Standard)	5	
10	Steel Drum medium 9*11 SS (Stainless steel , ISO Standard)	4	

Ophthalmic Equipment For SDH

Sr. No	Name of Item & Specification	Purchase Qty	Per Qty Rate
1	Trial Lens Set WITH Trial Frame- Ophthalmic Trial lens set & frame Reduced aperture lens Metal rim with imported ophthalmic lens , upto 20D spherical concave & convex in standard steps. Standard ACCESSORIES LIKE Prisms- 10 pcs, Black ,Red, Green ,Slit, Pinhole , White Maddox. Box Portable Operation Mode- Manual Trial Frame adjustable adult & Child, Material- Stainless steel , Adjustable Individually - yes.	7	

Sr. No	Name of Item & Specification	Purchase Qty	Per Qty Rate
2	Direct Ophthalmoscope Specification- Ophthalmoscope self luminous battery operated with standard lens wheel & other accessories. It should have Indian standard specification or internationally accepted certification 1(one) spare bulb will be included in supply. Should have features of small pupil, macular grid, green filter, blue filter Diopter shift in steps of 1 D Range from -20D to +20D Rechargeable Battery Handle	1	
3	TONOMETER SPECIFICATION: High Quality Agate Bearing For An Extremely Long Service Life, Precision Measurement On A Scale Of 0 To 20 Subdivision & 0 to -1 Sub-Division, Perfect Reading Of The Scale With Red Pointer, All Vital Parts Are Made Of Stainless Steel Other Parts Are Chrome-Plated, Complete In A Black Deluxe Case With Velvet Look Inserts, Supplied With Three Weights 95.5g, 7.5g, 10g & Conversion Table.	4	
4	LED Vision Chart Specification- Display Type - 22" LED Backlight, Resolution - 1920*1080@Hz Chart Window size - 477*268mm, Contrast Brightness - 3000:1/250Cd/m2, Standard Chart Type- Letters, Numbers, Pediatric, languages, logMAR, Snellan Chart, Crowding row Range- 1ft to 20Ft (1Mtr to 6 Mtr) worth 4dot, Dot chart Chart change over device- Remote Control Distance of Installation - 1Mtr to 6 Mtr (step 1 Mtr)	4	
5	TOURCH STANDARD size, LED, RECHARGEABLE	13	

दरपत्रकांच्या अटी व शर्ती

- १) दरपत्रकात नमुद दर हे सर्व करांसहीत असावेत.
- २) दिनांक ०१/०१/२०२६ रोजी सायंकाळी ५.४५ पूर्वी जिल्हा सामान्य रुग्णालय, सिंधुदुर्ग येथील एनएचएम विभाग यांचेकडे सीलबंद लखोट्यामध्ये त्यावर राष्ट्रीय अंधत्व व दृष्टि क्षीणता नियंत्रण कार्यक्रमकरिता उपकरणाबाबतचे दरपत्रक असे नमुद करून सादर करावेत.
- ३) सीलबंद लखोटा दिनांक ०२/०१/२०२६ रोजी सकाळी ११.०० वाजता जिल्हा शल्य चिकित्सक, सिंधुदुर्ग यांचे कक्षात उघडणेत येतील. सादर करणारे दरपत्रक संबंधित पुरवठाधारकाने स्वतः अथवा आपले प्रतिनिधी यांनी दिलेल्या दिवशी वेळेत उपस्थित रहावे. उपस्थित राहिले नाही तर दरपत्रक उघडून न्युनतम दराने पुरवठा करणारे व्यावसायिक यांचे दर निश्चित करून स्विकृत करणेत येतील.
- ४) सदर स्पेसिफिकेशन प्रमाणेच सर्व उपकरणे व साहित्य असणे बंधनकारक आहे. तसेच साहित्य उत्तम दर्जा व गुणवत्ता पूर्ण असली पाहिजे.
- ५) पुरवठा आदेश प्राप्त झालेपासून दहा दिवसांच्या कालावधीमध्ये उपकरणे जिल्हा रुग्णालय, सिंधुदुर्ग राष्ट्रीय अंधत्व व दृष्टि क्षीणता नियंत्रण कार्यक्रम कक्ष येथे पुरवठा करणेचे आहे तसे न केल्यास प्रति आठवडा ०.५: दंड आकारण्यात येईल व तसे न केल्यास संबंधित पुरवठा आदेश रद्द होतील व स्विकृत दर आदेश संपुष्टात येतील.
- ६) पुरवठा केलेले उपकरणे साहित्य निकृष्ट दर्जाचे असल्यास त्याकरीता साहित्य पुरवठा विनाखर्च करावा लागेल.
- ७) उपकरणे साहित्य प्राप्त झाल्यानंतर संबंधित विभागाचा उपकरणे उत्तम दर्जा व गुणवत्ता पूर्ण असलेबाबतचा दाखला दिल्यानंतर देयकांची अदायगी करणेत येईल.
- ८) दरपत्रकासोबत व्यवसाय नोंदणी प्रमाणपत्र /GST Registration Certificate व पॅन कार्ड इ. कागदपत्रे जोडणेत यावीत.


 Civil Surgeon
 जिल्हा रुग्णालय, सिंधुदुर्ग

