

दुरध्वनी क्रमांक कार्यालय: ०२५३ २५७२०३८, २५७६३६८ औषध भांडार-२३१९३६५ फॅक्स: २५७७९४९इ मेल medstoreghn@gmail.com	कार्यालय - जिल्हा सामान्य रुग्णालय ,त्र्यंबक मार्ग,हॉटेल राजदुत समोर, नाशिक ०२.
आरोग्य सेवा	जा.क्र.जि.रु.नाशिक/ओभा/जिरुना/ई दरपत्रके/ 10401 /२६ जिल्हा शल्य चिकित्सक,जिल्हा रुग्णालय,नाशिक दिनांक- 103/2026 23 MAR 2026

Web Site Quotation Notice No. 02/2026

Date-23.03.2026.

Open Notice

Civil Surgeon, Nashik invites quotations (**2 Envelope System**) to purchase of following Items from eligible Supplier. The Supplier who is interested please see terms & Conditions attached with & Submit the Quotation in prescribed Manner. Use Separate Envelopes for Technical Bid & Financial Bid.

Sr.No	Name & Description of Medicine	Unit
01	Ophthalmic Operation Table Steel	01
02	Distant LED Vision Chart	01
03	Phaco tip with sleeve AMO compact Sovereign silver 19G 30degree	01
04	Mini Trial Box for refraction	01
05	Hydro canulae angled 26G	01
06	Lieberman speculum	01
07	Curved Scissor 4-inch SS	01
08	Simcoe canulae 22 guage direct	01
09	Corneal scissors	01
10	Lims forceps	01
11	Autoref with Keratometer	01

Submission :-

1	Submission of quotation by Hand Delivery or her/his own Risk by post or Courier before Last Date	Last Date. 30.03.2026 Time Before- 4 PM
2	Opening of Quotation	Date- 30.03.2026. Time -5 To 6 PM Place-Civil Surgeon Office G.H. Nashik

Terms & Condition :-

1	Rates	Including all Taxes
2	Delivery	Medical Store, Civil Hospital, Nashik
3	Acceptance of Rates	Minimum 3 Quotations are required for comparative Rates. Lowest rates will be accepted.
4	Delivery	10 Days from the date supply order If Unable to supply within stipulated period penalty will be deducted as per Govt. Rule
5	Payment	CMP/NEFT/Cheque

6	Self-Attested Document (Technical Bid)	Supplier Should submit 1.Shop Act License 2.GST registration certificate 3.PAN Card Copy 4.Annexure 1,2 on 5.Drug license
7	Rate Format (Price Bid)	To Be Prepare on Letter Pad only Duly Signed by supplier with Name & Rubber Stamp. DO Not write rate in handwriting or overtyping or use of Whitener & Use Separate Envelope For Price Bid

Disqualification of Quotation: -

- 1.Failure of Required Supplier Technical qualification.
2. Late Receipt of quotation envelope.
- 3.Rate format submission not in proper format.
- 4.If Quotations received only in one envelope then will be disqualified.

(Use 2 Seperate Envelopes for Technical Documents & Price Bid & Mention it on envelopes)

Summay
Civil Surgeon, Nashik

(On Self letter head)

Annexure -1

DECLARATION BY SUPPLIER

I/we herewith declare that, I/we have not quoted rate in this quotation greater than MRP or market rate. I/we have not quoted blacklisted Mfg. Company in this quotation. I/we or our firm employee are not related with Civil Surgeon, Nashik or their organizational Person.

मी/आम्ही असे जाहीर करतो की या दरपत्रकामध्ये किमान मुल्यापेक्षा अधिक दर नमुद केलेले नाहीत अथवा बाजारभावा पेक्षा अधिक दर नमुद केलेले नाहीत. या दरपत्रकात नमुद करणेत आलेली उत्पादक कंपनी ही काळ्या यादीतील नाही. मी किंवा माझे व्यवसायातील नोकर वर्ग यांचा जिल्हा शल्यचिकित्सक, नाशिक किंवा त्यांचे आधिपत्याखालील संस्था या मध्ये कोणतेही नाते वा हितसंबंध नाहीत.

Place-

Date-

Name, Signature Of Supplier

Seal & Rubber Stamp

(On Self letter head)

Annexure -2

निविदाकाराचे हमीपत्र

मी/आम्ही मे. या हमीपत्रा द्वारे लिहून देतो की, दरपत्रक मागविणाऱ्या खरेदी प्राधिकार्या बरोबर कोणत्याही प्रकारचे हितसंबंध नसून हितसंबंधा बाबत संघर्ष नाही. तसेच खरेदी प्राधिकार्याकडे सादर करण्यात करण्यात आलेले दरपत्रक हे एकल असून दुसऱ्या कोणत्याही संस्थेसोबत संयुक्तरीत्या किंवा संगणमताने साखळी करून दरपत्रक भरलेले नाही, असे आढळून आल्यास नियमानुसार योग्य त्या दंडात्मक कार्यवाहीस पात्र राहिल.

दिनांक:

ठिकाण:

दरपत्रक सादर कर्त्याची स्वाक्षरी व शिक्का

Rate Format

To,
The Civil Surgeon,
Nashik

Sub-Submission of Quotation

10401

Ref-Your Office Quotation Notice No. जा. क्र.जिरुना/औ.भांडार/दरपत्रके/ /२०२६
दि.२३/०३/२०२६.

Resp Sir/Madam

With ref to above subject I/We are here with submitting quotation for Govt.
Hospital Purchase.

Sr.No	Name & Description of Medicine	Unit	Unit Rate	Mfg By
01	Ophthalmic Operation Table Steel	No.		
02	Distant LED Vision Chart	No.		
03	Phaco tip with sleeve AMO compact Sovereign silver 19G 30degree	No.		
04	Mini Trial Box for refraction	No.		
05	Hydro canulae angled 26G	No.		
06	Lieberman speculum	No.		
07	Curved Scissor 4-inch SS	No.		
08	Simcoe canulae 22 guage direct	No.		
09	Corneal scissors	No.		
10	Lims forceps	No.		
11	Autoref with Keratometer	No.		

1. Rate with Inclusive all taxes
2. For Destination
3. Delivery 10 Days
4. Payment 100 % after Supply of Item
5. As per terms and conditions attached.

Name, Signature of Supplier

Seal & Rubber Stamp


WEB SITE DOCUMENT UPLOAD FORM (WDU)

Name of Program	DISTRICT HOSPITAL NASHIK		
Name & Designation of Officer	CIVIL SURGEON NASHIK		
Date of Publication	23/03/2026		
Displayed on Website (Please { } checkbox)	<input type="checkbox"/> www.nrhm.maharashtra.gov.in	<input type="checkbox"/> www.arogya.maharashtra.gov.in	<input checked="" type="checkbox"/>
Published on section/tab of website			
Document Title should be displayed as (in Marathi)	DPDC TRIBLE		
Document Title should be displayed as (in English)	Quotation for DPDC Ophthalmic Instrument		
Document Size (Document should be in PDF format and size will not exceed more than 20MB)			

Declaration

I hereby declare that all information provided in this website document upload form (WDU) for the purpose of uploading/updating document on website only and correct to the best of my knowledge. All documents responsibility will be on concern program department only; IT NHM department is not responsible for any breach cause to content of uploaded documents. IT NHM deparment is responsible only for uploading/updating documents on websites

Date: 23/3/2026


Signature of Program Authority

Sr. No. 03

SPECIFICATION PHACO TIP

SR NO	NAME	Technical Specification
1	phaco tip with sleeve AMO compact sovereign	PHACO TIP JOHNSONG AND JOHNSON ^{2.2} 2.8MM WITH ^{2.2} 2.8 SLEEV 6 PIECES FIT TO AMO COMPACT MACHINE



Medical Officer (C-1)
Civil Hospital Nashik


Additional Civil Surgeon
Civil Hospital Nashik


Civil Surgeon
Civil Hospital, Nashik

Sr. No. 05

SPECIFICATION HYDRO CANULAE 26G		
SR NO	NAME	Technical Specification
1	Hydro canulae angled 26g	HYDRO CANULAE ANGLED 26 GUAGE STAINLESS STEEL REUSABLE AUTOCLAVBLE SAND BLASTED


Medical Officer (CI-1)
C. H. Nasik
Medical Officer Class -1
Civil hospital Nasik


Additional Civil Surgeon
Civil Hospital Nashik


Civil Surgeon
Civil Surgeon, Nashik
Civil Hospital, Nashik

Sr. No. 06

SPECIFICATION LIBERMAN SPECULAM

SR NO	NAME	Technical Specification
1	Liberma speculum	LIBERMAN EYE SPECULAUM STAINLESS STEEL AUTOCLAVABLE NONRUSTING SANDBLASTED 3 INCH SIZE WITH SCREW


Medical Officer (CI-1)
Medical Officer Class - I
Civil Hospital Nashik


Additional Civil Surgeon
Additional Civil Surgeon Nashik
Civil Hospital Nashik


Civil Surgeon
Civil Surgeon Nashik
Civil Hospital, Nashik

Sr. No. 08

SPECIFICATION SIMCOE CANULAE 22

SR NO	NAME	Technical Specification
1	simcoe canulae 22, guage direct	Simcoe canulae direct 21guage stainless steel with silicon tube autoclavable rust free sand blasted

Medical Officer (CI-1)
C. H. Nashik

Additional Civil Surgeon
Civil Hospital Nashik

Civil Surgeon
Civil Hospital, Nashik

Sr. No. 10

SPECIFICATION LIMS FORCEPS

SR NO	NAME	Technical Specification
1	Lims forceps	lims or corneal forcep 3inch stainless steel2*1 teeth, antirust sandblasted

Medical Officer (CI-1)
C. H. Nashik

Additional Civil Surgeon
Civil Hospital Nashik

Civil Surgeon
Civil Hospital, Nashik

Sr. No. 11

SPECIFICATION AUTO KERATOMETER

SR NO	NAME	Technical Specification
1	Auto Keratometer	<p>Refraction measurement :- Sphere (Sph):- -30D to +60D,0.12D/0.25D Steps Cylinder (cyl) :- -80D to +8D,0.12D/0.25D Steps Axis :- 0 to 180degree,1 steps Pupil distance 45 ~ 85mm,0.5mm steps Vertical mdistance 0m,12mm,13.75mm Minimum pupil diameter 02.0mm Charter Automatic foggingmethod Display 5.7inch true color LCD Print important thermal printer <i>K Reading 40D To 60D</i> Power supply :- 100~ 120v/200~240V,50/60HZ Dimention 275(W) *509(D)*430-460(H)mm</p>

sh

Medical Officer (GL-1)
C. H. Nashik

[Signature]

Additional Civil Surgeon
Civil Hospital Nashik

[Signature]

Civil Surgeon
Civil Hospital, Nashik

Technical Specification of Simple Operation Table / Ordinary Operation Table

1	Clinical purpose	An operating table, sometimes called operating room table, is the table on which the patient lies during a surgical operation. This surgical equipment is usually found inside the surgery room of a hospital.
2	Used by clinical department/ward	Operation theatre
3	Technical characteristics specific to this type of device)	<p>01) Should have Stainless steel top 18 SWG. SS 304 grade.</p> <p>02) Over all Size 1830mm x 560mm x 760mm (with variation of 50mm)</p> <p>03) Kidney Bridge.</p> <p>04) Removable & folding foot end with S.S. drainage tray.</p> <p>05) Height adjustable & head low / up (T & TR) position bt screw & lever mechanism.</p> <p>06) Lateral tilt position with belts to hold the position.</p> <p>07) Lithotomy Crutches.</p> <p>08) Gear up type angular arrangement for lithotomy bars.</p> <p>09) Suitable mattress 25 mm thick should be tear proof covered with non pinching Rexine, seamless joint, washable and water-proof, having welcrow for fixing on table top.</p> <p>10) Aluminum top antiskid step stool having two steps. & Arm Boards with cushion.</p> <p>11) Stainless steel removable side railings.</p> <p>12) Should have Castor wheel for easy mobility, foot lock to floor.</p>
4	User's interface	Manual
5	Weight (lbs, kg)	Should be able to bear patient having weight up to 180 kg
6	Accessories (mandatory)	<p>1) Side rail clamp.- 01 No</p> <p>2) Shoulder support.</p> <p>3) Arm support (Qty 2 Nos)</p> <p>4) I V pole</p> <p>5) Body restraining belt</p> <p>6) Leg supports :(Qty 2 Nos)</p> <p>7) Lateral supports</p>
7	Atmosphere / Ambiance (air conditioning, humidity, dust ...)	<p>1) Operating condition: Capable of operating continuously in ambient temperature of 10 to 40° C and relative humidity of 15 to 90% in ideal circumstances.</p> <p>2) Storage condition: Capable of being stored continuously in ambient temperature of 0 to 50° C and relative humidity of 15 to 90%.</p>
8	User's care, Cleaning, Disinfection & Sterility issues	<p>1) Disinfection: Parts of the Device that are designed to come into contact with the patient or the operator should either be capable of easy disinfection or be protected by a single use/disposable cover.</p> <p>2) Sterilization not required.</p>

Ramesh
13/6/19
K. Ganesh
K. Ganesh

Dr. Vinayak
Dr. Vinayak
Dr. Vinayak
13/6/19.

Amol Wadke
Amol Wadke
Biomedical Engineer
R.R.G.H. Amal

Pradeep
Pradeep
विद्यार्थी विकास, संशोधन एवं
विभागीय प्रबंध सेवा समिति

9	Certificates	1) Should be FDA/ CE approved product. 2) Manufacturer should have ISO certificate for quality standard.
10	Training of staff (medical, paramedical, technicians)	1) Training of users on operation and basic maintenance. 2) Advanced maintenance tasks required shall be documented
11	Warranty	Three Years
12	Operating manuals, service manuals, other manuals	Should provide 2 set of User, technical and maintenance manuals to be supplied in English/Marathi language along with diagrams.
13	Recommendations or warnings	Any warning signs would be adequately displayed

[Signature]
Dr. N. K. Kulkarni
Thane Civil Hosp
13/6/19.

[Signature]
Amol Wadkar
Biomedical Engineer
R. R. S. H. Amravati

[Signature]
डॉ. बाबुल गोडसे
मुख्य वैद्य, अमरावती तह
विभागीय संदर्भ सेवा केंद्राला
अमरावती.

[Signature]
Dr. V. N. N. N.
A. M. C. Mumbai

[Signature]
Dr. G. G. G.
12/6/19
K. S. Hospital
Mumbai