

GOVERNMENT OF MAHARASHTRA



**CIVIL SURGEON,
GENERAL HOSPITAL, CHANDRAPUR
NEAR CHOTA BAZAR, MAIN ROAD, CHANDRAPUR-442402**

E-mail – cschandrapur@gmail.com

Phone : 07172-264640, 252103

Fax : 7172-264640

Quotation for supply of
EMERGENCY INSTRUMENTS & MATERIALS

Not Transferable


Civil Surgeon
General Hospital, Chandrapur

Quotation reference No:

CSC/Quotation/Emergency Instruments & Materials/E-Quotation

CSC/Quotation/ Emergency instruments & Materials /E-Quotation

**CIVIL SURGEON, GENERAL HOSPITAL, CHANDRAPUR
(MAHARASHTRA STATE)**

CSC/Quotation/ Emergency Instruments & Materials /E-Quotation

Civil Surgeon, General Hospital, Chandrapur invites Online E-QUOTATION in two envelope system from the reputed and experienced Manufacturers / supplier / distributor for purchase of Eme Medicine

Interested eligible Supplier may obtain further information of name of items technical specifications, required quantities and other terms and conditions applicable for procurement of above items from Office Of The Civil Surgeon, General Hospital, Chandrapur.

QUOTATION SCHEDULE

All bid related activities (Process) like Download Quotation Document, Bid submission & other documents will be governed by the time schedule given under Key Dates below:

S. No.	Activity	Date & Time
1	Date of Release & uploading of Quotation Documents	08/05/2026
2	Last date of submission of Technical & Financial Proposals (Only Submitted Physically)	15/05/2026 Upto 5:00pm
3	Date of opening of Technical Bid	15/05/2026 5:30Pm

Address for communication

:Civil Surgeon,
General Hospital, Chandrapur-442402

Phone : 07172-264640, 252103

E-mail – chandrapurmedstore@gmail.com

The Purchase Commette at Civil Surgeon, General Hospital, Chandrapur level reserves the right to increase or decrease the quantity to be purchased and also reserves the right to cancel or revise or any of the all the Quotation or part of Quotation without giving any reasons.

**CIVIL SURGEON
GENERAL HOSPITAL, CHANDRAPUR**


Civil Surgeon
General Hospital, Chandrapur

TERMS AND CONDITIONS

1. Introduction :-

CIVIL SURGEON, CHANDRAPUR Hereinafter referred to as a "Purchaser" invites two Envelope systems for supply of item specified in **Annexure-A** Schedule of Requirements (**EMERGENCY Emergency MEDICINE**), for the use in health Institute in the District Chandrapur in State of Maharashtra.

1.1 Interested eligible Supplier may obtain further information of technical specification, required quantities and other terms and conditions applicable for procurement of item.

1.2 All Quotation related activities (Process) like Quotation Document, Quotation submission will be governed by the time schedule.

1.3 Purchase Committee, Civil hospital, Chandrapur reserves the right about the quantities to be purchased without assigning any reason thereof.

1.4 If any Supplier wishes to lodge any complaint against the other Supplier regarding submission of false documents, information etc. the Supplier Has to Deposit Rs 10,000/- in the form of Demand Draft drawn in favour of Administrative Officer, General Hospital, Chandrapur payable at Chandrapur in Term of Deposit. The Issue will be Submitted to Purchase Committee along with fact. The Amount so deposited shall be refund if after Scrutiny the Complaint is found to be true by the Purchase Committee However, if the Complaint Found to be false & Malicious the Deposit will be Forfeited, No Interest Shall be paid Against this deposit. The Committee decision will be Final.

2. Eligibility Criteria : Quotation shall have Submit Bids in two Parts:-

A) Technical Bid (Envelope-1)

B) Financial Bid (Envelope-2)

A) Technical Bid :- (Eligibility Criteria for Quotation Bid) :- Suppliers should submit following documents in Quotation Bid.

1. Manufacturer/Sole Distributors, Wholesale Supplier who qualify the Quotation conditions and qualification criteria, are eligible to participate and submit their offer against this Quotation invitation.

2. Manufacturer in all the case must have own factory and Manufacturing in their own premises / Sole Distributors, Wholesale Supplier license or shop establishment registration issued by respective statutory authority.

3 PAN registration.

4 GST Registration Documents Attested Copies

5 Drug Licence

Note: Quotation are not allowed from manufacturer / Sole Distributors and Whole Sale Supplier, for the item(s) for which the firm found guilty of malpractice, misconduct, or blacklisted/debarred either by Public Health Department, Govt. of Maharashtra or by any local authority, Other State Government/Central Government's organizations.

CSC/Quotation/ Emergency instruments & Materials /E-Quotation

B) Financial Bid:-

1 Cost of bidding:-

The Supplier shall bear all costs associated with the preparation and submission of their Quotation and the Purchaser will in no case be responsible or liable for those costs, regardless of the conduct or outcome of the Quotation process.

2 Clarification of Quotation document:-

A prospective Supplier requiring any clarification of the Quotation document shall contact the Purchaser **by letter, or email 4 days prior to last date & time of closing sale of Quotation.** Email ID - chandrapurmedstore@gmail.com

3 Submission of Quotation:-

Supplier should be submitted on or before last date of submission.

Quotation should be submitted in two envelopes i.e. Technical Bid in envelop no.1 & Commercial bid in Envelop no. 2.

6.1 Late Quotation offers:-

Late Quotation on any count shall be rejected summarily. Delay due to Transport reason (for e.g. :electricity/internet/etc) will not be condoned.

6.2 Envelope No. 1 (Technical Bid):

Technical offer must be submitted along with following documents.

1. **Quotation Form** as per Annexure-1.(Duly Sign By the Supplier)
2. Attested photocopy of factory license / Manufacturing license / Distributor license/shoestablishment registration / Whole Sale Supplier license/NDPS Licence issued by respective State Government.
- 4 GST Registration Certificate
- 2 Attested Photo copy of PAN Card .
- 2 Affidavit on non-judicial stamp paper of Rs. 500/- that rates quoted in the Quotation are not higher than the rates quoted to other Govt. Departments /Govt. Undertakings or any prevailing rate contracts. And Also mentioning therein that the rates Quoted are not Exceeding the Rates Approved under DPCO/N.P.P.A. & M.R.P. if in future the rates are brought down by the Government or Competent authority the Affidavit Should Be Notrized
- 7 Affidavit on non-judicial stamp paper of Rs. 500/- regarding the firm has not been found guilty of malpractice, misconduct, Punished or blacklisted/debarred either by Public Health Department, Govt. of Maharashtra or by any local authority and other State Government/Central Government's organizations in the past Two years. Affidavit Should Be Notrized
- 2 Affidavit on non-judicial stamp paper of Rs.500/- regarding the Supply Should be done in Mentioned Period . Affidavit Should Be Notrized
- 2 Affidavit On Non-Judicial Stamp Paper of Rs.500/- That "I Accepted All terms & Condition Without Any Complaint & Submitted All Information & Document are True" affidavit Should be Notrized

12 Envelope No. 2 (Price bid):-

- a) All Commercial offers must be submitted.
- b) Rates should be quoted in the Price Schedule only.

- c) All Rates Must be Inclusive of All Taxes.
- d) Supplier are strictly prohibited to change/alter specifications or unit size given in Schedule of requirements while quoting.

13 Deadline for submission of Quotations

14 For Submission of Quotation Supplier must complete the Quotation submission stage as per schedule of the Quotation.

15 **Opening of Quotation:** On the date and time specified in the Quotation notice following procedure will be adopted for opening of Quotation for which Supplier is free to attend himself or not or depute an authorized officer as his representative.

15 Opening of Envelope No.1 (Technical bid) Envelope No.1 (Technical bid) of the Supplier will be opened in the presence of Quotation opening authority

16 Opening of Envelope No.2

This envelope will be opened as per e-Quotation procedure after opening of Envelope No.1 (Technical bid). Date and time of price bid opening will be after completion of technical scrutiny Financial Bid Will Open

17 Period of Validity of Quotation:

The Quotation shall remain **valid for a period of 180 days** from date of Opening Of Financial Bid.

18 Prices

- a. The prices quoted and accepted will be binding on the Supplier and valid for a period of Six Month from the date of Opening Of Financial Envelop.
- b. Rates should be quoted for each of the required item separately on **door delivery basis** according to the unit asked for strictly as per the format of price schedule. The rate should be inclusive of all taxes with GST ,All incidental Charges Packing & Forwarding ,etc. Quotation for the supply of item quoted in the bid with conditions like '**AT CURRENT MARKET RATES**' shall not be accepted. The Purchaser shall not be responsible for damages, handling, clearing, transport charges etc. will not be paid. The deliveries should be made as stipulated in the purchase order placed with successful Supplier.
- c. If at any time during the period of contract, the price of Quotation items is reduced or brought down by any Law or Act of the Central or State Government or by the Supplier himself, the Supplier shall be morally and statutorily bound to inform the Purchaser immediately about such reduction in the contracted prices. The Purchaser is empowered to reduce the rates accordingly.
- d. To ensure sustained supply without any interruption the Purchaser reserves the right to split orders for supplying the requirements amongst more than one Supplier provided that, the rates and other conditions of supply are same.

19 SPECIFICATIONS

The Supplier shall carefully read and understand the specifications, quality requirements Packing applicable standards, acts and rules.

20 Evaluation of Quotations:

- 13.1 After opening of Envelope No. 1 (Technical bid), on the scheduled date, time and venue, the **Purchase committee shall examine** the contents of the Quotations received along with all prescribed mandatory documents.
- 13.2 The Purchase committee shall scrutinize the documents mentioned above for its eligibility, validity, applicability, compliance and substantiation including post Qualification criteria stipulated in Quotation document.
- 13.3 The Purchase committee shall also analyse that there is no collusive or fraudulent CSC/Quotation/ Emergency instruments & Materials /E-Quotation

practice Involved in the entire Quotationing process amongst all the Quotation received.

13.4 Any Quotation during the evaluation process do not meet the Quotation conditions laid down in the Quotation document will be declared as not acceptable and such Quotations shall not be considered for further evaluation. However, the Supplier can check their Quotation evaluation status at office of The Civil Surgeon, Chandrapur.

13.5 The technical scrutiny shall be on the basis of submitted substantiation documents and rules.

13.6 **The Purchase Commettee reserves the right about the quantity to be purchased and also reserves the right to cancel or revise or any of the all the Quotations or part of Quotations without giving any reasons thereto with no cost to the Purchaser.The Purchaser also reserves the right to wave off some terms & conditions in the benefit of the patient & office.**

17. Delivery Period &Place of delivery :

The goods should be delivered within 30 to 40 days Or As Stipulated in the Supply Order On the basis of emergency from the date of receipt of supply order .

18. Liquidated damages:

If the Supplier fails to deliver any or all of the goods within the period(s) specified in the Contract, the Purchaser shall, without prejudice to its other remedies under the Contract, deduct from the Contract Price, as liquidated damages, a sum equivalent to 0.5% of the delivered price of the delayed goods for each week or part thereof of delay until actual delivery, up to a maximum deduction of 10%. Once the maximum is reached, the Purchaser may consider for termination of the Contract.

19. Default Clause / Cancellation on failure to supply :

If the supplier fails to commence delivery as scheduled or to deliver the quantities ordered to him within the delivery period stipulated in the contract/Supply order it shall be discretion of the purchaser either. (a) to extend the delivery period or (b) to cancel the contract in whole or in part for the unsupplied quantities without any show cause notice. In the event of extension, liquidated damages, will be applicable. If the purchaser decides to cancel the contract, the mode of repurchase will be at the discretion of the purchaser. The supplier shall be liable to pay any loss by way of extra expenditure or other incidental expenses, which the purchaser may sustain on account of such repurchase at the risk and cost of the supplier. In addition to action above, the purchaser may debar the defaulting supplier from future orders, for maximum period of 3 years.

The supplies will be deemed to be completed only upon receipt of the quality certificates from the laboratories wherever necessary.

The Purchaser shall be the final authority to reject full or any part of the supply, which is not confirming to the specifications and other terms and conditions. No payment shall be made for rejected stores. Rejected items must be removed by the Supplier within two weeks of the date of rejection at their own cost and replaced immediately. In case rejected items are not removed it will be destroys at the risk, responsibility & cost of Manufacturer/Sole Distributors/Wholesale Supplier.

20. Warranty(If Applicable)

20.1 The warranty shall remain valid for 24to 36months from the date of supply at consignee destination.

20.2 The Supplier should submit the written warranty that all goods supplied under the Contract are of the most recent or current models and that they incorporate all recent improvements in design

and materials provided in the

- 20.3** The Purchaser shall have the right to make claims under the above warranty after the Goods have been delivered to the final destination indicated in the Contract. Upon receipt of a written notice from the Purchaser, the Supplier shall, within the period of 15 days replace the defective Goods without cost to the Purchaser. The Supplier will be entitled to remove, at his own risk and cost, the defective Goods once the replacement Goods have been delivered.
- 20.4** If, after being notified that the defect has been confirmed pursuant to above clause, the Supplier fails to replace the defective Goods within the period of 15 days the Purchaser may proceed to take such remedial action as may be necessary, including removal and disposal, at the Supplier's risk and expense and without prejudice to any other rights that the Purchaser may have against the Supplier under the Contract. The Purchaser will also be entitled to claim for storage, in respect of the defective Goods for the period following notification and deduct the sum from payments due to the Supplier under this Contract.

21. Payment :

The Payment shall be made by following Procedure :

- a) After acceptance of Quotation, The Supply Order of goods will be given by the Civil Surgeon, General Hospital, CHANDRAPUR respectively towards their individual requirement and payment will be done accordingly.
 - b) After confirming the supply, the invoice will be certified by hospital authorities, after satisfaction, they will submit the Invoice along with order copy, delivery challan and test the Civil Surgeon, General Hospital, Chandrapur will make final payment(s) to vender.
- 21.1** The purchaser shall have every right to deduct the pending dues on account of loss, compensation, or any remedial action in monetary terms from the said payment. The supplier shall not agitate the said issue in future.
- 21 Filling Of Quotation Rate – Prescribed Formet on Supplier Letter pad With Duly Signature & Rubber Stamp. If Same Rates are Found, Equal Quantity will be Fix For Purchase
- 22 Disqualification – 1) Rates Over M.R.P., 2) Overwriting in Rates, 3) Not In Prescribed Formet 4) Non Submission Of Documents

CIVIL SURGEON,
CHANDRAPUR

**INFORMATION & DOCUMENTS REQUIRED FOR CMPS PAYMENT PROCEDURE
FOR GOVT PAYEE REGISTRATION APPLICATION
INFORMATION TO BE PREPARED ON BUSINESS LETTERPAD ONLY**

Sr. No.	Details	Information
1	Full Name Of Company / Business / Supplier / Firm	M/s _____
2	Full Address Including Pin Code, State, Country	PIN CODE - _____
3	Contact Telephone No with STD Code	_____
4	Contact Mobile No	_____
5	E-Mail Address	_____
6	PAN Card No	_____
7	TAN No	_____
8	Nationalised Bank Name	_____
9	Branch Name & Address	_____
10	Bank Account No	_____
11	Bank Account Type	_____
12	Bank IFSC Code No	_____
13	Cancelled Cheque No	_____

Declaration:

Herewith certify that above information is correct & true.

Attached Attested Documents:

- 1. PAN Card Xerox
- 2. Bank Pass Book Front Page Xerox
- 3. Blank Cancelled Cheque

(Authorized Signatory)

Name & Rubber Stamp/Seal

Copy To:

1. The Civil Surgeon, Chandrapur
2. Central Medical Store, T.B. Hosp. Campus, Ramnagar, Chandrapur

CSC/Quotation/Emergency Instruments & Materials/E-Quotation

**FORMET FOR QUOTATION
(TO BE TYPE ON SUPPLIER LETTER PAD)**

DATE-.....

To,
The Civil Surgeon,
Chandrapur

Sub :- Submission Of Quotation of EMERGENCY Emergency Medicine
Ref :- Web Site Quotation Notice CSC/Quotation/ Emergency Medicine /E-Quotation

with ref. to Above Sub, We are Herewith Submitting Following Items Rate For Govt Purchase

Sr. No	Name Of Particulars with Specification	Unit	Rate (Without Taxes & Transpotation)	Taxeble Amount	Total Rate (With Taxes)
1	Ophthalmoscope	Each			
2	Retinoscope Streak	Each			
3	Digital Vision Drum LED	Each			
4	A Scan Biometer	Each			
5	Refracto-Keratometer	Each			
6	Fogger Machine	Each			

Your Faithfully

Supplier Stamp & Sign.



महाराष्ट्र शासन आरोग्य सेवा
जिल्हा सामान्य रुग्णालय चंद्रपूर “छोटाबाजार” चंद्रपूर पिन ४४२४०२

रुग्णालय दुरध्वनी कार्यालय दुरध्वनी जिल्हा शल्य चिकित्सक (वैयक्तिक)	07172 250400 07172 264640 07172 252103 07172 253992	Email : cs_chandrapur@rediffmail.com cschandrapur@gmail.com Fax No. 07172 253992
--	--	--

टिप :

- १) सदर खरेदी ही उपलब्ध अनुदानानुसार करण्यात येत असल्यामुळे ई-दरपत्रकामध्ये मध्ये नमुद बाबींची संख्यानुसार खरेदी करणे या कार्यालयाला बंधनकारक नाही याची सर्व पुरवठादारांनी नोंद घ्यावी.
- २) दर सादर करतांना सर्व करांसह जिल्हा सामान्य रुग्णालय चंद्रपूर चे औषधी भांडार पर्यंत पुरवठा करण्याचे दर द्यावे लागतील.
- ३) दरपत्रक कार्यवाही मध्ये कोणतेही कारण न देता रद्द करण्याचा अधिकार जिल्हा शल्य चिकित्सक, सामान्य रुग्णालय, चंद्रपूर यांचे स्तरावरील जिल्हा खरेदी समितीला राखून ठेवलेला आहे.
- ५) अटी व शर्ती निश्चित करण्याचे व शिथिल करण्याचे अधिकार खरेदी समिती, सामान्य रुग्णालय, चंद्रपूर यांनी राखून ठेवण्यात येत आहे.
- ६) महाराष्ट्र शासन, उद्योगउर्जा व कामगारविभाग यांचे शासननिर्णय दि. १.१२.२०१६ मधील व दरपत्रकामध्ये नमुद सर्व अटी व शर्ती पुरवठादारास लागू राहतील याची नोंद घ्यावी.
- ७) आर्थिक दरपत्रक ज्या मध्ये किंमत आणी अन्य आर्थिक तपशिल असेल तसेच त्यामध्ये वस्तुची मुळ किंमत, केंद्रीय मुल्यवर्धितकर/महाराष्ट्र मुल्यवर्धित कर व वाहतुक खर्चस्वतंत्र दर्शविण्यात यावा.

जिल्हा शल्य चिकित्सक
सामान्य रुग्णालय, चंद्रपूर

BETA 200 LED Ophthalmoscope – Technical Specifications

General Information

Product Variant: BETA 200 LED Ophthalmoscope

Materials: Synthetics, metal, glass

Compliance: REACH | RoHS compliant

Biocompatibility: Compliant

Surface: Synthetics, metal, glass

Warranty: 5 years

Environmental Conditions - Operation

Temperature: +10°C to +35°C

Humidity: 30% to 75%

Air Pressure: 700 hPa to 1060 hPa

Environmental Conditions - Storage

Temperature: +5°C to +45°C

Humidity: 45% to 80%

Air Pressure: 500 hPa to 1060 hPa

Environmental Conditions - Transport

Temperature: -20°C to +50°C

Humidity: 45% to 80%

Air Pressure: 500 hPa to 1060 hPa

Operating Features

Controls: Lens wheel, Aperture wheel, Filter switch

[Signature]
CIVIL SURGEON
District General Hospital
Chandrapur

[Signature]
जिल्हा वरिष्ठ आर्य वैदिक शास्त्रज्ञ
आर्य शास्त्रज्ञ संस्थान (अ.वि.सं.)
सायबान्द इन्स्टीट्यूट चंद्रपूर

[Signature]
Assistant Professor
Dept. of Ophthalmology
GMCH Chandrapur

Display: Indirect illuminated index of refraction

Power Supply: HEINE Rechargeable Handles (3.5 V), Battery Handles (2.5 V), EN 200 Wall Transformer


Mechanical Specifications

Weight: 73 g (device), 114-116 g (with packaging)


Dimensions: 98 × 48 × 28 mm

Packaging: 108 × 42 × 68 mm


Connection: AV for rechargeable handle



जिल्हा नेत्र शल्य चिकित्सक वर्ग-१
जिल्हा कार्यक्रम व्यवस्थापक (अनिका)
सामान्य रुग्णालय चंद्रपूर



Assistant Professor
Dept. of Ophthalmology
GMCH Chandrapur



CIVIL SURGEON
District General Hospital
Chandrapur

BETA 200 LED Retinoscope – Simplified Notes

MECHANICAL

- Weight: 90 g (200 g with packaging)
- Dimensions (product): 129 x 37 x 43 mm
- Protection class: IP40

GENERAL

- Materials: Metal, glass, plastic
- Biocompatibility: Conform
- Environmental conditions:
 - Operation: +10°C to +35°C
 - Storage: -10°C to +55°C
 - Transport: -40°C to +70°C
- Warranty: 5 years

FUNCTIONAL FEATURES

- Light intensity control (rheostat)
- ParaStop for parallel beam selection
- Single control for vergence & rotation
- Detachable brow rest

ELECTRICAL

- Input voltage: 1.8-3.7V

[Signature]
 Assistant Professor
 Dept. of Ophthalmology
 GMCH Chandrapur


[Signature]
CIVIL SURGEON
 District General Hospital
 Chandrapur

[Signature]
 जिल्हा सहायक आरोग्य अधिकारी
 (आरोग्य) जिल्हा प्रशासन कार्यालय
 चंद्रपूर

- Power consumption: ~390 mA
- Power source: Rechargeable (3.5 V) or battery (2.5 V)


OPTICAL

- Light source: LED (HQ)
- Color temperature: ~3000 K (warm light)
- Streak length: ~35 mm at 500 mm distance
- Streak width: ~1.1 mm at 500 mm distance
- Working distance: 500 mm



जिल्हा वंश शल्य चिकित्सक वर्ग-१
तथा जिल्हा कार्यक्रम व्यवस्थापक (अभि-१),
सामान्य रुग्णालय चंद्रपूर


Assistant Professor
Dept. of Ophthalmology
GMCH Chandrapur


CIVIL SURGEON
District General Hospital
Chandrapur

LED Vision Chart Specifications

Features:

- Compact and lightweight design
- Supports multiple regional languages
- Exclusive image gallery for various eye diseases
- Reverse mode available for all charts
- Compensation for red-green filter
- Suitable for all refraction rooms
- Easy wall mounting
- Cordless remote operation from any direction
- Charts offered: C/E Chart, Dot Chart, Snellen Chart, Ishihara Chart, Astigmatic Fan, LogMAR Chart, Spot Light, Duochrome Test Light

Specifications:

Color: Black

Material: LED

Display Size: 18.5 inches

Gross Weight: Approximately 4 kg

[Signature]
CIVIL SURGEON
 District General Hospital
 Chandrapur

[Signature]

डिप्टी मेडिकल ऑफिसर (आयुर्वेद)
 जिला आयुर्वेद अस्पताल (आयुर्वेद)
 सामाजिक कल्याण नगर

[Signature]
 Assistant Professor
 Dept. of Ophthalmology
 GMCH Chandrapur

A-scan Biometry

TECHNICAL SPECIFICATIONS

[BIOMETRY]

Adjustable gain: 20 to 110 dB

Time Gain Control (TGC): 0 to 30 dB

11 MHz probe

Transducer frequency: 11 MHz

Tip diameter: 7 mm (0.28")

Electronic resolution: 0.03 mm (0.0012")

Depth: 60 mm (2.4") for 1,536 points

Contact and immersion techniques compatible

Aiming beam: LED or laser pointer Probeam™

Axial length measurements

Ultrasound propagation velocity adjustable per segment (anterior chamber, lens, vitreous) and IOL and vitreous material

Built-in pattern recognition: phakic, aphakic, PMMA, acrylic and silicone for pseudophakic eyes

Automatic calculation of standard deviation and average total length (series of 10 measurements)

Acquisition modes: automatic, auto + save, manual

Automatic detection of scleral spike

IOL calculation

SRK-T, SRK-II, HOLLADAY, BINKHORST-II, HOFFER-Q, HAIGIS

Post-op refractive calculation:

- Pre-op and post-op refraction, pre-op and post-op keratometry
- 6 different methods for keratometry correction and implant calculation: history derived, refraction derived, contact lens method, Rosa regression, Shammas regression, Double K/SRK-T (Dr Aramberr) formula)

[Signature]
CIVIL SURGEON
 District General Hospital
 Chandrapur

[Signature]
 ...
 ...
 ...

[Signature]
Assistant Professor
 Dept. of Ophthalmology
 GMCCH Chandrapur

9 values bracketed for desired ametropia for each IOL (IOL increment steps: 0.25 D or 0.50 D)

Simultaneous display of 4 different IOL calculations

DATA MANAGEMENT

Built-in physician and patient database

Exportation of still images

Customisable digital and printed reports


DICOM compatible (Worklist, Storage, Print)

EMR compatible


Compatible with PC and USB video printers



जिल्हा नव शल्य चिकित्सक वर्ग-१
नया जिल्हा कार्यालय व्यवस्थापक (अनिका)
सामान्य रुग्णालय चंद्रपूर



Assistant Professor
Dept. of Ophthalmology
GMCH Chandrapur



CHIEF SURGEON
District General Hospital
Chandrapur

AUTO- REFRACTOMETER - KERATOMETER

Technical Specifications

Refractive Power Measuring Range

Spherical Refractive Power: -25D to +22D (0.01D / 0.12D / 0.25D steps)

Cylindrical Refractive Power: 0D to +10D (0.12D / 0.25D steps)

Astigmatic Axial Angle: 0° to 180° (1° / 5° steps)

Minimal Measurable Pupil Diameter: Ø 2 mm

Corneal Curvature Measuring Range

Corneal Curvature Radius: 5.00 mm to 10.00 mm (0.01 mm steps)

Corneal Refractive Power: 67.50D to 33.75D (0.12D / 0.25D steps)

(Where corneal refractive power = 1.3375)

Corneal Astigmatic Refractive Power: 0D to ±10D (0.12D / 0.25D steps)

Corneal Astigmatic Axial Angle: 0° to 180° (1° / 5° steps)

PD Measurement Range

20 to 85 mm (0.5 mm step)

Data Transport Terminal

USB (Import) / RS-232C (Export) / LAN (Export)

Dimensions

317-341 mm (W) × 521-538 mm (D) × 447-477 mm (H)

Weight

15 kg

Power Supply

100-240 V AC, 50-60 Hz, 30-70 VA

hll
CIVIL BURGÉON
District General Hospital
Chandrapur


M. K. Mishra
जिल्हा स्तर शाखा चिकित्सक (अभिज्ञ)
जेष्ठ कार्यक्रम व्यवस्थापक (अभिज्ञ)
सार्वजनिक रुग्णालय चंद्रपूर


Dr. ...
Assistant Professor
Dept. of Ophthalmology
GMCH Chandrapur


FOGGING MACHINE SPECIFICATION

(For Aerial Disinfection)

- 1. **Quantity:** 2 Fogging Machines
- 2. **Type:** Lightweight, portable model
- 3. **Particle Size Generation:** 0.3 to 1 micron
- 4. **Coverage Area:** Horizontal reach: 40-45 ft
Vertical reach: 15-20 ft
- 5. **Suitable Volume:** 12,000 to 15,000 cubic feet
- 6. **Chemical Compatibility:** Suitable for all water-based disinfectants
- 7. **Motor Specifications:** Speed: 26,000 RPM
Type: Double-stage motor with tapered fans
- 8. **Motor Protection:** Thermal overload protection
- 9. **Nozzle Assembly:** Material: Stainless Steel (SS 304)
Design: Non-clogging vortex design
- 10. **Power Head:** Material: High-grade engineering plastic (unbreakable)
- 11. **Solution Tank:** Capacity: 5 liters
Material: HDPE
Feature: Graduated marking
- 12. **Strainer and Handle:** Material: Stainless Steel (SS 304)
- 13. **Discharge Capacity:** 0 to 90 ml/min
- 14. **Intake Air Filter:** Double-layer filtration
- 15. **Weight:** Empty weight: 5 to 6 kg


CIVIL SURGEON
 District General Hospital
 Chandrapur


 दिवाडा गेट पास विद्विजलक वार्ड-१
 शिवा मंगलम अस्पताल (अति)
 सामाजिक कल्याण चंद्रपूर


 Assistant Professor
 Dept. of Ophthalmology
 GMCH Chandrapur