सत्यमेव जयते
महाराष्ट्रशासन

उपसंचालक आरोग्य सेवा
राज्य सार्वजनिक आरोग्य प्रयोगशाळा, पुणे
लष्कर पाणी पुरवठा आवार, सेंट मेरी शाळेजवळ, पुलगेट, स्टॅन्ह्ले मार्ग, पुणे ४११००१

दूरध्वनी - कार्यालय :- (०२०) २६३३०२२४/२६३३०२२५

Email : statehealthlab@gmail.com aosphlpune@gmail.com

आरोग्य सेवा

जा .क्र.उसंआसे/रासाआप्र/भांडार/ कॅलिब्रेशन दरपत्रके मागवीणे/२०२६ / 7740-41

/ दिनांक : 13/ ११/२०२६

प्रति,

.....

.....

.....

विषय —राज्य सार्वजनिक आरोग्य प्रयोगशाळेतील उपकरणांचे मुल्यांकन (कॅलिब्रेशन) करण्याकरिता दरपत्रके सादर करणेबाबत.

आपणास कळविण्यात येते की, उपसंचालक, आरोग्य सेवा, राज्य सार्वजनिक आरोग्य प्रयोगशाळा पुणे या प्रयोगशाळेतील उपकरणांचे मुल्यांकन (कॅलिब्रेशन) करुन घेणेसाठी दरपत्रके सोबत जोडलेल्या अटी व शर्तीनुसार मागविण्यात येत आहेत. मुल्यांकन (कॅलिब्रेशन) करुन घ्यावयाच्या एकूण ७९ उपकरणांची यादी सोबत जोडलेली आहे.

सदर दरपत्रक सिलबंद लिफाफयात दिनांक १९/११/२०२६ रोजी पोहचतील अशा रितीने पाठवावीत. दरपत्रक स्विकारणे अथवा नाकारणे तसेच सदर खरेदी प्रक्रिया पुर्णपणे रद्द करण्याचे अधिकार मा.उपसंचालक आरोग्य सेवा, राज्य सार्वजनिक आरोग्य प्रयोगशाळा, पुणे यांनी राखुन ठेवले आहेत.

(श्री.मनोहर.द.धुम)
उपसंचालक आरोग्य सेवा,
राज्य सार्वजनिक आरोग्य प्रयोगशाळा पुणे-१

सोबत- अटी,शर्ती
प्रत -नोटीस बोर्ड/संकेतस्थळ

अटी व शर्ती –

१) दरपत्रकाच्या लिफाफयावर " Quotation for Calibration" असा स्पष्ट उल्लेख असावा.

२) दरपत्रे पुरवठादाराच्या लेटर हेडवर सादर करावी व त्यावर शिक्यासह स्वाक्षरी असावी.

३) दर नमुद करतांना मुळ किंमत (Basic price), जीएसटीचे दर व रक्कम (GST rate & amount) सवलतीचे रक्कम (Discount amount) व इतर खर्च स्पष्टपणे वेगवेगळे दाखवावे.मान्य केलेल्या दराव्यतीरिक्त इतर कोणताही खर्च अदा केले जाणार नाही.

४) दरपत्रकावर जीएसटी क्रमांक व पॅन नंबर याचा उल्लेख असावा.

५) दरपत्रकासोबत पॅन कार्ड व जीएसटी नोंदणी प्रमाणपत्राच्या स्वसांक्षाकीत छायांकित प्रती (Self attested xerox copy) सादर केले असतील तरच दरपत्रक वैध मान्यात येईल. जीएसटीचे प्रोव्हीजनल नोंदणी प्रमाणपत्र जोडलेले दरपत्रक अवैध समजण्यात येतील.

६) दरपत्रके दि.३०/११/२०२६ रोजी सकाळी ११ वाजता उघडण्यात येतील.दरपत्रके उघडण्याच्या वेळी संबंधीत पुरवठादार / त्यांचे प्रतिनीधी कार्यालयामध्ये उपस्थित राहू शकतात.

७) ई-मेल मार्गाने सादर केलेल्या दरपत्रकांचा स्वीकार केला जाणार नाही.

८) दरपत्रक स्विकारणे अथवा नाकारणे तसेच सादर प्रक्रिया पुर्ण पणे रद्द करण्याचे अधिकार उपसंचालक आरोग्य सेवा, राज्य सार्वजनिक आरोग्य प्रयोगशाळा, पुणे यांनी राखून ठेवले आहेत.

९) स्थानिक बाजारापेठेतील पुरवठादारास प्राधान्य देण्यात येईल.

१०) दरपञके सादर करताना दरपञकांसोबत एन.ए.बी.एल ॲक्रीडीटेशन प्रमाणपञ जोडणे आवश्यक आहे.

११) ज्या उपकरणांचे मुल्यांकन (Calibration) करण्याचे पॉईन्ट दरपञक मागणी पञात नमुद नसतील त्याची पाहणी अथवा त्या संबंधीचे काही अडचण असतील तर दरपञक सादर करणा-या संस्थेचे प्रतिनिधी कार्यालयात येऊन पाहणी करु शकतात.

१२) पुरवठादार संस्थेचे नोंदणी प्रमाणपञ सादर करणे बंधनकारक राहील.

१३) पुरवठादाराने मागील तिन वर्षाचे आयकर भरणा केल्याचे प्रमाणपञ जोडणे बंधनकारक राहिल.

१४) दरपत्रके मागणी पत्रातील उपकरणांचे कॅलिब्रेशन (मुल्यांकन) हे याच कार्यालयात येऊन करणे बंधनकारक राहील.

(श्री.मनोहर.द.धुम)
उपसंचालक आरोग्य सेवा,
राज्य सार्वजनिक आरोग्य प्रयोगशाळा पुणे-१

State Public Health Laboratory

List Of Instruments in Microbiology Section for calibration in 2026

A) Food Microbiology Section

Sr. No	Name of Equipment / Instrument	Model	ID Number	Due date of calibration	Calibration range / point
1	Balance	Contech	SPHL/BIO/BAL/08	26/11/2026	Weight 1,2,510,20.50,,100,200,500, ,1000 gm (gms)
2	Balance	Star	SPHL/BIO/BAL/07	26/11/2026	Weight 1,5,10,20,.50,100,200,500, 1000,2000,5000 gm
3	Refrigerator	LG	SPHL/BIO/REF/08	26/11/2026	Temp: Top 2°C, Middle: 5°C , Bottom: 8°C
4	Refrigerator	LG	SPHL/BIO/REF/09	26/11/2026	Temp: Top 2°C, Middle: 5°C , Bottom: 8°C
5	Refrigerator	Whirlpool	SPHL/BIO/REF/06	26/11/2026	Temp: Top 2°C, Middle: 5°C , Bottom: 8°C
6	Refrigerator	Westing House	SPHL/BIO/REF/07	26/11/2026	Temp: Top 2°C, Middle: 5°C , Bottom: 8°C
7	Refrigerator	Elanpro	SPHL/BIO/REF/10	---	Temp: Top 2°C, Middle: 5°C , Bottom: 8°C
8	Autoclave Portable	Equitron	SPHL/BIO/ACL/01	26/11/2026	Temp: 50,100,115,121,150,°C Pressure: 0,1,3,5,10,15,20,25,30 (lb) Timer -1,5,10,15,20,30,45 Minutes
9	Autoclave Portable	Equitron	SPHL/BIO/ACL/02	26/11/2026	Temp: 50,100,115,121,150,°C Pressure: 0,1,3,5,10,15,20,25,30 (lb) Timer -1,5,10,15,20,30,45 Minutes
10	Autoclave Portable	Equitron	SPHL/BIO/ACL/03	26/11/2026	Temp: 50,100,115,121,150,°C Pressure: 0,1,3,5,10,15,20,25,30 (lb) Timer -1,5,10,15,20,30,45 Minutes
11	Autoclave Vertical	Smita Scientific	SPHL/BIO/ACL/04	26/11/2026	Temp: 50,100,115,121,150,°C Pressure: 0,1,3,5,10,15,20,25,30 (lb) Timer -1,5,10,15,20,30,45 Minutes
12	Autoclave Vertical	Smita Scientific	SPHL/BIO/ACL/05	26/11/2026	Temp: 50,100,115,121,150,°C Pressure: 0,1,3,5,10,15,20,25,30 (lb) Timer -1,5,10,15,20,30,45 Minutes

13	Incubator	Smita Scientific/ PPI	SPHL/BIO/INCU/02	26/11/2026	Temp: 20,25,37,40,50,60 (°C)
14	Incubator	Esteem/ itherm	SPHL/BIO/INCU/03	26/11/2026	Temp: 20,25,42,50,60 (°C)
15	Incubator	Bio Tech/ itherm	SPHL/BIO/INCU/04	26/11/2026	Temp:20, 25,40,55,60 (°C)
16	Incubator	Smita Scientific/ PPI	SPHL/BIO/INCU/07	26/11/2026	Temp: 20,25,37,40,50,60 (°C)
17	Hot Air Oven	Smita Scientific/ itherm	SPHL/BIO/OVEN/09	--	Temp :50,100,160,180,200°C
18	Serological Water Bath	Biotechnics India/ itherm	SPHL/BIO/SWB/01	26/11/2026	Temp: 20,30,44,60,100 (°C)
19	Biosafety Cabinet	IMSET	SPHL/BIO/BSC/01	26/11/2026	1.Positive Pressure Test 2.Air Velocity Test 3 Air Changes per hour 4.Manometer Pressure test 5.Particale Count Test 6.Tempreture and Humidity Test 7. HEPA Filter Integrity Test By PAO(Poly Alpha Olyfin
20	Digital Thermometer Indicator with Sensor	Multi	SPHL/BIO/DTHM/01	26/11/2026	Temp: -2,5,10,50,100,150,200 (°C)
21	Digital Thermometer Indicator with Sensor	Multi	SPHL/BIO/DTHM/07	26/11/2026	Temp: -2,5,10,50,100,150,200 (°C)
22	Digital Thermometer Indicator with Sensor	Multi	SPHL/BIO/DTHM/08	26/11/2026	Temp: -2,5,10,50,100,150,200 (°C)
23	Digital ThermoHygrometer	HTC	SS/THM/08	26/11/2026	Temp: -20,25,30,35,40,45 (°C) Humidity (% RH) – 20,30,50,70,90
24	Digital ThermoHygrometer	HTC	SS/THM/09	26/11/2026	Temp: -20,25,30,35,40,45 (°C) Humidity (% RH) – 20,30,50,70,90
25	Digital ThermoHygrometer	ACUTEQ - KT 908	HygrometerDTH -14	26/11/2026	Temp: -20,25,30,35,40,45 (°C) Humidity (% RH) – 20,30,50,70,90
26	BOD Incubator	Spire Automation & Innovation India /Itherm	SPHL/BIO/BOD/02	26/11/2026	Temp: -5,15,25,30,40 (°C)

27	Laminar Air Flow	Bio-Technics India	SPHL/BIO/LF/02	26/11/2026	1.Positive Pressure Test 2.Air Velocity Test 3 Air Changes per hour 4.Manometer Pressure test 5.Particale Count Test 6.Tempreture and Humidity Test 7. HEPA Filter Integrity Test
28	Micro Pipette (100 to 1000µl)	Thermo Scientific	SPHL/BIO/ Micropipette/01	26/11/2026	Nominal Value Of Volume (µL) : 100,500,1000
29	Micro Pipette (1000µl)	Kasablanka	SPHL/BIO/ Micropipette/02	---	Nominal Value Of Volume (µL) : 1000
30	Micro Pipette (100 to 1000µl)	Lab Quest Borosil	SPHL/BIO/ Micropipette/03	---	Nominal Value Of Volume (µL) : 100,500,1000
31	Stomacher	Bioxia/Selec	SPHL/BIO/ STOMACHER/01	26/11/2026	Timer -1,3,5,10,15,20,30, Minutes
32	Analytical Weight Box	Lab Line	SPHL/BIO/ WTS/01	26/11/2026	200g,100g,50g,20g,20g,10g, 5g,2g,2g,1g,500mg,200mg, 200mg,100mg,50mg, 20mg, 10mg,5mg,2mg,1mg
33	Deep Freezer Location –Food Bact Section	CELLFROST	SPHL/BIO/DPFT/03	26/11/2026	(-20) °C

B) Water Microbiology Section

Sr. No	Name of Equipment / Instrument	Model	ID Number	Due date of calibration	Calibration range / point
1	Incubator	Electrometer Corporation	SPHL/BIO/INCU/05	26/11/2026	Temp: 20,25,37,40,50,60 (°C)
2	Serological Water Bath	Biotechnics India	SPHL/BIO/SWB/02	26/11/2026	Temp: 20,30,44,60,100 (°C)
3	Hot Air Oven	Smita Scientific	SPHL/BIO/OVEN/10	26/11/2026	Temp :50,100,160,180,200°C
4	Refrigerator	Godrej Cold Gold	SPHL/BIO/REF/10	26/11/2026	Temp: Top 2°C, Middle: 5°C , Bottom: 8°C

C) ENT Microbiology Section

Sr. No	Name of Equipment / Instrument	Model	ID Number	Due date of calibration	Calibration range / point
1	Incubator	Spire Automation & Innovation Ind	SPHL/BIO/INCU/06	26/11/2026	Temp: 20,25,37,40,50,60 (°C)
2	Biosafety Cabinet	ACCO	SPHL/BIO/BSC/02	26/11/2026	1.Positive Pressure Test 2.Air Velocity Test 3 Air Changes per hour 4.Manometer Pressure test 5.Particale Count Test 6.Tempreture and Humidity Test 7. HEPA Filter Integrity Test
3	Biosafety Cabinet	KIM Micro Systems	SPHL/BIO/BSC/03	---	1.Positive Pressure Test 2.Air Velocity Test 3 Air Changes per hour 4.Manometer Pressure test 5.Particale Count Test 6.Tempreture and Humidity Test 7. HEPA Filter Integrity Test
4	Refrigerator	Godrej Cold Gold	SPHL/BIO/REF/11	26/11/2026	Temp: Top 2°C, Middle: 5°C , Bottom: 8°C
5	Refrigerator	Celfrost	SPHL/BIO/REF/12	26/11/2026	Temp: Top 2°C, Middle: 5°C , Bottom: 8°C

Food Section

List of Equipment to be Calibrated

Sr. No	Location	Instrument Name	Make	Instrument ID	Range	Due Date
1	Food Section	Analytical Balance	Aczet	SPHL/FOOD/BAL/03	0 –200gm	26.11.2026
2		Standard Weight Box	Labfit	SPHL/FOOD/WTS/01	1 mg – 200gm	30.11.2026
3		Digital Thermo-Hygrometer	Acuteq (KT-908)	SPHL/FOOD/DTH-6	-10–50 °C 20-99%RH	30.11.2026

Misc- I&III Section

List of Equipment to be Calibrated

Sr. No	Location	Instrument Name	Make	Instrument ID	Range	Due Date
1	Misc I&III Section	Hot Air Oven	Pathak Electronics	SPHL/MI&MIII/OVEN/01	0°C – 300°C	26.11.2026
2		Muffle Furnace	Bio Tech. india (I – Therm)	SPHL/MI&MIII/MF/02	50 °C – 1200 °C	25.02.2027

MISC II section

List of Equipment to be Calibrated

Sr. No	Location	Instrument Name	Make	Instrument ID	Range	Due Date
1	Misc- II Section	Muffle Furnace	Mahavir Electronics	SPHL/MII/MF/01	50 °C – 1100 °C	27.11.2026
2		Hot Air Oven	International Arora Scientific & Engg	SPHL/MII/OVEN/01	0 – 250 °C	26.11.2026
3		Digital Thermo-Hygrometer	Temptec	SPHL/MII/DTH-15	0– 50 °C 20-90RH	19.01.2027

Oil &Milk section

List of Equipment to be Calibrated

Sr. No	Location	Instrument Name	Make	Instrument ID	Range	Due Date
1	Oil & Milk Section	Hot Air Oven	Yorco	SPHL/O&M/OVEN/01	0 – 300 °C	27.11.2026

Instrument section

List of Equipment to be Calibrated

Sr. No	Location	Instrument Name	Make	Instrument ID	Range	Due Date
1	**Instrument section**	Analytical Balance	Aczet	SPHL/INST/BAL/01	0– 80gm	26.11.2026
2		Orbital Incubator	Neolab Instruments	SPHL/INST/INC-SH/01	0–60 °C	26.11.2026
3		Deep Freezer	Vest Frost	SPHL/INST/DF/01	-20°C – 0 °C	26.11.2026
4		Centrifuge	Remi Electronics	SPHL/INST/CFG/02	0-16000RPM	30.11.2026
5		Micropipette	(Lab Quest) Borosil	SPHL/INST/MICPP/01	10–100µl	28.11.2026
6		Micropipette	(Lab Quest) Borosil	SPHL/INST/MICPP/02	20–200µl	28.11.2026
7		Micropipette	(Lab Quest) Borosil	SPHL/INST/MICPP/03	100–1000µl	28.11.2026
8		Micropipette	(Lab Quest) Borosil	SPHL/INST/MICPP/04	100–1000µl	-
9		Temp Indicator With Sensor	I- therm	SPHL/FOOD/DTI/03	50–1200°C 0–300 °C	30.11.2026
10		Digital Thermo-Hygrometer	Acuteq (KT-908)	SPHL/INST/DTH-8	-10–50 °C 20-99%RH	30.11.2026
11		Digital Thermo-Hygrometer	Acuteq (KT-908)	SPHL/INST/DTH-9	-10–50 °C 20-99%RH	30.11.2026
12		Digital Thermo-Hygrometer	Acuteq (KT-908)	SPHL/INST/DTH-10	-10–50 °C 20-99%RH	30.11.2026
13		Digital Thermo-Hygrometer	Acuteq (KT-908)	SPHL/INST/DTH-11	-10–50 °C 20–99%RH	30.11.2026
14		GCMS-MS	Agilent	SPHL/INST/GCMS/01	60 °C- 350 °C	-
15		LCMS-MS	Agilent	SPHL/INST/LCMS/01	Ambient to 50 °C	-
16		HPLC Waters	Waters	SPHL/INST/HPLC/01	Ambient to 40 °C	-
17		GC-2010	Shimadzu	SPHL/INST/GC/01	Ambient to 250 °C	-

CFL/RFL Section

List of Equipment to be Calibrated

Sr. No	Location	Instrument Name	Make	Instrument ID	Range	Due Date
1	**CFL/RFL Section**	Analytical balanace	Sartorius	SPHL/CFL/BAL/01	0 – 220g	26.11.2026
2		Digital Thermo-hygrometer	Acuteq (KT-908)	SPHL/CFL/DTH-12	-10– 50 °C 20-99%RH	30.11.2026
3		Hot Air Oven	Memmert	SPHL/CFL/OVEN/01	0 – 250 °C	27.11.2026
4		Glass Thermometer	JRM	SPHL/CFL/GTM/01	-10 – 50 °C	30.11.2026
5		Glass Thermometer	GH Zeal	SPHL/CFL/GTM/02	-10 –50 °C	30.11.2026
6		Weight Box	Universal Electric	SPHL/CFL/WTS/01	1 mg – 200g	30.11.2026
7		Muffle Furnace	Bio tech. India	SPHL/CFL/MF/02	50 –1200 °C	25.02.2027
8		Analytical balanace	RADWAG	SPHL/CFL/BAL/02	0-220 g	26.11.2026

Water Chemical section

List of Equipment to be Calibrated

Sr. No	Location	Instrument Name	Make	Instrument ID	Range	Due Date
1	Water Chemical section	Digital Thermo-Hygrometer	Acuteq (KT-908)	SPHL/WTR/DTH-7	-10– 50 °C 20-99%RH	30.11.2026
2		Hot Air Oven	International Arora Scientific & Engg	SPHL/WTR/OVEN/01	0 – 250 °C	26.11.2026
3		Conductivity Meter	Lab Quest (Borosil)	SPHL/WTR/CON/02	Upto 200 ms	-