राष्ट्रीय ग्रामीण आरोग्य अभियान,
जिल्हा एकात्मिक आरोग्य व कुटुंब कल्याण सोसायटी, सांगली
रजि. नं. महाराष्ट्र/६८२/२००५ - सांगली

पत्ता :- जिल्हा शल्य चिकित्सक, सांगली
फोन नं. :- ०२३३ - २३७४८१७,
फॅक्स नं. :- ०२३३ - २३७४८१७
Email Address :- cssangli10@gmail.com

आरोग्य सेवा
जाहिर नोटीस
दर मागणी पत्र

जा.क्र/CSOffice/कोटेशन/साहित्य साधन सामुग्री खरेदी / 2982 /२०२६
दिनांक - १३/१/२०२६

प्रति,

.....

.....

विषय :- राष्ट्रीय असंसर्गजन्य रोग नियंत्रण कार्यक्रमांतर्गत येणाऱ्या एन.पी.एन.सी.डी. कार्यक्रमांतर्गत CHC एन.सी.डी क्लिनीक करिता आवश्यक साधनसामग्री खरेदी करणेबाबत...

महोदय,

उपरोक्त विषयास अनुसरुन, राष्ट्रीय आरोग्य अभियानांतर्गत सांगली जिल्हयात असंसर्गजन्य आजार कक्ष (NCD) स्थापन करणेत आलेला आहे. सदर असंसर्गजन्य रोग नियंत्रण कार्यक्रमांतर्गत येणाऱ्या एन.पी.एन.सी.डी.कार्यक्रमांतर्गत उपजिल्हा व ग्रामीण रुग्णालयाकरिता आवश्यक उपकरणे/साधनसामुग्री खरेदी करणेची आहे. सदर उपकरणे/साधनसामुग्रीचा पुरवठा करणेबाबत आपणांकडून ई-दरपत्रके मागविणेत येत आहेत. सदर ई-दरपत्रके बंद लखोटयातून दिनांक २३/११/२०२६ रोजी दुपारी ३ वाजेपर्यंत मा. जिल्हा शल्य चिकित्सक, सांगली यांचे नांवे स्विकारणेत येतील. ई-दरपत्रके मंजुरीचा अंतिम अधिकार मा. जिल्हा शल्य चिकित्सक, सांगली यांना राहतील.

खालील प्रमाणे उपकरणे/साधनसामुग्रीचा पुरवठा मा. जिल्हा शल्य चिकित्सक कार्यालय, सांगली येथे पोहोच करणेसह ई-दरपत्रके दर मागविणेत येत आहेत.

अ. क्र	उपकरणे/साहित्य सामग्री तपशील	दर
१)	सोबत सविस्तर तपशील वार यादी जोडणेत आले आहे.	

(सर्व बाबींचे दर देणे बंधनकारक आहे.)

वरील उपकरणे/साधनसामुग्रीची ई-दरपत्रके वर दिलेल्या तारखेच्या/वेळेच्या आंत जिल्हा शल्य चिकित्सक कार्यालय, सांगली येथे पोहोच करणेत याव्यात.

- **सदर दर हे सर्व शासकिय करांसहित व जीएसटीसह दर देणेत यावेत.**

जिल्हा शल्य चिकित्सक, सांगली

राष्ट्रीय असंसर्गजन्य रोग नियंत्रण कार्यक्रमाअंतर्गत एन.पी.सी.डी.सी.एस अंतर्गत उपकरणे/साहित्य सामग्री खरेदी सन २०२६-२०२७

अ. क्र.	तपशील	नग	दर प्रती नग (सर्व शासकीय करासहित)
1	**डायमंड बी.पी. ॲपरॅटस् (डिजीटल व्हिथ मरक्युरी)** 1)Technical Specifications 2)Specification Parameter Standard Technical Details 3)Measurement Principle Auscultatory method (using a stethoscope) 4)Measurement Location Upper Arm 5)Pressure Measurement Range0 to 300 mmHg 6)Pulse Measurement Range 30 to 200 beats per minute (on models with pulse tracking) 7)Pressure Accuracy ±3 mmHg 8)Pulse Accuracy ±5% of the reading 9)Display Type Vertical LCD or LED digital column (mimicking a glass mercury tube) 10)Inflation Method Manual, using a premium rubber/latex bulb 11)Deflation Method Manual, via an air release control valve 12)Cuff Size / Type Standard Adult (22–42 cm) with secure Velcro fastening 13)Power Source Dual Operation: 3x or 4x AA Alkaline Batteries OR a USB Type-C/AC Adapter 14)Storage Capacity Some models save past readings (e.g., 30 to 90 memory slots) Body Material Durable ABS Plastic or Die-Cast Aluminum casing	1	
2	**Cuscos Specullum** **Specifications-** 1)Material-German Stainless Steel 2)Curvature-Angled Single Ended 3)Sterility-Non –Sterile 4)Usage-Reusable 5)Size-Small,Medium,Large	1	
3	**Acetic Acid 5% solution** • **Specifications-** Chemical Name – CHO3COOH 500 ml solution bottle	1 (bottle)	
4	**Sterile Cotton Swab Stick.** **Specification-** 1) Cotton tipped application single tip 6" with Wooden Hanle.	1 (100 Stick pack)	
5	टॉर्च (स्टिल २ सेल)	1	
6	मेजरींग टेप (राऊंड टेप)	1	

राष्ट्रीय असंसर्गजन्य रोग नियंत्रण कार्यक्रमाअंतर्गत एन.पी.सी.डी.सी.एस अंतर्गत
उपकरणे/साहित्य सामग्री खरेदी सन २०२६-२०२७

अ. क्र.	तपशील	नग	दर प्रती नग (सर्व शासकीय करासहित)
7	**1. Professional Adult Weighing Scales (Digital Platform)** * Capacity & Load: Up to 200 kg – 250 kg (Clinical) or up to 500 kg (Heavy-Duty Platform). * Accuracy & Resolution: Class-III accuracy with a graduation of 50 g to 100 g. * Platform Build: Minimum footprint of 330 mm x 340 mm or greater, constructed with heavy-duty cast iron or reinforced non-slip stainless steel. * Display & Readout: High-visibility 5-digit LED/LCD display (minimum digit height of 14 mm) with a swift reading time of under 5 seconds. * Key Functions: Automatic power-on auto-zeroing, manual/automatic tare provision, and an integrated telescopic height measuring rod (110 cm to 200 cm range with 1 mm graduations). * Power Supply: Dual-mode functionality operating on standard 220–240V AC power with an internal rechargeable battery backup providing at least 4 to 8 hours of continuous operational runtime.	1	
8	स्टेस्थोस्कोप (लाईफ टोन) (Adult)	1	
9	**Glucometer** **Key Specifications & Features** * Test Time: 5 seconds * Blood Sample Size: 0.5 µL (microlitres) * Measuring Range: 20 to 800 mg/dL * Coding: No coding required * Memory Capacity: Stores up to 500 test results * Display Type: LCD screen * Hygienic Strip Ejection: Built-in strip and lancet ejection button to remove used items without direct contact * Connectivity / Sync: Manual log to mobile application via a free companion app * Warranty: Lifetime warranty on the glucometer	1	
10	**Glucostrip,** **Specification – for Strips.** * No chip or coding required. * Single use strips. * 100 strips work with control-D machine. * Blood glucose test strips for use in whole blood. Each strips contains the enzyme glucose oxidase > 1 IU	1	
11	हाईट मेजरिंग स्केल	1	
12	एम.व्ही.ए. सिरिंज / एम.आर. सिरिंज	1	
13	**Pipelle Biopsy**	1	

CIVIL SURGEON
District Hospital, Sangli.