

Government of Maharashtra

## District Water Testing Laboratory, Latur

Office of the Senior Geologist,

Groundwater Surveys and Development Agency,

Latur-413512

Email: - [dllltr24@gmail.com](mailto:dllltr24@gmail.com)

No. DIST/LAB/GSDA/LTR/CHEM/

2023

Date: -

रासायनिक पृथःकरण अहवाल

पाणी नमुना पाठवनाऱ्याचे नाव व पत्ता:- Principal, KV BSF School, Ta. Chakur Dist Latur

पाठवनाऱ्याचे पत्र क्र. व दिनांक २९.११.२०२३

स्त्रोताचा स्थळ: KV BSF School, Chakur

जिल्हा प्रयोगशाळा संदर्भ क्र. व दिनांक- C/23/1839

परीक्षण सुरु केल्याचा दिनांक- ३०.११.२०२३

नमुना घेतल्याचा दिनांक २९.११.२०२३

स्त्रोताचा प्रकार आर ओ

Parameter	PH	TD S	TURBIDITY	TOTAL HARDNESS	ALKALINITY	CHLORIDE	CALCIUM	IRON	NITRATE	FLOURID	SULPHATE	REMARK
Unit	--	PPM	NTU	Mg/l	Mg/l	Mg/l	Mg/l	Mg/l	Mg/l	Mg/l	Mg/l	-----
Analysis Report	7.76	256	0	44	80	22	16	0.09	10.45	0.21	20.25	Fit (पिणयास योग्य)
Maximum Permissible Limit(BIS-IS10500-2012)	8.5	2000	5.0	600	600	1000	200	1.00	45.00	1.5	400	-----

NOTE: - 1) This Analysis Report Which Was Sample Provided By - Principal, KV BSF School, 1, Ta. Chakur Dist. Latur

2) The Quality of Water Is Dynamic. Therefore, it is Advised to Test the Sample at Regular Intervals.

3) As Per the analysis of this sample, the water is Fit for Drinking.

Chemist

Dist Lab, Latur

Senior Geologist

G.S.D. A, Latur